


Data Path : Z:\HPCHEM1\BNA F\DATA\BF122316\
 Data File : BF091941.D
 Acq On : 24 Dec 2016 00:32
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 21 16:17:51 2016



AutoFind: Scans 825, 826, 827; Background Corrected with Scan 821

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.7	12143	PASS
68	69	0.00	2	1.5	187	PASS
69	198	0.00	100	34.8	12199	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	44.9	15741	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	35042	PASS
199	198	5	9	6.2	2186	PASS
275	198	10	60	23.9	8376	PASS
365	198	1	100	2.3	808	PASS
441	442	0.01	24	15.4	4472	PASS
442	198	50	100	82.9	29044	PASS
443	442	15	24	18.3	5320	PASS