




Data Path : Z:\HPCHEM1\BNA F\Data\BF122415\
Data File : BF084073.D
Acq On : 24 Dec 2015 12:11
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 28 02:53:52 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 24 16:48:39 2015
Response via : Initial Calibration



TIC: BF084073.D

(69) Pentachlorophenol (C)




11.128min (-0.011) 2408.30ng

response 15929

Ion	Exp%	Act%
265.70	100	100
268.00	65.50	65.74
264.00	62.70	67.14
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF122415\
Data File : BF084073.D
Acq On : 24 Dec 2015 12:11
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 28 02:53:52 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 24 16:48:39 2015
Response via : Initial Calibration



TIC: BF084073.D

(76) Benzidine

12.648min (+0.001) 616.50ng

response 157614

Ion	Exp%	Act%
184.00	100	100
185.00	15.20	13.75
183.00	12.30	11.77
0.00	0.00	0.00