

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122420\
 Data File : BF122848.D
 Acq On : 24 Dec 2020 9:38
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 24 10:56:20 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 18:14:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	80	0.00
2	1,4-Dioxane	40.000	34.894	12.8	70	0.02
3	Pyridine	40.000	36.380	9.0	71	0.01
4	n-Nitrosodimethylamine	40.000	35.274	11.8	69	0.02
5 S	2-Fluorophenol	80.000	79.268	0.9	79	0.00
6	Aniline	40.000	39.540	1.2	78	0.00
7 S	Phenol-d6	80.000	75.372	5.8	75	0.00
8	2-Chlorophenol	40.000	38.438	3.9	76	0.00
9	Benzaldehyde	40.000	36.687	8.3	83	0.00
10 C	Phenol	40.000	38.189	4.5	77	0.00
11	bis(2-Chloroethyl)ether	40.000	38.106	4.7	75	0.00
12	1,3-Dichlorobenzene	40.000	37.205	7.0	74	0.00
13 C	1,4-Dichlorobenzene	40.000	36.631	8.4	74	0.00
14	1,2-Dichlorobenzene	40.000	34.975	12.6	73	0.00
15	Benzyl Alcohol	40.000	37.644	5.9	73	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	31.581	21.0	62	0.00
17	2-Methylphenol	40.000	37.011	7.5	72	0.00
18	Hexachloroethane	40.000	35.902	10.2	71	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.052	9.9	72	0.00
20	3+4-Methylphenols	40.000	36.739	8.2	75	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	81	0.00
22	Acetophenone	40.000	37.784	5.5	77	0.00
23 S	Nitrobenzene-d5	80.000	70.993	11.3	70	0.00
24	Nitrobenzene	40.000	38.072	4.8	76	0.00
25	Isophorone	40.000	37.496	6.3	74	0.00
26 C	2-Nitrophenol	40.000	40.829	-2.1	79	0.00
27	2,4-Dimethylphenol	40.000	43.717	-9.3	87	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.586	8.5	73	0.00
29 C	2,4-Dichlorophenol	40.000	37.037	7.4	73	0.00
30	1,2,4-Trichlorobenzene	40.000	36.559	8.6	74	0.00
31	Naphthalene	40.000	37.593	6.0	76	0.00
32	Benzoic acid	40.000	32.232	19.4	60	0.01
33	4-Chloroaniline	40.000	38.429	3.9	77	0.00
34 C	Hexachlorobutadiene	40.000	35.297	11.8	71	0.00
35	Caprolactam	40.000	38.058	4.9	75	0.01
36 C	4-Chloro-3-methylphenol	40.000	35.668	10.8	70	0.00
37	2-Methylnaphthalene	40.000	35.819	10.5	72	0.00
38	1-Methylnaphthalene	40.000	35.727	10.7	72	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	78	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.084	2.3	75	0.00
41 P	Hexachlorocyclopentadiene	40.000	33.614	16.0	61	0.00
42 S	2,4,6-Tribromophenol	80.000	78.960	1.3	75	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.944	7.6	70	0.00
44	2,4,5-Trichlorophenol	40.000	39.424	1.4	75	0.00
45 S	2-Fluorobiphenyl	80.000	69.950	12.6	73	0.00
46	1,1'-Biphenyl	40.000	39.226	1.9	76	0.00
47	2-Chloronaphthalene	40.000	37.310	6.7	73	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122420\
 Data File : BF122848.D
 Acq On : 24 Dec 2020 9:38
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 24 10:56:20 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 18:14:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	34.984	12.5	65	0.00
49	Acenaphthylene	40.000	37.674	5.8	73	0.00
50	Dimethylphthalate	40.000	37.572	6.1	73	0.00
51	2,6-Dinitrotoluene	40.000	39.698	0.8	75	0.00
52 C	Acenaphthene	40.000	37.077	7.3	71	0.00
53	3-Nitroaniline	40.000	36.237	9.4	68	0.00
54 P	2,4-Dinitrophenol	40.000	37.464	6.3	68	0.00
55	Dibenzofuran	40.000	37.175	7.1	73	0.00
56 P	4-Nitrophenol	40.000	34.679	13.3	63	0.01
57	2,4-Dinitrotoluene	40.000	39.019	2.5	73	0.00
58	Fluorene	40.000	36.426	8.9	75	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.660	8.4	69	0.00
60	Diethylphthalate	40.000	37.363	6.6	72	0.00
61	4-Chlorophenyl-phenylether	40.000	36.232	9.4	72	0.00
62	4-Nitroaniline	40.000	36.864	7.8	70	0.00
63	Azobenzene	40.000	37.150	7.1	71	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	80	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.420	-8.6	81	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.360	6.6	74	0.00
67	4-Bromophenyl-phenylether	40.000	35.932	10.2	71	0.00
68	Hexachlorobenzene	40.000	35.589	11.0	70	0.00
69	Atrazine	40.000	36.641	8.4	73	0.00
70 C	Pentachlorophenol	40.000	34.795	13.0	63	0.00
71	Phenanthrene	40.000	36.147	9.6	73	0.00
72	Anthracene	40.000	38.168	4.6	77	0.00
73	Carbazole	40.000	37.593	6.0	75	0.00
74	Di-n-butylphthalate	40.000	36.455	8.9	73	0.00
75 C	Fluoranthene	40.000	36.641	8.4	73	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	81	0.00
77	Benzydine	40.000	46.722	-16.8	83	0.00
78	Pyrene	40.000	36.898	7.8	74	0.00
79 S	Terphenyl-d14	80.000	69.319	13.4	73	0.00
80	Butylbenzylphthalate	40.000	35.893	10.3	71	0.00
81	Benzo(a)anthracene	40.000	37.289	6.8	76	0.00
82	3,3'-Dichlorobenzidine	40.000	42.426	-6.1	85	0.00
83	Chrysene	40.000	37.531	6.2	76	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	36.998	7.5	76	0.00
85 c	Di-n-octyl phthalate	40.000	37.031	7.4	75	0.00
86 I	Perylene-d12	20.000	20.000	0.0	86	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.276	6.8	81	0.01
88	Benzo(b)fluoranthene	40.000	36.653	8.4	75	0.00
89	Benzo(k)fluoranthene	40.000	38.475	3.8	83	0.00
90 C	Benzo(a)pyrene	40.000	36.986	7.5	78	0.00
91	Dibenzo(a,h)anthracene	40.000	38.662	3.3	85	0.01
92	Benzo(g,h,i)perylene	40.000	38.170	4.6	85	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122420\
Data File : BF122848.D
Acq On : 24 Dec 2020 9:38
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 24 10:56:20 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 15 18:14:23 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0