

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF122617\
 Data File : BF101730.D
 Acq On : 26 Dec 2017 22:54
 Operator : SJ/JU
 Sample : I6966-05
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DB-EW-3-0-4

Manual Integrations
 APPROVED

Sohil
 12/27/2017 4:15:05 PM

Quant Time: Dec 27 06:22:17 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 26 10:48:01 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	138306	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	504183	20.00	ng	0.00
38) Acenaphthene-d10	9.83	164	193258	20.00	ng	0.00
63) Phenanthrene-d10	11.32	188	328470	20.00	ng	0.00
75) Chrysene-d12	13.97	240	265660	20.00	ng	0.00
86) Perylene-d12	15.45	264	238167	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	315568	33.99	ng	0.00
7) Phenol-d6	6.43	99	586340	50.74	ng	0.00
23) Nitrobenzene-d5	7.36	82	343986	37.98	ng	0.00
41) 2,4,6-Tribromophenol	10.62	330	171302	91.99	ng	0.00
44) 2-Fluorobiphenyl	9.15	172	744161	59.73	ng	0.00
78) Terphenyl-d14	12.90	244	702186	63.72	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.55	163	38249	2.461	ng	Qvalue # 97
57) Fluorene	10.38	166	26847	2.007	ng	97
70) Phenanthrene	11.35	178	254695	13.943	ng	99
71) Anthracene	11.40	178	57573m	3.086	ng	
74) Fluoranthene	12.54	202	361071	18.832	ng	98
77) Pyrene	12.77	202	322954	16.198	ng	99
80) Benzo(a)anthracene	13.97	228	150394	8.573	ng	97
82) Chrysene	14.00	228	160230	9.674	ng	97
85) Indeno(1,2,3-cd)pyrene	16.94	276	75011	5.086	ng	92
87) Benzo(b)fluoranthene	15.03	252	169307m	11.663	ng	
88) Benzo(k)fluoranthene	15.05	252	64659m	4.581	ng	
89) Benzo(a)pyrene	15.39	252	119563	8.934	ng	# 83
91) Benzo(g,h,i)perylene	17.39	276	73684	6.476	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF122617\
Data File : BF101730.D
Acq On : 26 Dec 2017 22:54
Operator : SJ/JU
Sample : I6966-05
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DB-EW-3-0-4

Manual Integrations
APPROVED

Sohil
12/27/2017 4:15:05 PM

Quant Time: Dec 27 06:22:17 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 26 10:48:01 2017
Response via : Initial Calibration

