



Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF122619\
Data File : BF118308.D
Acq On : 26 Dec 2019 11:04
Operator : JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 27 07:13:22 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF122419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 24 15:58:52 2019
Response via : Initial Calibration



TIC: BF118308.D

(77) Benzidine



12.733min (-0.000) 0.00ng

response 713540

Ion	Exp%	Act%
184.00	100	100
185.00	15.40	14.64
183.00	11.80	11.93
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF122619\
Data File : BF118308.D
Acq On : 26 Dec 2019 11:04
Operator : JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 27 07:13:22 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF122419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 24 15:58:52 2019
Response via : Initial Calibration



TIC: BF118308.D

(70) Pentachlorophenol (C)

11.198min (-0.000) 34872.85ng

response 114009

Ion	Exp%	Act%
265.70	100	100
268.00	63.30	63.65
264.00	62.40	64.16
0.00	0.00	0.00