

Data Path : Z:\HPCHEM1\BNA F\DATA\BF122717\
 Data File : BF101779.D
 Acq On : 27 Dec 2017 22:49
 Operator : SJ/JU
 Sample : I7064-05 2X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SAMPLE-3

Manual Integrations
 APPROVED

Sohil
 12/28/2017 2:03:05 PM

Quant Time: Dec 28 05:48:48 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 26 10:48:01 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	145005	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	548216	20.00	ng	0.00
38) Acenaphthene-d10	9.83	164	217573	20.00	ng	0.00
63) Phenanthrene-d10	11.32	188	357298	20.00	ng	0.00
75) Chrysene-d12	13.97	240	315008	20.00	ng	0.00
86) Perylene-d12	15.44	264	268164	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	525493	53.98	ng	0.01
7) Phenol-d6	6.44	99	638490	52.70	ng	0.00
23) Nitrobenzene-d5	7.36	82	375896	38.17	ng	0.00
41) 2,4,6-Tribromophenol	10.63	330	110687	52.80	ng	0.00
44) 2-Fluorobiphenyl	9.15	172	660627	47.10	ng	0.00
78) Terphenyl-d14	12.90	244	483713	37.02	ng	0.00
Target Compounds						
10) Phenol	6.46	94	30825	2.232	ng	Qvalue # 54
49) Dimethylphthalate	9.55	163	60730	3.470	ng	98
70) Phenanthrene	11.34	178	150077	7.553	ng	97
71) Anthracene	11.40	178	46282	2.280	ng	97
74) Fluoranthene	12.54	202	294012	14.097	ng	98
77) Pyrene	12.77	202	270087	11.424	ng	99
80) Benzo(a)anthracene	13.96	228	169316	8.140	ng	97
82) Chrysene	14.00	228	157312	8.010	ng	98
87) Benzo(b)fluoranthene	15.02	252	176690m	10.810	ng	
88) Benzo(k)fluoranthene	15.04	252	72437m	4.558	ng	
89) Benzo(a)pyrene	15.39	252	121818	8.085	ng	# 92
91) Benzo(a,h,i)perylene	17.40	276	64454	5.031	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF122717\
Data File : BF101779.D
Acq On : 27 Dec 2017 22:49
Operator : SJ/JU
Sample : I7064-05 2X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SAMPLE-3

Manual Integrations
APPROVED

Sohil
12/28/2017 2:03:05 PM

Quant Time: Dec 28 05:48:48 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 26 10:48:01 2017
Response via : Initial Calibration

