

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF122719\  
 Data File : BF118355.D  
 Acq On : 27 Dec 2019 23:24  
 Operator : JU  
 Sample : K6443-01 5X  
 Misc :  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 PH-1

Quant Time: Dec 28 03:08:24 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF122419.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Dec 24 15:58:52 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.85  | 152  | 125350   | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.14  | 136  | 467637   | 20.00 | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.90  | 164  | 219423   | 20.00 | ng    | 0.00     |
| 64) Phenanthrene-d10        | 11.39 | 188  | 394143   | 20.00 | ng    | 0.00     |
| 76) Chrysene-d12            | 14.05 | 240  | 338917   | 20.00 | ng    | 0.00     |
| 87) Perylene-d12            | 15.54 | 264  | 265390   | 20.00 | ng    | 0.01     |
| System Monitoring Compounds |       |      |          |       |       |          |
| 5) 2-Fluorophenol           | 5.47  | 112  | 127979   | 17.43 | ng    | 0.00     |
| 7) Phenol-d6                | 6.49  | 99   | 159720   | 17.53 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.42  | 82   | 87440    | 10.69 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 10.70 | 330  | 40127    | 14.81 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 9.22  | 172  | 191732   | 13.34 | ng    | 0.00     |
| 79) Terphenyl-d14           | 12.98 | 244  | 211563   | 10.36 | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF122719\  
Data File : BF118355.D  
Acq On : 27 Dec 2019 23:24  
Operator : JU  
Sample : K6443-01 5X  
Misc :  
ALS Vial : 23 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
PH-1

Quant Time: Dec 28 03:08:24 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF122419.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Dec 24 15:58:52 2019  
Response via : Initial Calibration

