

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122723\
 Data File : BF136757.D
 Acq On : 27 Dec 2023 14:26
 Operator : CG\JU
 Sample : 05967-13MSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COLTOMSD

Quant Time: Dec 27 15:08:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 21 01:54:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.781	152	60631	20.000	ng	-0.01
21) Naphthalene-d8	8.063	136	224935	20.000	ng	0.00
39) Acenaphthene-d10	9.816	164	104762	20.000	ng	-0.01
64) Phenanthrene-d10	11.310	188	179003	20.000	ng	0.00
76) Chrysene-d12	13.963	240	113608	20.000	ng	0.00
86) Perylene-d12	15.445	264	143829	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	128896	33.170	ng	0.00
7) Phenol-d6	6.428	99	90922	18.057	ng	-0.01
23) Nitrobenzene-d5	7.345	82	292736	66.595	ng	-0.01
42) 2,4,6-Tribromophenol	10.616	330	104816	84.953	ng	0.00
45) 2-Fluorobiphenyl	9.139	172	543396	69.764	ng	-0.01
79) Terphenyl-d14	12.904	244	478790	58.895	ng	-0.01
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.557	88	23791	14.694	ng	98
3) Pyridine	3.322	79	61537	15.577	ng	96
4) n-Nitrosodimethylamine	3.263	42	29039	13.765	ng	# 96
6) Aniline	6.445	93	73874	14.674	ng	# 86
8) 2-Chlorophenol	6.575	128	88950	20.917	ng	93
9) Benzaldehyde	6.340	77	39367	13.745	ng	93
10) Phenol	6.440	94	40025	7.446	ng	91
11) bis(2-Chloroethyl)ether	6.522	93	135857	33.236	ng	95
12) 1,3-Dichlorobenzene	6.722	146	150555	32.454	ng	99
13) 1,4-Dichlorobenzene	6.804	146	152199	32.110	ng	95
14) 1,2-Dichlorobenzene	6.951	146	145875	33.073	ng	99
15) Benzyl Alcohol	6.928	79	75637	22.266	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.057	45	219776	32.885	ng	99
17) 2-Methylphenol	7.045	107	58616	16.639	ng	98
18) Hexachloroethane	7.292	117	56077	32.576	ng	92
19) n-Nitroso-di-n-propyla...	7.192	70	99366	32.555	ng	99
20) 3+4-Methylphenols	7.198	107	63357	14.395	ng	# 85
22) Acetophenone	7.192	105	206777	36.676	ng	# 88
24) Nitrobenzene	7.363	77	143339	32.552	ng	96
25) Isophorone	7.598	82	266417	33.308	ng	98
26) 2-Nitrophenol	7.681	139	52621	24.977	ng	97
27) 2,4-Dimethylphenol	7.722	122	72148	28.129	ng	97
28) bis(2-Chloroethoxy)met...	7.810	93	164051	33.742	ng	98
29) 2,4-Dichlorophenol	7.928	162	78178	23.676	ng	96
30) 1,2,4-Trichlorobenzene	8.004	180	123406	33.542	ng	99
31) Naphthalene	8.081	128	407225	32.954	ng	99
32) Benzoic acid	7.798	122	10079	4.061	ng	98
33) 4-Chloroaniline	8.134	127	55594	12.185	ng	96
34) Hexachlorobutadiene	8.198	225	74070	33.137	ng	98
35) Caprolactam	8.481	113	5629	5.167	ng	# 79
36) 4-Chloro-3-methylphenol	8.622	107	74257	19.976	ng	99
37) 2-Methylnaphthalene	8.775	142	254182	31.961	ng	100
38) 1-Methylnaphthalene	8.875	142	243652	31.228	ng	100
40) 1,2,4,5-Tetrachloroben...	8.939	216	126770	39.073	ng	99
41) Hexachlorocyclopentadiene	8.928	237	93529	89.775	ng	99
43) 2,4,6-Trichlorophenol	9.057	196	53155	25.547	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122723\
 Data File : BF136757.D
 Acq On : 27 Dec 2023 14:26
 Operator : CG\JU
 Sample : 05967-13MSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COLTOMSD

Quant Time: Dec 27 15:08:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 21 01:54:25 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.104	196	55054	23.744	ng	99
46) 1,1'-Biphenyl	9.239	154	323735	36.825	ng	100
47) 2-Chloronaphthalene	9.263	162	238792	35.730	ng	100
48) 2-Nitroaniline	9.363	65	71210	34.026	ng	98
49) Acenaphthylene	9.681	152	384372	37.628	ng	99
50) Dimethylphthalate	9.539	163	257303	33.625	ng	99
51) 2,6-Dinitrotoluene	9.604	165	56827	32.722	ng	# 85
52) Acenaphthene	9.851	154	213613	33.735	ng	99
53) 3-Nitroaniline	9.775	138	30488	16.573	ng	95
54) 2,4-Dinitrophenol	9.886	184	28588	32.050	ng	# 75
55) Dibenzofuran	10.028	168	334255	34.823	ng	97
56) 4-Nitrophenol	9.957	139	13628	10.974	ng	# 84
57) 2,4-Dinitrotoluene	10.010	165	71032	31.507	ng	98
58) Fluorene	10.369	166	258811	33.982	ng	98
59) 2,3,4,6-Tetrachlorophenol	10.145	232	39105	21.909	ng	# 98
60) Diethylphthalate	10.239	149	257086	32.381	ng	99
61) 4-Chlorophenyl-phenyle...	10.363	204	126319	34.122	ng	92
62) 4-Nitroaniline	10.392	138	54279	30.606	ng	92
63) Azobenzene	10.522	77	239654	32.318	ng	99
65) 4,6-Dinitro-2-methylph...	10.422	198	21301	20.764	ng	94
66) n-Nitrosodiphenylamine	10.480	169	218848	37.636	ng	99
67) 4-Bromophenyl-phenylether	10.851	248	72006	36.221	ng	97
68) Hexachlorobenzene	10.916	284	79878	36.056	ng	96
69) Atrazine	11.016	200	65950	44.343	ng	96
70) Pentachlorophenol	11.116	266	40189	38.919	ng	95
71) Phenanthrene	11.333	178	328970	35.817	ng	98
72) Anthracene	11.386	178	326406	35.091	ng	99
73) Carbazole	11.545	167	286173	32.632	ng	99
74) Di-n-butylphthalate	11.874	149	374919	35.035	ng	100
75) Fluoranthene	12.527	202	308353	31.386	ng	99
77) Benzidine	12.657	184	98185	29.311	ng	99
78) Pyrene	12.757	202	312622	28.031	ng	100
80) Butylbenzylphthalate	13.386	149	131044	32.292	ng	95
81) Benzo(a)anthracene	13.957	228	271532	34.420	ng	98
82) 3,3'-Dichlorobenzidine	13.916	252	49633	22.154	ng	# 99
83) Chrysene	13.992	228	254554	32.997	ng	98
84) Bis(2-ethylhexyl)phtha...	13.945	149	180235	35.300	ng	97
85) Di-n-octyl phthalate	14.563	149	325170	41.171	ng	98
87) Indeno(1,2,3-cd)pyrene	16.951	276	361990	34.858	ng	99
88) Benzo(b)fluoranthene	15.010	252	306630	31.919	ng	98
89) Benzo(k)fluoranthene	15.039	252	267047	29.419	ng	99
90) Benzo(a)pyrene	15.380	252	267596	33.000	ng	98
91) Dibenzo(a,h)anthracene	16.968	278	303792	35.658	ng	99
92) Benzo(g,h,i)perylene	17.404	276	277775	31.834	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122723\
Data File : BF136757.D
Acq On : 27 Dec 2023 14:26
Operator : CG\JU
Sample : 05967-13MSD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
COLTOMSD

Quant Time: Dec 27 15:08:49 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 21 01:54:25 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.781 min Scan# 79
Delta R.T. -0.011 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD

Tgt Ion:152 Resp: 60631
Ion Ratio Lower Upper
152 100
150 157.8 127.0 190.6
115 61.3 49.0 73.4



#2
1,4-Dioxane
Concen: 14.694 ng
RT: 2.557 min Scan# 80
Delta R.T. 0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion: 88 Resp: 23791
Ion Ratio Lower Upper
88 100
58 81.9 67.4 101.0
43 33.6 27.7 41.5








#3
Pyridine
Concen: 15.577 ng
RT: 3.322 min Scan# 21
Delta R.T. 0.018 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion: 79 Resp: 61537
Ion Ratio Lower Upper
79 100
52 83.4 67.7 101.5
51 36.4 34.0 51.0



#4
n-Nitrosodimethylamine
Concen: 13.765 ng
RT: 3.263 min Scan# 200
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion: 42 Resp: 29039
Ion Ratio Lower Upper
42 100
74 115.2 96.0 144.0
44 8.7 5.4 8.2#









#5
2-Fluorophenol
Concen: 33.170 ng
RT: 5.422 min Scan# 50
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion:112 Resp: 128896
Ion Ratio Lower Upper
112 100
64 63.6 51.1 76.7
63 32.5 26.1 39.1



#6
Aniline
Concen: 14.674 ng
RT: 6.445 min Scan# 741
Delta R.T. -0.018 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion: 93 Resp: 73874
Ion Ratio Lower Upper
93 100
66 51.9 34.6 51.8#
65 29.0 17.8 26.8#









#7
Phenol-d6
Concen: 18.057 ng
RT: 6.428 min Scan# 71
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion: 99 Resp: 90922
Ion Ratio Lower Upper
99 100
42 18.3 16.1 24.1
71 31.8 27.4 41.2



#8
2-Chlorophenol
Concen: 20.917 ng
RT: 6.575 min Scan# 763
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion: 128 Resp: 88950
Ion Ratio Lower Upper
128 100
130 31.5 11.3 51.3
64 45.1 32.5 72.5









#9
Benzaldehyde
Concen: 13.745 ng
RT: 6.340 min Scan# 71
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD

Tgt Ion: 77 Resp: 39367
Ion Ratio Lower Upper
77 100
105 108.7 84.8 124.8
106 107.3 77.8 117.8



#10
Phenol
Concen: 7.446 ng
RT: 6.440 min Scan# 740
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion: 94 Resp: 40025
Ion Ratio Lower Upper
94 100
65 37.1 16.6 56.6
66 63.0 32.6 72.6





#11

bis(2-Chloroethyl)ether

Concen: 33.236 ng

RT: 6.522 min Scan# 71

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion: 93 Resp: 135857

Ion Ratio Lower Upper

93 100

63 75.3 61.2 101.2

95 32.2 11.6 51.6



#12

1,3-Dichlorobenzene

Concen: 32.454 ng

RT: 6.722 min Scan# 788

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26




Tgt Ion:146 Resp: 150555


Ion Ratio Lower Upper

146 100

148 62.7 50.5 75.7

75 31.9 26.2 39.4





#13

1,4-Dichlorobenzene

Concen: 32.110 ng

RT: 6.804 min Scan# 80

Delta R.T. -0.006 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion:146 Resp: 152199

Ion Ratio Lower Upper

146 100

148 65.0 49.2 73.8

111 40.8 35.2 52.8



#14

1,2-Dichlorobenzene

Concen: 33.073 ng

RT: 6.951 min Scan# 827

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26




Tgt Ion:146 Resp: 145875

Ion Ratio Lower Upper

146 100

148 62.7 49.6 74.4

111 44.2 36.2 54.2









#15
Benzyl Alcohol
Concen: 22.266 ng
RT: 6.928 min Scan# 81
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion: 79 Resp: 75637
Ion Ratio Lower Upper
79 100
108 76.9 61.0 91.6
77 60.4 49.8 74.8



#16
2,2'-oxybis(1-Chloropropane)
Concen: 32.885 ng
RT: 7.057 min Scan# 845
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion: 45 Resp: 219776
Ion Ratio Lower Upper
45 100
77 14.2 0.0 34.3
79 12.2 0.0 31.4









#17
2-Methylphenol
Concen: 16.639 ng
RT: 7.045 min Scan# 84
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD

Tgt Ion:	107	Resp:	58616
Ion	Ratio	Lower	Upper
107	100		
108	110.9	90.3	135.5
77	50.6	38.7	58.1
79	49.3	37.8	56.6



#18
Hexachloroethane
Concen: 32.576 ng
RT: 7.292 min Scan# 885
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:	117	Resp:	56077
Ion	Ratio	Lower	Upper
117	100		
119	96.4	75.8	113.6
201	100.9	69.1	103.7









#19
n-Nitroso-di-n-propylamine
Concen: 32.555 ng
RT: 7.192 min Scan# 868
Delta R.T. -0.018 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion:	70	Resp:	99366
Ion Ratio	Lower	Upper	
70	100		
42	66.4	54.2	81.4
101	9.0	7.6	11.4
130	21.1	17.1	25.7



#20
3+4-Methylphenols
Concen: 14.395 ng
RT: 7.198 min Scan# 869
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:	107	Resp:	63357
Ion Ratio	Lower	Upper	
107	100		
108	87.2	66.0	106.0
77	102.3	114.2	154.2#
79	25.6	7.2	47.2









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.063 min Scan# 1016
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt	Ion:136	Resp:	224935
Ion	Ratio	Lower	Upper
136	100		
137	10.3	9.0	13.4
54	7.3	7.2	10.8
68	5.1	4.9	7.3



#22
Acetophenone
Concen: 36.676 ng
RT: 7.192 min Scan# 868
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt	Ion:105	Resp:	206777
Ion	Ratio	Lower	Upper
105	100		
71	7.2	4.8	7.2#
51	29.1	32.8	49.2#
120	22.4	17.0	25.6





#23

Nitrobenzene-d5

Concen: 66.595 ng

RT: 7.345 min Scan# 894

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion: 82 Resp: 292736

Ion Ratio Lower Upper

82 100

128 45.4 35.3 52.9

54 56.4 44.5 66.7



#24

Nitrobenzene

Concen: 32.552 ng

RT: 7.363 min Scan# 897

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26




Tgt Ion: 77 Resp: 143339


Ion Ratio Lower Upper

77 100

123 48.7 36.7 55.1

65 14.2 11.8 17.6








#25
Isophorone
Concen: 33.308 ng
RT: 7.598 min Scan# 91
Delta R.T. -0.018 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion: 82 Resp: 266417
Ion Ratio Lower Upper
82 100
95 7.0 5.5 8.3
138 17.1 14.5 21.7



#26
2-Nitrophenol
Concen: 24.977 ng
RT: 7.681 min Scan# 951
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:139 Resp: 52621
Ion Ratio Lower Upper
139 100
109 28.3 21.9 32.9
65 51.1 38.6 58.0





#27

2,4-Dimethylphenol

Concen: 28.129 ng

RT: 7.722 min Scan# 91

Delta R.T. -0.006 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion:122 Resp: 72148

Ion Ratio Lower Upper

122 100

107 110.9 92.0 138.0

121 59.3 47.5 71.3



#28

bis(2-Chloroethoxy)methane

Concen: 33.742 ng

RT: 7.810 min Scan# 973

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26




Tgt Ion: 93 Resp: 164051

Ion Ratio Lower Upper

93 100

95 33.4 25.5 38.3

123 11.1 9.2 13.8









#29
2,4-Dichlorophenol
Concen: 23.676 ng
RT: 7.928 min Scan# 994
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion:162 Resp: 78178
Ion Ratio Lower Upper
162 100
164 62.6 41.7 81.7
98 34.2 19.5 59.5



#30
1,2,4-Trichlorobenzene
Concen: 33.542 ng
RT: 8.004 min Scan# 1006
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:180 Resp: 123406
Ion Ratio Lower Upper
180 100
182 94.7 75.4 113.0
145 29.8 23.1 34.7





#31

Naphthalene

Concen: 32.954 ng

RT: 8.081 min Scan# 1019

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion:128 Resp: 407225

Ion Ratio Lower Upper

128 100

129 11.0 8.7 13.1

127 13.2 10.8 16.2



#32

Benzoic acid

Concen: 4.061 ng

RT: 7.798 min Scan# 971

Delta R.T. -0.059 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26




Tgt Ion:122 Resp: 10079


Ion Ratio Lower Upper

122 100

105 122.1 100.8 140.8

77 87.9 70.9 110.9








#33
4-Chloroaniline
Concen: 12.185 ng
RT: 8.134 min Scan# 1030
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD

Tgt Ion:	127	Resp:	55594
Ion	Ratio	Lower	Upper
127	100		
129	32.1	26.3	39.5
65	27.5	25.8	38.8
92	19.2	15.8	23.6



#34
Hexachlorobutadiene
Concen: 33.137 ng
RT: 8.198 min Scan# 1039
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:	225	Resp:	74070
Ion	Ratio	Lower	Upper
225	100		
223	62.8	50.6	76.0
227	61.7	51.9	77.9








#35
 Caprolactam
 Concen: 5.167 ng
 RT: 8.481 min Scan# 1087
 Delta R.T. -0.041 min
 Lab File: BF136757.D
 Acq: 27 Dec 2023 14:26




Instrument :
 BNA_F
 ClientSampleId :
 COLTOMSD


Tgt Ion:113 Resp: 5629
 Ion Ratio Lower Upper
 113 100
 55 146.2 163.1 203.1#
 56 124.7 121.7 161.7



#36
 4-Chloro-3-methylphenol
 Concen: 19.976 ng
 RT: 8.622 min Scan# 1111
 Delta R.T. -0.012 min
 Lab File: BF136757.D
 Acq: 27 Dec 2023 14:26

Tgt Ion:107 Resp: 74257
 Ion Ratio Lower Upper
 107 100
 144 24.1 20.2 30.2
 142 78.0 62.3 93.5





#37

2-Methylnaphthalene

Concen: 31.961 ng

RT: 8.775 min Scan# 1137

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :





COLTOMSD

Tgt Ion:142 Resp: 254182

Ion Ratio Lower Upper

142 100

141 88.9 71.1 106.7



#38

1-Methylnaphthalene

Concen: 31.228 ng

RT: 8.875 min Scan# 1154

Delta R.T. -0.012 min

Lab File: BF136757.D




Acq: 27 Dec 2023 14:26


Tgt Ion:142 Resp: 243652

Ion Ratio Lower Upper

142 100

141 91.9 73.6 110.4





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.816 min Scan# 1165

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion:164 Resp: 104762

Ion Ratio Lower Upper

164 100

162 101.2 81.6 122.4

160 45.4 34.9 52.3



#40

1,2,4,5-Tetrachlorobenzene

Concen: 39.073 ng

RT: 8.939 min Scan# 1165

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Tgt Ion:216 Resp: 126770




Ion Ratio Lower Upper


216 100

214 78.6 62.3 93.5

179 20.6 16.4 24.6

108 18.4 14.0 21.0








#41
Hexachlorocyclopentadiene
Concen: 89.775 ng
RT: 8.928 min Scan# 1163
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD

Tgt Ion:237 Resp: 93529
Ion Ratio Lower Upper
237 100
235 62.1 43.2 83.2
272 14.1 0.0 34.2



#42
2,4,6-Tribromophenol
Concen: 84.953 ng
RT: 10.616 min Scan# 1450
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:330 Resp: 104816
Ion Ratio Lower Upper
330 100
332 97.2 77.4 116.0
141 28.9 23.4 35.2





#43

2,4,6-Trichlorophenol

Concen: 25.547 ng

RT: 9.057 min Scan# 1185

Delta R.T. -0.006 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion:196 Resp: 53155

Ion Ratio Lower Upper

196 100

198 94.0 75.1 112.7

200 30.5 22.5 33.7



#44

2,4,5-Trichlorophenol

Concen: 23.744 ng

RT: 9.104 min Scan# 1193

Delta R.T. -0.006 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Tgt Ion:196 Resp: 55054




Ion Ratio Lower Upper


196 100

198 97.7 77.0 115.6

97 45.4 36.1 54.1

132 26.4 21.8 32.6





#45

2-Fluorobiphenyl

Concen: 69.764 ng

RT: 9.139 min Scan# 1199

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion:172 Resp: 543396

Ion Ratio Lower Upper

172 100

171 36.0 29.0 43.6

170 23.2 19.0 28.4



#46

1,1'-Biphenyl

Concen: 36.825 ng

RT: 9.239 min Scan# 1216

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26




Tgt Ion:154 Resp: 323735


Ion Ratio Lower Upper

154 100

153 41.0 21.1 61.1

76 11.8 0.0 31.7





#47

2-Chloronaphthalene

Concen: 35.730 ng

RT: 9.263 min Scan# 1

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion:162 Resp: 238792

Ion Ratio Lower Upper

162 100

127 42.1 33.6 50.4

164 32.2 25.7 38.5



#48

2-Nitroaniline

Concen: 34.026 ng

RT: 9.363 min Scan# 1237

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26




Tgt Ion: 65 Resp: 71210


Ion Ratio Lower Upper

65 100

92 64.9 53.2 79.8

138 100.0 81.1 121.7





#49

Acenaphthylene

Concen: 37.628 ng

RT: 9.681 min Scan# 11

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion:152 Resp: 384372

Ion Ratio Lower Upper

152 100

151 20.9 16.2 24.4

153 13.1 10.2 15.2



#50

Dimethylphthalate

Concen: 33.625 ng

RT: 9.539 min Scan# 1267

Delta R.T. -0.018 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26




Tgt Ion:163 Resp: 257303


Ion Ratio Lower Upper

163 100

194 3.7 3.3 4.9

164 10.1 7.8 11.6









#51
2,6-Dinitrotoluene
Concen: 32.722 ng
RT: 9.604 min Scan# 11
Delta R.T. -0.018 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion:165 Resp: 56827
Ion Ratio Lower Upper
165 100
63 61.1 39.8 59.6#
89 57.5 38.8 58.2



#52
Acenaphthene
Concen: 33.735 ng
RT: 9.851 min Scan# 1320
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:154 Resp: 213613
Ion Ratio Lower Upper
154 100
153 115.0 90.7 136.1
152 55.2 43.7 65.5









#53
3-Nitroaniline
Concen: 16.573 ng
RT: 9.775 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion:138 Resp: 30488
Ion Ratio Lower Upper
138 100
108 11.5 8.6 12.8
92 113.1 95.2 142.8



#54
2,4-Dinitrophenol
Concen: 32.050 ng
RT: 9.886 min Scan# 1326
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:184 Resp: 28588
Ion Ratio Lower Upper
184 100
63 74.5 40.9 61.3#
154 65.6 42.7 64.1#









#55
Dibenzenofuran
Concen: 34.823 ng
RT: 10.028 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion:168 Resp: 334255
Ion Ratio Lower Upper
168 100
139 34.6 29.3 43.9
169 13.2 11.0 16.4



#56
4-Nitrophenol
Concen: 10.974 ng
RT: 9.957 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:139 Resp: 13628
Ion Ratio Lower Upper
139 100
109 66.9 45.8 85.8
65 123.0 77.8 117.8#









#57
2,4-Dinitrotoluene
Concen: 31.507 ng
RT: 10.010 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD

Tgt Ion:165	Resp:	71032
Ion	Ratio	Lower Upper
165	100	
63	48.9	41.0 61.6
89	74.2	59.9 89.9
182	2.7	2.0 3.0



#58
Fluorene
Concen: 33.982 ng
RT: 10.369 min Scan# 1408
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:166	Resp:	258811
Ion	Ratio	Lower Upper
166	100	
165	94.2	73.7 110.5
167	13.5	10.4 15.6





#59

2,3,4,6-Tetrachlorophenol

Concen: 21.909 ng

RT: 10.145 min Scan# 1

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD

Tgt Ion:232 Resp: 39105





Ion Ratio Lower Upper

232 100

131 44.2 35.1 52.7

130 2.4 1.6 2.4#

166 31.6 23.6 35.4



#60

Diethylphthalate

Concen: 32.381 ng

RT: 10.239 min Scan# 1386

Delta R.T. -0.018 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26




Tgt Ion:149 Resp: 257086


Ion Ratio Lower Upper

149 100

177 20.9 16.9 25.3

150 12.5 9.3 13.9









#61
4-Chlorophenyl-phenylether
Concen: 34.122 ng
RT: 10.363 min Scan# 1407
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion:204 Resp: 126319
Ion Ratio Lower Upper
204 100
206 30.6 25.6 38.4
141 52.7 48.4 72.6



#62
4-Nitroaniline
Concen: 30.606 ng
RT: 10.392 min Scan# 1412
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:138 Resp: 54279
Ion Ratio Lower Upper
138 100
92 45.9 32.1 72.1
108 63.5 50.1 90.1









#63
Azobenzene
Concen: 32.318 ng
RT: 10.522 min Scan# 1436
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion: 77	Resp: 239654
Ion Ratio	Lower Upper
77 100	
182 24.2	6.0 46.0
105 22.1	2.3 42.3
51 31.1	11.1 51.1



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.310 min Scan# 1568
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:188	Resp: 179003
Ion Ratio	Lower Upper
188 100	
94 8.4	7.5 11.3
80 8.9	8.1 12.1









#65
4,6-Dinitro-2-methylphenol
Concen: 20.764 ng
RT: 10.422 min Scan# 1419
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD

Tgt Ion:198 Resp: 21301
Ion Ratio Lower Upper
198 100
51 50.6 34.1 74.1
105 43.0 27.3 67.3



#66
n-Nitrosodiphenylamine
Concen: 37.636 ng
RT: 10.480 min Scan# 1427
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:169 Resp: 218848
Ion Ratio Lower Upper
169 100
168 66.4 53.8 80.6
167 35.1 28.7 43.1









#67
4-Bromophenyl-phenylether
Concen: 36.221 ng
RT: 10.851 min Scan# 1490
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD

Tgt Ion:248 Resp: 72006
Ion Ratio Lower Upper
248 100
250 102.2 77.8 116.6
141 74.0 59.2 88.8



#68
Hexachlorobenzene
Concen: 36.056 ng
RT: 10.916 min Scan# 1501
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:284 Resp: 79878
Ion Ratio Lower Upper
284 100
142 32.1 23.4 35.0
249 29.3 22.3 33.5





#69

Atrazine

Concen: 44.343 ng

RT: 11.016 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion:200 Resp: 65950

Ion Ratio Lower Upper

200 100

173 26.2 7.6 47.6

215 51.3 28.3 68.3



#70

Pentachlorophenol

Concen: 38.919 ng

RT: 11.116 min Scan# 1535

Delta R.T. -0.006 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26




Tgt Ion:266 Resp: 40189


Ion Ratio Lower Upper

266 100

268 70.1 52.6 78.8

264 63.8 49.0 73.4





#71

Phenanthrene

Concen: 35.817 ng

RT: 11.333 min Scan# 1572

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion:178 Resp: 328970

Ion Ratio Lower Upper

178 100

176 19.9 15.3 22.9

179 15.6 11.9 17.9



#72

Anthracene

Concen: 35.091 ng

RT: 11.386 min Scan# 1581

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26




Tgt Ion:178 Resp: 326406


Ion Ratio Lower Upper

178 100

176 19.0 14.9 22.3

179 14.7 11.8 17.6





#73

Carbazole

Concen: 32.632 ng

RT: 11.545 min Scan# 1608

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion:167 Resp: 286173

Ion Ratio Lower Upper

167 100

166 21.1 17.3 25.9

139 12.1 9.8 14.6



#74

Di-n-butylphthalate

Concen: 35.035 ng

RT: 11.874 min Scan# 1664

Delta R.T. -0.012 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26




Tgt Ion:149 Resp: 374919

Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.2

104 4.3 3.6 5.4









#75
Fluoranthene
Concen: 31.386 ng
RT: 12.527 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion:202 Resp: 308353
Ion Ratio Lower Upper
202 100
101 10.3 0.0 29.3
203 17.3 0.0 37.1



#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.963 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:240 Resp: 113608
Ion Ratio Lower Upper
240 100
120 11.0 10.3 15.5
236 24.0 19.4 29.2









#77
Benzidine
Concen: 29.311 ng
RT: 12.657 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion:184 Resp: 98185
Ion Ratio Lower Upper
184 100
185 13.3 11.2 16.8
183 11.5 9.3 13.9



#78
Pyrene
Concen: 28.031 ng
RT: 12.757 min Scan# 1814
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26


Tgt Ion:202 Resp: 312622
Ion Ratio Lower Upper
202 100
200 19.7 15.9 23.9
203 17.4 13.8 20.6








#79
 Terphenyl-d14
 Concen: 58.895 ng
 RT: 12.904 min Scan# 1839
 Delta R.T. -0.012 min
 Lab File: BF136757.D
 Acq: 27 Dec 2023 14:26

Instrument :
 BNA_F
 ClientSampleId :
 COLTOMSD







Tgt Ion:244 Resp: 478790
 Ion Ratio Lower Upper
 244 100
 212 6.5 5.0 7.4
 122 11.2 9.0 13.4



#80
 Butylbenzylphthalate
 Concen: 32.292 ng
 RT: 13.386 min Scan# 1921
 Delta R.T. -0.006 min
 Lab File: BF136757.D
 Acq: 27 Dec 2023 14:26

Tgt Ion:149 Resp: 131044
 Ion Ratio Lower Upper
 149 100
 91 62.6 54.1 81.1
 206 21.2 17.1 25.7









#81
Benzo(a)anthracene
Concen: 34.420 ng
RT: 13.957 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion:228 Resp: 271532
Ion Ratio Lower Upper
228 100
226 27.4 21.4 32.0
229 20.0 14.9 22.3



#82
3,3'-Dichlorobenzidine
Concen: 22.154 ng
RT: 13.916 min Scan# 2011
Delta R.T. -0.012 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:252 Resp: 49633
Ion Ratio Lower Upper
252 100
254 65.2 52.2 78.4
126 12.4 8.0 12.0#









#83
Chrysene
Concen: 32.997 ng
RT: 13.992 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD

Tgt Ion:228 Resp: 254554
Ion Ratio Lower Upper
228 100
226 29.9 24.7 37.1
229 19.9 15.4 23.0



#84
Bis(2-ethylhexyl)phthalate
Concen: 35.300 ng
RT: 13.945 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:149 Resp: 180235
Ion Ratio Lower Upper
149 100
167 27.8 20.9 31.3
279 4.1 2.9 4.3









#85
Di-n-octyl phthalate
Concen: 41.171 ng
RT: 14.563 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion:149 Resp: 325170
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 11.8 8.9 13.3



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.445 min Scan# 2271
Delta R.T. 0.000 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:264 Resp: 143829
Ion Ratio Lower Upper
264 100
260 24.8 18.3 27.5
265 22.6 17.0 25.6





#87

Indeno(1,2,3-cd)pyrene

Concen: 34.858 ng

RT: 16.951 min Scan# 2197

Delta R.T. 0.000 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion:276 Resp: 361990

Ion Ratio Lower Upper

276 100

138 22.1 18.7 28.1

277 25.2 20.2 30.2



#88

Benzo(b)fluoranthene

Concen: 31.919 ng

RT: 15.010 min Scan# 2197

Delta R.T. -0.006 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26




Tgt Ion:252 Resp: 306630


Ion Ratio Lower Upper

252 100

253 22.1 17.4 26.0

125 10.5 7.4 11.2









#89
Benzo(k)fluoranthene
Concen: 29.419 ng
RT: 15.039 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26




Instrument :
BNA_F
ClientSampleId :
COLTOMSD


Tgt Ion:252 Resp: 267047
Ion Ratio Lower Upper
252 100
253 21.9 17.0 25.6
125 10.9 8.2 12.4



#90
Benzo(a)pyrene
Concen: 33.000 ng
RT: 15.380 min Scan# 2260
Delta R.T. -0.006 min
Lab File: BF136757.D
Acq: 27 Dec 2023 14:26

Tgt Ion:252 Resp: 267596
Ion Ratio Lower Upper
252 100
253 22.1 17.0 25.6
125 11.7 8.8 13.2





#91

Dibenzo(a,h)anthracene

Concen: 35.658 ng

RT: 16.968 min Scan# 21

Delta R.T. 0.000 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Instrument :

BNA_F

ClientSampleId :

COLTOMSD





Tgt Ion:278 Resp: 303792

Ion Ratio Lower Upper

278 100

139 15.3 12.2 18.4

279 23.1 17.8 26.6



#92

Benzo(g,h,i)perylene

Concen: 31.834 ng

RT: 17.404 min Scan# 2604

Delta R.T. -0.006 min

Lab File: BF136757.D

Acq: 27 Dec 2023 14:26

Tgt Ion:276 Resp: 277775

Ion Ratio Lower Upper

276 100

277 21.9 18.3 27.5

138 19.6 15.6 23.4

