

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122723\
 Data File : BP019111.D
 Acq On : 27 Dec 2023 12:33
 Operator : MA/JU
 Sample : SST00570
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Dec 28 02:26:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 28 02:21:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.811	152	140229	20.000	ng/u1	0.00
20) Naphthalene-d8	10.605	136	528956	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.457	164	320414	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.210	188	749145	20.000	ng/u1	0.00
79) Chrysene-d12	21.316	240	843236	20.000	ng/u1	0.00
88) Perylene-d12	23.686	264	998578	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.240	96	9728	2.379	ng/uL	0.00
4) Pyridine-d5	3.664	84	62434	6.058	ng/u1	0.00
7) Phenol-d5	6.987	99	72152	5.738	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.146	67	45786	6.294	ng/u1	0.00
11) 2-Chlorophenol-d4	7.340	132	55682	5.492	ng/u1	0.00
15) 4-Methylphenol-d8	8.528	113	52750	5.388	ng/u1	0.00
21) Nitrobenzene-d5	8.975	128	24461	5.350	ng/u1	0.00
24) 2-Nitrophenol-d4	9.693	143	24232	4.904	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.234	165	53274	5.339	ng/u1	0.00
31) 4-Chloroaniline-d4	10.752	131	75299	6.135	ng/u1	0.00
46) Dimethylphthalate-d6	13.875	166	160777	5.681	ng/u1	0.00
49) Acenaphthylene-d8	14.145	160	170295	5.546	ng/u1	0.00
54) 4-Nitrophenol-d4	14.675	143	21879	4.578	ng/u1	0.00
60) Fluorene-d10	15.451	176	141127	5.772	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.581	200	20901	4.256	ng/u1	0.00
73) Anthracene-d10	17.310	188	220741	5.642	ng/u1	0.00
81) Pyrene-d10	19.557	212	293437	5.710	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.533	264	316389	5.512	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.275	88	10797	2.392	ng/uL	98
5) Pyridine	3.681	79	66479	6.355	ng/u1	96
6) Benzaldehyde	6.952	77	23544	4.032	ng/u1	99
8) Phenol	7.011	94	74774	6.020	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.240	93	60441	6.014	ng/u1	97
12) 2-Chlorophenol	7.375	128	59110	5.697	ng/u1	96
13) 2-Methylphenol	8.263	108	49738	5.413	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.340	45	73949	6.238	ng/u1	99
16) Acetophenone	8.640	105	88080	6.007	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.622	70	40646	5.559	ng/u1	99
18) 4-Methylphenol	8.593	108	56390	5.701	ng/u1	96
19) Hexachloroethane	8.881	117	24946	5.732	ng/u1	93
22) Nitrobenzene	9.016	77	63332	5.669	ng/u1	100
23) Isophorone	9.540	82	109940	5.458	ng/u1	98
25) 2-Nitrophenol	9.728	139	27095	5.135	ng/u1	98
26) 2,4-Dimethylphenol	9.793	107	44801	4.823	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.028	93	76061	5.713	ng/u1	96
29) 2,4-Dichlorophenol	10.257	162	54008	5.538	ng/u1	99
30) Naphthalene	10.657	128	187227	5.871	ng/u1	100
32) 4-Chloroaniline	10.775	127	73329	6.226	ng/u1	97
33) Hexachlorobutadiene	10.940	225	43072	5.739	ng/u1	98
34) Caprolactam	11.557	113	11079	4.341	ng/u1	94
35) 4-Chloro-3-methylphenol	11.910	107	49989	5.222	ng/u1	99
36) 2-Methylnaphthalene	12.275	142	122864	5.714	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122723\
 Data File : BP019111.D
 Acq On : 27 Dec 2023 12:33
 Operator : MA/JU
 Sample : SST00570
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Dec 28 02:26:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 28 02:21:23 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.493	142	124665	5.749	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.640	216	76058	5.764	ng/ul	99
40) Hexachlorocyclopentadiene	12.616	237	36543	4.989	ng/ul	99
41) 2,4,6-Trichlorophenol	12.893	196	41895	5.320	ng/ul	92
42) 2,4,5-Trichlorophenol	12.963	196	43153	5.162	ng/ul	99
43) 1,1'-Biphenyl	13.287	154	164352	5.880	ng/ul	99
44) 2-Chloronaphthalene	13.328	162	130115	5.802	ng/ul	98
45) 2-Nitroaniline	13.546	65	27229	5.057	ng/ul	88
47) Dimethylphthalate	13.922	163	158250	5.665	ng/ul	99
48) 2,6-Dinitrotoluene	14.040	165	24443	4.761	ng/ul	92
50) Acenaphthylene	14.175	152	196088	5.675	ng/ul	99
51) 3-Nitroaniline	14.375	138	20617	4.108	ng/ul	93
52) Acenaphthene	14.522	153	137020	5.851	ng/ul	99
53) 2,4-Dinitrophenol	14.581	184	12601	3.810	ng/ul	95
55) 4-Nitrophenol	14.687	109	18512	4.847	ng/ul	88
56) Dibenzofuran	14.857	168	201160	6.015	ng/ul	100
57) 2,4-Dinitrotoluene	14.834	165	36171	4.791	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.087	232	37849	5.182	ng/ul	97
59) Diethylphthalate	15.287	149	150834	5.568	ng/ul	98
61) Fluorene	15.510	166	159091	5.944	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.504	204	85770	5.868	ng/ul	96
63) 4-Nitroaniline	15.540	138	18117	3.685	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.592	198	21801	4.193	ng/ul#	85
67) N-Nitrosodiphenylamine	15.722	169	131149	5.736	ng/ul	98
68) 4-Bromophenyl-phenylether	16.404	248	50006	5.370	ng/ul	95
69) Hexachlorobenzene	16.510	284	65487	5.755	ng/ul	97
70) Atrazine	16.681	200	42566	6.597	ng/ul	98
71) Pentachlorophenol	16.863	266	29802	4.461	ng/ul	97
72) Phenanthrene	17.257	178	270636	5.961	ng/ul	99
74) Anthracene	17.345	178	262441	5.807	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.251	216	73590	5.650	ng/ul	99
76) Pentachlorobenzene	14.775	250	74627	5.766	ng/ul	96
77) Carbazole	17.628	167	230823	6.140	ng/ul	99
78) Di-n-butylphthalate	18.181	149	248661	5.333	ng/ul	99
80) Fluoranthene	19.233	202	341399	5.789	ng/ul	100
82) Pyrene	19.586	202	366017	5.924	ng/ul	100
83) Butylbenzylphthalate	20.469	149	109218	4.949	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.245	252	100855	4.781	ng/ul	97
85) Benzo(a)anthracene	21.304	228	382202	5.737	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.233	149	161818	4.979	ng/ul	97
87) Chrysene	21.351	228	364522	5.868	ng/ul	99
89) Di-n-octyl phthalate	22.145	149	272040	4.764	ng/ul	100
90) Benzo(b)fluoranthene	22.963	252	387927	5.531	ng/ul	99
91) Benzo(k)fluoranthene	23.010	252	400624	5.813	ng/ul	99
93) Benzo(a)pyrene	23.580	252	377299	5.715	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.151	276	482676	5.703	ng/ul	99
95) Dibenzo(a,h)anthracene	26.163	278	404839	5.750	ng/ul	99
96) Benzo(g,h,i)perylene	26.898	276	386659	5.701	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122723\
Data File : BP019111.D
Acq On : 27 Dec 2023 12:33
Operator : MA/JU
Sample : SSTD00570
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :


Quant Time: Dec 28 02:26:18 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 28 02:21:23 2023
Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 2.392 ng/uL
RT: 3.275 min Scan# 31
Delta R.T. 0.006 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Instrument :
BNA_F
ClientSampleId :



Tgt Ion: 88 Resp: 10797




Ion	Ratio	Lower	Upper
88	100		
43	29.5	21.0	31.4
58	70.4	56.6	84.8




#5
Pyridine
Concen: 6.355 ng/uL
RT: 3.681 min Scan# 101
Delta R.T. 0.006 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Tgt Ion: 79 Resp: 66479

Ion	Ratio	Lower	Upper
79	100		
52	68.5	52.4	78.6
51	34.2	25.9	38.9









#6
Benzaldehyde
Concen: 4.032 ng/ul
RT: 6.952 min Scan# 611
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33




Instrument :
BNA_F
ClientSampleId :


Tgt Ion: 77 Resp: 23544
Ion Ratio Lower Upper
77 100
105 97.6 78.6 117.8
106 91.3 74.8 112.2



#8
Phenol
Concen: 6.020 ng/ul
RT: 7.011 min Scan# 667
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Tgt Ion: 94 Resp: 74774
Ion Ratio Lower Upper
94 100
65 26.3 21.1 31.7
66 35.7 29.3 43.9









#10
 Bis(2-Chloroethyl)ether
 Concen: 6.014 ng/ul
 RT: 7.240 min Scan# 706
 Delta R.T. -0.000 min
 Lab File: BP019111.D
 Acq: 27 Dec 2023 12:33

Instrument :
 BNA_F
 ClientSampleId :

Tgt Ion: 93 Resp: 60441




Ion	Ratio	Lower	Upper
93	100		
63	77.5	59.4	89.0
95	33.4	26.6	40.0




#12
 2-Chlorophenol
 Concen: 5.697 ng/ul
 RT: 7.375 min Scan# 729
 Delta R.T. -0.000 min
 Lab File: BP019111.D
 Acq: 27 Dec 2023 12:33

Tgt Ion: 128 Resp: 59110

Ion	Ratio	Lower	Upper
128	100		
64	48.8	35.9	53.9
130	31.1	25.3	37.9





#13

2-Methylphenol

Concen: 5.413 ng/ul

RT: 8.263 min Scan# 880

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F





ClientSampleId :

Tgt Ion:108 Resp: 49738

Ion Ratio Lower Upper

108 100

107 94.0 72.6 109.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 6.238 ng/ul

RT: 8.340 min Scan# 893

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion: 45 Resp: 73949


Ion Ratio Lower Upper

45 100

77 16.9 13.6 20.4

79 12.9 9.7 14.5









#16
Acetophenone
Concen: 6.007 ng/ul
RT: 8.640 min Scan# 944
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:	105	Resp:	88080
Ion	Ratio	Lower	Upper
105	100		
77	75.1	59.8	89.6
51	23.9	19.3	28.9



#17
N-Nitroso-di-n-propylamine
Concen: 5.559 ng/ul
RT: 8.622 min Scan# 941
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Tgt Ion:	70	Resp:	40646
Ion	Ratio	Lower	Upper
70	100		
42	50.7	40.6	61.0
101	8.7	7.3	10.9
130	21.1	16.1	24.1









#18
4-Methylphenol
Concen: 5.701 ng/ul
RT: 8.593 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:108 Resp: 56390
Ion Ratio Lower Upper
108 100
107 106.1 88.1 132.1



#19
Hexachloroethane
Concen: 5.732 ng/ul
RT: 8.881 min Scan# 985
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Tgt Ion:117 Resp: 24946
Ion Ratio Lower Upper
117 100
201 123.3 92.5 138.7
199 76.7 56.9 85.3









#22
Nitrobenzene
Concen: 5.669 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33




Instrument :
BNA_F
ClientSampleId :


Tgt Ion: 77 Resp: 63332
Ion Ratio Lower Upper
77 100
123 44.2 35.5 53.3
65 12.9 10.3 15.5



#23
Isophorone
Concen: 5.458 ng/ul
RT: 9.540 min Scan# 1097
Delta R.T. -0.006 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Tgt Ion: 82 Resp: 109940
Ion Ratio Lower Upper
82 100
95 7.2 5.9 8.9
138 17.1 14.7 22.1









#25
2-Nitrophenol
Concen: 5.135 ng/ul
RT: 9.728 min Scan# 1129
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:139 Resp: 27095
Ion Ratio Lower Upper
139 100
65 51.3 38.8 58.2
109 32.7 26.2 39.2



#26
2,4-Dimethylphenol
Concen: 4.823 ng/ul
RT: 9.793 min Scan# 1140
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Tgt Ion:107 Resp: 44801
Ion Ratio Lower Upper
107 100
121 46.6 39.8 59.6
122 82.9 66.4 99.6





#27

Bis(2-Chloroethoxy)methane

Concen: 5.713 ng/ul

RT: 10.028 min Scan# 1180

Delta R.T. -0.000 min

Lab File: BP019111.D





Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :

Tgt Ion: 93	Resp: 76061
Ion Ratio	Lower Upper
93 100	
95 30.5	26.7 40.1
123 9.8	8.5 12.7



#29

2,4-Dichlorophenol

Concen: 5.538 ng/ul




RT: 10.257 min Scan# 1219


Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Tgt Ion: 162	Resp: 54008
Ion Ratio	Lower Upper
162 100	
164 66.6	52.6 78.8
98 33.9	26.7 40.1





#30

Naphthalene

Concen: 5.871 ng/ul

RT: 10.657 min Scan# 1287

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNF

ClientSampleId :





Tgt Ion:128 Resp: 187227

Ion Ratio Lower Upper

128 100

129 10.7 8.6 13.0

127 13.0 10.6 15.8



#32

4-Chloroaniline

Concen: 6.226 ng/ul

RT: 10.775 min Scan# 1307

Delta R.T. -0.006 min

Lab File: BP019111.D




Acq: 27 Dec 2023 12:33


Tgt Ion:127 Resp: 73329

Ion Ratio Lower Upper

127 100

129 30.6 25.6 38.4





#33

Hexachlorobutadiene

Concen: 5.739 ng/ul

RT: 10.940 min Scan# 11

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:225 Resp: 43072

Ion Ratio Lower Upper

225 100

223 60.3 50.5 75.7

227 66.2 53.0 79.4



#34

Caprolactam

Concen: 4.341 ng/ul

RT: 11.557 min Scan# 1440

Delta R.T. -0.012 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion:113 Resp: 11079


Ion Ratio Lower Upper

113 100

55 163.4 122.6 184.0

56 119.7 93.0 139.4









#35
4-Chloro-3-methylphenol
Concen: 5.222 ng/ul
RT: 11.910 min Scan# 1500
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:107 Resp: 49989
Ion Ratio Lower Upper
107 100
144 25.9 21.0 31.6
142 82.5 64.8 97.2



#36
2-Methylnaphthalene
Concen: 5.714 ng/ul
RT: 12.275 min Scan# 1562
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Tgt Ion:142 Resp: 122864
Ion Ratio Lower Upper
142 100
141 89.4 71.7 107.5









#37
1-Methylnaphthalene
Concen: 5.749 ng/ul
RT: 12.493 min Scan# 1599
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:142 Resp: 124665
Ion Ratio Lower Upper
142 100
141 91.4 74.4 111.6
116 3.8 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 5.764 ng/ul
RT: 12.640 min Scan# 1624
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Tgt Ion:216 Resp: 76058
Ion Ratio Lower Upper
216 100
214 77.2 62.6 94.0
179 19.7 15.8 23.8
108 15.2 11.9 17.9





#40

Hexachlorocyclopentadiene

Concen: 4.989 ng/ul

RT: 12.616 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:237 Resp: 36543

Ion Ratio Lower Upper

237 100

235 64.4 51.2 76.8

272 12.8 10.9 16.3



#41

2,4,6-Trichlorophenol

Concen: 5.320 ng/ul

RT: 12.893 min Scan# 1667

Delta R.T. 0.006 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion:196 Resp: 41895


Ion Ratio Lower Upper

196 100

198 86.8 77.4 116.0

200 29.1 24.6 36.8





#42

2,4,5-Trichlorophenol

Concen: 5.162 ng/ul

RT: 12.963 min Scan# 1678

Delta R.T. 0.006 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:196 Resp: 43153

Ion Ratio Lower Upper

196 100

198 97.0 77.3 115.9

200 30.0 25.7 38.5



#43

1,1'-Biphenyl

Concen: 5.880 ng/ul

RT: 13.287 min Scan# 1734

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion:154 Resp: 164352


Ion Ratio Lower Upper

154 100

153 39.6 31.9 47.9

76 12.4 9.5 14.3





#44

2-Chloronaphthalene

Concen: 5.802 ng/ul

RT: 13.328 min Scan# 1741

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:162 Resp: 130115

Ion Ratio Lower Upper

162 100

164 32.6 26.2 39.4

127 37.9 28.8 43.2



#45

2-Nitroaniline

Concen: 5.057 ng/ul

RT: 13.546 min Scan# 1778

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion: 65 Resp: 27229


Ion Ratio Lower Upper

65 100

92 68.1 60.2 90.4


138 100.4 92.9 139.3








#47
Dimethylphthalate
Concen: 5.665 ng/ul
RT: 13.922 min Scan# 1842
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:163 Resp: 158250
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.2
164 10.7 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 4.761 ng/ul
RT: 14.040 min Scan# 1862
Delta R.T. -0.006 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Tgt Ion:165 Resp: 24443
Ion Ratio Lower Upper
165 100
89 63.0 45.1 67.7
121 22.9 16.7 25.1









#50
Acenaphthylene
Concen: 5.675 ng/ul
RT: 14.175 min Scan# 1885
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:152 Resp: 196088
Ion Ratio Lower Upper
152 100
151 20.3 15.9 23.9
153 13.4 10.6 15.8



#51
3-Nitroaniline
Concen: 4.108 ng/ul
RT: 14.375 min Scan# 1919
Delta R.T. 0.006 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Tgt Ion:138 Resp: 20617
Ion Ratio Lower Upper
138 100
108 13.9 9.8 14.6
92 123.1 91.9 137.9





#52

Acenaphthene

Concen: 5.851 ng/ul

RT: 14.522 min Scan# 1944

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:153 Resp: 137020

Ion Ratio Lower Upper

153 100

152 48.2 38.1 57.1

154 88.2 70.1 105.1



#53

2,4-Dinitrophenol

Concen: 3.810 ng/ul

RT: 14.581 min Scan# 1954

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion:184 Resp: 12601


Ion Ratio Lower Upper

184 100

63 56.5 48.5 72.7

154 64.8 49.3 73.9









#55
4-Nitrophenol
Concen: 4.847 ng/ul
RT: 14.687 min Scan# 1972
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:109 Resp: 18512
Ion Ratio Lower Upper
109 100
139 137.4 102.6 154.0
65 135.2 92.9 139.3



#56
Dibenzofuran
Concen: 6.015 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Tgt Ion:168 Resp: 201160
Ion Ratio Lower Upper
168 100
139 34.1 27.4 41.0









#57
2,4-Dinitrotoluene
Concen: 4.791 ng/ul
RT: 14.834 min Scan# 1997
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:165 Resp: 36171
Ion Ratio Lower Upper
165 100
63 41.2 31.5 47.3
182 3.3 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 5.182 ng/ul
RT: 15.087 min Scan# 2040
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Tgt Ion:232 Resp: 37849
Ion Ratio Lower Upper
232 100
131 38.4 29.3 43.9
130 2.1 1.4 2.2
166 29.4 22.6 33.8





#59

Diethylphthalate

Concen: 5.568 ng/ul

RT: 15.287 min Scan# 2074

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:149 Resp: 150834

Ion Ratio Lower Upper

149 100

177 21.9 18.7 28.1

150 12.5 9.8 14.8



#61

Fluorene

Concen: 5.944 ng/ul

RT: 15.510 min Scan# 2112

Delta R.T. 0.006 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion:166 Resp: 159091


Ion Ratio Lower Upper

166 100

165 96.2 79.3 118.9

167 12.8 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 5.868 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:204 Resp: 85770
Ion Ratio Lower Upper
204 100
206 32.1 26.3 39.5
141 56.4 42.2 63.2



#63
4-Nitroaniline
Concen: 3.685 ng/ul
RT: 15.540 min Scan# 2117
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Tgt Ion:138 Resp: 18117
Ion Ratio Lower Upper
138 100
92 56.7 42.2 63.4
108 87.8 66.2 99.4









#66
4,6-Dinitro-2-methylphenol
Concen: 4.193 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:198 Resp: 21801
Ion Ratio Lower Upper
198 100
182 8.5 4.0 6.0#
77 29.2 17.4 26.2#



#67
N-Nitrosodiphenylamine
Concen: 5.736 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. -0.000 min
Lab File: BP019111.D
Acq: 27 Dec 2023 12:33

Tgt Ion:169 Resp: 131149
Ion Ratio Lower Upper
169 100
168 64.5 53.8 80.8
167 36.7 29.5 44.3





#68

4-Bromophenyl-phenylether

Concen: 5.370 ng/ul

RT: 16.404 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:248 Resp: 50006

Ion Ratio Lower Upper

248 100

250 102.6 78.3 117.5

141 64.5 48.6 72.8



#69

Hexachlorobenzene

Concen: 5.755 ng/ul

RT: 16.510 min Scan# 2282

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion:284 Resp: 65487


Ion Ratio Lower Upper

284 100

142 30.2 22.9 34.3

249 30.8 23.7 35.5





#70

Atrazine

Concen: 6.597 ng/ul

RT: 16.681 min Scan# 2311

Delta R.T. -0.000 min

Lab File: BP019111.D





Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :

Tgt Ion:	200	Resp:	42566
Ion Ratio	Lower	Upper	
200	100		
173	24.1	20.2	30.2
215	49.1	39.9	59.9



#71

Pentachlorophenol

Concen: 4.461 ng/ul




RT: 16.863 min Scan# 2342


Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Tgt Ion:	266	Resp:	29802
Ion Ratio	Lower	Upper	
266	100		
264	58.8	50.0	75.0
268	66.3	52.8	79.2





#72

Phenanthrene

Concen: 5.961 ng/ul

RT: 17.257 min Scan# 2409

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:178 Resp: 270636

Ion Ratio Lower Upper

178 100

179 15.5 11.9 17.9

176 18.2 15.0 22.6



#74

Anthracene

Concen: 5.807 ng/ul

RT: 17.345 min Scan# 2424

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion:178 Resp: 262441


Ion Ratio Lower Upper

178 100

179 14.9 12.2 18.4

176 18.4 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 5.650 ng/uL

RT: 13.251 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:216 Resp: 73590

Ion Ratio Lower Upper

216 100

214 77.7 61.6 92.4

218 48.3 38.9 58.3



#76

Pentachlorobenzene

Concen: 5.766 ng/uL

RT: 14.775 min Scan# 1987

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion:250 Resp: 74627


Ion Ratio Lower Upper

250 100

248 61.0 50.4 75.6


252 60.9 52.1 78.1








#77
 Carbazole
 Concen: 6.140 ng/ul
 RT: 17.628 min Scan# 2471
 Delta R.T. 0.006 min
 Lab File: BP019111.D
 Acq: 27 Dec 2023 12:33

Instrument :
 BNA_F
 ClientSampleId :







Tgt Ion:167 Resp: 230823
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.4 26.0
 139 13.0 9.8 14.6



#78
 Di-n-butylphthalate
 Concen: 5.333 ng/ul
 RT: 18.181 min Scan# 2566
 Delta R.T. -0.000 min
 Lab File: BP019111.D
 Acq: 27 Dec 2023 12:33

Tgt Ion:149 Resp: 248661
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.3 10.9
 104 5.3 3.9 5.9





#80

Fluoranthene

Concen: 5.789 ng/ul

RT: 19.233 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:202 Resp: 341399

Ion Ratio Lower Upper

202 100

101 9.9 7.8 11.6

100 7.6 6.0 9.0



#82

Pyrene

Concen: 5.924 ng/ul

RT: 19.586 min Scan# 2805

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion:202 Resp: 366017


Ion Ratio Lower Upper

202 100

101 11.5 9.2 13.8

100 9.9 7.8 11.6





#83

Butylbenzylphthalate

Concen: 4.949 ng/ul

RT: 20.469 min Scan# 2955

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:149 Resp: 109218

Ion Ratio Lower Upper

149 100

91 71.1 55.6 83.4

206 21.8 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 4.781 ng/ul

RT: 21.245 min Scan# 3087

Delta R.T. 0.006 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion:252 Resp: 100855


Ion Ratio Lower Upper

252 100

254 62.3 51.5 77.3

126 8.9 7.6 11.4





#85

Benzo(a)anthracene

Concen: 5.737 ng/ul

RT: 21.304 min Scan# 3097

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:228 Resp: 382202

Ion Ratio Lower Upper

228 100

229 19.4 15.5 23.3

226 26.2 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 4.979 ng/ul

RT: 21.233 min Scan# 3085

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion:149 Resp: 161818


Ion Ratio Lower Upper

149 100

167 26.3 22.5 33.7

279 4.3 4.0 6.0





#87

Chrysene

Concen: 5.868 ng/ul

RT: 21.351 min Scan# 3106

Delta R.T. -0.006 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:228 Resp: 364522

Ion Ratio Lower Upper

228 100

226 30.1 23.4 35.2

229 19.3 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 4.764 ng/ul




RT: 22.145 min Scan# 3240


Delta R.T. -0.006 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Tgt Ion:149 Resp: 272040





#90

Benzo(b)fluoranthene

Concen: 5.531 ng/ul

RT: 22.963 min Scan# 3379

Delta R.T. -0.006 min


Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :






Tgt Ion:252 Resp: 387927

Ion Ratio Lower Upper

252 100

253 22.3 17.8 26.8

125 9.4 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 5.813 ng/ul

RT: 23.010 min Scan# 3387

Delta R.T. -0.006 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion:252 Resp: 400624


Ion Ratio Lower Upper

252 100

253 21.5 17.4 26.0

125 8.9 6.7 10.1





#93

Benzo(a)pyrene

Concen: 5.715 ng/ul

RT: 23.580 min Scan# 3485

Delta R.T. -0.006 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:252 Resp: 377299

Ion Ratio Lower Upper

252 100

253 23.2 17.5 26.3

125 10.2 7.4 11.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 5.703 ng/ul

RT: 26.151 min Scan# 3921

Delta R.T. -0.000 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33




Tgt Ion:276 Resp: 482676


Ion Ratio Lower Upper

276 100

138 20.9 16.6 25.0

277 25.5 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 5.750 ng/ul

RT: 26.163 min Scan# 3924

Delta R.T. -0.006 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:278 Resp: 404839

Ion Ratio Lower Upper

278 100

139 14.4 11.0 16.4

279 23.6 19.2 28.8



#96

Benzo(g,h,i)perylene

Concen: 5.701 ng/ul

RT: 26.898 min Scan# 4048

Delta R.T. -0.006 min

Lab File: BP019111.D

Acq: 27 Dec 2023 12:33

Tgt Ion:276 Resp: 386659

Ion Ratio Lower Upper

276 100

138 19.2 15.2 22.8

277 24.0 19.4 29.0

