

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122723\
 Data File : BP019119.D
 Acq On : 27 Dec 2023 17:49
 Operator : MA/JU
 Sample : 06019-10
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Dec 28 03:44:27 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 28 03:15:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.805	152	156873	20.000	ng/u1	0.00
20) Naphthalene-d8	10.605	136	575750	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	363502	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.210	188	863936	20.000	ng/u1	0.00
79) Chrysene-d12	21.316	240	1002120	20.000	ng/u1	0.00
88) Perylene-d12	23.680	264	1207939	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.223	96	60	0.013	ng/uL	-0.01
4) Pyridine-d5	3.646	84	839	0.073	ng/u1	-0.01
7) Phenol-d5	6.999	99	23	0.002	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.146	67	44	0.005	ng/u1	0.00
11) 2-Chlorophenol-d4	7.334	132	17	0.001	ng/u1	0.00
15) 4-Methylphenol-d8	8.516	113	34	0.003	ng/u1	-0.01
21) Nitrobenzene-d5	8.981	128	33	0.007	ng/u1	0.00
24) 2-Nitrophenol-d4	9.687	143	12	0.002	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	46	0.004	ng/u1	0.00
31) 4-Chloroaniline-d4	10.734	131	30	0.002	ng/u1	-0.02
46) Dimethylphthalate-d6	13.851	166	17	0.001	ng/u1	-0.02
49) Acenaphthylene-d8	14.146	160	153	0.004	ng/u1	0.00
54) 4-Nitrophenol-d4	14.651	143	20	0.004	ng/u1	-0.02
60) Fluorene-d10	15.422	176	8	0.000	ng/u1	-0.03
65) 4,6-Dinitro-2-methylph...	15.593	200	20	0.004	ng/u1	0.01
73) Anthracene-d10	17.316	188	2501	0.055	ng/u1	0.00
81) Pyrene-d10	19.563	212	2970	0.049	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.680	264	1121930	16.159	ng/u1	0.15

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122723\
Data File : BP019119.D
Acq On : 27 Dec 2023 17:49
Operator : MA/JU
Sample : 06019-10
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Dec 28 03:44:27 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 28 03:15:29 2023
Response via : Initial Calibration

