

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122723\
 Data File : BF136751.D
 Acq On : 27 Dec 2023 11:14
 Operator : CG\JU
 Sample : PB157937BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB157937BL

Quant Time: Dec 27 11:52:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 21 01:54:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.781	152	83156	20.000	ng	-0.01
21) Naphthalene-d8	8.057	136	319270	20.000	ng	-0.01
39) Acenaphthene-d10	9.816	164	157296	20.000	ng	-0.01
64) Phenanthrene-d10	11.310	188	289311	20.000	ng	0.00
76) Chrysene-d12	13.962	240	163908	20.000	ng	0.00
86) Perylene-d12	15.445	264	160858	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.439	112	661636	124.145	ng	0.01
7) Phenol-d6	6.434	99	827801	119.871	ng	0.00
23) Nitrobenzene-d5	7.345	82	504999	80.939	ng	-0.01
42) 2,4,6-Tribromophenol	10.616	330	210176	113.455	ng	0.00
45) 2-Fluorobiphenyl	9.139	172	974164	83.297	ng	-0.01
79) Terphenyl-d14	12.910	244	907175	77.345	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122723\
Data File : BF136751.D
Acq On : 27 Dec 2023 11:14
Operator : CG\JU
Sample : PB157937BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB157937BL

Quant Time: Dec 27 11:52:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 21 01:54:25 2023
Response via : Initial Calibration

