

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122822\
 Data File : BF131854.D
 Acq On : 28 Dec 2022 10:59
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 29 07:03:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 26 05:13:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	103	0.00
2	1,4-Dioxane	40.000	41.539	-3.8	104	0.00
3	Pyridine	40.000	43.287	-8.2	105	0.00
4	n-Nitrosodimethylamine	40.000	43.269	-8.2	108	0.00
5 S	2-Fluorophenol	80.000	83.705	-4.6	105	-0.01
6	Aniline	40.000	41.643	-4.1	105	0.00
7 S	Phenol-d6	80.000	83.231	-4.0	106	0.00
8	2-Chlorophenol	40.000	41.260	-3.1	104	-0.01
9	Benzaldehyde	40.000	36.068	9.8	100	-0.01
10 C	Phenol	40.000	41.958	-4.9	106	0.00
11	bis(2-Chloroethyl)ether	40.000	39.780	0.5	101	0.00
12	1,3-Dichlorobenzene	40.000	40.784	-2.0	104	0.00
13 C	1,4-Dichlorobenzene	40.000	40.860	-2.1	103	0.00
14	1,2-Dichlorobenzene	40.000	40.997	-2.5	105	-0.01
15	Benzyl Alcohol	40.000	41.564	-3.9	104	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.587	1.0	100	-0.01
17	2-Methylphenol	40.000	41.827	-4.6	106	0.00
18	Hexachloroethane	40.000	39.745	0.6	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.304	1.7	102	0.00
20	3+4-Methylphenols	40.000	40.638	-1.6	103	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	40.878	-2.2	104	0.00
23 S	Nitrobenzene-d5	80.000	82.690	-3.4	104	0.00
24	Nitrobenzene	40.000	41.891	-4.7	105	0.00
25	Isophorone	40.000	39.996	0.0	101	-0.01
26 C	2-Nitrophenol	40.000	41.837	-4.6	104	-0.01
27	2,4-Dimethylphenol	40.000	41.881	-4.7	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.921	0.2	101	-0.01
29 C	2,4-Dichlorophenol	40.000	41.129	-2.8	104	0.00
30	1,2,4-Trichlorobenzene	40.000	40.255	-0.6	101	-0.01
31	Naphthalene	40.000	40.812	-2.0	103	-0.01
32	Benzoic acid	40.000	43.886	-9.7	103	0.00
33	4-Chloroaniline	40.000	40.019	-0.0	100	-0.01
34 C	Hexachlorobutadiene	40.000	39.107	2.2	99	-0.01
35	Caprolactam	40.000	40.325	-0.8	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.966	-2.4	102	0.00
37	2-Methylnaphthalene	40.000	40.326	-0.8	103	-0.01
38	1-Methylnaphthalene	40.000	39.746	0.6	101	-0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.178	-5.4	103	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.986	7.5	91	0.00
42 S	2,4,6-Tribromophenol	80.000	80.647	-0.8	95	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.563	-6.4	101	0.00
44	2,4,5-Trichlorophenol	40.000	42.320	-5.8	100	0.00
45 S	2-Fluorobiphenyl	80.000	82.064	-2.6	100	0.00
46	1,1'-Biphenyl	40.000	41.628	-4.1	101	0.00
47	2-Chloronaphthalene	40.000	41.634	-4.1	100	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122822\
 Data File : BF131854.D
 Acq On : 28 Dec 2022 10:59
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 29 07:03:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 26 05:13:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.294	-5.7	100	0.00
49	Acenaphthylene	40.000	41.312	-3.3	98	0.00
50	Dimethylphthalate	40.000	39.732	0.7	95	0.00
51	2,6-Dinitrotoluene	40.000	40.814	-2.0	96	0.00
52 C	Acenaphthene	40.000	41.651	-4.1	100	0.00
53	3-Nitroaniline	40.000	41.908	-4.8	98	0.00
54 P	2,4-Dinitrophenol	40.000	40.388	-1.0	96	0.00
55	Dibenzofuran	40.000	41.152	-2.9	99	0.00
56 P	4-Nitrophenol	40.000	41.554	-3.9	95	0.00
57	2,4-Dinitrotoluene	40.000	40.971	-2.4	96	0.00
58	Fluorene	40.000	40.491	-1.2	97	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.942	-4.9	99	0.00
60	Diethylphthalate	40.000	38.835	2.9	93	0.00
61	4-Chlorophenyl-phenylether	40.000	40.269	-0.7	97	0.00
62	4-Nitroaniline	40.000	41.644	-4.1	98	0.00
63	Azobenzene	40.000	40.024	-0.1	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	95	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.070	-5.2	94	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.104	-2.8	97	0.00
67	4-Bromophenyl-phenylether	40.000	40.352	-0.9	96	0.00
68	Hexachlorobenzene	40.000	40.754	-1.9	96	0.00
69	Atrazine	40.000	40.562	-1.4	93	0.00
70 C	Pentachlorophenol	40.000	42.548	-6.4	92	0.00
71	Phenanthrene	40.000	40.129	-0.3	95	0.00
72	Anthracene	40.000	40.786	-2.0	96	0.00
73	Carbazole	40.000	40.059	-0.1	94	0.00
74	Di-n-butylphthalate	40.000	38.114	4.7	89	0.00
75 C	Fluoranthene	40.000	39.329	1.7	92	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	89	0.00
77	Benzidine	40.000	41.358	-3.4	93	0.00
78	Pyrene	40.000	44.153	-10.4	93	0.00
79 S	Terphenyl-d14	80.000	85.829	-7.3	90	0.00
80	Butylbenzylphthalate	40.000	42.902	-7.3	89	0.00
81	Benzo(a)anthracene	40.000	43.536	-8.8	92	0.00
82	3,3'-Dichlorobenzidine	40.000	41.871	-4.7	89	0.00
83	Chrysene	40.000	42.881	-7.2	90	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.250	-5.6	89	0.00
85 c	Di-n-octyl phthalate	40.000	42.903	-7.3	91	0.00
86 I	Perylene-d12	20.000	20.000	0.0	92	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.262	-8.2	97	0.00
88	Benzo(b)fluoranthene	40.000	40.160	-0.4	89	0.00
89	Benzo(k)fluoranthene	40.000	41.389	-3.5	95	0.00
90 C	Benzo(a)pyrene	40.000	40.830	-2.1	94	0.00
91	Dibenzo(a,h)anthracene	40.000	43.435	-8.6	97	0.00
92	Benzo(g,h,i)perylene	40.000	38.298	4.3	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122822\
Data File : BF131854.D
Acq On : 28 Dec 2022 10:59
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 29 07:03:36 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 26 05:13:34 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0