




Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF122823\  
Data File : BF136795.D  
Acq On : 28 Dec 2023 15:36  
Operator : CG\JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 28 15:57:04 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF122023.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Dec 21 01:54:25 2023  
Response via : Initial Calibration



TIC: BF136795.D\data.ms

**(70) Pentachlorophenol (C)**




11.116min (-0.006) 34129.26 ng

response 125415

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 65.70  | 66.49  |
| 264.00 | 61.20  | 61.73  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF122823\  
Data File : BF136795.D  
Acq On : 28 Dec 2023 15:36  
Operator : CG\JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 28 16:45:23 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF122023.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Dec 21 01:54:25 2023  
Response via : Initial Calibration



TIC: BF136795.D\data.ms

**(77) Benzidine**

12.663min (+ 0.000) 0.00 ng

response 630906

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 14.00  | 13.56  |
| 183.00 | 11.60  | 11.18  |
| 0.00   | 0.00   | 0.00   |