


Data Path : Z:\HPCHEM1\BNA F\DATA\BF122916\
 Data File : BF092015.D
 Acq On : 29 Dec 2016 10:37
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 28 17:32:10 2016



AutoFind: Scans 821, 822, 823; Background Corrected with Scan 818

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.3	50075	PASS
68	69	0.00	2	1.6	792	PASS
69	198	0.00	100	42.9	50715	PASS
70	69	0.00	2	0.3	159	PASS
127	198	10	80	53.0	62655	PASS
197	198	0.00	2	0.5	549	PASS
198	198	100	100	100.0	118304	PASS
199	198	5	9	6.8	8057	PASS
275	198	10	60	23.7	28065	PASS
365	198	1	100	3.0	3592	PASS
441	442	0.01	24	15.9	17453	PASS
442	198	50	100	92.6	109516	PASS
443	442	15	24	18.4	20170	PASS