

Data Path : Z:\HPCHEM1\BNA F\DATA\BF123016\
 Data File : BF092057.D
 Acq On : 30 Dec 2016 20:40
 Operator : UM/SJ
 Sample : H6318-10
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-08

Manual Integrations
 APPROVED

mohammad
 1/3/2017 9:14:34 AM

Quant Time: Dec 30 22:34:35 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 28 17:32:10 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.68	152	177386	20.00	ng	-0.02
21) Naphthalene-d8	7.97	136	741305	20.00	ng	-0.02
38) Acenaphthene-d10	9.72	164	395126	20.00	ng	-0.03
63) Phenanthrene-d10	11.20	188	477159	20.00	ng	-0.03
75) Chrysene-d12	13.83	240	383082	20.00	ng	-0.05
86) Perylene-d12	15.21	264	417847	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.29	112	847467	79.77	ng	-0.03
7) Phenol-d6	6.34	99	590086	45.49	ng	-0.05
23) Nitrobenzene-d5	7.25	82	979039	73.44	ng	-0.03
41) 2,4,6-Tribromophenol	10.51	330	381262	116.60	ng	-0.03
44) 2-Fluorobiphenyl	9.05	172	1570196	80.78	ng	-0.03
78) Terphenyl-d14	12.79	244	1173822	74.31	ng	-0.03
Target Compounds						
10) Phenol	6.35	94	46331	3.13	ng	Qvalue # 77
31) Naphthalene	8.00	128	674408m	19.88	ng	
37) 2-Methylnaphthalene	8.68	142	589089	24.51	ng	99
49) Dimethylphthalate	9.45	163	143143	5.66	ng	# 95
83) Bis(2-ethylhexyl)phthalate	13.83	149	48390	3.21	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF123016\
Data File : BF092057.D
Acq On : 30 Dec 2016 20:40
Operator : UM/SJ
Sample : H6318-10
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-08

Manual Integrations
APPROVED

mohammad
1/3/2017 9:14:34 AM

Quant Time: Dec 30 22:34:35 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 28 17:32:10 2016
Response via : Initial Calibration

