

Data Path : Z:\HPCHEM1\BNA F\DATA\BF123115\
 Data File : BF084215.D
 Acq On : 1 Jan 2016 00:40
 Operator : UM/SJ
 Sample : PB87590BL
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB87590BL

Quant Time: Jan 02 00:52:12 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF123115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 31 14:15:57 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	305328	20.00	ng	0.00
21) Naphthalene-d8	8.18	136	1197973	20.00	ng	-0.01
38) Acenaphthene-d10	9.94	164	579837	20.00	ng	-0.01
63) Phenanthrene-d10	11.42	188	1085577	20.00	ng	0.00
75) Chrysene-d12	14.05	240	785323	20.00	ng	-0.01
86) Perylene-d12	15.49	264	585193	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	1916775	101.54	ng	0.01
7) Phenol-d6	6.52	99	2501590	105.52	ng	-0.01
23) Nitrobenzene-d5	7.46	82	1505323	69.93	ng	-0.01
41) 2,4,6-Tribromophenol	10.73	330	590881	100.94	ng	-0.01
44) 2-Fluorobiphenyl	9.26	172	2633215	79.48	ng	0.00
78) Terphenyl-d14	13.01	244	2410114	68.50	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF123115\
Data File : BF084215.D
Acq On : 1 Jan 2016 00:40
Operator : UM/SJ
Sample : PB87590BL
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB87590BL

Quant Time: Jan 02 00:52:12 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF123115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 31 14:15:57 2015
Response via : Initial Calibration

