

Data File : H:\VOL1\GCMSDATA\J\120718\BJ001014.D
Acq On : 18 Jul 2012 13:13
Sample : P4-ICV
Misc : 1.0ug/mL BNA J HLM
MS Integration Params: RTEINT.P
Quant Time: Jul 18 14:02 2012

Vial: 9
Operator: HLM
Inst : BNA J
Multiplr: 1.00

Quant Results File: BJ071812.RES

Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ071812.M (RTE Integrator)
Title : P4 by SOP M.5
Last Update : Wed Jul 18 14:02:57 2012
Response via : Initial Calibration
DataAcq Meth : P4

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) 1,4-Dichlorobenzene-d4	4.65	150	480437	2.50	ug/ml	0.01
System Monitoring Compounds						
2) Tripropylphosphate	0.00	99	0	0.00	ug/ml	
Spiked Amount	2.500	Range	20 - 150	Recovery	=	0.00%#
Target Compounds						
3) White Phosphorus	5.05	124	394810	1.00	ug/ml	Ovalue 99

Data File : H:\VOL1\GCMSDATA\J\120718\BJ001014.D

Vial: 9

Acq On : 18 Jul 2012 13:13

Operator: HLM

Sample : P4-ICV

Inst : BNA J

Misc : 1.0uq/mL BNA J HLM

Multiplr: 1.00

MS Integration Params: RTEINT.P

Quant Time: Jul 18 14:02 2012


Quant Results File: BJ071812.RES


Method : H:\VOL1\GCMSDATA\J\METHODS\BJ071812.M (RTE Integrator)

Title : P4 by SOP M.5

Last Update : Wed Jul 18 14:03:32 2012

Response via : Initial Calibration





#3

White Phosphorus

Concen: 1.00 ug/ml

RT: 5.05 min Scan# 45

Delta R.T. 0.01 min

Lab File: BJ001014.D

Acq: 18 Jul 2012 13:13



Tgt Ion:124 Resp: 394810

Ion Ratio Lower Upper

124 100

93 7.9 0.0 27.8

62 19.5 0.0 39.0



Abundance Ion 124.00 (123.70 to 124.70): E

Ion 93.00 (92.70 to 93.70): BJ0

Ion 62.00 (61.70 to 62.70): BJ0

