

Data File : H:\VOL1\GCMSDATA\J\120718\BJ001016.D

Vial: 11

Acq On : 18 Jul 2012 14:36

Operator: HLM

Sample : PB64274BS

Inst : BNA J

Misc : WATER BNA J HLM

Multiplr: 1.00

MS Integration Params: RTEINT.P

Quant Time: Jul 18 16:06 2012

Quant Results File: BJ071812.RES

Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ071812.M (RTE Integrator)

Title : P4 by SOP M.5

Last Update : Wed Jul 18 14:03:32 2012

Response via : Initial Calibration

DataAcq Meth : P4

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) 1,4-Dichlorobenzene-d4	4.65	150	386175	2.50	ug/ml	0.01
System Monitoring Compounds						
2) Tripropylphosphate	6.72	99	757595	2.45	ug/ml	0.00
Spiked Amount	2.500	Range	20 - 150	Recovery	=	98.00%
Target Compounds						
3) White Phosphorus	5.05	124	168900	0.53	ug/ml	Ovalue 98

Quantitation Report

Data File : H:\VOL1\GCMSDATA\J\120718\BJ001016.D

Vial: 11

Acq On : 18 Jul 2012 14:36

Operator: HLM

Sample : PB64274BS

Inst : BNA J

Misc : WATER BNA J HLM

Multiplr: 1.00

MS Integration Params: RTEINT.P

Quant Time: Jul 18 16:06 2012


Quant Results File: BJ071812.RES


Method : H:\VOL1\GCMSDATA\J\METHODS\BJ071812.M (RTE Integrator)

Title : P4 by SOP M.5

Last Update : Wed Jul 18 14:03:32 2012

Response via : Initial Calibration





#3

White Phosphorus

Concen: 0.53 ug/ml

RT: 5.05 min Scan# 45

Delta R.T. 0.01 min

Lab File: BJ001016.D

Acq: 18 Jul 2012 14:36

Tgt Ion:124 Resp: 168900

Ion Ratio Lower Upper

124 100

93 7.6 0.0 27.8

62 20.0 0.0 39.0

