

Data File : H:\VOL1\GCMSDATA\J\120718\BJ001017.D
Acq On : 18 Jul 2012 14:50
Sample : D3369-01
Misc : WATER BNA J HLM
MS Integration Params: RTEINT.P
Quant Time: Jul 18 16:06 2012

Vial: 12
Operator: HLM
Inst : BNA J
Multiplr: 1.00

Quant Results File: BJ071812.RES

Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ071812.M (RTE Integrator)
Title : P4 by SOP M.5
Last Update : Wed Jul 18 14:03:32 2012
Response via : Initial Calibration
DataAcq Meth : P4

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) 1,4-Dichlorobenzene-d4	4.65	150	392467	2.50	ug/ml	0.00
System Monitoring Compounds						
2) Tripropylphosphate	6.72	99	642624	2.04	ug/ml	0.00
Spiked Amount	2.500	Range	20 - 150	Recovery	=	81.60%
Target Compounds						
3) White Phosphorus	5.34	124	11041	0.03	ug/ml#	Ovalue 51


Quantitation Report


Data File : H:\VOL1\GCMSDATA\J\120718\BJ001017.D
 Acq On : 18 Jul 2012 14:50
 Sample : D3369-01
 Misc : WATER BNA J HLM
 MS Integration Params: RTEINT.P
 Quant Time: Jul 18 16:06 2012

Vial: 12
 Operator: HLM
 Inst : BNA J
 Multiplr: 1.00

Quant Results File: BJ071812.RES

Method : H:\VOL1\GCMSDATA\J\METHODS\BJ071812.M (RTE Integrator)
 Title : P4 by SOP M.5
 Last Update : Wed Jul 18 14:03:32 2012
 Response via : Initial Calibration





#3

White Phosphorus

Concen: 0.03 ug/ml

RT: 5.34 min Scan# 71

Delta R.T. 0.30 min

Lab File: BJ001017.D

Acq: 18 Jul 2012 14:50


Tgt Ion:124 Resp: 11041

Ion Ratio Lower Upper

124 100

93 0.0 0.0 27.8

62 46.4 0.0 39.0#



Abundance

Ion 124.00 (123.70 to 124.70): E

Ion 93.00 (92.70 to 93.70): BJ0

Ion 62.00 (61.70 to 62.70): BJ0

