

Data Path : Z:\HPCHEM1\BNA_J\DATA\BJ071812\
Data File : BJ001018.D
Acq On : 18 Jul 2012 15:03
Operator : HLM
Sample : D3369-01MS
Misc : WATER BNA_J HLM
ALS Vial : 13 Sample Multiplier: 1

Quant Time: Jul 23 01:41:01 2012
Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ071812.M
Quant Title : P4 by SOP M.5
QLast Update : Wed Jul 18 14:03:32 2012
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	4.65	150	375591	2.50	ug/ml	0.00

System Monitoring Compounds

2) Tripropylphosphate	6.72	99	674711	2.24	ug/ml	0.00
Spiked Amount	2.500	Range	20 - 150	Recovery	=	89.60%

Target Compounds					Qvalue
3) White Phosphorus	5.04	124	148847	0.48	ug/ml 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_J\DATA\BJ071812\
Data File : BJ001018.D
Acq On : 18 Jul 2012 15:03
Operator : HLM
Sample : D3369-01MS
Misc : WATER BNA_J HLM
ALS Vial : 13 Sample Multiplier: 1

Quant Time: Jul 23 01:41:01 2012
Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ071812.M
Quant Title : P4 by SOP M.5
QLast Update : Wed Jul 18 14:03:32 2012
Response via : Initial Calibration

