

Method Path : Z:\HPCHEM1\BNA_J\method\
Method File : BJ071812.M
Title : P4 by SOP M.5
Last Update : Wed Jul 18 14:02:57 2012
Response Via : Initial Calibration

Calibration Files

1 =BJ001010.D 0.1 =BJ001013.D 10 =BJ001007.D 5 =BJ001008.D 2.5 =BJ001009.D 0.5 =BJ001011.D 0.25=BJ001012.D

	Compound	1	0.1	10	5	2.5	0.5	0.25	Avg	%RSD
1) I	1,4-Dichlorobenzen...									
2) S	Tripropylphosp...	2.042	2.238	2.025	1.993	1.745	1.910	2.071	2.003	7.55
3) T	White Phosphorus	2.077	2.096	2.004	2.071	2.036	2.003	2.156	2.063	2.63

(#) = Out of Range