

Data File : H:\VOL1\GCMSDATA\J\120720\BJ001042.D
Acq On : 20 Jul 2012 14:52
Sample : PB64247BL
Misc : WATER BNA J HLM
MS Integration Params: rteint.p
Quant Time: Jul 20 16:26 2012

Vial: 9
Operator: HLM
Inst : BNA J
Multiplr: 1.00

Quant Results File: BJ072012.RES

Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ072012.M (RTE Integrator)
Title : CWA degradation products by GCMS
Last Update : Fri Jul 20 14:08:00 2012
Response via : Initial Calibration
DataAcq Meth : CWAJ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	4.67	152	249273	2.00	ug/mL	0.00
6) Naphthalene-d8	5.96	136	1005908	2.00	ug/mL	0.00
System Monitoring Compounds						
10) Tripropylphosphate	7.03	99	144213	0.39	ug/mL	0.02
Spiked Amount 0.500	Range	70 - 130	Recovery	=	78.00%	
12) 2-Bromothioanisole	7.26	204	358679	4.56	ug/mL	0.00
Spiked Amount 5.000	Range	70 - 130	Recovery	=	91.20%	
Target Compounds						
14) p-Chlorophenylmethanesulfon	8.18	111	4491	0.03	ug/mL#	Ovalue 39

Quantitation Report

Data File : H:\VOL1\GCMSDATA\J\120720\BJ001042.D

Vial: 9

Acq On : 20 Jul 2012 14:52

Operator: HLM

Sample : PB64247BL

Inst : BNA J

Misc : WATER BNA J HLM

Multiplr: 1.00

MS Integration Params: rteint.p

Quant Time: Jul 20 16:26 2012


Quant Results File: BJ072012.RES


Method : H:\VOL1\GCMSDATA\J\METHODS\BJ072012.M (RTE Integrator)

Title : CWA degradation products by GCMS

Last Update : Fri Jul 20 14:08:00 2012

Response via : Initial Calibration





#14
 p-Chlorophenylmethylsulfone
 Concen: 0.03 ug/mL
 RT: 8.18 min Scan# 411
 Delta R.T. 0.05 min
 Lab File: BJ001042.D
 Acq: 20 Jul 2012 14:52

Tgt Ion:	111	Resp:	4491
Ion Ratio	Lower	Upper	
111	100		
75	0.0	22.6	62.6#
127	0.0	17.2	57.2#
190	0.0	2.0	42.0#

