

Data File : H:\VOL1\GCMSDATA\J\120801\BJ001059.D

Vial: 2

Acq On : 1 Aug 2012 14:01

Operator: HLM

Sample : P4-CCV

Inst : BNA J

Misc : 1.0PPM BNA J HLM

Multiplr: 1.00

MS Integration Params: RTEINT.P

Quant Time: Aug 1 16:26 2012

Quant Results File: BJ071812.RES

Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ071812.M (RTE Integrator)

Title : P4 by SOP M.5

Last Update : Wed Aug 01 16:27:38 2012

Response via : Initial Calibration

DataAcq Meth : P4

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	4.62	150	841514	2.50	ug/ml	0.00
System Monitoring Compounds						
2) Tripropylphosphate	6.70	99	521481	0.77	ug/ml	0.00
Spiked Amount	2.500	Range	20 - 150	Recovery	=	30.80%
Target Compounds						Ovalue
3) White Phosphorus	5.01	124	737365	1.06	ug/ml	100

Quantitation Report

Data File : H:\VOL1\GCMSDATA\J\120801\BJ001059.D

Vial: 2

Acq On : 1 Aug 2012 14:01

Operator: HLM

Sample : P4-CCV

Inst : BNA J

Misc : 1.0PPM BNA J HLM

Multiplr: 1.00

MS Integration Params: RTEINT.P

Quant Time: Aug 1 16:26 2012


Quant Results File: BJ071812.RES


Method : H:\VOL1\GCMSDATA\J\METHODS\BJ071812.M (RTE Integrator)

Title : P4 by SOP M.5

Last Update : Wed Aug 01 16:27:38 2012

Response via : Initial Calibration





#3

White Phosphorus

Concen: 1.06 ug/ml

RT: 5.01 min Scan# 42

Delta R.T. 0.00 min

Lab File: BJ001059.D

Acq: 1 Aug 2012 14:01


Tgt Ion:124 Resp: 737365

Ion Ratio Lower Upper

124 100

93 8.4 0.0 28.4

62 20.6 0.6 40.6



Abundance Ion 124.00 (123.70 to 124.70): E

Ion 93.00 (92.70 to 93.70): BJ0

Ion 62.00 (61.70 to 62.70): BJ0

