

Data Path : S:\HPCHEM1\BNA_J\data\BJ081612\
 Data File : BJ001080.D
 Acq On : 16 Aug 2012 17:02
 Operator : HLM
 Sample : D3733-02DL 100X
 Misc : WATER BNA_J HLM
 ALS Vial : 7 Sample Multiplier: 1

Quant Time: Aug 20 04:32:33 2012
 Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ072012.M
 Quant Title : CWA degradation products by GCMS
 QLast Update : Fri Aug 17 10:52:26 2012
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	4.626	152	436345	2.00	ug/mL	0.00
6) Naphthalene-d8	5.936	136	1809993	2.00	ug/mL	0.00
System Monitoring Compounds						
10) Tripropylphosphate	0.000	99	0d	0.00	ug/mL	
Spiked Amount 0.500	Range	70 - 130	Recovery	=	0.00%#	
12) 2-Bromothioanisole	0.000	204	0d	0.00	ug/mL	
Spiked Amount 5.000	Range	70 - 130	Recovery	=	0.00%#	
Target Compounds						
						Qvalue
5) 1,4-Oxathiane	3.616	46	138476	0.68	ug/mL#	77
7) 1,4-Dithiane	5.082	46	511348	3.01	ug/mL#	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : S:\HPCHEM1\BNA_J\data\BJ081612\
Data File : BJ001080.D
Acq On : 16 Aug 2012 17:02
Operator : HLM
Sample : D3733-02DL 100X
Misc : WATER BNA_J HLM
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Aug 20 04:32:33 2012
Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ072012.M
Quant Title : CWA degradation products by GCMS
QLast Update : Fri Aug 17 10:52:26 2012
Response via : Initial Calibration

