

Data Path : S:\HPCHEM1\BNA_J\data\BJ081612\
 Data File : BJ001081.D
 Acq On : 16 Aug 2012 17:20
 Operator : HLM
 Sample : D3733-02
 Misc : WATER BNA_J HLM
 ALS Vial : 8 Sample Multiplier: 1

Quant Time: Aug 20 04:32:38 2012
 Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ072012.M
 Quant Title : CWA degradation products by GCMS
 QLast Update : Fri Aug 17 10:52:26 2012
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	4.639	152	438091	2.00	ug/mL	0.00
6) Naphthalene-d8	5.920	136	1841442	2.00	ug/mL	0.00
System Monitoring Compounds						
10) Tripropylphosphate	6.987	99	387975	0.58	ug/mL	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery	=	116.00%	
12) 2-Bromothioanisole	7.229	204	634531	4.41	ug/mL	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery	=	88.20%	
Target Compounds						
						Qvalue
5) 1,4-Oxathiane	3.629	46	9462595	46.41	ug/mL	94
7) 1,4-Dithiane	5.109	46	19759498	114.28	ug/mL#	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : S:\HPCHEM1\BNA_J\data\BJ081612\
Data File : BJ001081.D
Acq On : 16 Aug 2012 17:20
Operator : HLM
Sample : D3733-02
Misc : WATER BNA_J HLM
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Aug 20 04:32:38 2012
Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ072012.M
Quant Title : CWA degradation products by GCMS
QLast Update : Fri Aug 17 10:52:26 2012
Response via : Initial Calibration

