

Method Path : S:\HPCHEM1\BNA_J\method\
 Method File : BJ100412.M
 Title : P4 by SOP M.5
 Last Update : Thu Oct 04 17:45:25 2012
 Response Via : Initial Calibration

Calibration Files

1 =BJ001169.D 0.1 =BJ001172.D 10 =BJ001166.D 5 =BJ001167.D 2.5 =BJ001168.D 0.5 =BJ001170.D 0.25=BJ001171.D

	Compound	1	0.1	10	5	2.5	0.5	0.25	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2) S	Tripropylphosp...	2.183	2.442	1.655	2.049	2.024	2.256	2.211	2.117	11.65
3) T	White Phosphorus	2.417	2.434	2.271	2.329	2.316	2.369	2.394	2.361	2.49

(#) = Out of Range