

Method Path : Z:\HPCHEM1\BNA_J\METHOD\
Method File : BJ111212.M
Title : P4 by SOP M.5
Last Update : Mon Nov 12 15:57:06 2012
Response Via : Initial Calibration

Calibration Files

1 =BJ001200.D 0.1 =BJ001203.D 10 =BJ001197.D
5 =BJ001198.D 2.5 =BJ001199.D 0.5 =BJ001201.D

Compound	1	0.1	10	5	2.5	0.5	Avg	%RSD
----------	---	-----	----	---	-----	-----	-----	------

1) I	1,4-Dichlorobenzene-d	-----ISTD-----							
2) S	Tripropylphosphat	2.005	2.053	2.092	2.015	1.969	1.988	2.016	2.12
3) T	White Phosphorus	2.375	2.425	2.374	2.358	2.324	2.345	2.371	1.42

(#) = Out of Range ### Number of calibration levels exceeded format ###