

Data Path : Z:\HPCHEM1\BNA_J\DATA\BJ112612\
Data File : BJ001274.D
Acq On : 26 Nov 2012 13:33
Operator : HLM
Sample : P4-CCV
Misc : P4-CCV BNA_J HLM
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_J
LabSampleId :
P4-CCV

Quant Time: Nov 28 23:46:19 2012
Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ111212.M
Quant Title : P4 by SOP M.5
QLast Update : Mon Nov 26 16:48:06 2012
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	48#	-0.04
2 S	Tripropylphosphate	2.016	2.117	-5.0	51	-0.05
3 T	White Phosphorus	2.371	2.010	15.2	41#	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0