

Data Path : Z:\HPCHEM1\BNA_J\data\BJ042612\
 Data File : BJ000798.D
 Acq On : 26 Apr 2012 19:55
 Operator : HLM
 Sample : D2292-09
 Misc : WATER BNA_J HLM
 ALS Vial : 14 Sample Multiplier: 1

Quant Time: May 03 01:30:30 2012
 Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ042612.M
 Quant Title : CWA degradation products by GCMS
 QLast Update : Fri Apr 27 11:32:02 2012
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	4.755	152	293571	2.00	ug/mL	0.00
6) Naphthalene-d8	6.050	136	1149047	2.00	ug/mL	0.00
System Monitoring Compounds						
10) Tripropylphosphate	7.117	99	212033	0.54	ug/mL	0.02
Spiked Amount 0.500	Range 70 - 130		Recovery	=	108.00%	
12) 2-Bromothioanisole	7.359	204	327316	4.34	ug/mL	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery	=	86.80%	
Target Compounds						
						Qvalue
5) 1,4-Oxathiane	3.744	46	2133250	20.49	ug/mL#	76
7) 1,4-Dithiane	5.210	46	3133132	41.28	ug/mL	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_J\data\BJ042612\
Data File : BJ000798.D
Acq On : 26 Apr 2012 19:55
Operator : HLM
Sample : D2292-09
Misc : WATER BNA_J HLM
ALS Vial : 14 Sample Multiplier: 1

Quant Time: May 03 01:30:30 2012
Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ042612.M
Quant Title : CWA degradation products by GCMS
QLast Update : Fri Apr 27 11:32:02 2012
Response via : Initial Calibration

