

Data Path : Z:\HPCHEM1\BNA_J\data\BJ042612\
 Data File : BJ000808.D
 Acq On : 26 Apr 2012 22:56
 Operator : HLM
 Sample : D2330-08
 Misc : WATER BNA_J HLM
 ALS Vial : 24 Sample Multiplier: 1

Quant Time: May 03 01:35:10 2012
 Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ042612.M
 Quant Title : CWA degradation products by GCMS
 QLast Update : Fri Apr 27 11:32:02 2012
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	4.751	152	245880	2.00	ug/mL	-0.01
6) Naphthalene-d8	6.046	136	1046638	2.00	ug/mL	-0.01
System Monitoring Compounds						
10) Tripropylphosphate	7.128	99	187716	0.53	ug/mL	0.03
Spiked Amount 0.500	Range	70 - 130	Recovery	=	106.00%	
12) 2-Bromothioanisole	7.355	204	281326	4.10	ug/mL	0.00
Spiked Amount 5.000	Range	70 - 130	Recovery	=	82.00%	
Target Compounds						
5) 1,4-Oxathiane	3.740	46	627875	7.20	ug/mL	93
7) 1,4-Dithiane	5.206	46	2030698	29.37	ug/mL	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_J\data\BJ042612\
Data File : BJ000808.D
Acq On : 26 Apr 2012 22:56
Operator : HLM
Sample : D2330-08
Misc : WATER BNA_J HLM
ALS Vial : 24 Sample Multiplier: 1

Quant Time: May 03 01:35:10 2012
Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ042612.M
Quant Title : CWA degradation products by GCMS
QLast Update : Fri Apr 27 11:32:02 2012
Response via : Initial Calibration

