

Data Path : Z:\HPCHEM1\BNA_J\data\BJ042712\
 Data File : BJ000821.D
 Acq On : 27 Apr 2012 12:42
 Operator : HLM
 Sample : D2330-09DL 5X
 Misc : WATER BNA_J HLM
 ALS Vial : 6 Sample Multiplier: 1

Quant Time: May 02 07:13:58 2012
 Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ042612.M
 Quant Title : CWA degradation products by GCMS
 QLast Update : Fri Apr 27 15:25:19 2012
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	4.749	152	272963	2.00	ug/mL	0.01
6) Naphthalene-d8	6.044	136	1101764	2.00	ug/mL	0.00
System Monitoring Compounds						
10) Tripropylphosphate	7.098	99	40100	0.11	ug/mL	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery	=	22.00%#	
12) 2-Bromothioanisole	7.340	204	59778	0.83	ug/mL	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery	=	16.60%#	
Target Compounds						
						Qvalue
5) 1,4-Oxathiane	3.725	46	621404	6.42	ug/mL#	37
7) 1,4-Dithiane	5.191	46	1333514	18.32	ug/mL	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_J\data\BJ042712\
Data File : BJ000821.D
Acq On : 27 Apr 2012 12:42
Operator : HLM
Sample : D2330-09DL 5X
Misc : WATER BNA_J HLM
ALS Vial : 6 Sample Multiplier: 1

Quant Time: May 02 07:13:58 2012
Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ042612.M
Quant Title : CWA degradation products by GCMS
QLast Update : Fri Apr 27 15:25:19 2012
Response via : Initial Calibration

