

Data Path : Z:\HPCHEM1\BNA_J\data\BJ050112\
 Data File : BJ000882.D
 Acq On : 1 May 2012 14:50
 Operator : HLM
 Sample : D2367-01DL 20X
 Misc : WATER BNA_J HLM
 ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 03 01:58:02 2012
 Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ042612.M
 Quant Title : CWA degradation products by GCMS
 QLast Update : Tue May 01 18:03:04 2012
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	4.751	152	260345	2.00	ug/mL	0.00
6) Naphthalene-d8	6.060	136	1094966	2.00	ug/mL	0.00
System Monitoring Compounds						
10) Tripropylphosphate	7.085	99	10088	0.03	ug/mL	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery	=	6.00%#	
12) 2-Bromothioanisole	7.355	204	8580	0.12	ug/mL	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery	=	2.40%#	
Target Compounds						
						Qvalue
5) 1,4-Oxathiane	3.726	46	699684	7.58	ug/mL	93
7) 1,4-Dithiane	5.206	46	1316967	18.21	ug/mL#	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_J\data\BJ050112\
Data File : BJ000882.D
Acq On : 1 May 2012 14:50
Operator : HLM
Sample : D2367-01DL 20X
Misc : WATER BNA_J HLM
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 03 01:58:02 2012
Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ042612.M
Quant Title : CWA degradation products by GCMS
QLast Update : Tue May 01 18:03:04 2012
Response via : Initial Calibration

