

Data File : H:\VOL1\GCMSDATA\J\120828\BJ001085.D

Vial: 2

Acq On : 28 Aug 2012 14:53

Operator: HLM

Sample : WP-7

Inst : BNA J

Misc : WHITE PHOS STD BNA J HLM

Multiplr: 1.00

MS Integration Params: RTEINT.P

Quant Time: Aug 28 17:23 2012

Quant Results File: BJ082812.RES

Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ082812.M (RTE Integrator)

Title : P4 by SOP M.5

Last Update : Tue Aug 28 17:25:08 2012

Response via : Initial Calibration

DataAcq Meth : P4

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) 1,4-Dichlorobenzene-d4	4.59	150	761298	2.50	ug/ml	0.00
System Monitoring Compounds						
2) Tripropylphosphate	6.67	99	7353364	12.05	ug/ml	0.00
Spiked Amount	2.500	Range	20 - 150	Recovery	=	482.00%#
Target Compounds						
3) White Phosphorus	4.98	124	7570634	12.05	ug/ml	Ovalue 99


Quantitation Report


Data File : H:\VOL1\GCMSDATA\J\120828\BJ001085.D
 Acq On : 28 Aug 2012 14:53
 Sample : WP-7
 Misc : WHITE PHOS STD BNA J HLM
 MS Integration Params: RTEINT.P
 Quant Time: Aug 28 17:23 2012

Vial: 2
 Operator: HLM
 Inst : BNA J
 Multiplr: 1.00

Quant Results File: BJ082812.RES

Method : H:\VOL1\GCMSDATA\J\METHODS\BJ082812.M (RTE Integrator)
 Title : P4 by SOP M.5
 Last Update : Tue Aug 28 17:25:08 2012
 Response via : Initial Calibration





#3

White Phosphorus

Concen: 12.05 ug/ml

RT: 4.98 min Scan# 39

Delta R.T. 0.01 min

Lab File: BJ001085.D

Acq: 28 Aug 2012 14:53



Tgt Ion:124 Resp: 7570634

Ion Ratio Lower Upper

124 100

93 7.4 0.0 27.4

62 16.5 0.0 36.0



Abundance Ion 124.00 (123.70 to 124.70): E

Ion 93.00 (92.70 to 93.70): BJ0

Ion 62.00 (61.70 to 62.70): BJ0

