

Data File : H:\VOL1\GCMSDATA\J\120828\BJ001087.D

Vial: 4

Acq On : 28 Aug 2012 15:20

Operator: HLM

Sample : WP-5

Inst : BNA J

Misc : WHITE PHOS STD BNA J HLM

Multiplr: 1.00

MS Integration Params: RTEINT.P

Quant Time: Aug 28 17:23 2012

Quant Results File: BJ082812.M

Instrument :

BNA_J

ClientSampleId :


WP-5


Method : H:\VOL1\GCMSDATA\J\METHODS\BJ082812.M (RTE Integrator)

Title : P4 by SOP M.5

Last Update : Tue Aug 28 17:26:49 2012

Response via : Initial Calibration





#3

White Phosphorus

Concen: 2.73 ug/ml

RT: 4.98 min Scan# 39

Delta R.T. -0.00 min

Lab File: BJ001087.D

Acq: 28 Aug 2012 15:20

Instrument :

BNA_J

ClientSampleId :

WP-5


Tgt Ion:124 Resp: 2047656

Ion Ratio Lower Upper

124 100

93 7.2 0.0 27.4

62 16.3 0.0 36.0





Abundance

Ion 124.00 (123.70 to 124.70): E

Ion 93.00 (92.70 to 93.70): BJ0

Ion 62.00 (61.70 to 62.70): BJ0



Data File : H:\VOL1\GCMSDATA\J\120828\BJ001087.D
Acq On : 28 Aug 2012 15:20
Sample : WP-5
Misc : WHITE PHOS STD BNA J HLM
MS Integration Params: RTEINT.P
Quant Time: Aug 28 17:23 2012

Vial: 4
Operator: HLM
Inst : BNA J
Multiplr: 1.00

Instrument :
BNA_J
ClientSampleId :
WP-5

Quant Results File: BJ082812.M

Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ082812.M (RTE Integrator)
Title : P4 by SOP M.5
Last Update : Tue Aug 28 17:25:08 2012
Response via : Initial Calibration
DataAcq Meth : P4

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	4.59	150	909357	2.50	ug/ml	0.00
System Monitoring Compounds						
2) Tripropylphosphate	6.68	99	1785675	2.45	ug/ml	0.00
Spiked Amount	2.500	Range	20 - 150	Recovery	=	98.00%
Target Compounds						
3) White Phosphorus	4.98	124	2047656	2.73	ug/ml	Ovalue 99