

Data File : H:\VOL1\GCMSDATA\J\120828\BJ001090.D

Vial: 7

Acq On : 28 Aug 2012 16:00

Operator: HLM

Sample : WP-2

Inst : BNA J

Misc : WHITE PHOS STD BNA J HLM

Multiplr: 1.00

MS Integration Params: RTEINT.P

Quant Time: Aug 28 17:23 2012

Quant Results File: BJ082812.M

Instrument :

BNA_J

ClientSampleId :


WP-2




Method : H:\VOL1\GCMSDATA\J\METHODS\BJ082812.M (RTE Integrator)

Title : P4 by SOP M.5

Last Update : Tue Aug 28 17:26:49 2012

Response via : Initial Calibration





#3

White Phosphorus

Concen: 0.25 ug/ml

RT: 4.98 min Scan# 39

Delta R.T. -0.00 min

Lab File: BJ001090.D

Acq: 28 Aug 2012 16:00

Instrument :

BNA_J

ClientSampleId :

WP-2


Tgt Ion:124 Resp: 172226

Ion Ratio Lower Upper

124 100

93 7.6 0.0 27.4

62 16.1 0.0 36.0



Data File : H:\VOL1\GCMSDATA\J\120828\BJ001090.D
Acq On : 28 Aug 2012 16:00
Sample : WP-2
Misc : WHITE PHOS STD BNA J HLM
MS Integration Params: RTEINT.P
Quant Time: Aug 28 17:23 2012

Vial: 7
Operator: HLM
Inst : BNA J
Multiplr: 1.00

Instrument :
BNA_J
ClientSampleId :
WP-2

Quant Results File: BJ082812.M

Quant Method : H:\VOL1\GCMSDATA\J\METHODS\BJ082812.M (RTE Integrator)
Title : P4 by SOP M.5
Last Update : Tue Aug 28 17:25:08 2012
Response via : Initial Calibration
DataAcq Meth : P4

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	4.59	150	837703	2.50	ug/ml	0.00
System Monitoring Compounds						
2) Tripropylphosphate	6.68	99	162018	0.24	ug/ml	0.00
Spiked Amount	2.500	Range	20 - 150	Recovery	=	9.60%#
Target Compounds						
3) White Phosphorus	4.98	124	172226	0.25	ug/ml	Ovalue 100