



Quant Results File: BL072512.RES

AHPatel
9/6/2012 6:41:31 AM









#3
N-Nitrosodimethylamine
Concen: 61.85 ug/ml
RT: 3.36 min Scan# 29
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Instrument :
BNA_L
ClientSampleId :
SSTD050

Tgt Ion: 74 Resp: 684463
Ion Ratio Lower Upper
74 100
43 21.1 1.1 41.1




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:31 AM



#4
Pyridine
Concen: 60.36 ug/ml
RT: 3.46 min Scan# 38
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Tgt Ion: 79 Resp: 982769
Ion Ratio Lower Upper
79 100
52 64.0 44.0 84.0





#6

Benzaldehyde

Concen: 65.18 ug/ml

RT: 9.19 min Scan# 546

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L

ClientSampleId :

SSTD050

Tgt Ion: 77 Resp: 239530

Ion Ratio Lower Upper

77 100


105 90.5 70.5 110.5

106 92.3 72.3 112.3

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:31 AM






Abundance

Ion 77.00 (76.70 to 77.70): BL0

Ion 105.00 (104.70 to 105.70): B

Ion 106.00 (105.70 to 106.70): B



#7

Aniline

Concen: 50.93 ug/ml

RT: 9.41 min Scan# 565

Delta R.T. 0.00 min

Lab File: BL000632.D


Acq: 25 Jul 2012 13:43

Tgt Ion: 93 Resp: 956862

Ion Ratio Lower Upper

93 100



66 37.7 30.2 45.2




Abundance

Ion 93.00 (92.70 to 93.70): BL0

Ion 66.00 (65.70 to 66.70): BL0









#8
bis(2-Chloroethyl)ether
Concen: 49.65 ug/ml
RT: 9.52 min Scan# 575
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Instrument :
BNA_L
ClientSampled :
SSTD050

Tgt Ion: 93 Resp: 643002
Ion Ratio Lower Upper
93 100
63 76.5 56.5 96.5
95 31.6 11.6 51.6




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:31 AM



#10
Phenol
Concen: 52.38 ug/ml
RT: 9.36 min Scan# 561
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Tgt Ion: 94 Resp: 888897
Ion Ratio Lower Upper
94 100
65 20.9 0.9 40.9
66 23.3 3.3 43.3









#11
 2-Chlorophenol
 Concen: 43.55 ug/ml
 RT: 9.62 min Scan# 584
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampled :
 SST050

Tgt Ion:128 Resp: 552433
 Ion Ratio Lower Upper
 128 100
 64 53.6 33.6 73.6
 130 31.8 11.8 51.8




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:31 AM



#12
 1,3-Dichlorobenzene
 Concen: 40.36 ug/ml
 RT: 9.91 min Scan# 609
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion:146 Resp: 558583
 Ion Ratio Lower Upper
 146 100
 148 64.7 44.7 84.7
 113 12.4 0.0 32.4









#13
 1,4-Dichlorobenzene
 Concen: 40.71 ug/ml
 RT: 10.04 min Scan# 621
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampled :
 SST050

Tgt Ion:146 Resp: 562975
 Ion Ratio Lower Upper
 146 100
 148 64.1 44.1 84.1
 113 11.5 0.0 31.5




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:31 AM



#14
 1,2-Dichlorobenzene
 Concen: 40.22 ug/ml
 RT: 10.31 min Scan# 645
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion:146 Resp: 513975
 Ion Ratio Lower Upper
 146 100
 148 64.2 44.2 84.2
 113 13.3 0.0 33.3





#15

Benzyl alcohol

Concen: 43.86 ug/ml

RT: 10.26 min Scan# 640

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L

ClientSampled :

SSTD050

Tgt Ion:108 Resp: 343722

Ion Ratio Lower Upper

108 100


79 111.7 91.7 131.7

77 63.1 43.1 83.1

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:31 AM






Abundance

Ion 108.00 (107.70 to 108.70): E

Ion 79.00 (78.70 to 79.70): BL0

Ion 77.00 (76.70 to 77.70): BL0



#16

2,2-oxybis(1-chloropropane)

Concen: 42.73 ug/ml

RT: 10.49 min Scan# 661

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43


Tgt Ion: 45 Resp: 942182

Ion Ratio Lower Upper

45 100

77 10.5 0.0 30.5

79 9.0 0.0 29.0






Abundance

Ion 45.00 (44.70 to 45.70): BL0

Ion 77.00 (76.70 to 77.70): BL0

Ion 79.00 (78.70 to 79.70): BL0





#17

2-Methylphenol

Concen: 45.48 ug/ml

RT: 10.45 min Scan# 657

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L

ClientSampleId :

SSTD050

Tgt Ion:108 Resp: 489343

Ion Ratio Lower Upper





108 100

107 88.3 68.3 108.3

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:31 AM



#18

Acetophenone

Concen: 40.86 ug/ml

RT: 10.73 min Scan# 682

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43




Tgt Ion:105 Resp: 540111


Ion Ratio Lower Upper

105 100

71 7.2 0.0 27.2

51 34.1 14.1 54.1









#19
Hexachloroethane
Concen: 40.20 ug/ml
RT: 10.93 min Scan# 700
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Instrument :
BNA_L
ClientSampleId :
SSTD050

Tgt Ion: 117 Resp: 211504
Ion Ratio Lower Upper
117 100
200 0.0 0.0 20.0
199 51.1 31.1 71.1




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:31 AM



#20
N-Nitroso-di-n-propylamine
Concen: 59.35 ug/ml
RT: 10.74 min Scan# 683
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Tgt Ion: 70 Resp: 332800
Ion Ratio Lower Upper
70 100
42 50.9 30.9 70.9
130 24.9 4.9 44.9





#21

4-Methylphenol

Concen: 42.06 ug/ml

RT: 10.73 min Scan# 682

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L

ClientSampled :

SSTD050

Tgt Ion:108 Resp: 455575

Ion Ratio Lower Upper

108 100


107 109.6 89.6 129.6

Manual Integrations

APPROVED




AHPatel

9/6/2012 6:41:31 AM



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E



#24

Nitrobenzene

Concen: 58.34 ug/ml

RT: 11.03 min Scan# 709

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43


Tgt Ion: 77 Resp: 479327

Ion Ratio Lower Upper

77 100

123 51.0 31.0 71.0



65 12.3 0.0 32.3




Abundance Ion 77.00 (76.70 to 77.70): BL0

Ion 123.00 (122.70 to 123.70): E

Ion 65.00 (64.70 to 65.70): BL0





#25

Isophorone

Concen: 50.40 ug/ml

RT: 11.47 min Scan# 748

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L

ClientSampled :

SSTD050

Tgt Ion: 82 Resp: 869972

Ion Ratio Lower Upper


82 100

138 15.2 0.0 35.2

Manual Integrations
APPROVED




AHPatel

9/6/2012 6:41:31 AM



Abundance Ion 82.00 (81.70 to 82.70): BL0

Ion 138.00 (137.70 to 138.70): B



#26

2-Nitrophenol

Concen: 37.58 ug/ml

RT: 11.61 min Scan# 760

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43


Tgt Ion: 139 Resp: 245206

Ion Ratio Lower Upper

139 100

65 41.0 21.0 61.0



109 13.4 0.0 33.4




Abundance Ion 139.00 (138.70 to 139.70): B

Ion 65.00 (64.70 to 65.70): BL0

Ion 109.00 (108.70 to 109.70): B






#27
2,4-Dimethylphenol
Concen: 42.71 ug/ml
RT: 11.69 min Scan# 767
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Instrument :
BNA_L
ClientSampleId :
SSTD050




Tgt Ion:107 Resp: 372694
Ion Ratio Lower Upper
107 100
121 54.9 34.9 74.9
122 96.4 76.4 116.4

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:31 AM




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 121.00 (120.70 to 121.70): E
Ion 122.00 (121.70 to 122.70): E






#28
Benzoic Acid
Concen: 69.22 ug/ml
RT: 11.96 min Scan# 791
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Tgt Ion:105 Resp: 477392
Ion Ratio Lower Upper
105 100
77 70.9 50.9 90.9
122 85.1 65.1 105.1



Abundance Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BL0
Ion 122.00 (121.70 to 122.70): E









#29
bis(2-Chloroethoxy)methane
Concen: 45.80 ug/ml
RT: 11.86 min Scan# 782
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Instrument :
BNA_L
ClientSampled :
SSTD050

Tgt Ion: 93 Resp: 548409
Ion Ratio Lower Upper
93 100
95 31.7 11.7 51.7
123 11.0 0.0 31.0




Manual Integrations
APPROVED




AHPatel
9/6/2012 6:41:31 AM



#30
2,4-Dichlorophenol
Concen: 35.51 ug/ml
RT: 12.05 min Scan# 799
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Tgt Ion:162 Resp: 298916
Ion Ratio Lower Upper
162 100
164 62.6 42.6 82.6
98 37.8 17.8 57.8









#32
1,2,4-Trichlorobenzene
Concen: 41.30 ug/ml
RT: 12.21 min Scan# 813
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Tgt Ion:180 Resp: 338953
Ion Ratio Lower Upper
180 100
182 94.6 74.6 114.6
145 27.7 7.7 47.7

Instrument :
BNA_L
ClientSampleId :
SSTD050


Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:31 AM



#33
Naphthalene
Concen: 39.11 ug/ml
RT: 12.36 min Scan# 826
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Tgt Ion:128 Resp: 1077175
Ion Ratio Lower Upper
128 100
129 10.5 0.0 30.5
127 12.7 0.0 32.7






#34
4-Chloroaniline
Concen: 36.33 ug/ml
RT: 12.45 min Scan# 834
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Instrument :
BNA_L
ClientSampled :
SSTD050




Tgt Ion:127 Resp: 418133
Ion Ratio Lower Upper
127 100
129 32.3 12.3 52.3

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:31 AM




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E






#35
Hexachlorobutadiene
Concen: 72.62 ug/ml
RT: 12.59 min Scan# 847
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Tgt Ion:225 Resp: 144442
Ion Ratio Lower Upper
225 100
223 61.6 41.6 81.6
227 64.8 44.8 84.8



Abundance Ion 224.80 (224.50 to 225.50): E
Ion 222.70 (222.40 to 223.40): E
Ion 226.60 (226.30 to 227.30): E









#36
 Caprolactam
 Concen: 33.39 ug/ml
 RT: 13.10 min Scan# 892
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampled :
 SST050

Tgt Ion:113 Resp: 97433
 Ion Ratio Lower Upper
 113 100
 55 167.4 147.4 187.4
 56 110.3 90.3 130.3




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:31 AM



#37
 4-Chloro-3-methylphenol
 Concen: 39.29 ug/ml
 RT: 13.34 min Scan# 913
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion:107 Resp: 263595
 Ion Ratio Lower Upper
 107 100
 144 29.1 9.1 49.1
 142 91.7 71.7 111.7









#38
 2-Methylnaphthalene
 Concen: 34.95 ug/ml
 RT: 13.62 min Scan# 938
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampled :
 SSTD050

Tgt Ion:142 Resp: 574536
 Ion Ratio Lower Upper
 142 100
 141 84.4 64.4 104.4




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:31 AM



#40
 Hexachlorocyclopentadiene
 Concen: 84.23 ug/ml
 RT: 13.93 min Scan# 965
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion:237 Resp: 127617
 Ion Ratio Lower Upper
 237 100
 235 62.7 42.7 82.7
 272 14.2 0.0 34.2









#41
 2,4,6-Trichlorophenol
 Concen: 53.90 ug/ml
 RT: 14.14 min Scan# 984
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampled :
 SSTD050

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	150211		
198	95.3	75.3	115.3	
200	30.2	10.2	50.2	




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:31 AM



#42
 2,4,5-Trichlorophenol
 Concen: 53.65 ug/ml
 RT: 14.21 min Scan# 990
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion	Ratio	Lower	Upper
196	100		
198	93.2	73.2	113.2
200	30.9	10.9	50.9






#44
 1,1-Biphenyl
 Concen: 41.29 ug/ml
 RT: 14.48 min Scan# 1014
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampled :
 SST050

Tgt Ion:154 Resp: 590097
 Ion Ratio Lower Upper
 154 100
 153 40.3 20.3 60.3
 76 29.7 9.7 49.7

Manual Integrations
 APPROVED

AHPatel
 9/6/2012 6:41:31 AM






Abundance

Ion 154.00 (153.70 to 154.70): E


Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BL0



#45
 2-Chloronaphthalene
 Concen: 41.24 ug/ml
 RT: 14.51 min Scan# 1017
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion:162 Resp: 480031
 Ion Ratio Lower Upper
 162 100
 164 31.9 11.9 51.9
 127 34.7 14.7 54.7






Abundance

Ion 162.00 (161.70 to 162.70): E

Ion 164.00 (163.70 to 164.70): E

Ion 127.00 (126.70 to 127.70): E









#46
 2-Nitroaniline
 Concen: 57.82 ug/ml
 RT: 14.70 min Scan# 1034
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampled :
 SSTD050

Tgt Ion: 65 Resp: 133070
 Ion Ratio Lower Upper
 65 100
 92 71.3 51.3 91.3
 138 128.3 108.3 148.3




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:31 AM



#47
 Acenaphthylene
 Concen: 37.50 ug/ml
 RT: 15.29 min Scan# 1086
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion: 152 Resp: 670504
 Ion Ratio Lower Upper
 152 100
 151 19.7 0.0 39.7
 153 12.5 0.0 32.5









#48
 Dimethylphthalate
 Concen: 34.63 ug/ml
 RT: 15.05 min Scan# 1065
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampleId :
 SST050

Tgt Ion	Ratio	Lower	Upper
163	100		
194	5.8	0.0	25.8
164	10.4	0.0	30.4




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:31 AM



#49
 2,6-Dinitrotoluene
 Concen: 32.62 ug/ml
 RT: 15.16 min Scan# 1074
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion	Ratio	Lower	Upper
165	100		
89	38.2	18.2	58.2
121	15.8	0.0	35.8





#50

Acenaphthene

Concen: 37.92 ug/ml

RT: 15.61 min Scan# 1114

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L

ClientSampled :

SSTD050

Tgt Ion:153 Resp: 403699

Ion Ratio Lower Upper

153 100


152 45.2 25.2 65.2

154 95.2 75.2 115.2

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:31 AM






Abundance

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E



#51

3-Nitroaniline

Concen: 31.02 ug/ml

RT: 15.47 min Scan# 1102

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43


Tgt Ion:138 Resp: 103163

Ion Ratio Lower Upper

138 100

108 9.0 0.0 29.0

92 94.3 74.3 114.3






Abundance

Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BL0









#52
 2,4-Dinitrophenol
 Concen: 34.88 ug/ml
 RT: 15.66 min Scan# 1119
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampleId :
 SST050

Tgt Ion:184 Resp: 53337
 Ion Ratio Lower Upper
 184 100
 63 27.6 7.6 47.6
 154 26.0 6.0 46.0




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:31 AM



#53
 Dibenzofuran
 Concen: 38.13 ug/ml
 RT: 15.94 min Scan# 1143
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion:168 Resp: 513646
 Ion Ratio Lower Upper
 168 100
 139 27.3 7.3 47.3





#54

2,4-Dinitrotoluene

Concen: 29.34 ug/ml

RT: 15.91 min Scan# 1141

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L





ClientSampled :

SSTD050

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	117187		
63	27.6	7.6	47.6	
182	6.1	0.0	26.1	

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:31 AM



#55

4-Nitrophenol

Concen: 26.85 ug/ml




RT: 15.79 min Scan# 1130


Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Tgt Ion	Ratio	Lower	Upper
139	100		
65	57.5	37.5	77.5
109	27.5	7.5	47.5









#56
 2,3,4,6-Tetrachlorophenol
 Concen: 69.07 ug/ml
 RT: 16.17 min Scan# 1164
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampleId :
 SST050

Tgt Ion:	232	Resp:	86590
Ion	Ratio	Lower	Upper
232	100		
230	79.8	63.8	95.8
131	34.7	27.8	41.6
166	25.7	20.6	30.8




Manual Integrations
 APPROVED




AHPatel
 9/6/2012 6:41:31 AM



#57
 Fluorene
 Concen: 35.56 ug/ml
 RT: 16.57 min Scan# 1199
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion:	166	Resp:	361605
Ion	Ratio	Lower	Upper
166	100		
165	92.2	72.2	112.2
167	13.3	0.0	33.3









#58
 4-Chlorophenyl-phenylether
 Concen: 59.80 ug/ml
 RT: 16.58 min Scan# 1200
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampleId :
 SST050

Tgt Ion	Ratio	Lower	Upper
204	100		
206	32.0	12.0	52.0
141	45.2	25.2	65.2


Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:31 AM



#59
 Diethylphthalate
 Concen: 28.94 ug/ml
 RT: 16.38 min Scan# 1182
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion	Ratio	Lower	Upper
149	100		
177	21.9	1.9	41.9
150	11.3	0.0	31.3









#60
4-Nitroaniline
Concen: 21.45 ug/ml
RT: 16.60 min Scan# 1202
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Instrument :
BNA_L
ClientSampleId :
SSTD050

Tgt Ion:138 Resp: 68410
Ion Ratio Lower Upper
138 100
92 42.1 22.1 62.1
108 31.1 11.1 51.1




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:31 AM



#63
1,2,4,5-Tetrachlorobenzene
Concen: 85.71 ug/ml
RT: 13.94 min Scan# 966
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Tgt Ion:216 Resp: 229820
Ion Ratio Lower Upper
216 100
214 79.3 63.4 95.2
179 17.9 14.3 21.5
218 47.4 37.9 56.9





#64

4,6-Dinitro-2-methylphenol

Concen: 40.05 ug/ml

RT: 16.67 min Scan# 1208

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L


ClientSampled :

SSTD050




Tgt Ion:198 Resp: 59103

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:31 AM



Abundance Ion 198.00 (197.70 to 198.70): E



#65

n-Nitrosodiphenylamine

Concen: 37.35 ug/ml

RT: 16.79 min Scan# 1219

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43


Tgt Ion:169 Resp: 295310

Ion Ratio Lower Upper

169 100

168 62.0 42.0 82.0



167 31.9 11.9 51.9




Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E









#66
 1,2-Diphenylhydrazine
 Concen: 56.86 ug/ml
 RT: 16.87 min Scan# 1226
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampleId :
 SST050

Tgt Ion: 77 Resp: 452565
 Ion Ratio Lower Upper
 77 100
 182 20.5 0.5 40.5
 105 14.6 0.0 34.6




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:31 AM



#67
 4-Bromophenyl-phenylether
 Concen: 76.24 ug/ml
 RT: 17.49 min Scan# 1281
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion: 248 Resp: 84613
 Ion Ratio Lower Upper
 248 100
 250 100.2 80.2 120.2
 141 54.0 34.0 74.0





#68

Hexachlorobenzene

Concen: 66.11 ug/ml

RT: 17.62 min Scan# 1292

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L





ClientSampleId :

SSTD050

Tgt Ion	Ratio	Lower	Upper
284	100		
142	32.2	12.2	52.2
249	26.6	6.6	46.6

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:31 AM



#69

Atrazine

Concen: 39.71 ug/ml




RT: 17.82 min Scan# 1310




Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Tgt Ion	Ratio	Lower	Upper
200	100		
173	31.1	11.1	51.1
115	0.0	0.0	20.0









#70
 Pentachlorophenol
 Concen: 50.63 ug/ml
 RT: 17.98 min Scan# 1324
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampleId :
 SST050

Tgt Ion	Ratio	Resp	Lower	Upper
266	100	35305		
264	64.4		44.4	84.4
268	62.0		42.0	82.0


Manual Integrations
 APPROVED




AHPatel
 9/6/2012 6:41:31 AM



#72
 Benzidine
 Concen: 15.23 ug/ml
 RT: 20.34 min Scan# 1533
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion:184 Resp: 17686









#73
Phenanthrene
Concen: 33.28 ug/ml
RT: 18.36 min Scan# 1358
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Instrument :
BNA_L
ClientSampleId :
SSTD050

Tgt Ion	Ratio	Lower	Upper
178	100		
179	14.8	0.0	34.8
176	17.7	0.0	37.7


Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:31 AM



#74
Anthracene
Concen: 32.87 ug/ml
RT: 18.45 min Scan# 1366
Delta R.T. 0.00 min
Lab File: BL000632.D
Acq: 25 Jul 2012 13:43

Tgt Ion	Ratio	Lower	Upper
178	100		
179	15.1	0.0	35.1
176	17.9	0.0	37.9









#75
 Carbazole
 Concen: 23.54 ug/ml
 RT: 18.72 min Scan# 1390
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampleId :
 SST050

Tgt Ion:167 Resp: 234846
 Ion Ratio Lower Upper
 167 100
 167 100.0 80.0 120.0




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:31 AM



#76
 Di-n-butylphthalate
 Concen: 22.66 ug/ml
 RT: 19.30 min Scan# 1441
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion:149 Resp: 396788
 Ion Ratio Lower Upper
 149 100
 150 8.9 0.0 28.9
 104 3.0 0.0 23.0









#77
Fluoranthene
 Concen: 37.63 ug/ml
 RT: 20.16 min Scan# 1517
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Instrument :
 BNA_L
 ClientSampleId :
 SST0050

Tgt Ion: 202 Resp: 299364
 Ion Ratio Lower Upper
 202 100
 101 25.1 5.1 45.1




Manual Integrations
APPROVED


AHPatel
 9/6/2012 6:41:31 AM



#79
Pyrene
 Concen: 27.23 ug/ml
 RT: 20.45 min Scan# 1543
 Delta R.T. 0.00 min
 Lab File: BL000632.D
 Acq: 25 Jul 2012 13:43

Tgt Ion: 202 Resp: 298721
 Ion Ratio Lower Upper
 202 100
 101 32.5 12.5 52.5
 100 26.4 6.4 46.4





#81

Butylbenzylphthalate

Concen: 14.88 ug/ml

RT: 21.22 min Scan# 1611

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L





ClientSampleId :

SSTD050

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	157836		
91	77.9	57.9	97.9	
206	15.0	0.0	35.0	

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:31 AM



#82

3,3-Dichlorobenzidine

Concen: 30.78 ug/ml




RT: 21.85 min Scan# 1667




Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Tgt Ion	Ratio	Lower	Upper
252	100		
254	63.9	43.9	83.9
126	18.1	0.0	38.1





#83

Benzo[a]anthracene

Concen: 37.36 ug/ml

RT: 21.90 min Scan# 1671

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L

ClientSampled :

SSTD050

Tgt Ion:228 Resp: 283574

Ion Ratio Lower Upper





228 100

229 18.6 0.0 38.6

226 24.7 4.7 44.7

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:31 AM



#84

Chrysene

Concen: 37.81 ug/ml

RT: 21.94 min Scan# 1675

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43


Tgt Ion:228 Resp: 272520


Ion Ratio Lower Upper

228 100

226 27.4 7.4 47.4

229 18.6 0.0 38.6





#85

bis(2-Ethylhexyl)phthalate

Concen: 16.53 ug/ml

RT: 21.87 min Scan# 1669

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L





ClientSampleId :

SSTD050

Tgt Ion:	149	Resp:	259596
Ion Ratio	Lower	Upper	
149	100		
167	32.5	12.5	52.5
279	7.1	0.0	27.1

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:31 AM



#87

Di-n-octylphthalate

Concen: 16.30 ug/ml

RT: 22.69 min Scan# 1741

Delta R.T. 0.00 min

Lab File: BL000632.D




Acq: 25 Jul 2012 13:43


Tgt Ion:149 Resp: 479606

Ion Ratio Lower Upper

149 100

150 9.7 0.0 29.7





#88

Benzo[b]fluoranthene

Concen: 41.75 ug/ml

RT: 23.39 min Scan# 1803

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L





ClientSampled :

SSTD050

Tgt Ion:	252	Resp:	343696
Ion Ratio	Lower	Upper	
252	100		
253	21.5	1.5	41.5
125	21.5	1.5	41.5

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:31 AM



#89

Benzo[k]fluoranthene

Concen: 38.90 ug/ml




RT: 23.43 min Scan# 1807


Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Tgt Ion:	252	Resp:	282182
Ion Ratio	Lower	Upper	
252	100		
253	21.4	1.4	41.4
125	21.5	1.5	41.5





#90

Benzo[a]pyrene

Concen: 40.93 ug/ml

RT: 23.94 min Scan# 1852

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L

ClientSampleId :

SSTD050

Tgt Ion:252 Resp: 306885
Ion Ratio Lower Upper

252 100


253 22.0 2.0 42.0

125 28.8 8.8 48.8

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:31 AM






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#91

Indeno[1,2,3-cd]pyrene

Concen: 41.69 ug/ml

RT: 26.35 min Scan# 2065

Delta R.T. 0.00 min

Lab File: BL000632.D


Acq: 25 Jul 2012 13:43

Tgt Ion:276 Resp: 342770

Ion Ratio Lower Upper

276 100



138 70.5 50.5 90.5




Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E





#92

Dibenzo[a,h]anthracene

Concen: 41.96 ug/ml

RT: 26.38 min Scan# 2068

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43

Instrument :

BNA_L





ClientSampled :

SSTD050

Tgt Ion:	278	Resp:	282211
Ion Ratio	Lower	Upper	
278	100		
139	40.5	20.5	60.5
279	23.5	3.5	43.5

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:31 AM



#93

Benzo[g,h,i]perylene

Concen: 43.01 ug/ml

RT: 27.07 min Scan# 2129

Delta R.T. 0.00 min

Lab File: BL000632.D

Acq: 25 Jul 2012 13:43




Tgt Ion:276 Resp: 304436

Ion Ratio Lower Upper

276 100

138 49.8 29.8 69.8

277 23.6 3.6 43.6



Data File : H:\VOL1\GCMSDATA\L\120725\BL000632.D
 Acq On : 25 Jul 2012 13:43
 Sample : SSTD050
 Misc : SSTD050 HLM
 MS Integration Params: rteint.p
 Quant Time: Jul 25 15:39 2012

Vial: 2
 Operator: HLM
 Inst : BNA L
 Multiplr: 1.00

Instrument :
 BNA_L
 ClientSampleId :
 SSTD050

Quant Results File: BL072512.M

Quant Method : H:\VOL1\GCMSDATA\L\METHODS\BL072512.M (RTE Integrator)
 Title : SW-846 Method 8270
 Last Update : Wed Jul 25 15:39:43 2012
 Response via : Initial Calibration
 DataAcq Meth : 8270L

Manual Integrations
 APPROVED

AHPatel
 9/6/2012 6:41:31 AM

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	10.01	152	401289	40.00	ug/ml	0.00
22) Naphthalene-d8	12.31	136	1152449	40.00	ug/ml	0.00
39) Acenaphthene-d10	15.55	164	390969	40.00	ug/ml	0.00
62) Phenanthrene-d10	18.32	188	400264	40.00	ug/ml	0.00
78) Chrysene-d12	21.91	240	292494	40.00	ug/ml	0.00
86) Perylene-d12	24.04	264	298927	40.00	ug/ml	0.00

System Monitoring Compounds

5) 2-Fluorophenol	7.51	112	685444	50.87	ug/ml	0.00
Spiked Amount	200.000	Range	20 - 110	Recovery	=	25.44%
9) Phenol-d5	9.34	99	762202	48.44	ug/ml	0.00
Spiked Amount	200.000	Range	10 - 115	Recovery	=	24.22%
23) Nitrobenzene-d5	11.00	82	471640	56.24	ug/ml	0.00
Spiked Amount	100.000	Range	40 - 110	Recovery	=	56.24%
43) 2-Fluorobiphenyl	14.30	172	528770	44.86	ug/ml	0.00
Spiked Amount	100.000	Range	50 - 110	Recovery	=	44.86%#
61) 2,4,6-Tribromophenol	17.02	330	39132	46.21	ug/ml	0.00
Spiked Amount	200.000	Range	40 - 125	Recovery	=	23.11%#
80) Terphenyl-d14	20.64	244	161217	39.82	ug/ml	0.00
Spiked Amount	100.000	Range	50 - 135	Recovery	=	39.82%#

Target Compounds

						Qvalue
3) N-Nitrosodimethylamine	3.36	74	684463	61.85	ug/ml	100
4) Pyridine	3.46	79	982769	60.36	ug/ml	100
6) Benzaldehyde	9.19	77	239530	65.18	ug/ml	100
7) Aniline	9.41	93	956862	50.93	ug/ml	100
8) bis(2-Chloroethyl)ether	9.52	93	643002	49.65	ug/ml	100
10) Phenol	9.36	94	888897	52.38	ug/ml	100
11) 2-Chlorophenol	9.62	128	552433	43.55	ug/ml	100
12) 1,3-Dichlorobenzene	9.91	146	558583	40.36	ug/ml	100
13) 1,4-Dichlorobenzene	10.04	146	562975	40.71	ug/ml	100
14) 1,2-Dichlorobenzene	10.31	146	513975	40.22	ug/ml	100
15) Benzyl alcohol	10.26	108	343722	43.86	ug/ml	100
16) 2,2-oxybis(1-chloropropane)	10.49	45	942182	42.73	ug/ml	100
17) 2-Methylphenol	10.45	108	489343	45.48	ug/ml	100
18) Acetophenone	10.73	105	540111	40.86	ug/ml	100
19) Hexachloroethane	10.93	117	211504	40.20	ug/ml	100
20) N-Nitroso-di-n-propylamine	10.74	70	332800	59.35	ug/ml	100
21) 4-Methylphenol	10.73	108	455575	42.06	ug/ml	100
24) Nitrobenzene	11.03	77	479327	58.34	ug/ml	100
25) Isophorone	11.47	82	869972	50.40	ug/ml	100
26) 2-Nitrophenol	11.61	139	245206	37.58	ug/ml	100
27) 2,4-Dimethylphenol	11.69	107	372694	42.71	ug/ml	100
28) Benzoic Acid	11.96	105	477392	69.22	ug/ml	100
29) bis(2-Chloroethoxy)methane	11.86	93	548409	45.80	ug/ml	100
30) 2,4-Dichlorophenol	12.05	162	298916	35.51	ug/ml	100
32) 1,2,4-Trichlorobenzene	12.21	180	338953	41.30	ug/ml	100
33) Naphthalene	12.36	128	1077175	39.11	ug/ml	100
34) 4-Chloroaniline	12.45	127	418133	36.33	ug/ml	100

(#) = qualifier out of range (m) = manual integration

Data File : H:\VOL1\GCMSDATA\L\120725\BL000632.D
 Acq On : 25 Jul 2012 13:43
 Sample : SSTD050
 Misc : SSTD050 HLM
 MS Integration Params: rteint.p
 Quant Time: Jul 25 15:39 2012

Vial: 2
 Operator: HLM
 Inst : BNA L
 Multiplr: 1.00

Instrument :
 BNA_L
 ClientSampleId :
 SSTD050

Quant Results File: BL072512.M

Quant Method : H:\VOL1\GCMSDATA\L\METHODS\BL072512.M (RTE Integrator)
 Title : SW-846 Method 8270
 Last Update : Wed Jul 25 15:39:43 2012
 Response via : Initial Calibration
 DataAcq Meth : 8270L

Manual Integrations
 APPROVED

AHPatel
 9/6/2012 6:41:31 AM

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
35) Hexachlorobutadiene	12.59	225	144442	72.62	ug/ml	100
36) Caprolactam	13.10	113	97433	33.39	ug/ml	100
37) 4-Chloro-3-methylphenol	13.34	107	263595	39.29	ug/ml	100
38) 2-Methylnaphthalene	13.62	142	574536	34.95	ug/ml	100
40) Hexachlorocyclopentadiene	13.93	237	127617	84.23	ug/ml	100
41) 2,4,6-Trichlorophenol	14.14	196	150211	53.90	ug/ml	100
42) 2,4,5-Trichlorophenol	14.21	196	167909	53.65	ug/ml	100
44) 1,1-Biphenyl	14.48	154	590097	41.29	ug/ml	100
45) 2-Chloronaphthalene	14.51	162	480031	41.24	ug/ml	100
46) 2-Nitroaniline	14.70	65	133070	57.82	ug/ml	100
47) Acenaphthylene	15.29	152	670504	37.50	ug/ml	100
48) Dimethylphthalate	15.05	163	421885	34.63	ug/ml	100
49) 2,6-Dinitrotoluene	15.16	165	108665	32.62	ug/ml	100
50) Acenaphthene	15.61	153	403699	37.92	ug/ml	100
51) 3-Nitroaniline	15.47	138	103163	31.02	ug/ml	100
52) 2,4-Dinitrophenol	15.66	184	53337	34.88	ug/ml	100
53) Dibenzofuran	15.94	168	513646	38.13	ug/ml	100
54) 2,4-Dinitrotoluene	15.91	165	117187	29.34	ug/ml	100
55) 4-Nitrophenol	15.79	139	80734	26.85	ug/ml	100
56) 2,3,4,6-Tetrachlorophenol	16.17	232	86590	69.07	ug/ml	100
57) Fluorene	16.57	166	361605	35.56	ug/ml	100
58) 4-Chlorophenyl-phenylether	16.58	204	166689	59.80	ug/ml	100
59) Diethylphthalate	16.38	149	332637	28.94	ug/ml	100
60) 4-Nitroaniline	16.60	138	68410	21.45	ug/ml	100
63) 1,2,4,5-Tetrachlorobenzene	13.94	216	229820	85.71	ug/ml	100
64) 4,6-Dinitro-2-methylphenol	16.67	198	59103	40.05	ug/ml	100
65) n-Nitrosodiphenylamine	16.79	169	295310	37.35	ug/ml	100
66) 1,2-Diphenylhydrazine	16.87	77	452565	56.86	ug/ml	100
67) 4-Bromophenyl-phenylether	17.49	248	84613	76.24	ug/ml	100
68) Hexachlorobenzene	17.62	284	89774	66.11	ug/ml	100
69) Atrazine	17.82	200	53712	39.71	ug/ml	100
70) Pentachlorophenol	17.98	266	35305	50.63	ug/ml	100
72) Benzidine	20.34	184	17686	15.23	ug/ml	100
73) Phenanthrene	18.36	178	365311	33.28	ug/ml	100
74) Anthracene	18.45	178	374387	32.87	ug/ml	100
75) Carbazole	18.72	167	234846	23.54	ug/ml	100
76) Di-n-butylphthalate	19.30	149	396788	22.66	ug/ml	100
77) Fluoranthene	20.16	202	299364	37.63	ug/ml	100
79) Pyrene	20.45	202	298721	27.23	ug/ml	100
81) Butylbenzylphthalate	21.22	149	157836	14.88	ug/ml	100
82) 3,3-Dichlorobenzidine	21.85	252	58953	30.78	ug/ml	100
83) Benzo[a]anthracene	21.90	228	283574	37.36	ug/ml	100
84) Chrysene	21.94	228	272520	37.81	ug/ml	100
85) bis(2-Ethylhexyl)phthalate	21.87	149	259596	16.53	ug/ml	100
87) Di-n-octylphthalate	22.69	149	479606	16.30	ug/ml	100
88) Benzo[b]fluoranthene	23.39	252	343696	41.75	ug/ml	100
89) Benzo[k]fluoranthene	23.43	252	282182	38.90	ug/ml	100
90) Benzo[a]pyrene	23.94	252	306885	40.93	ug/ml	100
91) Indeno[1,2,3-cd]pyrene	26.35	276	342770	41.69	ug/ml	100

(#) = qualifier out of range (m) = manual integration
 BL000632.D BL072512.M Wed Jul 25 15:40:51 2012

Data File : H:\VOL1\GCMSDATA\L\120725\BL000632.D

Vial: 2

Acq On : 25 Jul 2012 13:43

Operator: HLM

Sample : SSTD050

Inst : BNA L

Misc : SSTD050 HLM

Multiplr: 1.00

MS Integration Params: rteint.p

Quant Time: Jul 25 15:39 2012

Quant Results File: BL072512.F

Instrument :

BNA_L

ClientSampleId :

SSTD050

Manual Integrations
APPROVEDAHPatel
9/6/2012 6:41:31 AM

Quant Method : H:\VOL1\GCMSDATA\L\METHODS\BL072512.M (RTE Integrator)
Title : SW-846 Method 8270
Last Update : Wed Jul 25 15:39:43 2012
Response via : Initial Calibration
DataAcq Meth : 8270L

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
92) Dibenz[a,h]anthracene	26.38	278	282211	41.96	ug/ml	100
93) Benzo[g,h,i]perylene	27.07	276	304436	43.01	ug/ml	100