

#3





N-Nitrosodimethylamine
Concen: 112.60 ug/ml m
RT: 3.628 min Scan# 53
Delta R.T. 0.267 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Instrument :
BNA_L
LabSampleId :
SSTD160

Tgt Ion: 74 Resp: 2001470
Ion Ratio Lower Upper
74 100
43 21.7 2.3 42.3

Manual Integrations
APPROVED




AHPatel
9/6/2012 6:41:42 AM




#4

Pyridine
Concen: 116.21 ug/ml m
RT: 3.719 min Scan# 61
Delta R.T. 0.256 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Tgt Ion: 79 Resp: 3169466
Ion Ratio Lower Upper
79 100
52 67.6 47.6 87.6





#7

Aniline

Concen: 120.17 ug/ml

RT: 9.444 min Scan# 56

Delta R.T. 0.030 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L

LabSampleId :

SSTD160

Tgt Ion: 93 Resp: 2934189

Ion Ratio Lower Upper





93 100

66 41.4 30.6 46.0

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:42 AM



#8

bis(2-Chloroethyl)ether

Concen: 111.88 ug/ml

RT: 9.568 min Scan# 579

Delta R.T. 0.041 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14




Tgt Ion: 93 Resp: 1919494


Ion Ratio Lower Upper

93 100

63 80.1 58.9 98.9

95 32.4 11.9 51.9





#10

Phenol

Concen: 106.44 ug/ml

RT: 9.421 min Scan# 56

Delta R.T. 0.053 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L

LabSampleId :

SSTD160

Tgt Ion: 94 Resp: 2377505

Ion Ratio Lower Upper

94 100





65 28.9 1.6 41.6

66 39.4 4.3 44.3

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:42 AM



#11

2-Chlorophenol

Concen: 106.56 ug/ml

RT: 9.670 min Scan# 588

Delta R.T. 0.041 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14




Tgt Ion:128 Resp: 1509016


Ion Ratio Lower Upper

128 100

64 56.1 37.3 77.3

130 32.3 12.4 52.4





#12

1,3-Dichlorobenzene

Concen: 109.78 ug/ml

RT: 9.929 min Scan# 611

Delta R.T. 0.019 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L





LabSampleId :

SSTD160

Tgt Ion	Ratio	Lower	Upper
146	100		
148	63.6	44.0	84.0
113	12.9	0.0	32.5

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:42 AM



#13

1,4-Dichlorobenzene

Concen: 107.76 ug/ml




RT: 10.065 min Scan# 623


Delta R.T. 0.030 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Tgt Ion	Ratio	Lower	Upper
146	100		
148	64.2	44.6	84.6
113	12.5	0.0	32.8





#14

1,2-Dichlorobenzene

Concen: 106.46 ug/ml

RT: 10.347 min Scan# 64

Delta R.T. 0.030 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L

LabSampleId :

SSTD160

Tgt Ion:146 Resp: 1377845

Ion Ratio Lower Upper

146 100





148 65.1 44.0 84.0

113 13.0 0.0 33.6

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:42 AM



#15

Benzyl alcohol

Concen: 110.16 ug/ml

RT: 10.291 min Scan# 643

Delta R.T. 0.030 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14




Tgt Ion:108 Resp: 977995


Ion Ratio Lower Upper

108 100

79 116.7 94.5 134.5

77 67.9 43.7 83.7





#16

2,2-oxybis(1-chloropropane)

Concen: 113.69 ug/ml

RT: 10.528 min Scan# 66

Delta R.T. 0.030 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L

LabSampleId :

SSTD160

Tgt Ion: 45 Resp: 2886043

Ion Ratio Lower Upper

45 100





77 11.1 0.0 30.0

79 10.0 0.0 28.9

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:42 AM



#17

2-Methylphenol

Concen: 107.29 ug/ml

RT: 10.483 min Scan# 660

Delta R.T. 0.030 min

Lab File: BL000641.D




Acq: 25 Jul 2012 19:14


Tgt Ion: 108 Resp: 1393529

Ion Ratio Lower Upper

108 100

107 87.8 67.8 107.8





#19

Hexachloroethane

Concen: 113.18 ug/ml

RT: 10.946 min Scan# 700

Delta R.T. 0.019 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L





LabSampleId :

SSTD160

Tgt Ion	Ratio	Resp Lower	Resp Upper
117	100		
200	0.0	0.0	20.0
199	46.0	21.2	61.2

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:42 AM



#20

N-Nitroso-di-n-propylamine

Concen: 119.74 ug/ml




RT: 10.810 min Scan# 689


Delta R.T. 0.064 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Tgt Ion	Ratio	Resp Lower	Resp Upper
70	100		
42	49.4	32.8	72.8
130	23.8	3.0	43.0





#21

4-Methylphenol

Concen: 116.16 ug/ml

RT: 10.776 min Scan# 68

Delta R.T. 0.041 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L

LabSampleId :

SSTD160

Tgt Ion:108 Resp: 1370833

Ion Ratio Lower Upper





108 100

107 105.3 86.9 126.9

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:42 AM



#24

Nitrobenzene

Concen: 109.17 ug/ml

RT: 11.070 min Scan# 712

Delta R.T. 0.030 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14




Tgt Ion: 77 Resp: 1504902


Ion Ratio Lower Upper

77 100

123 52.2 26.9 66.9

65 12.0 0.0 32.1





#25

Isophorone

Concen: 113.21 ug/ml

RT: 11.522 min Scan# 75

Delta R.T. 0.041 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L

LabSampleId :

SSTD160

Tgt Ion: 82 Resp: 2982708

Ion Ratio Lower Upper





82 100

138 15.7 0.0 35.5

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:42 AM



#26

2-Nitrophenol

Concen: 113.58 ug/ml

RT: 11.634 min Scan# 762

Delta R.T. 0.019 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14




Tgt Ion: 139 Resp: 807316


Ion Ratio Lower Upper

139 100

65 39.3 20.8 60.8

109 13.8 0.0 33.5









#27
2,4-Dimethylphenol
Concen: 112.15 ug/ml
RT: 11.725 min Scan# 77
Delta R.T. 0.030 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Instrument :
BNA_L
LabSampleId :
SSTD160

Tgt Ion:107 Resp: 1151087
Ion Ratio Lower Upper
107 100
121 54.6 33.2 73.2
122 96.6 75.9 115.9




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:42 AM



#28
Benzoic Acid
Concen: 134.39 ug/ml
RT: 12.041 min Scan# 798
Delta R.T. 0.064 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Tgt Ion:105 Resp: 990624
Ion Ratio Lower Upper
105 100
77 69.6 52.0 92.0
122 83.9 61.6 101.6









#29
 bis(2-Chloroethoxy)methane
 Concen: 106.77 ug/ml
 RT: 11.883 min Scan# 78
 Delta R.T. 0.019 min
 Lab File: BL000641.D
 Acq: 25 Jul 2012 19:14

Instrument :
 BNA_L
 LabSampleId :
 SSTD160

Tgt Ion	Ratio	Lower	Upper
93	100		
95	32.5	11.6	51.6
123	10.9	0.0	30.8




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:42 AM



#30
 2,4-Dichlorophenol
 Concen: 109.91 ug/ml
 RT: 12.086 min Scan# 802
 Delta R.T. 0.030 min
 Lab File: BL000641.D
 Acq: 25 Jul 2012 19:14

Tgt Ion	Ratio	Lower	Upper
162	100		
164	64.1	43.5	83.5
98	34.0	19.5	59.5





#32

1,2,4-Trichlorobenzene

Concen: 105.02 ug/ml

RT: 12.233 min Scan# 81

Delta R.T. 0.019 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L

LabSampleId :

SSTD160

Tgt Ion:180 Resp: 992368
Ion Ratio Lower Upper

180 100





182 95.5 75.6 115.6

145 27.8 8.3 48.3

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:42 AM



#33

Naphthalene

Concen: 101.86 ug/ml

RT: 12.368 min Scan# 827

Delta R.T. 0.008 min

Lab File: BL000641.D




Acq: 25 Jul 2012 19:14


Tgt Ion:128 Resp: 3069158
Ion Ratio Lower Upper

128 100

129 11.2 0.0 30.7

127 13.7 0.0 33.1









#34
4-Chloroaniline
Concen: 116.57 ug/ml
RT: 12.459 min Scan# 834
Delta R.T. 0.008 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Instrument :
BNA_L
LabSampleId :
SSTD160

Tgt Ion:127 Resp: 1448499
Ion Ratio Lower Upper
127 100
129 32.0 12.1 52.1




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:42 AM



#35
Hexachlorobutadiene
Concen: 104.07 ug/ml
RT: 12.606 min Scan# 848
Delta R.T. 0.008 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Tgt Ion:225 Resp: 416221
Ion Ratio Lower Upper
225 100
223 63.4 41.7 81.7
227 65.0 45.2 85.2









#37
 4-Chloro-3-methylphenol
 Concen: 110.74 ug/ml
 RT: 13.351 min Scan# 914
 Delta R.T. 0.008 min
 Lab File: BL000641.D
 Acq: 25 Jul 2012 19:14

Instrument :
 BNA_L
 LabSampleId :
 SSTD160

Tgt Ion	Ratio	Resp	Lower	Upper
107	100	871412		
144	29.4	9.7	49.7	
142	91.3	71.3	111.3	




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:42 AM



#38
 2-Methylnaphthalene
 Concen: 103.67 ug/ml
 RT: 13.633 min Scan# 939
 Delta R.T. 0.008 min
 Lab File: BL000641.D
 Acq: 25 Jul 2012 19:14

Tgt Ion	Ratio	Resp	Lower	Upper
142	100	1723293		
141	85.7	64.5	104.5	









#40
Hexachlorocyclopentadiene
Concen: 126.61 ug/ml
RT: 13.938 min Scan# 966
Delta R.T. 0.008 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Instrument :
BNA_L
LabSampleId :
SSTD160

Tgt Ion	Ratio	Resp Lower	Upper
237	100		
235	62.5	42.1	82.1
272	14.3	0.0	35.7




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:42 AM



#41
2,4,6-Trichlorophenol
Concen: 109.16 ug/ml
RT: 14.164 min Scan# 986
Delta R.T. 0.019 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Tgt Ion	Ratio	Resp Lower	Upper
196	100		
198	95.5	73.1	113.1
200	31.7	9.7	49.7





#42

2,4,5-Trichlorophenol

Concen: 111.72 ug/ml

RT: 14.220 min Scan# 991

Delta R.T. 0.008 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L





LabSampleId :

SSTD160

Tgt Ion:	196	Resp:	560838
Ion Ratio	Lower	Upper	
196	100		
198	93.2	74.5	114.5
200	30.5	10.8	50.8

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:42 AM



#45

2-Chloronaphthalene

Concen: 122.63 ug/ml




RT: 14.536 min Scan# 1019


Delta R.T. 0.007 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Tgt Ion:	162	Resp:	1743123
Ion Ratio	Lower	Upper	
162	100		
164	32.7	12.8	52.8
127	35.0	10.9	50.9









#46
2-Nitroaniline
Concen: 114.82 ug/ml
RT: 14.728 min Scan# 1036
Delta R.T. 0.019 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Instrument :
BNA_L
LabSampleId :
SSTD160

Tgt Ion	Ratio	Lower	Upper
65	100		
92	71.4	49.2	89.2
138	129.1	102.4	142.4




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:42 AM



#47
Acenaphthylene
Concen: 104.74 ug/ml
RT: 15.316 min Scan# 1088
Delta R.T. 0.019 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.1	0.0	39.9
153	12.9	0.0	32.6









#48
Dimethylphthalate
Concen: 112.21 ug/ml
RT: 15.101 min Scan# 1069
Delta R.T. 0.030 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Instrument :
BNA_L
LabSampleId :
SSTD160

Tgt Ion	Ratio	Resp	Lower	Upper
163	100	1583798		
194	5.8	0.0	25.9	
164	10.1	0.0	29.6	




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:42 AM



#49
2,6-Dinitrotoluene
Concen: 114.08 ug/ml
RT: 15.191 min Scan# 1077
Delta R.T. 0.030 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	421467		
89	39.0	19.2	59.2	
121	16.4	0.0	36.2	





#50

Acenaphthene

Concen: 105.79 ug/ml

RT: 15.632 min Scan# 111

Delta R.T. 0.008 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L

LabSampleId :

SSTD160

Tgt Ion:153 Resp: 1298708

Ion Ratio Lower Upper

153 100





152 46.0 25.1 65.1

154 95.3 75.9 115.9

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:42 AM



#51

3-Nitroaniline

Concen: 141.26 ug/ml

RT: 15.508 min Scan# 1105

Delta R.T. 0.030 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14




Tgt Ion:138 Resp: 466505


Ion Ratio Lower Upper

138 100

108 9.2 0.0 30.0

92 101.5 81.9 121.9





#52

2,4-Dinitrophenol

Concen: 145.60 ug/ml

RT: 15.688 min Scan# 11

Delta R.T. 0.019 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L





LabSampleId :

SSTD160

Tgt Ion:	184	Resp:	271286
Ion Ratio	Lower	Upper	
184	100		
63	42.1	11.7	51.7
154	28.0	6.0	46.0

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:42 AM



#53

Dibenzofuran

Concen: 102.45 ug/ml




RT: 15.959 min Scan# 1145


Delta R.T. 0.019 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Tgt Ion:	168	Resp:	1630737
Ion Ratio	Lower	Upper	
168	100		
139	28.5	7.3	47.3





#54

2,4-Dinitrotoluene

Concen: 110.97 ug/ml

RT: 15.948 min Scan# 11

Delta R.T. 0.030 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L





LabSampleId :

SSTD160

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	441084		
63	46.5	13.1	53.1	
182	5.7	0.0	26.1	

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:42 AM



#55

4-Nitrophenol

Concen: 126.61 ug/ml




RT: 15.824 min Scan# 1133


Delta R.T. 0.030 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Tgt Ion	Ratio	Resp	Lower	Upper
139	100	356035		
65	70.1	44.1	84.1	
109	29.4	6.8	46.8	









#56
 2,3,4,6-Tetrachlorophenol
 Concen: 117.58 ug/ml
 RT: 16.185 min Scan# 1164
 Delta R.T. 0.008 min
 Lab File: BL000641.D
 Acq: 25 Jul 2012 19:14

Instrument :
 BNA_L
 LabSampleId :
 SSTD160

Tgt Ion	Ratio	Resp Lower	Resp Upper
232	100		
230	79.5	64.6	96.8
131	37.1	29.2	43.8
166	26.2	20.6	30.8




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:42 AM



#57
 Fluorene
 Concen: 99.58 ug/ml
 RT: 16.592 min Scan# 1201
 Delta R.T. 0.008 min
 Lab File: BL000641.D
 Acq: 25 Jul 2012 19:14

Tgt Ion	Ratio	Resp Lower	Resp Upper
166	100		
165	93.2	72.3	112.3
167	13.8	0.0	33.1









#58
4-Chlorophenyl-phenylether
Concen: 95.11 ug/ml
RT: 16.592 min Scan# 1201
Delta R.T. 0.008 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Instrument :
BNA_L
LabSampleId :
SSTD160

Tgt Ion	Ratio	Resp	Lower	Upper
204	100	486856		
206	32.7	12.0	52.0	
141	54.2	26.4	66.4	




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:42 AM



#59
Diethylphthalate
Concen: 112.16 ug/ml
RT: 16.411 min Scan# 1185
Delta R.T. 0.030 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	1287078		
177	22.4	1.3	41.3	
150	11.4	0.0	31.6	









#60
4-Nitroaniline
Concen: 169.86 ug/ml
RT: 16.659 min Scan# 1207
Delta R.T. 0.041 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Instrument :
BNA_L
LabSampleId :
SSTD160

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	396897		
92	41.0	16.5	56.5	
108	28.3	8.1	48.1	




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:42 AM



#64
4,6-Dinitro-2-methylphenol
Concen: 127.66 ug/ml
RT: 16.704 min Scan# 1211
Delta R.T. 0.030 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Tgt Ion:198 Resp: 290684









#65
 n-Nitrosodiphenylamine
 Concen: 103.81 ug/ml
 RT: 16.829 min Scan# 1221
 Delta R.T. 0.019 min
 Lab File: BL000641.D
 Acq: 25 Jul 2012 19:14

Instrument :
 BNA_L
 LabSampleId :
 SSTD160

Tgt Ion	Ratio	Lower	Upper
169	100		
168	63.2	40.9	80.9
167	33.4	12.5	52.5




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:42 AM



#66
 1,2-Diphenylhydrazine
 Concen: 102.36 ug/ml
 RT: 16.896 min Scan# 1228
 Delta R.T. 0.019 min
 Lab File: BL000641.D
 Acq: 25 Jul 2012 19:14

Tgt Ion	Ratio	Lower	Upper
77	100		
182	21.3	0.0	39.4
105	14.9	0.0	34.5





#67

4-Bromophenyl-phenylether

Concen: 102.83 ug/ml

RT: 17.517 min Scan# 12

Delta R.T. 0.019 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L





LabSampleId :

SSTD160

Tgt Ion	Ratio	Resp	Lower	Upper
248	100	304527		
250	98.7	78.2	118.2	
141	47.8	34.7	74.7	

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:42 AM



#68

Hexachlorobenzene

Concen: 99.31 ug/ml




RT: 17.642 min Scan# 1294


Delta R.T. 0.019 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Tgt Ion	Ratio	Resp	Lower	Upper
284	100	313563		
142	31.3	14.3	54.3	
249	27.8	6.7	46.7	









#70
 Pentachlorophenol
 Concen: 131.66 ug/ml
 RT: 17.992 min Scan# 1325
 Delta R.T. 0.008 min
 Lab File: BL000641.D
 Acq: 25 Jul 2012 19:14

Instrument :
 BNA_L
 LabSampleId :
 SSTD160

Tgt Ion	Ratio	Resp	Lower	Upper
266	100	217002		
264	65.0		44.5	84.5
268	65.3		45.0	85.0




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:42 AM



#72
 Benzidine
 Concen: 612.82 ug/ml
 RT: 20.340 min Scan# 1533
 Delta R.T. 0.008 min
 Lab File: BL000641.D
 Acq: 25 Jul 2012 19:14

Tgt Ion:184 Resp: 642303









#73
Phenanthrene
Concen: 103.52 ug/ml
RT: 18.376 min Scan# 1359
Delta R.T. 0.008 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Instrument :
BNA_L
LabSampleId :
SSTD160

Tgt Ion:178 Resp: 1397368
Ion Ratio Lower Upper
178 100
179 15.3 0.0 35.3
176 18.5 0.0 38.6




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:42 AM



#74
Anthracene
Concen: 104.01 ug/ml
RT: 18.466 min Scan# 1367
Delta R.T. 0.008 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Tgt Ion:178 Resp: 1440511
Ion Ratio Lower Upper
178 100
179 15.6 0.0 34.9
176 18.4 0.0 37.0





#75

Carbazole

Concen: 109.10 ug/ml m

RT: 18.737 min Scan# 1390

Delta R.T. 0.008 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L

LabSampleId :

SSTD160

Tgt Ion:167 Resp: 1303325

Ion Ratio Lower Upper





167 100

167 90.8 80.0 120.0

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:42 AM



#76

Di-n-butylphthalate

Concen: 107.17 ug/ml

RT: 19.302 min Scan# 1441

Delta R.T. 0.008 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14




Tgt Ion:149 Resp: 1645944


Ion Ratio Lower Upper

149 100

150 9.4 0.0 29.4

104 3.7 0.0 23.7





#77

Fluoranthene

Concen: 98.78 ug/ml

RT: 20.171 min Scan# 1517

Delta R.T. 0.008 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L

LabSampleId :

SSTD160

Tgt Ion:202 Resp: 1138464

Ion Ratio Lower Upper





202 100

101 29.5 5.5 45.5

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:42 AM



#79

Pyrene

Concen: 104.69 ug/ml

RT: 20.465 min Scan# 1544

Delta R.T. 0.007 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14




Tgt Ion:202 Resp: 1213673


Ion Ratio Lower Upper

202 100

101 35.8 11.7 51.7

100 29.0 6.0 46.0









#81
Butylbenzylphthalate
Concen: 109.34 ug/ml
RT: 21.232 min Scan# 1612
Delta R.T. 0.008 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Instrument :
BNA_L
LabSampleId :
SSTD160

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
91	67.9	51.8	91.8
206	17.7	0.0	36.6




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:42 AM



#82
3,3-Dichlorobenzidine
Concen: 162.15 ug/ml
RT: 21.865 min Scan# 1668
Delta R.T. 0.008 min
Lab File: BL000641.D
Acq: 25 Jul 2012 19:14

Tgt Ion	Ratio	Resp Lower	Resp Upper
252	100		
254	64.4	47.7	87.7
126	15.5	0.0	35.4





#83

Benzo[a]anthracene

Concen: 105.21 ug/ml

RT: 21.910 min Scan# 16

Delta R.T. 0.019 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L





LabSampleId :

SSTD160

Tgt	Ion	Ratio	Resp	Lower	Upper
228	100		989324		
229	19.4	0.0	39.4		
226	25.1	5.2	45.2		

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:42 AM



#84

Chrysene

Concen: 89.77 ug/ml




RT: 21.955 min Scan# 1676


Delta R.T. 0.008 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Tgt	Ion	Ratio	Resp	Lower	Upper
228	100		828021		
226	28.0	6.9	46.9		
229	19.3	0.0	39.1		









#85
 bis(2-Ethylhexyl)phthalate
 Concen: 108.06 ug/ml
 RT: 21.876 min Scan# 1669
 Delta R.T. 0.008 min
 Lab File: BL000641.D
 Acq: 25 Jul 2012 19:14

Instrument :
 BNA_L
 LabSampleId :
 SSTD160

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
167	31.5	9.1	49.1
279	4.8	0.0	24.0




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:42 AM



#87
 Di-n-octylphthalate
 Concen: 118.99 ug/ml
 RT: 22.689 min Scan# 1741
 Delta R.T. -0.004 min
 Lab File: BL000641.D
 Acq: 25 Jul 2012 19:14

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
150	10.1	0.0	29.4





#88

Benzo[b]fluoranthene

Concen: 108.41 ug/ml

RT: 23.401 min Scan# 1803

Delta R.T. 0.008 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L





LabSampleId :

SSTD160

Tgt Ion:	252	Resp:	1000245
Ion Ratio	Lower	Upper	
252	100		
253	21.7	1.5	41.5
125	26.7	1.4	41.4

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:42 AM



#89

Benzo[k]fluoranthene

Concen: 111.05 ug/ml




RT: 23.457 min Scan# 1809


Delta R.T. 0.030 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Tgt Ion:	252	Resp:	900624
Ion Ratio	Lower	Upper	
252	100		
253	21.6	2.6	42.6
125	24.5	7.4	47.4





#90

Benzo[a]pyrene

Concen: 111.12 ug/ml

RT: 23.976 min Scan# 18

Delta R.T. 0.030 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L





LabSampleId :

SSTD160

Tgt Ion:	252	Resp:	942842
Ion Ratio	Lower	Upper	
252	100		
253	21.9	1.8	41.8
125	25.5	8.0	48.0

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:42 AM



#91

Indeno[1,2,3-cd]pyrene

Concen: 100.67 ug/ml




RT: 26.382 min Scan# 2068


Delta R.T. 0.030 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Tgt Ion:	276	Resp:	917763
Ion Ratio	Lower	Upper	
276	100		
138	57.1	37.7	77.7





#92

Dibenz[a,h]anthracene

Concen: 66.58 ug/ml

RT: 26.415 min Scan# 20

Delta R.T. 0.030 min

Lab File: BL000641.D

Acq: 25 Jul 2012 19:14

Instrument :

BNA_L

LabSampleId :

SSTD160

Tgt Ion:278 Resp: 487716
Ion Ratio Lower Upper

278 100





139 41.8 18.8 58.8

279 22.9 4.5 44.5

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:42 AM



#93

Benzo[g,h,i]perylene

Concen: 106.45 ug/ml

RT: 27.116 min Scan# 2133

Delta R.T. 0.053 min

Lab File: BL000641.D




Acq: 25 Jul 2012 19:14

Tgt Ion:276 Resp: 852163
Ion Ratio Lower Upper

276 100

138 55.6 36.8 76.8

277 22.9 4.5 44.5



Data Path : S:\HPCHEM1\BNA_L\data\BL072512\
 Data File : BL000641.D
 Acq On : 25 Jul 2012 19:14
 Operator : HLM
 Sample : SSTD160
 Misc : SSTD160 HLM
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_L
 LabSampleId :
 SSTD160

Manual Integrations
 APPROVED

AHPatel
 9/6/2012 6:41:42 AM

Quant Time: Sep 04 08:38:05 2012
 Quant Method : H:\VOL1\GCMSDATA\L\METHODS\BL072512.M
 Quant Title : SW-846 Method 8270
 QLast Update : Thu Jul 26 10:46:15 2012
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	10.031	152	420766	40.00	ug/ml	0.02
22) Naphthalene-d8	12.323	136	1324221	40.00	ug/ml	0.00
39) Acenaphthene-d10	15.564	164	485652	40.00	ug/ml	0.00
62) Phenanthrene-d10	18.330	188	591672	40.00	ug/ml	0.00
78) Chrysene-d12	21.921	240	382768	40.00	ug/ml	0.00
86) Perylene-d12	24.055	264	331035	40.00	ug/ml	0.00

System Monitoring Compounds

5) 2-Fluorophenol	7.592	112	1892517	109.48	ug/ml	0.08
Spiked Amount	200.000	Range	20 - 110	Recovery	=	54.74%
9) Phenol-d5	9.387	99	2117269	107.77	ug/ml	0.04
Spiked Amount	200.000	Range	10 - 115	Recovery	=	53.88%
23) Nitrobenzene-d5	11.036	82	1482818	109.76	ug/ml	0.03
Spiked Amount	100.000	Range	40 - 110	Recovery	=	109.76%
43) 2-Fluorobiphenyl	14.322	172	1632640	104.43	ug/ml	0.02
Spiked Amount	100.000	Range	50 - 110	Recovery	=	104.43%
61) 2,4,6-Tribromophenol	17.043	330	147454	107.99	ug/ml	0.02
Spiked Amount	200.000	Range	40 - 125	Recovery	=	53.99%
80) Terphenyl-d14	20.657	244	464538	77.44	ug/ml	0.00
Spiked Amount	100.000	Range	50 - 135	Recovery	=	77.44%

Target Compounds

						Qvalue
3) N-Nitrosodimethylamine	3.628	74	2001470m	112.60	ug/ml	
4) Pyridine	3.719	79	3169466m	116.21	ug/ml	
7) Aniline	9.444	93	2934189	120.17	ug/ml	95
8) bis(2-Chloroethyl)ether	9.568	93	1919494	111.88	ug/ml	99
10) Phenol	9.421	94	2377505	106.44	ug/ml	77
11) 2-Chlorophenol	9.670	128	1509016	106.56	ug/ml	99
12) 1,3-Dichlorobenzene	9.929	146	1572971	109.78	ug/ml	99
13) 1,4-Dichlorobenzene	10.065	146	1532736	107.76	ug/ml	99
14) 1,2-Dichlorobenzene	10.347	146	1377845	106.46	ug/ml	99
15) Benzyl alcohol	10.291	108	977995	110.16	ug/ml	97
16) 2,2-oxybis(1-chloropro...	10.528	45	2886043	113.69	ug/ml	97
17) 2-Methylphenol	10.483	108	1393529	107.29	ug/ml	100
19) Hexachloroethane	10.946	117	620711	113.18	ug/ml	92
20) N-Nitroso-di-n-propyla...	10.810	70	1128464	119.74	ug/ml	96
21) 4-Methylphenol	10.776	108	1370833	116.16	ug/ml	98
24) Nitrobenzene	11.070	77	1504902	109.17	ug/ml	94
25) Isophorone	11.522	82	2982708	113.21	ug/ml	99
26) 2-Nitrophenol	11.634	139	807316	113.58	ug/ml	98
27) 2,4-Dimethylphenol	11.725	107	1151087	112.15	ug/ml	99
28) Benzoic Acid	12.041	105	990624	134.39	ug/ml	97
29) bis(2-Chloroethoxy)met...	11.883	93	1715007	106.77	ug/ml	99
30) 2,4-Dichlorophenol	12.086	162	928849	109.91	ug/ml	96
32) 1,2,4-Trichlorobenzene	12.233	180	992368	105.02	ug/ml	100
33) Naphthalene	12.368	128	3069158	101.86	ug/ml	99
34) 4-Chloroaniline	12.459	127	1448499	116.57	ug/ml	100
35) Hexachlorobutadiene	12.606	225	416221	104.07	ug/ml	99
37) 4-Chloro-3-methylphenol	13.351	107	871412	110.74	ug/ml	100

Data Path : S:\HPCHEM1\BNA_L\data\BL072512\
 Data File : BL000641.D
 Acq On : 25 Jul 2012 19:14
 Operator : HLM
 Sample : SSTD160
 Misc : SSTD160 HLM
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_L
 LabSampleId :
 SSTD160

Manual Integrations
 APPROVED

AHPatel
 9/6/2012 6:41:42 AM

Quant Time: Sep 04 08:38:05 2012
 Quant Method : H:\VOL1\GCMSDATA\L\METHODS\BL072512.M
 Quant Title : SW-846 Method 8270
 QLast Update : Thu Jul 26 10:46:15 2012
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
38) 2-Methylnaphthalene	13.633	142	1723293	103.67	ug/ml	99
40) Hexachlorocyclopentadiene	13.938	237	430579	126.61	ug/ml	99
41) 2,4,6-Trichlorophenol	14.164	196	507494	109.16	ug/ml	97
42) 2,4,5-Trichlorophenol	14.220	196	560838	111.72	ug/ml	99
45) 2-Chloronaphthalene	14.536	162	1743123	122.63	ug/ml	96
46) 2-Nitroaniline	14.728	65	500864	114.82	ug/ml	95
47) Acenaphthylene	15.316	152	2138982	104.74	ug/ml	100
48) Dimethylphthalate	15.101	163	1583798	112.21	ug/ml	99
49) 2,6-Dinitrotoluene	15.191	165	421467	114.08	ug/ml	100
50) Acenaphthene	15.632	153	1298708	105.79	ug/ml	99
51) 3-Nitroaniline	15.508	138	466505	141.26	ug/ml	99
52) 2,4-Dinitrophenol	15.688	184	271286	145.60	ug/ml	88
53) Dibenzofuran	15.959	168	1630737	102.45	ug/ml	98
54) 2,4-Dinitrotoluene	15.948	165	441084	110.97	ug/ml	80
55) 4-Nitrophenol	15.824	139	356035	126.61	ug/ml	93
56) 2,3,4,6-Tetrachlorophenol	16.185	232	348152	117.58	ug/ml	99
57) Fluorene	16.592	166	1147452	99.58	ug/ml	99
58) 4-Chlorophenyl-phenyle...	16.592	204	486856	95.11	ug/ml	93
59) Diethylphthalate	16.411	149	1287078	112.16	ug/ml	98
60) 4-Nitroaniline	16.659	138	396897	169.86	ug/ml	95
64) 4,6-Dinitro-2-methylph...	16.704	198	290684	127.66	ug/ml	100
65) n-Nitrosodiphenylamine	16.829	169	1065810	103.81	ug/ml	98
66) 1,2-Diphenylhydrazine	16.896	77	1593513	102.36	ug/ml	97
67) 4-Bromophenyl-phenylether	17.517	248	304527	102.83	ug/ml	96
68) Hexachlorobenzene	17.642	284	313563	99.31	ug/ml	96
70) Pentachlorophenol	17.992	266	217002	131.66	ug/ml	100
72) Benzidine	20.340	184	642303	612.82	ug/ml	100
73) Phenanthrene	18.376	178	1397368	103.52	ug/ml	100
74) Anthracene	18.466	178	1440511	104.01	ug/ml	97
75) Carbazole	18.737	167	1303325m	109.10	ug/ml	
76) Di-n-butylphthalate	19.302	149	1645944	107.17	ug/ml	100
77) Fluoranthene	20.171	202	1138464	98.78	ug/ml	92
79) Pyrene	20.465	202	1213673	104.69	ug/ml	93
81) Butylbenzylphthalate	21.232	149	624588	109.34	ug/ml	96
82) 3,3-Dichlorobenzidine	21.865	252	304035	162.15	ug/ml	97
83) Benzo[a]anthracene	21.910	228	989324	105.21	ug/ml	100
84) Chrysene	21.955	228	828021	89.77	ug/ml	99
85) bis(2-Ethylhexyl)phtha...	21.876	149	887796	108.06	ug/ml	96
87) Di-n-octylphthalate	22.689	149	1595660	118.99	ug/ml	98
88) Benzo[b]fluoranthene	23.401	252	1000245	108.41	ug/ml	94
89) Benzo[k]fluoranthene	23.457	252	900624	111.05	ug/ml	96
90) Benzo[a]pyrene	23.976	252	942842	111.12	ug/ml	97
91) Indeno[1,2,3-cd]pyrene	26.382	276	917763	100.67	ug/ml	99
92) Dibenz[a,h]anthracene	26.415	278	487716	66.58	ug/ml	96
93) Benzo[g,h,i]perylene	27.116	276	852163	106.45	ug/ml	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed