

#3

N-Nitrosodimethylamine

Concen: 62.43 ug/ml

RT: 3.380 min Scan# 31

Delta R.T. 0.019 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Instrument :

BNA_L

LabSampleId :

SSTD080

Tgt Ion: 74 Resp: 1153605





Ion Ratio Lower Upper

74 100

43 22.1 2.3 42.3

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:47 AM



#4

Pyridine

Concen: 63.06 ug/ml m

RT: 3.448 min Scan# 37

Delta R.T. -0.015 min

Lab File: BL000645.D


Acq: 25 Jul 2012 21:43


Tgt Ion: 79 Resp: 1787962

Ion Ratio Lower Upper

79 100

52 66.4 47.6 87.6





#6

Benzaldehyde

Concen: 47.39 ug/ml

RT: 9.195 min Scan# 546

Delta R.T. -0.004 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Instrument :

BNA_L

LabSampleId :

SSTD080

Tgt Ion: 77 Resp: 288006
Ion Ratio Lower Upper

77 100





105 90.8 70.4 110.4

106 90.3 70.4 110.4

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:47 AM



#7

Aniline

Concen: 65.23 ug/ml

RT: 9.410 min Scan# 565

Delta R.T. -0.004 min




Lab File: BL000645.D


Acq: 25 Jul 2012 21:43

Tgt Ion: 93 Resp: 1655821
Ion Ratio Lower Upper

93 100

66 39.8 30.6 46.0





#8
 bis(2-Chloroethyl)ether
 Concen: 61.78 ug/ml
 RT: 9.534 min Scan# 576
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43





Instrument :
 BNA_L
 LabSampleId :
 SST080

Tgt Ion: 93 Resp: 1101867

Ion	Ratio	Lower	Upper
93	100		
63	77.2	58.9	98.9
95	32.0	11.9	51.9

Manual Integrations
 APPROVED




AHPatel
 9/6/2012 6:41:47 AM




#10
 Phenol
 Concen: 63.47 ug/ml
 RT: 9.376 min Scan# 562
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Tgt Ion: 94 Resp: 1473769

Ion	Ratio	Lower	Upper
94	100		
65	23.3	1.6	41.6
66	26.1	4.3	44.3









#11
2-Chlorophenol
Concen: 61.43 ug/ml
RT: 9.636 min Scan# 585
Delta R.T. 0.007 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Instrument :
BNA_L
LabSampleId :
SSTD080

Tgt Ion	Ratio	Resp Lower	Resp Upper
128	100		
64	52.7	37.3	77.3
130	32.3	12.4	52.4




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:47 AM



#12
1,3-Dichlorobenzene
Concen: 60.22 ug/ml
RT: 9.907 min Scan# 609
Delta R.T. -0.004 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp Lower	Resp Upper
146	100		
148	63.3	44.0	84.0
113	12.8	0.0	32.5









#13
 1,4-Dichlorobenzene
 Concen: 60.20 ug/ml
 RT: 10.042 min Scan# 621
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Instrument :
 BNA_L
 LabSampleId :
 SSTD080

Tgt Ion	Ratio	Resp Lower	Resp Upper
146	100		
148	64.2	44.6	84.6
113	12.5	0.0	32.8




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:47 AM



#14
 1,2-Dichlorobenzene
 Concen: 61.07 ug/ml
 RT: 10.324 min Scan# 646
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp Lower	Resp Upper
146	100		
148	64.8	44.0	84.0
113	12.9	0.0	33.6









#15
Benzyl alcohol
Concen: 65.72 ug/ml
RT: 10.268 min Scan# 641
Delta R.T. 0.007 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Instrument :
BNA_L
LabSampleId :
SSTD080

Tgt Ion	Ratio	Resp Lower	Resp Upper
108	100		
79	114.1	94.5	134.5
77	65.2	43.7	83.7




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:47 AM



#16
2,2-oxybis(1-chloropropane)
Concen: 64.33 ug/ml
RT: 10.505 min Scan# 662
Delta R.T. 0.007 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp Lower	Resp Upper
45	100		
77	10.3	0.0	30.0
79	9.5	0.0	28.9









#17
 2-Methylphenol
 Concen: 63.18 ug/ml
 RT: 10.460 min Scan# 658
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Instrument :
 BNA_L
 LabSampleId :
 SST080

Tgt Ion	Ratio	Resp	Lower	Upper
108	100	853082		
107	89.8	67.8	107.8	




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:47 AM



#18
 Acetophenone
 Concen: 60.83 ug/ml
 RT: 10.742 min Scan# 683
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp	Lower	Upper
105	100	928720		
71	5.5	0.0	27.0	
51	36.0	16.6	56.6	





#19

Hexachloroethane

Concen: 61.30 ug/ml

RT: 10.934 min Scan# 700

Delta R.T. 0.007 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Instrument :

BNA_L





LabSampleId :

SSTD080

Tgt Ion	Ratio	Resp Lower	Resp Upper
117	100		
200	0.0	0.0	20.0
199	49.5	21.2	61.2

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:47 AM



#20

N-Nitroso-di-n-propylamine

Concen: 63.24 ug/ml




RT: 10.753 min Scan# 684


Delta R.T. 0.007 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp Lower	Resp Upper
70	100		
42	53.4	32.8	72.8
130	22.7	3.0	43.0









#21
 4-Methylphenol
 Concen: 61.04 ug/ml
 RT: 10.742 min Scan# 683
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Instrument :
 BNA_L
 LabSampleId :
 SST080

Tgt Ion	Ratio	Resp	Lower	Upper
108	100	748804		
107	111.1	86.9	126.9	




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:47 AM



#24
 Nitrobenzene
 Concen: 61.35 ug/ml
 RT: 11.047 min Scan# 710
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp	Lower	Upper
77	100	877869		
123	50.2	26.9	66.9	
65	11.9	0.0	32.1	





#25

Isophorone

Concen: 61.51 ug/ml

RT: 11.487 min Scan# 749

Delta R.T. 0.007 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Instrument :

BNA_L

LabSampleId :

SSTD080

Tgt Ion: 82 Resp: 1682208

Ion Ratio Lower Upper





82 100

138 15.0 0.0 35.5

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:47 AM



#26

2-Nitrophenol

Concen: 60.68 ug/ml

RT: 11.623 min Scan# 761

Delta R.T. 0.007 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43




Tgt Ion: 139 Resp: 447773


Ion Ratio Lower Upper

139 100

65 37.5 20.8 60.8

109 13.3 0.0 33.5









#27
2,4-Dimethylphenol
Concen: 62.74 ug/ml
RT: 11.702 min Scan# 768
Delta R.T. 0.007 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Instrument :
BNA_L
LabSampleId :
SSTD080

Tgt Ion	Ratio	Resp	Lower	Upper
107	100	668427		
121	54.8	33.2	73.2	
122	95.2	75.9	115.9	




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:47 AM



#28
Benzoic Acid
Concen: 145.95 ug/ml
RT: 12.041 min Scan# 798
Delta R.T. 0.064 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp	Lower	Upper
105	100	1116776		
77	70.8	52.0	92.0	
122	83.1	61.6	101.6	









#29
 bis(2-Chloroethoxy)methane
 Concen: 60.40 ug/ml
 RT: 11.871 min Scan# 783
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Instrument :
 BNA_L
 LabSampleId :
 SST080

Tgt Ion	Ratio	Resp	Lower	Upper
93	100	1007162		
95	32.2		11.6	51.6
123	11.2		0.0	30.8




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:47 AM



#30
 2,4-Dichlorophenol
 Concen: 63.22 ug/ml
 RT: 12.063 min Scan# 800
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp	Lower	Upper
162	100	554606		
164	63.8		43.5	83.5
98	37.1		19.5	59.5





#32

1,2,4-Trichlorobenzene

Concen: 60.55 ug/ml

RT: 12.210 min Scan# 81

Delta R.T. -0.004 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Instrument :

BNA_L

LabSampleId :

SSTD080

Tgt Ion:180 Resp: 593938
Ion Ratio Lower Upper

180 100





182 94.6 75.6 115.6

145 29.2 8.3 48.3

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:47 AM



#33

Naphthalene

Concen: 59.07 ug/ml

RT: 12.357 min Scan# 826

Delta R.T. -0.004 min

Lab File: BL000645.D




Acq: 25 Jul 2012 21:43


Tgt Ion:128 Resp: 1847607
Ion Ratio Lower Upper

128 100

129 11.0 0.0 30.7

127 13.3 0.0 33.1









#34
4-Chloroaniline
Concen: 64.75 ug/ml
RT: 12.447 min Scan# 834
Delta R.T. -0.004 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Instrument :
BNA_L
LabSampleId :
SSTD080

Tgt Ion	Ratio	Resp	Lower	Upper
127	100	835137		
129	32.0	12.1	52.1	




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:47 AM



#35
Hexachlorobutadiene
Concen: 59.45 ug/ml
RT: 12.594 min Scan# 847
Delta R.T. -0.004 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp	Lower	Upper
225	100	246838		
223	63.5	41.7	81.7	
227	64.5	45.2	85.2	









#36
 Caprolactam
 Concen: 65.36 ug/ml
 RT: 13.159 min Scan# 897
 Delta R.T. 0.041 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Instrument :
 BNA_L
 LabSampleId :
 SSTD080

Tgt Ion	Ratio	Resp	Lower	Upper
113	100	221585		
55	158.6		144.2	184.2
56	108.8		92.3	132.3




Manual Integrations
 APPROVED




AHPatel
 9/6/2012 6:41:47 AM



#37
 4-Chloro-3-methylphenol
 Concen: 62.73 ug/ml
 RT: 13.351 min Scan# 914
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp	Lower	Upper
107	100	512418		
144	27.6		9.7	49.7
142	87.2		71.3	111.3





#38

2-Methylnaphthalene

Concen: 60.17 ug/ml

RT: 13.622 min Scan# 938

Delta R.T. -0.004 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Tgt Ion:142 Resp: 1038303

Ion Ratio Lower Upper

142 100

141 85.0 64.5 104.5

Instrument :

BNA_L





LabSampleId :

SSTD080

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:47 AM



#40

Hexachlorocyclopentadiene

Concen: 62.23 ug/ml

RT: 13.938 min Scan# 966

Delta R.T. 0.007 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43


Tgt Ion:237 Resp: 231261


Ion Ratio Lower Upper

237 100

235 62.2 42.1 82.1

272 15.4 0.0 35.7









#41
 2,4,6-Trichlorophenol
 Concen: 60.03 ug/ml
 RT: 14.152 min Scan# 985
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Instrument :
 BNA_L
 LabSampleId :
 SST080

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	304962		
198	93.1	73.1	113.1	
200	30.3	9.7	49.7	




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:47 AM



#42
 2,4,5-Trichlorophenol
 Concen: 59.35 ug/ml
 RT: 14.220 min Scan# 991
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	325557		
198	94.0	74.5	114.5	
200	30.4	10.8	50.8	









#44
 1,1-Biphenyl
 Concen: 58.07 ug/ml
 RT: 14.491 min Scan# 1015
 Delta R.T. -0.004 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Instrument :
 BNA_L
 LabSampleId :
 SST080

Tgt Ion	Ratio	Lower	Upper
154	100		
153	40.9	19.9	59.9
76	27.4	3.3	43.3




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:47 AM



#45
 2-Chloronaphthalene
 Concen: 58.50 ug/ml
 RT: 14.525 min Scan# 1018
 Delta R.T. -0.004 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Lower	Upper
162	100		
164	32.8	12.8	52.8
127	34.3	10.9	50.9









#46
 2-Nitroaniline
 Concen: 60.38 ug/ml
 RT: 14.717 min Scan# 1035
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Instrument :
 BNA_L
 LabSampleId :
 SSTD080

Tgt Ion	Ratio	Resp	Lower	Upper
65	100	287823		
92	70.3	49.2	89.2	
138	126.6	102.4	142.4	




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:47 AM



#47
 Acenaphthylene
 Concen: 59.58 ug/ml
 RT: 15.293 min Scan# 1086
 Delta R.T. -0.004 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp	Lower	Upper
152	100	1329347		
151	20.0	0.0	39.9	
153	13.2	0.0	32.6	









#48
Dimethylphthalate
Concen: 59.24 ug/ml
RT: 15.078 min Scan# 1067
Delta R.T. 0.007 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Instrument :
BNA_L
LabSampleId :
SSTD080

Tgt Ion	Ratio	Resp Lower	Resp Upper
163	100		
194	5.8	0.0	25.9
164	10.0	0.0	29.6




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:47 AM



#49
2,6-Dinitrotoluene
Concen: 61.19 ug/ml
RT: 15.180 min Scan# 1076
Delta R.T. 0.019 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp Lower	Resp Upper
165	100		
89	31.7	19.2	59.2
121	14.5	0.0	36.2









#50
Acenaphthene
Concen: 58.62 ug/ml
RT: 15.620 min Scan# 1115
Delta R.T. -0.004 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Instrument :
BNA_L
LabSampleId :
SSTD080

Tgt Ion	Ratio	Resp Lower	Resp Upper
153	100		
152	46.2	25.1	65.1
154	96.0	75.9	115.9




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:47 AM



#51
3-Nitroaniline
Concen: 73.14 ug/ml
RT: 15.485 min Scan# 1103
Delta R.T. 0.007 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp Lower	Resp Upper
138	100		
108	9.6	0.0	30.0
92	105.1	81.9	121.9





#52

2,4-Dinitrophenol

Concen: 72.91 ug/ml

RT: 15.677 min Scan# 11

Delta R.T. 0.007 min

Lab File: BL000645.D


Acq: 25 Jul 2012 21:43

Instrument :

BNA_L

LabSampleId :



SSTD080



Tgt Ion:	184	Resp:	148436
Ion Ratio	Lower	Upper	
184	100		
63	36.3	11.7	51.7
154	26.5	6.0	46.0

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:47 AM



#53

Dibenzofuran

Concen: 59.54 ug/ml




RT: 15.948 min Scan# 1144


Delta R.T. 0.007 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Tgt Ion:	168	Resp:	1035482
Ion Ratio	Lower	Upper	
168	100		
139	26.9	7.3	47.3









#54
 2,4-Dinitrotoluene
 Concen: 62.05 ug/ml
 RT: 15.925 min Scan# 1142
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Instrument :
 BNA_L
 LabSampleId :
 SST080

Tgt Ion	Ratio	Resp Lower	Resp Upper
165	100		
63	39.0	13.1	53.1
182	5.6	0.0	26.1




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:47 AM



#55
 4-Nitrophenol
 Concen: 66.90 ug/ml
 RT: 15.801 min Scan# 1131
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43


Tgt Ion	Ratio	Resp Lower	Resp Upper
139	100		
65	74.1	44.1	84.1
109	29.9	6.8	46.8





#56
 2,3,4,6-Tetrachlorophenol
 Concen: 65.12 ug/ml
 RT: 16.174 min Scan# 1164
 Delta R.T. -0.004 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43




Instrument :
 BNA_L
 LabSampleId :
 SST080




Tgt Ion	Ratio	Resp Lower	Resp Upper
232	100		
230	77.2	64.6	96.8
131	35.0	29.2	43.8
166	25.1	20.6	30.8

Manual Integrations
 APPROVED



AHPatel
 9/6/2012 6:41:47 AM






#57
 Fluorene
 Concen: 57.94 ug/ml
 RT: 16.580 min Scan# 1200
 Delta R.T. -0.004 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43



Tgt Ion	Ratio	Resp Lower	Resp Upper
166	100		
165	92.1	72.3	112.3
167	13.5	0.0	33.1









#58
4-Chlorophenyl-phenylether
Concen: 57.24 ug/ml
RT: 16.591 min Scan# 1201
Delta R.T. 0.007 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Instrument :
BNA_L
LabSampleId :
SSTD080

Tgt Ion	Ratio	Resp Lower	Resp Upper
204	100		
206	33.3	12.0	52.0
141	43.8	26.4	66.4


Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:47 AM



#59
Diethylphthalate
Concen: 59.96 ug/ml
RT: 16.388 min Scan# 1183
Delta R.T. 0.007 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
177	20.9	1.3	41.3
150	11.4	0.0	31.6









#60
4-Nitroaniline
Concen: 79.30 ug/ml
RT: 16.625 min Scan# 1204
Delta R.T. 0.007 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Instrument :
BNA_L
LabSampleId :
SSTD080

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	202470		
92	41.7	16.5	56.5	
108	30.3	8.1	48.1	




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:47 AM



#63
1,2,4,5-Tetrachlorobenzene
Concen: 54.70 ug/ml
RT: 13.949 min Scan# 967
Delta R.T. 0.007 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp	Lower	Upper
216	100	416047		
214	80.9	64.0	96.0	
179	18.4	14.6	22.0	
218	48.8	38.6	57.8	









#64
 4,6-Dinitro-2-methylphenol
 Concen: 65.07 ug/ml
 RT: 16.693 min Scan# 1210
 Delta R.T. 0.019 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Instrument :
 BNA_L
 LabSampleId :
 SST080

Tgt Ion:198 Resp: 161270

Manual Integrations
 APPROVED




AHPatel
 9/6/2012 6:41:47 AM




#65
 n-Nitrosodiphenylamine
 Concen: 57.85 ug/ml
 RT: 16.817 min Scan# 1221
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Tgt Ion:169 Resp: 646437

Ion	Ratio	Lower	Upper
169	100		
168	62.4	40.9	80.9
167	31.9	12.5	52.5









#66
 1,2-Diphenylhydrazine
 Concen: 57.20 ug/ml
 RT: 16.885 min Scan# 1227
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Instrument :
 BNA_L
 LabSampleId :
 SST080

Tgt Ion	Ratio	Resp Lower	Resp Upper
77	100		
182	21.7	0.0	39.4
105	14.8	0.0	34.5




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:47 AM



#67
 4-Bromophenyl-phenylether
 Concen: 57.63 ug/ml
 RT: 17.506 min Scan# 1282
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp Lower	Resp Upper
248	100		
250	97.0	78.2	118.2
141	47.9	34.7	74.7









#68
Hexachlorobenzene
Concen: 56.75 ug/ml
RT: 17.630 min Scan# 1293
Delta R.T. 0.007 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Instrument :
BNA_L
LabSampleId :
SSTD080

Tgt Ion	Ratio	Resp Lower	Resp Upper
284	100		
142	30.0	14.3	54.3
249	27.0	6.7	46.7




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:47 AM



#69
Atrazine
Concen: 62.80 ug/ml
RT: 17.833 min Scan# 1311
Delta R.T. 0.007 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp Lower	Resp Upper
200	100		
173	31.9	10.7	50.7
115	0.0	0.0	20.0









#70
 Pentachlorophenol
 Concen: 68.06 ug/ml
 RT: 17.991 min Scan# 1325
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Instrument :
 BNA_L
 LabSampleId :
 SST080

Tgt Ion	Ratio	Resp Lower	Resp Upper
266	100		
264	65.6	44.5	84.5
268	66.4	45.0	85.0




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:47 AM



#72
 Benzidine
 Concen: 47.74 ug/ml
 RT: 20.340 min Scan# 1533
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Tgt Ion:184 Resp: 54464









#73
Phenanthrene
Concen: 57.52 ug/ml
RT: 18.375 min Scan# 1359
Delta R.T. 0.007 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Instrument :
BNA_L
LabSampleId :
SSTD080

Tgt Ion	Ratio	Resp Lower	Resp Upper
178	100		
179	15.1	0.0	35.3
176	18.4	0.0	38.6




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:47 AM



#74
Anthracene
Concen: 58.09 ug/ml
RT: 18.454 min Scan# 1366
Delta R.T. -0.004 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp Lower	Resp Upper
178	100		
179	15.4	0.0	34.9
176	18.0	0.0	37.0









#75
 Carbazole
 Concen: 61.25 ug/ml m
 RT: 18.725 min Scan# 1390
 Delta R.T. -0.004 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Instrument :
 BNA_L
 LabSampleId :
 SST080

Tgt Ion	Ratio	Resp	Lower	Upper
167	100	796445		
167	90.7	80.0	120.0	




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:47 AM



#76
 Di-n-butylphthalate
 Concen: 59.00 ug/ml
 RT: 19.301 min Scan# 1441
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	986259		
150	9.1	0.0	29.4	
104	3.4	0.0	23.7	









#77
Fluoranthene
 Concen: 58.97 ug/ml
 RT: 20.160 min Scan# 1517
 Delta R.T. -0.004 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Instrument :
 BNA_L
 LabSampleId :
 SST080

Tgt	Ion	Ratio	Resp	Lower	Upper
202	100		739703		
101	31.5	5.5	45.5		




Manual Integrations
APPROVED


AHPatel
 9/6/2012 6:41:47 AM



#79
Pyrene
 Concen: 67.78 ug/ml
 RT: 20.464 min Scan# 1544
 Delta R.T. 0.007 min
 Lab File: BL000645.D
 Acq: 25 Jul 2012 21:43

Tgt	Ion	Ratio	Resp	Lower	Upper
202	100		747035		
101	24.5	11.7	51.7		
100	19.6	6.0	46.0		









#81
Butylbenzylphthalate
Concen: 67.42 ug/ml
RT: 21.232 min Scan# 1612
Delta R.T. 0.007 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Instrument :
BNA_L
LabSampleId :
SSTD080

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
91	56.5	51.8	91.8
206	20.7	0.0	36.6




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:47 AM



#82
3,3-Dichlorobenzidine
Concen: 68.96 ug/ml
RT: 21.853 min Scan# 1667
Delta R.T. -0.004 min
Lab File: BL000645.D
Acq: 25 Jul 2012 21:43

Tgt Ion	Ratio	Resp Lower	Resp Upper
252	100		
254	65.1	47.7	87.7
126	17.5	0.0	35.4





#83

Benzo[a]anthracene

Concen: 59.91 ug/ml

RT: 21.898 min Scan# 16

Delta R.T. 0.007 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Instrument :

BNA_L





LabSampleId :

SSTD080

Tgt Ion:	228	Resp:	535557
Ion Ratio	Lower	Upper	
228	100		
229	19.1	0.0	39.4
226	24.7	5.2	45.2

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:47 AM



#84

Chrysene

Concen: 58.49 ug/ml




RT: 21.944 min Scan# 1675


Delta R.T. -0.004 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Tgt Ion:	228	Resp:	512935
Ion Ratio	Lower	Upper	
228	100		
226	27.5	6.9	46.9
229	19.3	0.0	39.1





#85

bis(2-Ethylhexyl)phthalate

Concen: 66.94 ug/ml

RT: 21.876 min Scan# 1669

Delta R.T. 0.007 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Instrument :

BNA_L





LabSampleId :

SSTD080

Tgt	Ion:149	Resp:	522882
Ion	Ratio	Lower	Upper
149	100		
167	31.7	9.1	49.1
279	6.1	0.0	24.0

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:47 AM



#87

Di-n-octylphthalate

Concen: 76.47 ug/ml




RT: 22.689 min Scan# 1741


Delta R.T. -0.004 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Tgt	Ion:149	Resp:	790936
Ion	Ratio	Lower	Upper
149	100		
150	9.6	0.0	29.4





#88

Benzo[b]fluoranthene

Concen: 62.42 ug/ml

RT: 23.389 min Scan# 1803

Delta R.T. -0.004 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Instrument :

BNA_L





LabSampleId :

SSTD080

Tgt Ion:	252	Resp:	444186
Ion Ratio	Lower	Upper	
252	100		
253	20.8	1.5	41.5
125	24.1	1.4	41.4

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:47 AM



#89

Benzo[k]fluoranthene

Concen: 63.21 ug/ml




RT: 23.434 min Scan# 1807


Delta R.T. 0.007 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Tgt Ion:	252	Resp:	395389
Ion Ratio	Lower	Upper	
252	100		
253	21.7	2.6	42.6
125	24.2	7.4	47.4





#90

Benzo[a]pyrene

Concen: 61.57 ug/ml

RT: 23.954 min Scan# 18

Delta R.T. 0.007 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Instrument :

BNA_L





LabSampleId :

SSTD080

Tgt Ion:	252	Resp:	402922
Ion Ratio		Lower	Upper
252	100		
253	21.3	1.8	41.8
125	24.8	8.0	48.0

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:47 AM



#91

Indeno[1,2,3-cd]pyrene

Concen: 58.14 ug/ml




RT: 26.347 min Scan# 2065


Delta R.T. -0.004 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Tgt Ion:	276	Resp:	408830
Ion Ratio		Lower	Upper
276	100		
138	54.3	37.7	77.7





#92

Dibenz[a,h]anthracene

Concen: 58.92 ug/ml

RT: 26.381 min Scan# 20

Delta R.T. -0.004 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Instrument :

BNA_L





LabSampleId :

SSTD080

Tgt	Ion	278	Resp:	332901
Ion	Ratio	Lower	Upper	
278	100			
139	42.4	18.8	58.8	
279	24.0	4.5	44.5	

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:47 AM



#93

Benzo[g,h,i]perylene

Concen: 58.64 ug/ml




RT: 27.070 min Scan# 2129

Delta R.T. 0.007 min

Lab File: BL000645.D

Acq: 25 Jul 2012 21:43

Tgt	Ion	276	Resp:	362089
Ion	Ratio	Lower	Upper	
276	100			
138	53.9	36.8	76.8	
277	23.5	4.5	44.5	



Data Path : S:\HPCHEM1\BNA_L\data\BL072512\
 Data File : BL000645.D
 Acq On : 25 Jul 2012 21:43
 Operator : HLM
 Sample : SSTD080
 Misc : SSTD080 HLM
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_L
 LabSampleId :
 SSTD080

Manual Integrations
 APPROVED

AHPatel
 9/6/2012 6:41:47 AM

Quant Time: Sep 06 05:37:55 2012
 Quant Method : H:\VOL1\GCMSDATA\L\METHODS\BL072512.M
 Quant Title : SW-846 Method 8270
 QLast Update : Thu Jul 26 10:46:15 2012
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	10.008	152	437418	40.00	ug/ml	0.00
22) Naphthalene-d8	12.323	136	1374601	40.00	ug/ml	0.00
39) Acenaphthene-d10	15.564	164	530655	40.00	ug/ml	0.00
62) Phenanthrene-d10	18.330	188	643981	40.00	ug/ml	0.00
78) Chrysene-d12	21.910	240	363892	40.00	ug/ml	0.00
86) Perylene-d12	24.044	264	255323	40.00	ug/ml	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	7.513	112	1103011	61.38	ug/ml	0.00
Spiked Amount 200.000	Range 20	- 110	Recovery	=	30.69%	
9) Phenol-d5	9.353	99	1288917	63.11	ug/ml	0.00
Spiked Amount 200.000	Range 10	- 115	Recovery	=	31.56%	
23) Nitrobenzene-d5	11.013	82	856592	61.08	ug/ml	0.00
Spiked Amount 100.000	Range 40	- 110	Recovery	=	61.08%	
43) 2-Fluorobiphenyl	14.310	172	999077	58.48	ug/ml	0.00
Spiked Amount 100.000	Range 50	- 110	Recovery	=	58.48%	
61) 2,4,6-Tribromophenol	17.032	330	92946	62.30	ug/ml	0.00
Spiked Amount 200.000	Range 40	- 125	Recovery	=	31.15%#	
80) Terphenyl-d14	20.656	244	393248	68.95	ug/ml	0.00
Spiked Amount 100.000	Range 50	- 135	Recovery	=	68.95%	
Target Compounds						
						Qvalue
3) N-Nitrosodimethylamine	3.380	74	1153605	62.43	ug/ml	100
4) Pyridine	3.448	79	1787962m	63.06	ug/ml	
6) Benzaldehyde	9.195	77	288006	47.39	ug/ml	100
7) Aniline	9.410	93	1655821	65.23	ug/ml	98
8) bis(2-Chloroethyl)ether	9.534	93	1101867	61.78	ug/ml	99
10) Phenol	9.376	94	1473769	63.47	ug/ml	96
11) 2-Chlorophenol	9.636	128	904312	61.43	ug/ml	96
12) 1,3-Dichlorobenzene	9.907	146	896956	60.22	ug/ml	99
13) 1,4-Dichlorobenzene	10.042	146	890094	60.20	ug/ml	99
14) 1,2-Dichlorobenzene	10.324	146	821646	61.07	ug/ml	99
15) Benzyl alcohol	10.268	108	606545	65.72	ug/ml	99
16) 2,2-oxybis(1-chloropro...	10.505	45	1697561	64.33	ug/ml	99
17) 2-Methylphenol	10.460	108	853082	63.18	ug/ml	98
18) Acetophenone	10.742	105	928720	60.83	ug/ml	98
19) Hexachloroethane	10.934	117	349501	61.30	ug/ml	87
20) N-Nitroso-di-n-propyla...	10.753	70	619527	63.24	ug/ml	99
21) 4-Methylphenol	10.742	108	748804	61.04	ug/ml	96
24) Nitrobenzene	11.047	77	877869	61.35	ug/ml	96
25) Isophorone	11.487	82	1682208	61.51	ug/ml	99
26) 2-Nitrophenol	11.623	139	447773	60.68	ug/ml	96
27) 2,4-Dimethylphenol	11.702	107	668427	62.74	ug/ml	99
28) Benzoic Acid	12.041	105	1116776	145.95	ug/ml	98
29) bis(2-Chloroethoxy)met...	11.871	93	1007162	60.40	ug/ml	99
30) 2,4-Dichlorophenol	12.063	162	554606	63.22	ug/ml	98
32) 1,2,4-Trichlorobenzene	12.210	180	593938	60.55	ug/ml	99
33) Naphthalene	12.357	128	1847607	59.07	ug/ml	99
34) 4-Chloroaniline	12.447	127	835137	64.75	ug/ml	100

Data Path : S:\HPCHEM1\BNA_L\data\BL072512\
 Data File : BL000645.D
 Acq On : 25 Jul 2012 21:43
 Operator : HLM
 Sample : SSTD080
 Misc : SSTD080 HLM
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_L
 LabSampleId :
 SSTD080

Manual Integrations
 APPROVED

AHPatel
 9/6/2012 6:41:47 AM

Quant Time: Sep 06 05:37:55 2012
 Quant Method : H:\VOL1\GCMSDATA\L\METHODS\BL072512.M
 Quant Title : SW-846 Method 8270
 QLast Update : Thu Jul 26 10:46:15 2012
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) Hexachlorobutadiene	12.594	225	246838	59.45	ug/ml	98
36) Caprolactam	13.159	113	221585	65.36	ug/ml	96
37) 4-Chloro-3-methylphenol	13.351	107	512418	62.73	ug/ml	96
38) 2-Methylnaphthalene	13.622	142	1038303	60.17	ug/ml	99
40) Hexachlorocyclopentadiene	13.938	237	231261	62.23	ug/ml	100
41) 2,4,6-Trichlorophenol	14.152	196	304962	60.03	ug/ml	100
42) 2,4,5-Trichlorophenol	14.220	196	325557	59.35	ug/ml	99
44) 1,1-Biphenyl	14.491	154	1097697	58.07	ug/ml	96
45) 2-Chloronaphthalene	14.525	162	908545	58.50	ug/ml	97
46) 2-Nitroaniline	14.717	65	287823	60.38	ug/ml	97
47) Acenaphthylene	15.293	152	1329347	59.58	ug/ml	99
48) Dimethylphthalate	15.078	163	913686	59.24	ug/ml	99
49) 2,6-Dinitrotoluene	15.180	165	247008	61.19	ug/ml	90
50) Acenaphthene	15.620	153	786418	58.62	ug/ml	99
51) 3-Nitroaniline	15.485	138	263915	73.14	ug/ml	97
52) 2,4-Dinitrophenol	15.677	184	148436	72.91	ug/ml	95
53) Dibenzofuran	15.948	168	1035482	59.54	ug/ml	99
54) 2,4-Dinitrotoluene	15.925	165	269500	62.05	ug/ml	91
55) 4-Nitrophenol	15.801	139	205575	66.90	ug/ml	89
56) 2,3,4,6-Tetrachlorophenol	16.174	232	210671	65.12	ug/ml	97
57) Fluorene	16.580	166	729445	57.94	ug/ml	100
58) 4-Chlorophenyl-phenyle...	16.591	204	320162	57.24	ug/ml	97
59) Diethylphthalate	16.388	149	751764	59.96	ug/ml	99
60) 4-Nitroaniline	16.625	138	202470	79.30	ug/ml	93
63) 1,2,4,5-Tetrachloroben...	13.949	216	416047	54.70	ug/ml	99
64) 4,6-Dinitro-2-methylph...	16.693	198	161270	65.07	ug/ml	100
65) n-Nitrosodiphenylamine	16.817	169	646437	57.85	ug/ml	98
66) 1,2-Diphenylhydrazine	16.885	77	969115	57.20	ug/ml	97
67) 4-Bromophenyl-phenylether	17.506	248	185768	57.63	ug/ml	96
68) Hexachlorobenzene	17.630	284	195024	56.75	ug/ml	96
69) Atrazine	17.833	200	136539	62.80	ug/ml	98
70) Pentachlorophenol	17.991	266	122098	68.06	ug/ml	98
72) Benzidine	20.340	184	54464	47.74	ug/ml	100
73) Phenanthrene	18.375	178	845104	57.52	ug/ml	100
74) Anthracene	18.454	178	875629	58.09	ug/ml	98
75) Carbazole	18.725	167	796445m	61.25	ug/ml	
76) Di-n-butylphthalate	19.301	149	986259	59.00	ug/ml	99
77) Fluoranthene	20.160	202	739703	58.97	ug/ml	88
79) Pyrene	20.464	202	747035	67.78	ug/ml	87
81) Butylbenzylphthalate	21.232	149	366133	67.42	ug/ml	83
82) 3,3-Dichlorobenzidine	21.853	252	122924	68.96	ug/ml	96
83) Benzo[a]anthracene	21.898	228	535557	59.91	ug/ml	99
84) Chrysene	21.944	228	512935	58.49	ug/ml	99
85) bis(2-Ethylhexyl)phtha...	21.876	149	522882	66.94	ug/ml	95
87) Di-n-octylphthalate	22.689	149	790936	76.47	ug/ml	99
88) Benzo[b]fluoranthene	23.389	252	444186	62.42	ug/ml	96
89) Benzo[k]fluoranthene	23.434	252	395389	63.21	ug/ml	96
90) Benzo[a]pyrene	23.954	252	402922	61.57	ug/ml	96
91) Indeno[1,2,3-cd]pyrene	26.347	276	408830	58.14	ug/ml	95
92) Dibenz[a,h]anthracene	26.381	278	332901	58.92	ug/ml	96

Data Path : S:\HPCHEM1\BNA_L\data\BL072512\
Data File : BL000645.D
Acq On : 25 Jul 2012 21:43
Operator : HLM
Sample : SSTD080
Misc : SSTD080 HLM
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_L
LabSampleId :
SSTD080

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:47 AM

Quant Time: Sep 06 05:37:55 2012
Quant Method : H:\VOL1\GCMSDATA\L\METHODS\BL072512.M
Quant Title : SW-846 Method 8270
QLast Update : Thu Jul 26 10:46:15 2012
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) Benzo[g,h,i]perylene	27.070	276	362089	58.64	ug/ml	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed