

#3

N-Nitrosodimethylamine

Concen: 91.06 ug/ml

RT: 3.418 min Scan# 34

Delta R.T. 0.057 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L

LabSampleId :

SSTD120

Tgt Ion: 74 Resp: 1517591

Ion Ratio Lower Upper





74 100

43 21.9 2.3 42.3

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:49 AM



#4

Pyridine

Concen: 92.21 ug/ml m

RT: 3.486 min Scan# 40

Delta R.T. 0.023 min

Lab File: BL000646.D




Acq: 25 Jul 2012 22:21


Tgt Ion: 79 Resp: 2357992

Ion Ratio Lower Upper

79 100

52 65.1 47.6 87.6





#6

Benzaldehyde

Concen: 43.30 ug/ml

RT: 9.199 min Scan# 54

Delta R.T. 0.000 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L

LabSampleId :

SSTD120

Tgt Ion: 77 Resp: 237319

Ion Ratio Lower Upper

77 100





105 91.9 70.4 110.4

106 90.6 70.4 110.4

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:49 AM



#7

Aniline

Concen: 91.43 ug/ml

RT: 9.425 min Scan# 566

Delta R.T. 0.011 min

Lab File: BL000646.D




Acq: 25 Jul 2012 22:21


Tgt Ion: 93 Resp: 2093107

Ion Ratio Lower Upper

93 100

66 39.0 30.6 46.0





#8

bis(2-Chloroethyl)ether

Concen: 87.44 ug/ml

RT: 9.538 min Scan# 57

Delta R.T. 0.011 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L

LabSampleId :





SSTD120

Tgt Ion: 93 Resp: 1406645

Ion	Ratio	Lower	Upper
93	100		
63	80.3	58.9	98.9
95	32.2	11.9	51.9

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:49 AM



#10

Phenol

Concen: 86.95 ug/ml

RT: 9.391 min Scan# 563




Delta R.T. 0.023 min


Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Tgt Ion: 94 Resp: 1820984

Ion	Ratio	Lower	Upper
94	100		
65	25.8	1.6	41.6
66	32.4	4.3	44.3









#11
2-Chlorophenol
Concen: 86.24 ug/ml
RT: 9.640 min Scan# 58
Delta R.T. 0.011 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Instrument :
BNA_L
LabSampleId :
SSTD120

Tgt Ion	Ratio	Lower	Upper
128	100		
64	56.7	37.3	77.3
130	32.1	12.4	52.4




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:49 AM



#12
1,3-Dichlorobenzene
Concen: 85.59 ug/ml
RT: 9.911 min Scan# 609
Delta R.T. 0.000 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Lower	Upper
146	100		
148	62.9	44.0	84.0
113	13.0	0.0	32.5





#13

1,4-Dichlorobenzene

Concen: 86.93 ug/ml

RT: 10.046 min Scan# 621

Delta R.T. 0.011 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L

LabSampleId :

SSTD120

Tgt Ion:146 Resp: 1159291

Ion Ratio Lower Upper

146 100





148 64.1 44.6 84.6

113 12.3 0.0 32.8

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:49 AM



#14

1,2-Dichlorobenzene

Concen: 86.20 ug/ml

RT: 10.328 min Scan# 646

Delta R.T. 0.011 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21




Tgt Ion:146 Resp: 1045981


Ion Ratio Lower Upper

146 100

148 64.5 44.0 84.0

113 13.1 0.0 33.6









#15
Benzyl alcohol
Concen: 90.02 ug/ml
RT: 10.272 min Scan# 641
Delta R.T. 0.011 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Instrument :
BNA_L
LabSampleId :
SSTD120

Tgt Ion	Ratio	Resp	Lower	Upper
108	100	749340		
79	116.3	94.5	134.5	
77	66.2	43.7	83.7	




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:49 AM



#16
2,2-oxybis(1-chloropropane)
Concen: 88.20 ug/ml
RT: 10.509 min Scan# 662
Delta R.T. 0.011 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp	Lower	Upper
45	100	2099134		
77	10.5	0.0	30.0	
79	9.6	0.0	28.9	









#17
 2-Methylphenol
 Concen: 87.97 ug/ml
 RT: 10.464 min Scan# 658
 Delta R.T. 0.011 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Instrument :
 BNA_L
 LabSampleId :
 SSTD120

Tgt Ion:108 Resp: 1071314
 Ion Ratio Lower Upper
 108 100
 107 88.5 67.8 107.8




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:49 AM



#18
 Acetophenone
 Concen: 83.95 ug/ml
 RT: 10.746 min Scan# 683
 Delta R.T. 0.011 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Tgt Ion:105 Resp: 1155866
 Ion Ratio Lower Upper
 105 100
 71 4.6 0.0 27.0
 51 35.0 16.6 56.6





#19

Hexachloroethane

Concen: 87.63 ug/ml

RT: 10.938 min Scan# 700

Delta R.T. 0.011 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L





LabSampleId :

SSTD120

Tgt Ion	Ratio	Resp Lower	Resp Upper
117	100		
200	0.0	0.0	20.0
199	50.6	21.2	61.2

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:49 AM



#20

N-Nitroso-di-n-propylamine

Concen: 87.98 ug/ml




RT: 10.769 min Scan# 685


Delta R.T. 0.023 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp Lower	Resp Upper
70	100		
42	50.4	32.8	72.8
130	23.8	3.0	43.0









#21
4-Methylphenol
Concen: 82.44 ug/ml
RT: 10.758 min Scan# 68
Delta R.T. 0.023 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Instrument :
BNA_L
LabSampleId :
SSTD120

Tgt Ion: 108 Resp: 912096
Ion Ratio Lower Upper
108 100
107 105.6 86.9 126.9




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:49 AM



#24
Nitrobenzene
Concen: 86.27 ug/ml
RT: 11.062 min Scan# 711
Delta R.T. 0.023 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Tgt Ion: 77 Resp: 1099980
Ion Ratio Lower Upper
77 100
123 55.2 26.9 66.9
65 11.6 0.0 32.1





#25

Isophorone

Concen: 86.13 ug/ml

RT: 11.503 min Scan# 750

Delta R.T. 0.023 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L

LabSampleId :

SSTD120

Tgt Ion: 82 Resp: 2098919

Ion Ratio Lower Upper





82 100

138 15.9 0.0 35.5

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:49 AM



#26

2-Nitrophenol

Concen: 88.56 ug/ml

RT: 11.627 min Scan# 761

Delta R.T. 0.011 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21




Tgt Ion: 139 Resp: 582258


Ion Ratio Lower Upper

139 100

65 38.3 20.8 60.8

109 13.5 0.0 33.5









#27
2,4-Dimethylphenol
Concen: 88.78 ug/ml
RT: 11.706 min Scan# 76
Delta R.T. 0.011 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Instrument :
BNA_L
LabSampleId :
SSTD120

Tgt Ion	Ratio	Resp	Lower	Upper
107	100	842898		
121	53.9	33.2	73.2	
122	96.2	75.9	115.9	




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:49 AM



#28
Benzoic Acid
Concen: 211.06 ug/ml
RT: 12.090 min Scan# 802
Delta R.T. 0.113 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp	Lower	Upper
105	100	1439066		
77	69.9	52.0	92.0	
122	85.5	61.6	101.6	









#29
 bis(2-Chloroethoxy)methane
 Concen: 83.50 ug/ml
 RT: 11.875 min Scan# 783
 Delta R.T. 0.011 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Instrument :
 BNA_L
 LabSampleId :
 SSTD120

Tgt Ion:	93	Resp:	1240580
Ion	Ratio	Lower	Upper
93	100		
95	32.1	11.6	51.6
123	11.0	0.0	30.8




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:49 AM



#30
 2,4-Dichlorophenol
 Concen: 86.34 ug/ml
 RT: 12.067 min Scan# 800
 Delta R.T. 0.011 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Tgt Ion:	162	Resp:	674900
Ion	Ratio	Lower	Upper
162	100		
164	62.6	43.5	83.5
98	38.2	19.5	59.5





#32

1,2,4-Trichlorobenzene

Concen: 85.19 ug/ml

RT: 12.214 min Scan# 81

Delta R.T. 0.000 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L

LabSampleId :

SSTD120

Tgt Ion:180 Resp: 744525
Ion Ratio Lower Upper

180 100





182 94.2 75.6 115.6

145 29.3 8.3 48.3

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:49 AM



#33

Naphthalene

Concen: 82.46 ug/ml

RT: 12.361 min Scan# 826

Delta R.T. 0.000 min

Lab File: BL000646.D




Acq: 25 Jul 2012 22:21


Tgt Ion:128 Resp: 2298160
Ion Ratio Lower Upper

128 100

129 11.2 0.0 30.7

127 13.6 0.0 33.1









#34
4-Chloroaniline
Concen: 88.62 ug/ml
RT: 12.451 min Scan# 834
Delta R.T. 0.000 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Instrument :
BNA_L
LabSampleId :
SSTD120

Tgt Ion:127 Resp: 1018610
Ion Ratio Lower Upper
127 100
129 31.9 12.1 52.1




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:49 AM



#35
Hexachlorobutadiene
Concen: 83.86 ug/ml
RT: 12.598 min Scan# 847
Delta R.T. 0.000 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Tgt Ion:225 Resp: 310225
Ion Ratio Lower Upper
225 100
223 62.2 41.7 81.7
227 64.1 45.2 85.2









#36
 Caprolactam
 Concen: 85.01 ug/ml
 RT: 13.174 min Scan# 89
 Delta R.T. 0.057 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Instrument :
 BNA_L
 LabSampleId :
 SSTD120

Tgt Ion	Ratio	Resp	Lower	Upper
113	100	256807		
55	171.0	144.2	184.2	
56	111.4	92.3	132.3	




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:49 AM



#37
 4-Chloro-3-methylphenol
 Concen: 87.83 ug/ml
 RT: 13.355 min Scan# 914
 Delta R.T. 0.011 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp	Lower	Upper
107	100	639283		
144	27.4	9.7	49.7	
142	85.9	71.3	111.3	









#38
 2-Methylnaphthalene
 Concen: 83.45 ug/ml
 RT: 13.626 min Scan# 938
 Delta R.T. 0.000 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Instrument :
 BNA_L
 LabSampleId :
 SSTD120

Tgt Ion:142 Resp: 1283121
 Ion Ratio Lower Upper
 142 100
 141 84.8 64.5 104.5




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:49 AM



#40
 Hexachlorocyclopentadiene
 Concen: 94.37 ug/ml
 RT: 13.931 min Scan# 965
 Delta R.T. 0.000 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Tgt Ion:237 Resp: 287981
 Ion Ratio Lower Upper
 237 100
 235 63.4 42.1 82.1
 272 14.2 0.0 35.7









#41
 2,4,6-Trichlorophenol
 Concen: 88.79 ug/ml
 RT: 14.156 min Scan# 98
 Delta R.T. 0.011 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Instrument :
 BNA_L
 LabSampleId :
 SSTD120

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	370401		
198	95.5	73.1	113.1	
200	31.2	9.7	49.7	




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:49 AM



#42
 2,4,5-Trichlorophenol
 Concen: 87.42 ug/ml
 RT: 14.224 min Scan# 991
 Delta R.T. 0.011 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	393808		
198	95.6	74.5	114.5	
200	31.0	10.8	50.8	









#44
 1,1-Biphenyl
 Concen: 84.65 ug/ml
 RT: 14.495 min Scan# 1015
 Delta R.T. 0.000 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Instrument :
 BNA_L
 LabSampleId :
 SSTD120

Tgt Ion:154 Resp: 1313979
 Ion Ratio Lower Upper
 154 100
 153 40.4 19.9 59.9
 76 27.4 3.3 43.3




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:49 AM



#45
 2-Chloronaphthalene
 Concen: 84.59 ug/ml
 RT: 14.529 min Scan# 1018
 Delta R.T. 0.000 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Tgt Ion:162 Resp: 1078873
 Ion Ratio Lower Upper
 162 100
 164 33.1 12.8 52.8
 127 34.3 10.9 50.9









#46
2-Nitroaniline
Concen: 84.14 ug/ml
RT: 14.721 min Scan# 1035
Delta R.T. 0.011 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Instrument :
BNA_L
LabSampleId :
SSTD120

Tgt Ion	Ratio	Resp	Lower	Upper
65	100	329358		
92	70.5	49.2	89.2	
138	131.5	102.4	142.4	




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:49 AM



#47
Acenaphthylene
Concen: 85.20 ug/ml
RT: 15.297 min Scan# 1086
Delta R.T. 0.000 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp	Lower	Upper
152	100	1561321		
151	20.1	0.0	39.9	
153	13.4	0.0	32.6	





#48

Dimethylphthalate

Concen: 82.83 ug/ml

RT: 15.082 min Scan# 1067

Delta R.T. 0.011 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L





LabSampleId :

SSTD120

Tgt Ion:	163	Resp:	1049157
Ion Ratio		Lower	Upper
163	100		
194	5.6	0.0	25.9
164	9.9	0.0	29.6

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:49 AM



#49

2,6-Dinitrotoluene

Concen: 86.37 ug/ml




RT: 15.173 min Scan# 1075


Delta R.T. 0.011 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Tgt Ion:	165	Resp:	286334
Ion Ratio		Lower	Upper
165	100		
89	38.7	19.2	59.2
121	15.9	0.0	36.2









#50
Acenaphthene
Concen: 82.61 ug/ml
RT: 15.624 min Scan# 1115
Delta R.T. 0.000 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Instrument :
BNA_L
LabSampleId :
SSTD120

Tgt Ion	Ratio	Resp	Lower	Upper
153	100	910046		
152	45.6	25.1	65.1	
154	95.7	75.9	115.9	




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:49 AM



#51
3-Nitroaniline
Concen: 95.07 ug/ml
RT: 15.489 min Scan# 1103
Delta R.T. 0.011 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	281720		
108	9.3	0.0	30.0	
92	101.3	81.9	121.9	









#52
 2,4-Dinitrophenol
 Concen: 104.04 ug/ml
 RT: 15.681 min Scan# 111
 Delta R.T. 0.011 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Instrument :
 BNA_L
 LabSampleId :
 SSTD120

Tgt Ion	Ratio	Resp	Lower	Upper
184	100	173957		
63	32.6	11.7	51.7	
154	25.8	6.0	46.0	




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:49 AM



#53
 Dibenzofuran
 Concen: 84.22 ug/ml
 RT: 15.941 min Scan# 1143
 Delta R.T. 0.000 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp	Lower	Upper
168	100	1202867		
139	28.6	7.3	47.3	









#54
 2,4-Dinitrotoluene
 Concen: 83.38 ug/ml
 RT: 15.929 min Scan# 1142
 Delta R.T. 0.011 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Instrument :
 BNA_L
 LabSampleId :
 SSTD120

Tgt Ion	Ratio	Resp Lower	Resp Upper
165	100		
63	38.1	13.1	53.1
182	5.6	0.0	26.1




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:49 AM



#55
 4-Nitrophenol
 Concen: 88.66 ug/ml
 RT: 15.805 min Scan# 1131
 Delta R.T. 0.011 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp Lower	Resp Upper
139	100		
65	71.6	44.1	84.1
109	28.1	6.8	46.8









#56
 2,3,4,6-Tetrachlorophenol
 Concen: 87.89 ug/ml
 RT: 16.178 min Scan# 1164
 Delta R.T. 0.000 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Instrument :
 BNA_L
 LabSampleId :
 SSTD120

Tgt Ion	Ratio	Resp Lower	Resp Upper
232	100		
230	79.4	64.6	96.8
131	34.9	29.2	43.8
166	25.7	20.6	30.8




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:49 AM



#57
 Fluorene
 Concen: 76.43 ug/ml
 RT: 16.584 min Scan# 1200
 Delta R.T. 0.000 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp Lower	Resp Upper
166	100		
165	92.9	72.3	112.3
167	13.6	0.0	33.1









#58
4-Chlorophenyl-phenylether
Concen: 77.20 ug/ml
RT: 16.584 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Instrument :
BNA_L
LabSampleId :
SSTD120

Tgt Ion	Ratio	Resp Lower	Resp Upper
204	100		
206	32.7	12.0	52.0
141	52.1	26.4	66.4




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:49 AM



#59
Diethylphthalate
Concen: 80.23 ug/ml
RT: 16.392 min Scan# 1183
Delta R.T. 0.011 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
177	21.6	1.3	41.3
150	11.4	0.0	31.6









#60
4-Nitroaniline
Concen: 92.72 ug/ml
RT: 16.629 min Scan# 1204
Delta R.T. 0.011 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Instrument :
BNA_L
LabSampleId :
SSTD120

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	194416		
92	39.4	16.5	56.5	
108	28.0	8.1	48.1	




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:49 AM



#63
1,2,4,5-Tetrachlorobenzene
Concen: 90.32 ug/ml
RT: 13.953 min Scan# 967
Delta R.T. 0.011 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp	Lower	Upper
216	100	504378		
214	79.5	64.0	96.0	
179	18.4	14.6	22.0	
218	48.0	38.6	57.8	





#64

4,6-Dinitro-2-methylphenol

Concen: 98.53 ug/ml

RT: 16.686 min Scan# 1210

Delta R.T. 0.011 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L





LabSampleId :

SSTD120

Tgt Ion:198 Resp: 179279

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:49 AM



#65

n-Nitrosodiphenylamine

Concen: 87.68 ug/ml

RT: 16.810 min Scan# 1220




Delta R.T. 0.000 min


Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Tgt Ion:169 Resp: 719331

Ion	Ratio	Lower	Upper
169	100		
168	62.2	40.9	80.9
167	32.9	12.5	52.5









#66
 1,2-Diphenylhydrazine
 Concen: 87.19 ug/ml
 RT: 16.889 min Scan# 1227
 Delta R.T. 0.011 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Instrument :
 BNA_L
 LabSampleId :
 SSTD120

Tgt Ion: 77 Resp: 1084673
 Ion Ratio Lower Upper
 77 100
 182 24.2 0.0 39.4
 105 15.4 0.0 34.5




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:49 AM



#67
 4-Bromophenyl-phenylether
 Concen: 87.14 ug/ml
 RT: 17.499 min Scan# 1281
 Delta R.T. 0.000 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Tgt Ion: 248 Resp: 206234
 Ion Ratio Lower Upper
 248 100
 250 96.9 78.2 118.2
 141 59.8 34.7 74.7





#68

Hexachlorobenzene

Concen: 83.00 ug/ml

RT: 17.623 min Scan# 12

Delta R.T. 0.000 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L





LabSampleId :

SSTD120

Tgt Ion	Ratio	Resp	Lower	Upper
284	100	209409		
142	42.2	14.3	54.3	
249	28.9	6.7	46.7	

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:49 AM



#69

Atrazine

Concen: 91.63 ug/ml




RT: 17.838 min Scan# 1311


Delta R.T. 0.011 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp	Lower	Upper
200	100	146266		
173	30.1	10.7	50.7	
115	0.0	0.0	20.0	





#70

Pentachlorophenol

Concen: 103.85 ug/ml

RT: 17.984 min Scan# 1325

Delta R.T. 0.000 min

Lab File: BL000646.D


Acq: 25 Jul 2012 22:21

Instrument :

BNA_L

LabSampleId :




SSTD120



Tgt Ion:	266	Resp:	136771
Ion Ratio	Lower	Upper	
266	100		
264	63.9	44.5	84.5
268	63.8	45.0	85.0

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:49 AM



#72

Benzidine

Concen: 80.06 ug/ml




RT: 20.333 min Scan# 1532


Delta R.T. 0.000 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Tgt Ion:184 Resp: 67055









#73
 Phenanthrene
 Concen: 82.65 ug/ml
 RT: 18.368 min Scan# 1358
 Delta R.T. 0.000 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Instrument :
 BNA_L
 LabSampleId :
 SSTD120

Tgt Ion	Ratio	Resp Lower	Resp Upper
178	100		
179	15.2	0.0	35.3
176	18.0	0.0	38.6




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:49 AM



#74
 Anthracene
 Concen: 82.16 ug/ml
 RT: 18.459 min Scan# 1366
 Delta R.T. 0.000 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp Lower	Resp Upper
178	100		
179	15.4	0.0	34.9
176	17.8	0.0	37.0









#75
 Carbazole
 Concen: 86.33 ug/ml m
 RT: 18.730 min Scan# 1390
 Delta R.T. 0.000 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Instrument :
 BNA_L
 LabSampleId :
 SSTD120

Tgt Ion:167 Resp: 824103
 Ion Ratio Lower Upper
 167 100
 167 87.7 80.0 120.0




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:41:49 AM



#76
 Di-n-butylphthalate
 Concen: 86.64 ug/ml
 RT: 19.305 min Scan# 1441
 Delta R.T. 0.011 min
 Lab File: BL000646.D
 Acq: 25 Jul 2012 22:21

Tgt Ion:149 Resp: 1063365
 Ion Ratio Lower Upper
 149 100
 150 8.7 0.0 29.4
 104 2.9 0.0 23.7





#77

Fluoranthene

Concen: 81.87 ug/ml

RT: 20.164 min Scan# 1517

Delta R.T. 0.000 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L

LabSampleId :

SSTD120

Tgt Ion:202 Resp: 754003

Ion Ratio Lower Upper





202 100

101 28.8 5.5 45.5

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:49 AM



#79

Pyrene

Concen: 80.18 ug/ml

RT: 20.457 min Scan# 1543

Delta R.T. 0.000 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21




Tgt Ion:202 Resp: 784332


Ion Ratio Lower Upper

202 100

101 35.1 11.7 51.7

100 28.9 6.0 46.0





#81

Butylbenzylphthalate

Concen: 84.20 ug/ml

RT: 21.225 min Scan# 1612

Delta R.T. 0.000 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L

LabSampleId :

SSTD120

Tgt Ion:149 Resp: 405839
Ion Ratio Lower Upper

149 100





91 70.7 51.8 91.8

206 16.3 0.0 36.6

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:49 AM



#82

3,3-Dichlorobenzidine

Concen: 106.24 ug/ml

RT: 21.857 min Scan# 1667

Delta R.T. 0.000 min

Lab File: BL000646.D




Acq: 25 Jul 2012 22:21


Tgt Ion:252 Resp: 168085
Ion Ratio Lower Upper

252 100

254 66.0 47.7 87.7

126 15.5 0.0 35.4





#83

Benzo[a]anthracene

Concen: 84.19 ug/ml

RT: 21.903 min Scan# 16

Delta R.T. 0.011 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L





LabSampleId :

SSTD120

Tgt	Ion:228	Resp:	667982
Ion	Ratio	Lower	Upper
228	100		
229	19.4	0.0	39.4
226	25.3	5.2	45.2

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:49 AM



#84

Chrysene

Concen: 80.90 ug/ml




RT: 21.948 min Scan# 1675


Delta R.T. 0.000 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Tgt	Ion:228	Resp:	629637
Ion	Ratio	Lower	Upper
228	100		
226	27.7	6.9	46.9
229	19.5	0.0	39.1





#85

bis(2-Ethylhexyl)phthalate

Concen: 88.73 ug/ml

RT: 21.869 min Scan# 16

Delta R.T. 0.000 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L





LabSampleId :

SSTD120

Tgt Ion:	149	Resp:	615115
Ion Ratio		Lower	Upper
149	100		
167	29.9	9.1	49.1
279	4.4	0.0	24.0

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:49 AM



#87

Di-n-octylphthalate

Concen: 91.89 ug/ml




RT: 22.693 min Scan# 1741


Delta R.T. 0.000 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Tgt Ion:	149	Resp:	1029541
Ion Ratio		Lower	Upper
149	100		
150	9.7	0.0	29.4









#88
Benzo[b]fluoranthene
Concen: 86.70 ug/ml
RT: 23.393 min Scan# 1803
Delta R.T. 0.000 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Instrument :
BNA_L
LabSampleId :
SSTD120

Tgt Ion	Ratio	Resp Lower	Resp Upper
252	100		
253	22.0	1.5	41.5
125	25.9	1.4	41.4




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:41:49 AM



#89
Benzo[k]fluoranthene
Concen: 88.46 ug/ml
RT: 23.438 min Scan# 1807
Delta R.T. 0.011 min
Lab File: BL000646.D
Acq: 25 Jul 2012 22:21

Tgt Ion	Ratio	Resp Lower	Resp Upper
252	100		
253	21.7	2.6	42.6
125	26.8	7.4	47.4





#90

Benzo[a]pyrene

Concen: 88.66 ug/ml

RT: 23.958 min Scan# 18

Delta R.T. 0.011 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L





LabSampleId :

SSTD120

Tgt	Ion	252	Resp:	628542
Ion	Ratio	Lower	Upper	
252	100			
253	21.3	1.8	41.8	
125	26.9	8.0	48.0	

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:49 AM



#91

Indeno[1,2,3-cd]pyrene

Concen: 83.61 ug/ml




RT: 26.363 min Scan# 2066


Delta R.T. 0.011 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Tgt	Ion	276	Resp:	636821
Ion	Ratio	Lower	Upper	
276	100			
138	73.5	37.7	77.7	





#92

Dibenz[a,h]anthracene

Concen: 84.48 ug/ml

RT: 26.397 min Scan# 20

Delta R.T. 0.011 min

Lab File: BL000646.D

Acq: 25 Jul 2012 22:21

Instrument :

BNA_L

LabSampleId :

SSTD120

Tgt Ion:278 Resp: 517079
Ion Ratio Lower Upper

278 100





139 41.7 18.8 58.8

279 23.4 4.5 44.5

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:41:49 AM



#93

Benzo[g,h,i]perylene

Concen: 82.75 ug/ml

RT: 27.086 min Scan# 2130

Delta R.T. 0.023 min

Lab File: BL000646.D




Acq: 25 Jul 2012 22:21

Tgt Ion:276 Resp: 553503
Ion Ratio Lower Upper

276 100

138 55.2 36.8 76.8

277 23.3 4.5 44.5



Data Path : S:\HPCHEM1\BNA_L\data\BL072512\
 Data File : BL000646.D
 Acq On : 25 Jul 2012 22:21
 Operator : HLM
 Sample : SSTD120
 Misc : SSTD120 HLM
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_L
 LabSampleId :
 SSTD120

Manual Integrations
 APPROVED

AHPatel
 9/6/2012 6:41:49 AM

Quant Time: Sep 06 05:50:35 2012
 Quant Method : H:\VOL1\GCMSDATA\L\METHODS\BL072512.M
 Quant Title : SW-846 Method 8270
 QLast Update : Thu Jul 26 10:46:15 2012
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	10.012	152	394499	40.00	ug/ml	0.00
22) Naphthalene-d8	12.316	136	1224853	40.00	ug/ml	0.00
39) Acenaphthene-d10	15.557	164	435796	40.00	ug/ml	0.00
62) Phenanthrene-d10	18.323	188	472802	40.00	ug/ml	0.00
78) Chrysene-d12	21.914	240	322971	40.00	ug/ml	0.00
86) Perylene-d12	24.048	264	276588	40.00	ug/ml	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	7.528	112	1420018	87.61	ug/ml	0.01
Spiked Amount 200.000	Range 20	- 110	Recovery	=	43.81%	
9) Phenol-d5	9.369	99	1637877	88.92	ug/ml	0.02
Spiked Amount 200.000	Range 10	- 115	Recovery	=	44.46%	
23) Nitrobenzene-d5	11.017	82	1068296	85.49	ug/ml	0.01
Spiked Amount 100.000	Range 40	- 110	Recovery	=	85.49%	
43) 2-Fluorobiphenyl	14.315	172	1214388	86.56	ug/ml	0.01
Spiked Amount 100.000	Range 50	- 110	Recovery	=	86.56%	
61) 2,4,6-Tribromophenol	17.036	330	98279	80.21	ug/ml	0.01
Spiked Amount 200.000	Range 40	- 125	Recovery	=	40.10%	
80) Terphenyl-d14	20.649	244	413665	81.72	ug/ml	0.00
Spiked Amount 100.000	Range 50	- 135	Recovery	=	81.72%	
Target Compounds						
						Qvalue
3) N-Nitrosodimethylamine	3.418	74	1517591	91.06	ug/ml	99
4) Pyridine	3.486	79	2357992m	92.21	ug/ml	
6) Benzaldehyde	9.199	77	237319	43.30	ug/ml	99
7) Aniline	9.425	93	2093107	91.43	ug/ml	99
8) bis(2-Chloroethyl)ether	9.538	93	1406645	87.44	ug/ml	99
10) Phenol	9.391	94	1820984	86.95	ug/ml	87
11) 2-Chlorophenol	9.640	128	1145031	86.24	ug/ml	99
12) 1,3-Dichlorobenzene	9.911	146	1149811	85.59	ug/ml	99
13) 1,4-Dichlorobenzene	10.046	146	1159291	86.93	ug/ml	99
14) 1,2-Dichlorobenzene	10.328	146	1045981	86.20	ug/ml	99
15) Benzyl alcohol	10.272	108	749340	90.02	ug/ml	98
16) 2,2-oxybis(1-chloropro...	10.509	45	2099134	88.20	ug/ml	98
17) 2-Methylphenol	10.464	108	1071314	87.97	ug/ml	99
18) Acetophenone	10.746	105	1155866	83.95	ug/ml	97
19) Hexachloroethane	10.938	117	450573	87.63	ug/ml	85
20) N-Nitroso-di-n-propyla...	10.769	70	777384	87.98	ug/ml	97
21) 4-Methylphenol	10.758	108	912096	82.44	ug/ml	99
24) Nitrobenzene	11.062	77	1099980	86.27	ug/ml	90
25) Isophorone	11.503	82	2098919	86.13	ug/ml	99
26) 2-Nitrophenol	11.627	139	582258	88.56	ug/ml	97
27) 2,4-Dimethylphenol	11.706	107	842898	88.78	ug/ml	99
28) Benzoic Acid	12.090	105	1439066	211.06	ug/ml	96
29) bis(2-Chloroethoxy)met...	11.875	93	1240580	83.50	ug/ml	99
30) 2,4-Dichlorophenol	12.067	162	674900	86.34	ug/ml	98
32) 1,2,4-Trichlorobenzene	12.214	180	744525	85.19	ug/ml	98
33) Naphthalene	12.361	128	2298160	82.46	ug/ml	99
34) 4-Chloroaniline	12.451	127	1018610	88.62	ug/ml	100

Data Path : S:\HPCHEM1\BNA_L\data\BL072512\
 Data File : BL000646.D
 Acq On : 25 Jul 2012 22:21
 Operator : HLM
 Sample : SSTD120
 Misc : SSTD120 HLM
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_L
 LabSampleId :
 SSTD120

Manual Integrations
 APPROVED

AHPatel
 9/6/2012 6:41:49 AM

Quant Time: Sep 06 05:50:35 2012
 Quant Method : H:\VOL1\GCMSDATA\L\METHODS\BL072512.M
 Quant Title : SW-846 Method 8270
 QLast Update : Thu Jul 26 10:46:15 2012
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) Hexachlorobutadiene	12.598	225	310225	83.86	ug/ml	99
36) Caprolactam	13.174	113	256807	85.01	ug/ml	97
37) 4-Chloro-3-methylphenol	13.355	107	639283	87.83	ug/ml	95
38) 2-Methylnaphthalene	13.626	142	1283121	83.45	ug/ml	100
40) Hexachlorocyclopentadiene	13.931	237	287981	94.37	ug/ml	98
41) 2,4,6-Trichlorophenol	14.156	196	370401	88.79	ug/ml	97
42) 2,4,5-Trichlorophenol	14.224	196	393808	87.42	ug/ml	99
44) 1,1-Biphenyl	14.495	154	1313979	84.65	ug/ml	96
45) 2-Chloronaphthalene	14.529	162	1078873	84.59	ug/ml	97
46) 2-Nitroaniline	14.721	65	329358	84.14	ug/ml	94
47) Acenaphthylene	15.297	152	1561321	85.20	ug/ml	99
48) Dimethylphthalate	15.082	163	1049157	82.83	ug/ml	99
49) 2,6-Dinitrotoluene	15.173	165	286334	86.37	ug/ml	99
50) Acenaphthene	15.624	153	910046	82.61	ug/ml	100
51) 3-Nitroaniline	15.489	138	281720	95.07	ug/ml	99
52) 2,4-Dinitrophenol	15.681	184	173957	104.04	ug/ml	99
53) Dibenzofuran	15.941	168	1202867	84.22	ug/ml	97
54) 2,4-Dinitrotoluene	15.929	165	297400	83.38	ug/ml	92
55) 4-Nitrophenol	15.805	139	223716	88.66	ug/ml	93
56) 2,3,4,6-Tetrachlorophenol	16.178	232	233515	87.89	ug/ml	98
57) Fluorene	16.584	166	790301	76.43	ug/ml	99
58) 4-Chlorophenyl-phenyle...	16.584	204	354589	77.20	ug/ml	94
59) Diethylphthalate	16.392	149	826188	80.23	ug/ml	99
60) 4-Nitroaniline	16.629	138	194416	92.72	ug/ml	97
63) 1,2,4,5-Tetrachloroben...	13.953	216	504378	90.32	ug/ml	100
64) 4,6-Dinitro-2-methylph...	16.686	198	179279	98.53	ug/ml	100
65) n-Nitrosodiphenylamine	16.810	169	719331	87.68	ug/ml	99
66) 1,2-Diphenylhydrazine	16.889	77	1084673	87.19	ug/ml	93
67) 4-Bromophenyl-phenylether	17.499	248	206234	87.14	ug/ml	97
68) Hexachlorobenzene	17.623	284	209409	83.00	ug/ml	90
69) Atrazine	17.838	200	146266	91.63	ug/ml	99
70) Pentachlorophenol	17.984	266	136771	103.85	ug/ml	99
72) Benzidine	20.333	184	67055	80.06	ug/ml	100
73) Phenanthrene	18.368	178	891485	82.65	ug/ml	99
74) Anthracene	18.459	178	909225	82.16	ug/ml	98
75) Carbazole	18.730	167	824103m	86.33	ug/ml	
76) Di-n-butylphthalate	19.305	149	1063365	86.64	ug/ml	98
77) Fluoranthene	20.164	202	754003	81.87	ug/ml	93
79) Pyrene	20.457	202	784332	80.18	ug/ml	94
81) Butylbenzylphthalate	21.225	149	405839	84.20	ug/ml	99
82) 3,3-Dichlorobenzidine	21.857	252	168085	106.24	ug/ml	98
83) Benzo[a]anthracene	21.903	228	667982	84.19	ug/ml	100
84) Chrysene	21.948	228	629637	80.90	ug/ml	99
85) bis(2-Ethylhexyl)phtha...	21.869	149	615115	88.73	ug/ml	99
87) Di-n-octylphthalate	22.693	149	1029541	91.89	ug/ml	99
88) Benzo[b]fluoranthene	23.393	252	668383	86.70	ug/ml	95
89) Benzo[k]fluoranthene	23.438	252	599387	88.46	ug/ml	98
90) Benzo[a]pyrene	23.958	252	628542	88.66	ug/ml	98
91) Indeno[1,2,3-cd]pyrene	26.363	276	636821	83.61	ug/ml	79
92) Dibenz[a,h]anthracene	26.397	278	517079	84.48	ug/ml	96

Data Path : S:\HPCHEM1\BNA_L\data\BL072512\
Data File : BL000646.D
Acq On : 25 Jul 2012 22:21
Operator : HLM
Sample : SSTD120
Misc : SSTD120 HLM
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_L
LabSampleId :
SSTD120

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:41:49 AM

Quant Time: Sep 06 05:50:35 2012
Quant Method : H:\VOL1\GCMSDATA\L\METHODS\BL072512.M
Quant Title : SW-846 Method 8270
QLast Update : Thu Jul 26 10:46:15 2012
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) Benzo[g,h,i]perylene	27.086	276	553503	82.75	ug/ml	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed