

#3

N-Nitrosodimethylamine

Concen: 55.00 ug/ml

RT: 3.358 min Scan# 29

Delta R.T. 0.001 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L

LabSampleId :

SSTD050

Tgt Ion: 74 Resp: 697213

Ion Ratio Lower Upper





74 100

43 24.1 2.3 42.3

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:42:30 AM



#4

Pyridine

Concen: 56.21 ug/ml m

RT: 3.460 min Scan# 38

Delta R.T. 0.001 min

Lab File: BL000657.D




Acq: 26 Jul 2012 11:15


Tgt Ion: 79 Resp: 1087860

Ion Ratio Lower Upper

79 100

52 71.8 47.6 87.6





#6

Benzaldehyde

Concen: 110.92 ug/ml

RT: 9.208 min Scan# 54

Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L

LabSampleId :

SSTD050

Tgt Ion: 77 Resp: 315603
Ion Ratio Lower Upper

77 100





105 88.4 70.4 110.4

106 89.0 70.4 110.4

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:42:30 AM



#7

Aniline

Concen: 52.60 ug/ml

RT: 9.411 min Scan# 565

Delta R.T. 0.001 min




Lab File: BL000657.D


Acq: 26 Jul 2012 11:15

Tgt Ion: 93 Resp: 970919
Ion Ratio Lower Upper

93 100

66 40.8 30.6 46.0





#8

bis(2-Chloroethyl)ether

Concen: 52.38 ug/ml

RT: 9.535 min Scan# 57

Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L

LabSampleId :

SSTD050

Tgt Ion: 93 Resp: 648029
Ion Ratio Lower Upper

93 100





63 78.6 58.9 98.9

95 31.6 11.9 51.9

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:42:30 AM



#10

Phenol

Concen: 53.63 ug/ml

RT: 9.377 min Scan# 562

Delta R.T. 0.012 min

Lab File: BL000657.D




Acq: 26 Jul 2012 11:15


Tgt Ion: 94 Resp: 885939
Ion Ratio Lower Upper

94 100

65 21.2 1.6 41.6

66 23.4 4.3 44.3









#11
2-Chlorophenol
Concen: 51.20 ug/ml
RT: 9.637 min Scan# 58
Delta R.T. 0.012 min
Lab File: BL000657.D
Acq: 26 Jul 2012 11:15

Instrument :
BNA_L
LabSampleId :
SSTD050

Tgt Ion	Ratio	Resp Lower	Resp Upper
128	100		
64	58.9	37.3	77.3
130	33.0	12.4	52.4




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:42:30 AM



#12
1,3-Dichlorobenzene
Concen: 51.53 ug/ml
RT: 9.919 min Scan# 610
Delta R.T. 0.012 min
Lab File: BL000657.D
Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp Lower	Resp Upper
146	100		
148	64.7	44.0	84.0
113	13.0	0.0	32.5





#13

1,4-Dichlorobenzene

Concen: 51.78 ug/ml

RT: 10.043 min Scan# 621

Delta R.T. 0.001 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L





LabSampleId :

SSTD050

Tgt Ion:	146	Resp:	534673
Ion Ratio	Lower	Upper	
146	100		
148	63.2	44.6	84.6
113	12.9	0.0	32.8

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:42:30 AM



#14

1,2-Dichlorobenzene

Concen: 52.61 ug/ml




RT: 10.326 min Scan# 646


Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Tgt Ion:	146	Resp:	496874
Ion Ratio	Lower	Upper	
146	100		
148	64.5	44.0	84.0
113	13.7	0.0	33.6





#15

Benzyl alcohol

Concen: 53.31 ug/ml

RT: 10.269 min Scan# 641

Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L





LabSampleId :

SSTD050

Tgt Ion:	108	Resp:	352657
Ion Ratio	Lower	Upper	
108	100		
79	114.1	94.5	134.5
77	65.4	43.7	83.7

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:42:30 AM



#16

2,2-oxybis(1-chloropropane)

Concen: 55.83 ug/ml




RT: 10.506 min Scan# 662


Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Tgt Ion:	45	Resp:	1069654
Ion Ratio	Lower	Upper	
45	100		
77	9.4	0.0	30.0
79	8.4	0.0	28.9









#17
 2-Methylphenol
 Concen: 52.29 ug/ml
 RT: 10.461 min Scan# 658
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Instrument :
 BNA_L
 LabSampleId :
 SST050

Tgt Ion:108 Resp: 496362
 Ion Ratio Lower Upper
 108 100
 107 87.8 67.8 107.8




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:42:30 AM



#18
 Acetophenone
 Concen: 50.31 ug/ml
 RT: 10.743 min Scan# 683
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Tgt Ion:105 Resp: 559368
 Ion Ratio Lower Upper
 105 100
 71 8.2 0.0 27.0
 51 36.2 16.6 56.6









#19
Hexachloroethane
Concen: 50.76 ug/ml
RT: 10.935 min Scan# 700
Delta R.T. 0.001 min
Lab File: BL000657.D
Acq: 26 Jul 2012 11:15

Instrument :
BNA_L
LabSampleId :
SSTD050

Tgt Ion	Ratio	Resp Lower	Resp Upper
117	100		
200	0.0	0.0	20.0
199	41.3	21.2	61.2




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:42:30 AM



#20
N-Nitroso-di-n-propylamine
Concen: 52.54 ug/ml
RT: 10.755 min Scan# 684
Delta R.T. 0.012 min
Lab File: BL000657.D
Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp Lower	Resp Upper
70	100		
42	52.4	32.8	72.8
130	24.5	3.0	43.0









#21
 4-Methylphenol
 Concen: 52.57 ug/ml
 RT: 10.743 min Scan# 68
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Instrument :
 BNA_L
 LabSampleId :
 SST050

Tgt Ion	Ratio	Resp	Lower	Upper
108	100	467524		
107	108.8	86.9	126.9	




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:42:30 AM



#24
 Nitrobenzene
 Concen: 53.58 ug/ml
 RT: 11.048 min Scan# 710
 Delta R.T. 0.001 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp	Lower	Upper
77	100	518078		
123	48.5	26.9	66.9	
65	11.8	0.0	32.1	





#25

Isophorone

Concen: 51.53 ug/ml

RT: 11.489 min Scan# 749

Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L

LabSampleId :

SSTD050

Tgt Ion: 82 Resp: 959422

Ion Ratio Lower Upper





82 100

138 15.4 0.0 35.5

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:42:30 AM



#26

2-Nitrophenol

Concen: 51.83 ug/ml

RT: 11.624 min Scan# 761

Delta R.T. 0.001 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15




Tgt Ion: 139 Resp: 254246


Ion Ratio Lower Upper

139 100

65 42.3 20.8 60.8

109 13.8 0.0 33.5









#27
 2,4-Dimethylphenol
 Concen: 52.58 ug/ml
 RT: 11.703 min Scan# 76
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Instrument :
 BNA_L
 LabSampleId :
 SST050

Tgt Ion:	107	Resp:	390558
Ion Ratio	Lower	Upper	
107	100		
121	53.9	33.2	73.2
122	94.3	75.9	115.9




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:42:30 AM



#28
 Benzoic Acid
 Concen: 91.57 ug/ml
 RT: 11.974 min Scan# 792
 Delta R.T. -0.022 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Tgt Ion:	105	Resp:	526885
Ion Ratio	Lower	Upper	
105	100		
77	74.1	52.0	92.0
122	85.3	61.6	101.6









#29
bis(2-Chloroethoxy)methane
Concen: 52.17 ug/ml
RT: 11.872 min Scan# 783
Delta R.T. 0.012 min
Lab File: BL000657.D
Acq: 26 Jul 2012 11:15

Instrument :
BNA_L
LabSampleId :
SSTD050

Tgt Ion	Ratio	Resp	Lower	Upper
93	100	579505		
95	31.9	11.6	51.6	
123	10.9	0.0	30.8	




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:42:30 AM



#30
2,4-Dichlorophenol
Concen: 50.90 ug/ml
RT: 12.064 min Scan# 800
Delta R.T. 0.012 min
Lab File: BL000657.D
Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp	Lower	Upper
162	100	304163		
164	64.7	43.5	83.5	
98	38.0	19.5	59.5	





#32

1,2,4-Trichlorobenzene

Concen: 50.87 ug/ml

RT: 12.223 min Scan# 81

Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L

LabSampleId :

SSTD050

Tgt Ion:180 Resp: 334802
Ion Ratio Lower Upper

180 100





182 96.9 75.6 115.6

145 28.9 8.3 48.3

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:42:30 AM



#33

Naphthalene

Concen: 52.87 ug/ml

RT: 12.369 min Scan# 827

Delta R.T. 0.012 min

Lab File: BL000657.D




Acq: 26 Jul 2012 11:15


Tgt Ion:128 Resp: 1117388
Ion Ratio Lower Upper

128 100

129 10.6 0.0 30.7

127 13.1 0.0 33.1









#34
 4-Chloroaniline
 Concen: 52.61 ug/ml
 RT: 12.460 min Scan# 834
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Instrument :
 BNA_L
 LabSampleId :
 SST050

Tgt Ion:127 Resp: 471194
 Ion Ratio Lower Upper
 127 100
 129 32.6 12.1 52.1




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:42:30 AM



#35
 Hexachlorobutadiene
 Concen: 52.52 ug/ml
 RT: 12.606 min Scan# 848
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Tgt Ion:225 Resp: 147653
 Ion Ratio Lower Upper
 225 100
 223 61.0 41.7 81.7
 227 62.6 45.2 85.2





#36

Caprolactam

Concen: 50.06 ug/ml

RT: 13.126 min Scan# 89

Delta R.T. 0.001 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L





LabSampleId :

SSTD050

Tgt Ion:	113	Resp:	119664
Ion Ratio	Lower	Upper	
113	100		
55	159.7	144.2	184.2
56	107.9	92.3	132.3

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:42:30 AM



#37

4-Chloro-3-methylphenol

Concen: 52.10 ug/ml




RT: 13.352 min Scan# 914


Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Tgt Ion:	107	Resp:	290834
Ion Ratio	Lower	Upper	
107	100		
144	29.5	9.7	49.7
142	91.3	71.3	111.3









#38
 2-Methylnaphthalene
 Concen: 52.66 ug/ml
 RT: 13.634 min Scan# 938
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Instrument :
 BNA_L
 LabSampleId :
 SST050

Tgt Ion	Ratio	Resp	Lower	Upper
142	100	619057		
141	83.7	64.5	104.5	




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:42:30 AM



#40
 Hexachlorocyclopentadiene
 Concen: 48.35 ug/ml
 RT: 13.939 min Scan# 966
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp	Lower	Upper
237	100	123873		
235	63.3	42.1	82.1	
272	14.4	0.0	35.7	









#41
 2,4,6-Trichlorophenol
 Concen: 51.20 ug/ml
 RT: 14.153 min Scan# 98
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Instrument :
 BNA_L
 LabSampleId :
 SST050

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	166294		
198	94.4	73.1	113.1	
200	30.0	9.7	49.7	




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:42:30 AM



#42
 2,4,5-Trichlorophenol
 Concen: 50.15 ug/ml
 RT: 14.221 min Scan# 991
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	177666		
198	98.3	74.5	114.5	
200	31.3	10.8	50.8	









#44
 1,1-Biphenyl
 Concen: 51.71 ug/ml
 RT: 14.492 min Scan# 1015
 Delta R.T. 0.001 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Instrument :
 BNA_L
 LabSampleId :
 SST050

Tgt Ion	Ratio	Resp Lower	Resp Upper
154	100		
153	40.7	19.9	59.9
76	30.6	3.3	43.3




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:42:30 AM



#45
 2-Chloronaphthalene
 Concen: 51.36 ug/ml
 RT: 14.526 min Scan# 1018
 Delta R.T. 0.001 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp Lower	Resp Upper
162	100		
164	31.6	12.8	52.8
127	34.8	10.9	50.9









#46
 2-Nitroaniline
 Concen: 53.04 ug/ml
 RT: 14.718 min Scan# 1035
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Instrument :
 BNA_L
 LabSampleId :
 SST050

Tgt Ion:	65	Resp:	165903
Ion Ratio	Lower	Upper	
65	100		
92	67.4	49.2	89.2
138	119.6	102.4	142.4




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:42:30 AM



#47
 Acenaphthylene
 Concen: 51.40 ug/ml
 RT: 15.305 min Scan# 1087
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Tgt Ion:	152	Resp:	755559
Ion Ratio	Lower	Upper	
152	100		
151	19.2	0.0	39.9
153	12.7	0.0	32.6





#48

Dimethylphthalate

Concen: 49.22 ug/ml

RT: 15.079 min Scan# 1067

Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L





LabSampleId :

SSTD050

Tgt Ion:	163	Resp:	496523
Ion Ratio		Lower	Upper
163	100		
194	6.1	0.0	25.9
164	9.5	0.0	29.6

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:42:30 AM



#49

2,6-Dinitrotoluene

Concen: 49.26 ug/ml




RT: 15.170 min Scan# 1075


Delta R.T. 0.001 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Tgt Ion:	165	Resp:	130956
Ion Ratio		Lower	Upper
165	100		
89	38.6	19.2	59.2
121	16.1	0.0	36.2





#50

Acenaphthene

Concen: 52.26 ug/ml

RT: 15.633 min Scan# 1115

Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L





LabSampleId :

SSTD050

Tgt Ion	Ratio	Resp	Lower	Upper
153	100	455792		
152	45.9	25.1	65.1	
154	95.7	75.9	115.9	

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:42:30 AM



#51

3-Nitroaniline

Concen: 53.17 ug/ml




RT: 15.486 min Scan# 1103


Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	135861		
108	9.9	0.0	30.0	
92	100.6	81.9	121.9	





#52

2,4-Dinitrophenol

Concen: 43.13 ug/ml

RT: 15.678 min Scan# 11

Delta R.T. 0.013 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L





LabSampleId :

SSTD050

Tgt Ion	Ratio	Resp	Lower	Upper
184	100	60946		
63	30.8		11.7	51.7
154	27.0		6.0	46.0

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:42:30 AM



#53

Dibenzofuran

Concen: 50.84 ug/ml




RT: 15.949 min Scan# 1144


Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp	Lower	Upper
168	100	589202		
139	27.3		7.3	47.3





#54

2,4-Dinitrotoluene

Concen: 50.46 ug/ml

RT: 15.926 min Scan# 1142

Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L





LabSampleId :

SSTD050

Tgt Ion:	165	Resp:	148352
Ion Ratio	Lower	Upper	
165	100		
63	30.5	13.1	53.1
182	6.1	0.0	26.1

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:42:30 AM



#55

4-Nitrophenol

Concen: 47.47 ug/ml




RT: 15.802 min Scan# 1131


Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Tgt Ion:	139	Resp:	97009
Ion Ratio	Lower	Upper	
139	100		
65	62.7	44.1	84.1
109	27.3	6.8	46.8









#56
 2,3,4,6-Tetrachlorophenol
 Concen: 50.15 ug/ml
 RT: 16.186 min Scan# 1165
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Instrument :
 BNA_L
 LabSampleId :
 SST050

Tgt Ion	Ratio	Resp Lower	Resp Upper
232	100		
230	77.1	64.6	96.8
131	34.9	29.2	43.8
166	25.1	20.6	30.8




Manual Integrations
 APPROVED




AHPatel
 9/6/2012 6:42:30 AM



#57
 Fluorene
 Concen: 52.32 ug/ml
 RT: 16.581 min Scan# 1200
 Delta R.T. 0.001 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp Lower	Resp Upper
166	100		
165	91.2	72.3	112.3
167	13.3	0.0	33.1









#58
4-Chlorophenyl-phenylether
Concen: 51.43 ug/ml
RT: 16.592 min Scan# 1201
Delta R.T. 0.012 min
Lab File: BL000657.D
Acq: 26 Jul 2012 11:15

Instrument :
BNA_L
LabSampleId :
SSTD050

Tgt Ion	Ratio	Resp Lower	Resp Upper
204	100		
206	32.9	12.0	52.0
141	45.6	26.4	66.4


Manual Integrations
APPROVED


AHPatel
9/6/2012 6:42:30 AM



#59
Diethylphthalate
Concen: 48.29 ug/ml
RT: 16.389 min Scan# 1183
Delta R.T. 0.001 min
Lab File: BL000657.D
Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
177	21.4	1.3	41.3
150	11.6	0.0	31.6









#60
4-Nitroaniline
Concen: 49.87 ug/ml
RT: 16.615 min Scan# 1204
Delta R.T. 0.001 min
Lab File: BL000657.D
Acq: 26 Jul 2012 11:15

Instrument :
BNA_L
LabSampleId :
SSTD050

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	98492		
92	43.5	16.5	56.5	
108	29.5	8.1	48.1	




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:42:30 AM



#63
1,2,4,5-Tetrachlorobenzene
Concen: 55.91 ug/ml
RT: 13.950 min Scan# 967
Delta R.T. 0.012 min
Lab File: BL000657.D
Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp	Lower	Upper
216	100	246422		
214	78.7	64.0	96.0	
179	18.0	14.6	22.0	
218	47.8	38.6	57.8	





#64

4,6-Dinitro-2-methylphenol

Concen: 46.94 ug/ml

RT: 16.683 min Scan# 1210

Delta R.T. 0.001 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L





LabSampleId :

SSTD050

Tgt Ion:198 Resp: 72294

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:42:30 AM



#65

n-Nitrosodiphenylamine

Concen: 55.04 ug/ml

RT: 16.818 min Scan# 1221

Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15




Tgt Ion:169 Resp: 363354


Ion Ratio Lower Upper

169 100

168 61.8 40.9 80.9

167 31.7 12.5 52.5









#66
 1,2-Diphenylhydrazine
 Concen: 55.63 ug/ml
 RT: 16.886 min Scan# 1227
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Instrument :
 BNA_L
 LabSampleId :
 SST050

Tgt Ion	Ratio	Resp Lower	Resp Upper
77	100		
182	20.4	0.0	39.4
105	14.3	0.0	34.5




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:42:30 AM



#67
 4-Bromophenyl-phenylether
 Concen: 54.04 ug/ml
 RT: 17.507 min Scan# 1282
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp Lower	Resp Upper
248	100		
250	99.1	78.2	118.2
141	52.8	34.7	74.7









#68
Hexachlorobenzene
Concen: 54.54 ug/ml
RT: 17.631 min Scan# 1293
Delta R.T. 0.012 min
Lab File: BL000657.D
Acq: 26 Jul 2012 11:15

Instrument :
BNA_L
LabSampleId :
SSTD050

Tgt Ion	Ratio	Resp Lower	Resp Upper
284	100		
142	30.6	14.3	54.3
249	25.7	6.7	46.7




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:42:30 AM



#69
Atrazine
Concen: 48.49 ug/ml
RT: 17.835 min Scan# 1311
Delta R.T. 0.012 min
Lab File: BL000657.D
Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp Lower	Resp Upper
200	100		
173	30.0	10.7	50.7
115	0.0	0.0	20.0









#70
 Pentachlorophenol
 Concen: 45.31 ug/ml
 RT: 17.993 min Scan# 1325
 Delta R.T. 0.013 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Instrument :
 BNA_L
 LabSampleId :
 SST050

Tgt Ion	Ratio	Resp Lower	Resp Upper
266	100		
264	60.8	44.5	84.5
268	61.9	45.0	85.0




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:42:30 AM



#73
 Phenanthrene
 Concen: 52.05 ug/ml
 RT: 18.377 min Scan# 1359
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp Lower	Resp Upper
178	100		
179	14.9	0.0	35.3
176	18.3	0.0	38.6









#74
 Anthracene
 Concen: 52.87 ug/ml
 RT: 18.456 min Scan# 1366
 Delta R.T. 0.001 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Instrument :
 BNA_L
 LabSampleId :
 SST050

Tgt Ion	Ratio	Resp	Lower	Upper
178	100	479774		
179	15.6		0.0	34.9
176	17.4		0.0	37.0




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:42:30 AM



#75
 Carbazole
 Concen: 50.15 ug/ml m
 RT: 18.727 min Scan# 1390
 Delta R.T. 0.001 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp	Lower	Upper
167	100	390475		
167	91.1		80.0	120.0





#76

Di-n-butylphthalate

Concen: 47.51 ug/ml

RT: 19.303 min Scan# 14

Delta R.T. 0.001 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L





LabSampleId :

SSTD050

Tgt Ion:	149	Resp:	490452
Ion Ratio	Lower	Upper	
149	100		
150	9.2	0.0	29.4
104	3.5	0.0	23.7

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:42:30 AM



#77

Fluoranthene

Concen: 50.09 ug/ml




RT: 20.172 min Scan# 1518


Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Tgt Ion:	202	Resp:	373607
Ion Ratio	Lower	Upper	
202	100		
101	20.2	5.5	45.5









#79
 Pyrene
 Concen: 45.37 ug/ml
 RT: 20.466 min Scan# 1544
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Instrument :
 BNA_L
 LabSampleId :
 SST050

Tgt Ion	Ratio	Resp Lower	Resp Upper
202	100		
101	25.2	11.7	51.7
100	19.8	6.0	46.0




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:42:30 AM



#81
 Butylbenzylphthalate
 Concen: 47.37 ug/ml
 RT: 21.233 min Scan# 1612
 Delta R.T. 0.012 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
91	62.5	51.8	91.8
206	19.5	0.0	36.6





#82

3,3-Dichlorobenzidine

Concen: 57.94 ug/ml

RT: 21.854 min Scan# 1667

Delta R.T. 0.001 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L





LabSampleId :

SSTD050

Tgt Ion:	252	Resp:	81519
Ion Ratio	Lower	Upper	
252	100		
254	64.3	47.7	87.7
126	21.7	0.0	35.4

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:42:30 AM



#83

Benzo[a]anthracene

Concen: 50.59 ug/ml




RT: 21.900 min Scan# 1671


Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Tgt Ion:	228	Resp:	320113
Ion Ratio	Lower	Upper	
228	100		
229	19.3	0.0	39.4
226	25.0	5.2	45.2









#84
 Chrysene
 Concen: 53.04 ug/ml
 RT: 21.945 min Scan# 1675
 Delta R.T. 0.001 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Instrument :
 BNA_L
 LabSampleId :
 SST050

Tgt Ion	Ratio	Resp Lower	Resp Upper
228	100		
226	27.0	6.9	46.9
229	19.3	0.0	39.1




Manual Integrations
 APPROVED


AHPatel
 9/6/2012 6:42:30 AM



#85
 bis(2-Ethylhexyl)phthalate
 Concen: 48.88 ug/ml
 RT: 21.877 min Scan# 1669
 Delta R.T. 0.001 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
167	31.3	9.1	49.1
279	5.5	0.0	24.0





#87

Di-n-octylphthalate

Concen: 46.57 ug/ml

RT: 22.690 min Scan# 1741

Delta R.T. 0.001 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L

LabSampleId :

SSTD050

Tgt Ion:149 Resp: 493099

Ion Ratio Lower Upper





149 100

150 10.0 0.0 29.4

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:42:30 AM



#88

Benzo[b]fluoranthene

Concen: 48.48 ug/ml

RT: 23.390 min Scan# 1803

Delta R.T. 0.012 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15




Tgt Ion:252 Resp: 321707


Ion Ratio Lower Upper

252 100

253 22.2 1.5 41.5

125 27.2 1.4 41.4









#89
Benzo[k]fluoranthene
Concen: 54.23 ug/ml
RT: 23.435 min Scan# 1807
Delta R.T. 0.012 min
Lab File: BL000657.D
Acq: 26 Jul 2012 11:15

Instrument :
BNA_L
LabSampleId :
SSTD050

Tgt Ion	Ratio	Resp Lower	Resp Upper
252	100		
253	21.9	2.6	42.6
125	26.2	7.4	47.4




Manual Integrations
APPROVED


AHPatel
9/6/2012 6:42:30 AM



#90
Benzo[a]pyrene
Concen: 51.17 ug/ml
RT: 23.955 min Scan# 1853
Delta R.T. 0.012 min
Lab File: BL000657.D
Acq: 26 Jul 2012 11:15

Tgt Ion	Ratio	Resp Lower	Resp Upper
252	100		
253	22.0	1.8	41.8
125	27.7	8.0	48.0





#91

Indeno[1,2,3-cd]pyrene

Concen: 53.63 ug/ml

RT: 26.360 min Scan# 20

Delta R.T. 0.023 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15

Instrument :

BNA_L

LabSampleId :

SSTD050

Tgt Ion:276 Resp: 321847

Ion Ratio Lower Upper





276 100

138 56.4 37.7 77.7

Manual Integrations
APPROVED

AHPatel

9/6/2012 6:42:30 AM



#92

Dibenzo[a,h]anthracene

Concen: 52.83 ug/ml

RT: 26.394 min Scan# 2069

Delta R.T. 0.024 min

Lab File: BL000657.D

Acq: 26 Jul 2012 11:15




Tgt Ion:278 Resp: 265118


Ion Ratio Lower Upper

278 100

139 42.3 18.8 58.8


279 24.0 4.5 44.5





#93
 Benzo[g,h,i]perylene
 Concen: 52.73 ug/ml
 RT: 27.083 min Scan# 21
 Delta R.T. 0.023 min
 Lab File: BL000657.D
 Acq: 26 Jul 2012 11:15



Instrument :
 BNA_L
 LabSampleId :
 SST050



Tgt	Ion	276	Resp:	275566
	Ratio		Lower	Upper
276	100			
138	53.7	36.8	76.8	
277	23.0	4.5	44.5	

Manual Integrations
 APPROVED

AHPatel
 9/6/2012 6:42:30 AM



Data Path : S:\HPCHEM1\BNA_L\data\BL072612\
 Data File : BL000657.D
 Acq On : 26 Jul 2012 11:15
 Operator : HLM
 Sample : SSTD050
 Misc : SSTD050 HLM
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_L
 LabSampleId :
 SSTD050

Manual Integrations
 APPROVED

AHPatel
 9/6/2012 6:42:30 AM

Quant Time: Sep 06 06:42:24 2012
 Quant Method : H:\VOL1\GCMSDATA\L\METHODS\BL072512.M
 Quant Title : SW-846 Method 8270
 QLast Update : Thu Jul 26 11:02:15 2012
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	10.021	152	394373	40.00	ug/ml	0.01
22) Naphthalene-d8	12.324	136	1225601	40.00	ug/ml	0.01
39) Acenaphthene-d10	15.565	164	459925	40.00	ug/ml	0.01
62) Phenanthrene-d10	18.331	188	520957	40.00	ug/ml	0.01
78) Chrysene-d12	21.911	240	349728	40.00	ug/ml	0.00
86) Perylene-d12	24.056	264	314462	40.00	ug/ml	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	7.525	112	662414	53.59	ug/ml	0.01
Spiked Amount 200.000	Range 20	- 110	Recovery	=	26.80%	
9) Phenol-d5	9.354	99	755510	52.90	ug/ml	0.01
Spiked Amount 200.000	Range 10	- 115	Recovery	=	26.45%	
23) Nitrobenzene-d5	11.014	82	491585	52.72	ug/ml	0.01
Spiked Amount 100.000	Range 40	- 110	Recovery	=	52.72%	
43) 2-Fluorobiphenyl	14.312	172	568578	51.59	ug/ml	0.01
Spiked Amount 100.000	Range 50	- 110	Recovery	=	51.59%	
61) 2,4,6-Tribromophenol	17.033	330	50189	51.42	ug/ml	0.00
Spiked Amount 200.000	Range 40	- 125	Recovery	=	25.71%#	
80) Terphenyl-d14	20.658	244	197689	43.39	ug/ml	0.01
Spiked Amount 100.000	Range 50	- 135	Recovery	=	43.39%#	
Target Compounds						
						Qvalue
3) N-Nitrosodimethylamine	3.358	74	697213	55.00	ug/ml	96
4) Pyridine	3.460	79	1087860m	56.21	ug/ml	
6) Benzaldehyde	9.208	77	315603	110.92	ug/ml	98
7) Aniline	9.411	93	970919	52.60	ug/ml	96
8) bis(2-Chloroethyl)ether	9.535	93	648029	52.38	ug/ml	100
10) Phenol	9.377	94	885939	53.63	ug/ml	99
11) 2-Chlorophenol	9.637	128	522570	51.20	ug/ml	98
12) 1,3-Dichlorobenzene	9.919	146	532352	51.53	ug/ml	99
13) 1,4-Dichlorobenzene	10.043	146	534673	51.78	ug/ml	98
14) 1,2-Dichlorobenzene	10.326	146	496874	52.61	ug/ml	99
15) Benzyl alcohol	10.269	108	352657	53.31	ug/ml	99
16) 2,2-oxybis(1-chloropro...	10.506	45	1069654	55.83	ug/ml	98
17) 2-Methylphenol	10.461	108	496362	52.29	ug/ml	100
18) Acetophenone	10.743	105	559368	50.31	ug/ml	99
19) Hexachloroethane	10.935	117	203028	50.76	ug/ml	100
20) N-Nitroso-di-n-propyla...	10.755	70	375240	52.54	ug/ml	99
21) 4-Methylphenol	10.743	108	467524	52.57	ug/ml	98
24) Nitrobenzene	11.048	77	518078	53.58	ug/ml	98
25) Isophorone	11.489	82	959422	51.53	ug/ml	100
26) 2-Nitrophenol	11.624	139	254246	51.83	ug/ml	98
27) 2,4-Dimethylphenol	11.703	107	390558	52.58	ug/ml	99
28) Benzoic Acid	11.974	105	526885	91.57	ug/ml	97
29) bis(2-Chloroethoxy)met...	11.872	93	579505	52.17	ug/ml	100
30) 2,4-Dichlorophenol	12.064	162	304163	50.90	ug/ml	98
32) 1,2,4-Trichlorobenzene	12.223	180	334802	50.87	ug/ml	99
33) Naphthalene	12.369	128	1117388	52.87	ug/ml	100
34) 4-Chloroaniline	12.460	127	471194	52.61	ug/ml	99

Data Path : S:\HPCHEM1\BNA_L\data\BL072612\
 Data File : BL000657.D
 Acq On : 26 Jul 2012 11:15
 Operator : HLM
 Sample : SSTD050
 Misc : SSTD050 HLM
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_L
 LabSampleId :
 SSTD050

Manual Integrations
 APPROVED

AHPatel
 9/6/2012 6:42:30 AM

Quant Time: Sep 06 06:42:24 2012
 Quant Method : H:\VOL1\GCMSDATA\L\METHODS\BL072512.M
 Quant Title : SW-846 Method 8270
 QLast Update : Thu Jul 26 11:02:15 2012
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) Hexachlorobutadiene	12.606	225	147653	52.52	ug/ml	98
36) Caprolactam	13.126	113	119664	50.06	ug/ml	96
37) 4-Chloro-3-methylphenol	13.352	107	290834	52.10	ug/ml	100
38) 2-Methylnaphthalene	13.634	142	619057	52.66	ug/ml	99
40) Hexachlorocyclopentadiene	13.939	237	123873	48.35	ug/ml	98
41) 2,4,6-Trichlorophenol	14.153	196	166294	51.20	ug/ml	99
42) 2,4,5-Trichlorophenol	14.221	196	177666	50.15	ug/ml	97
44) 1,1-Biphenyl	14.492	154	645578	51.71	ug/ml	94
45) 2-Chloronaphthalene	14.526	162	523743	51.36	ug/ml	95
46) 2-Nitroaniline	14.718	65	165903	53.04	ug/ml	98
47) Acenaphthylene	15.305	152	755559	51.40	ug/ml	99
48) Dimethylphthalate	15.079	163	496523	49.22	ug/ml	100
49) 2,6-Dinitrotoluene	15.170	165	130956	49.26	ug/ml	99
50) Acenaphthene	15.633	153	455792	52.26	ug/ml	99
51) 3-Nitroaniline	15.486	138	135861	53.17	ug/ml	99
52) 2,4-Dinitrophenol	15.678	184	60946	43.13	ug/ml	98
53) Dibenzofuran	15.949	168	589202	50.84	ug/ml	100
54) 2,4-Dinitrotoluene	15.926	165	148352	50.46	ug/ml	96
55) 4-Nitrophenol	15.802	139	97009	47.47	ug/ml	98
56) 2,3,4,6-Tetrachlorophenol	16.186	232	107601	50.15	ug/ml	97
57) Fluorene	16.581	166	432000	52.32	ug/ml	99
58) 4-Chlorophenyl-phenyle...	16.592	204	193160	51.43	ug/ml	99
59) Diethylphthalate	16.389	149	404176	48.29	ug/ml	100
60) 4-Nitroaniline	16.615	138	98492	49.87	ug/ml	92
63) 1,2,4,5-Tetrachloroben...	13.950	216	246422	55.91	ug/ml	99
64) 4,6-Dinitro-2-methylph...	16.683	198	72294	46.94	ug/ml	100
65) n-Nitrosodiphenylamine	16.818	169	363354	55.04	ug/ml	99
66) 1,2-Diphenylhydrazine	16.886	77	554892	55.63	ug/ml	99
67) 4-Bromophenyl-phenylether	17.507	248	103728	54.04	ug/ml	98
68) Hexachlorobenzene	17.631	284	110157	54.54	ug/ml	95
69) Atrazine	17.835	200	69828	48.49	ug/ml	99
70) Pentachlorophenol	17.993	266	50193	45.31	ug/ml	96
73) Phenanthrene	18.377	178	458812	52.05	ug/ml	99
74) Anthracene	18.456	178	479774	52.87	ug/ml	99
75) Carbazole	18.727	167	390475m	50.15	ug/ml	
76) Di-n-butylphthalate	19.303	149	490452	47.51	ug/ml	99
77) Fluoranthene	20.172	202	373607	50.09	ug/ml	89
79) Pyrene	20.466	202	379325	45.37	ug/ml	88
81) Butylbenzylphthalate	21.233	149	187724	47.37	ug/ml	90
82) 3,3-Dichlorobenzidine	21.854	252	81519	57.94	ug/ml	94
83) Benzo[a]anthracene	21.900	228	320113	50.59	ug/ml	100
84) Chrysene	21.945	228	317963	53.04	ug/ml	100
85) bis(2-Ethylhexyl)phtha...	21.877	149	282815	48.88	ug/ml	96
87) Di-n-octylphthalate	22.690	149	493099	46.57	ug/ml	98
88) Benzo[b]fluoranthene	23.390	252	321707	48.48	ug/ml	93
89) Benzo[k]fluoranthene	23.435	252	332054	54.23	ug/ml	98
90) Benzo[a]pyrene	23.955	252	311796	51.17	ug/ml	99
91) Indeno[1,2,3-cd]pyrene	26.360	276	321847	53.63	ug/ml	98
92) Dibenz[a,h]anthracene	26.394	278	265118	52.83	ug/ml	96
93) Benzo[g,h,i]perylene	27.083	276	275566	52.73	ug/ml	96

Data Path : S:\HPCHEM1\BNA_L\data\BL072612\
Data File : BL000657.D
Acq On : 26 Jul 2012 11:15
Operator : HLM
Sample : SSTD050
Misc : SSTD050 HLM
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_L
LabSampleId :
SSTD050

Manual Integrations
APPROVED

AHPatel
9/6/2012 6:42:30 AM

Quant Time: Sep 06 06:42:24 2012
Quant Method : H:\VOL1\GCMSDATA\L\METHODS\BL072512.M
Quant Title : SW-846 Method 8270
QLast Update : Thu Jul 26 11:02:15 2012
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed