

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010121\
 Data File : BM028303.D
 Acq On : 31 Dec 2020 13:12
 Operator : JU/CG
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Dec 31 16:10:58 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-SIM-BM010121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 31 16:09:58 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.143	152	569	0.40	ng	0.00
7) Naphthalene-d8	10.959	136	2253	0.40	ng	0.00
13) Acenaphthene-d10	14.763	164	1190	0.40	ng	0.00
19) Phenanthrene-d10	17.480	188	2365	0.40	ng	# 0.00
27) Chrysene-d12	21.611	240	2357	0.40	ng	# 0.00
34) Perylene-d12	23.938	264	2231	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.646	112	692	0.41	ng	0.00
5) Phenol-d6	7.281	99	937	0.46	ng	0.00
8) Nitrobenzene-d5	9.299	82	722	0.44	ng	0.00
11) 2-Methylnaphthalene-d10	12.535	152	1363	0.35	ng	0.00
14) 2,4,6-Tribromophenol	16.237	330	215	0.24	ng	0.00
15) 2-Fluorobiphenyl	13.390	172	1766	0.37	ng	0.00
25) Fluoranthene-d10	19.491	212	2489	0.40	ng	0.00
29) Terphenyl-d14	20.071	244	2167	0.39	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.415	88	293	0.47	ng	88
3) n-Nitrosodimethylamine	3.770	42	429	0.64	ng	# 91
6) bis(2-Chloroethyl)ether	7.547	93	665	0.44	ng	94
9) Naphthalene	11.010	128	2297	0.40	ng	97
10) Hexachlorobutadiene	11.289	225	444	0.32	ng	# 96
12) 2-Methylnaphthalene	12.606	142	1532	0.36	ng	95
16) Acenaphthylene	14.491	152	1771	0.37	ng	99
17) Acenaphthene	14.824	154	1314	0.36	ng	94
18) Fluorene	15.804	166	1669	0.36	ng	98
20) Hexachlorobenzene	16.800	284	579	0.31	ng	94
21) Atrazine	19.516	200	532	0.32	ng	# 36
22) Pentachlorophenol	17.140	266	216	0.28	ng	96
23) Phenanthrene	17.523	178	2556	0.36	ng	100
24) Anthracene	17.619	178	2185	0.34	ng	99
26) Fluoranthene	19.516	202	2747	0.37	ng	# 96
28) Pyrene	19.878	202	2777	0.37	ng	98
30) Benzo(a)anthracene	21.590	228	2769	0.37	ng	98
31) Chrysene	21.643	228	2841	0.38	ng	98
32) Bis(2-ethylhexyl)phtha...	21.494	149	1292	0.34	ng	94
33) Indeno(1,2,3-cd)pyrene	26.313	276	3163	0.27	ng	# 91
35) Benzo(b)fluoranthene	23.232	252	2992	0.39	ng	# 87
36) Benzo(k)fluoranthene	23.277	252	2847	0.39	ng	# 86
37) Benzo(a)pyrene	23.835	252	2553	0.37	ng	# 83
38) Dibenzo(a,h)anthracene	26.327	278	2761	0.33	ng	# 86
39) Benzo(g,h,i)perylene	27.049	276	2794	0.29	ng	# 84


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010121\
Data File : BM028303.D
Acq On : 31 Dec 2020 13:12
Operator : JU/CG
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4

Quant Time: Dec 31 16:10:58 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-SIM-BM010121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 31 16:09:58 2020
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 8.143 min Scan# 635
 Delta R.T. -0.000 min
 Lab File: BM028303.D
 Acq: 31 Dec 2020 13:12




Instrument :
 BNA_M
 ClientSampled :
 SSTDICC0.4

Tgt Ion:152 Resp: 569
 Ion Ratio Lower Upper
 152 100
 150 145.8 116.6 174.8
 115 66.8 52.9 79.3






#2
 1,4-Dioxane
 Concen: 0.47 ng
 RT: 3.415 min Scan# 48
 Delta R.T. 0.000 min
 Lab File: BM028303.D
 Acq: 31 Dec 2020 13:12

Tgt Ion: 88 Resp: 293
 Ion Ratio Lower Upper
 88 100
 58 87.0 59.4 89.2
 43 43.7 32.3 48.5







Abundance Scan 92 (3.770 min): BM028303.D\data.ms (-84)




#3
n-Nitrosodimethylamine
Concen: 0.64 ng
RT: 3.770 min Scan# 92
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:	42	Resp:	429
Ion	Ratio	Lower	Upper
42	100		
74	88.6	64.4	96.6
44	9.1	2.8	4.2#




#4
2-Fluorophenol
Concen: 0.41 ng
RT: 5.646 min Scan# 325
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:	112	Resp:	692
Ion	Ratio	Lower	Upper
112	100		
64	65.3	49.5	74.3
63	36.4	32.0	48.0







Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4



#5
Phenol-d6
Concen: 0.46 ng
RT: 7.281 min Scan# 528
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	99	Resp:	937
Ion	Ratio	Lower	Upper
99	100		
42	27.9	22.4	33.6
71	38.4	37.8	56.6



#6
bis(2-Chloroethyl)ether
Concen: 0.44 ng
RT: 7.547 min Scan# 561
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:	93	Resp:	665
Ion	Ratio	Lower	Upper
93	100		
63	85.7	63.1	94.7
95	32.2	26.1	39.1









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.959 min Scan# 874
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	2253
Ion	Ratio	Lower	Upper
136	100		
137	14.2	10.6	15.8
54	10.7	8.6	12.8
68	7.8	5.7	8.5



#8
Nitrobenzene-d5
Concen: 0.44 ng
RT: 9.299 min Scan# 743
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:	82	Resp:	722
Ion	Ratio	Lower	Upper
82	100		
128	40.4	30.6	46.0
54	72.8	48.3	72.5









#9
Naphthalene
Concen: 0.40 ng
RT: 11.010 min Scan# 878
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	128	Resp:	2297
Ion Ratio	Lower	Upper	
128	100		
129	13.8	9.8	14.8
127	16.1	12.2	18.4



#10
Hexachlorobutadiene
Concen: 0.32 ng
RT: 11.289 min Scan# 900
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:	225	Resp:	444
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	66.7	51.0	76.6









#11
2-Methylnaphthalene-d10
Concen: 0.35 ng
RT: 12.535 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 1363
Ion Ratio Lower Upper
152 100
151 20.2 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.36 ng
RT: 12.606 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:142 Resp: 1532
Ion Ratio Lower Upper
142 100
141 89.8 69.4 104.2
115 39.9 35.7 53.5









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.763 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	164	Resp:	1190
Ion	Ratio	Lower	Upper
164	100		
162	100.0	80.6	120.8
160	47.5	39.5	59.3



#14
2,4,6-Tribromophenol
Concen: 0.24 ng
RT: 16.237 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:	330	Resp:	215
Ion	Ratio	Lower	Upper
330	100		
332	95.3	0.0	0.0#
141	35.8	34.4	51.6









#15
2-Fluorobiphenyl
Concen: 0.37 ng
RT: 13.390 min Scan# 1291
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion:172	Resp:	1766
Ion Ratio	Lower	Upper
172	100	
171	38.0	28.2 42.4
170	27.3	20.0 30.0



#16
Acenaphthylene
Concen: 0.37 ng
RT: 14.491 min Scan# 1400
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:152	Resp:	1771
Ion Ratio	Lower	Upper
152	100	
151	19.6	15.8 23.8
153	13.7	10.6 16.0









#17
Acenaphthene
Concen: 0.36 ng
RT: 14.824 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154	Resp:	1314	
Ion	Ratio	Lower	Upper
154	100		
153	113.1	83.6	125.4
152	56.0	43.9	65.9



#18
Fluorene
Concen: 0.36 ng
RT: 15.804 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:166	Resp:	1669	
Ion	Ratio	Lower	Upper
166	100		
165	96.0	78.6	118.0
167	13.4	10.7	16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.480 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion:188	Resp:	2365
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	14.0	5.0



#20
Hexachlorobenzene
Concen: 0.31 ng
RT: 16.800 min Scan# 1624
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:284	Resp:	579
Ion Ratio	Lower	Upper
284	100	
142	35.1	32.0
249	31.6	27.0









#21
Atrazine
Concen: 0.32 ng
RT: 19.516 min Scan# 2041
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion:200 Resp: 532
Ion Ratio Lower Upper
200 100
173 0.0 22.8 34.2#
215 0.0 38.1 57.1#



#22
Pentachlorophenol
Concen: 0.28 ng
RT: 17.140 min Scan# 1656
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:266 Resp: 216
Ion Ratio Lower Upper
266 100
264 59.7 50.9 76.3
268 66.7 51.5 77.3









#23
Phenanthrene
Concen: 0.36 ng
RT: 17.523 min Scan# 1692
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	178	Resp:	2556
Ion Ratio	Lower	Upper	
178	100		
176	18.2	14.8	22.2
179	15.5	12.4	18.6



#24
Anthracene
Concen: 0.34 ng
RT: 17.619 min Scan# 1701
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:	178	Resp:	2185
Ion Ratio	Lower	Upper	
178	100		
176	17.5	14.8	22.2
179	15.5	12.5	18.7









#25
Fluoranthene-d10
Concen: 0.40 ng
RT: 19.491 min Scan# 2036
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
212	100		
106	12.3	6.8	10.2#
104	7.3	4.1	6.1#



#26
Fluoranthene
Concen: 0.37 ng
RT: 19.516 min Scan# 2041
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion	Ratio	Lower	Upper
202	100		
101	10.8	5.8	8.6#
203	17.4	13.7	20.5









#27
Chrysene-d12
Concen: 0.40 ng
RT: 21.611 min Scan# 2291
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampled :
SSTDICCC0.4


Tgt Ion:240 Resp: 2357
Ion Ratio Lower Upper
240 100
120 10.2 5.3 7.9#
236 26.4 21.8 32.8



#28
Pyrene
Concen: 0.37 ng
RT: 19.878 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:202 Resp: 2777
Ion Ratio Lower Upper
202 100
200 19.8 16.6 24.8
203 17.8 13.8 20.8





#29

Terphenyl-d14

Concen: 0.39 ng

RT: 20.071 min Scan# 2146

Delta R.T. 0.000 min

Lab File: BM028303.D

Acq: 31 Dec 2020 13:12

Tgt Ion:244 Resp: 2167

Ion Ratio Lower Upper

244 100

212 7.8 7.3 10.9





122 7.0 7.3 10.9#

Instrument :

BNA_M

ClientSampleId :

SSTDICCC0.4



#30

Benzo(a)anthracene

Concen: 0.37 ng

RT: 21.590 min Scan# 2289

Delta R.T. 0.000 min

Lab File: BM028303.D

Acq: 31 Dec 2020 13:12




Tgt Ion:228 Resp: 2769


Ion Ratio Lower Upper

228 100

226 27.4 22.3 33.5

229 21.4 15.7 23.5









#31
Chrysene
Concen: 0.38 ng
RT: 21.643 min Scan# 2294
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion:228	Resp:	2841
Ion Ratio	Lower	Upper
228	100	
226	29.9	24.2 36.2
229	21.2	15.7 23.5



#32
Bis(2-ethylhexyl)phthalate
Concen: 0.34 ng
RT: 21.494 min Scan# 2280
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:149	Resp:	1292
Ion Ratio	Lower	Upper
149	100	
167	25.5	23.2 34.8
279	4.4	3.9 5.9









#33
Indeno(1,2,3-cd)pyrene
Concen: 0.27 ng
RT: 26.313 min Scan# 3507
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion:276 Resp: 3163
Ion Ratio Lower Upper
276 100
138 26.2 13.3 19.9#
277 25.4 20.2 30.4



#34
Perylene-d12
Concen: 0.40 ng
RT: 23.938 min Scan# 2813
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:264 Resp: 2231
Ion Ratio Lower Upper
264 100
260 25.7 19.8 29.6
265 74.1 51.4 77.0









#35
Benzo(b)fluoranthene
Concen: 0.39 ng
RT: 23.232 min Scan# 2607
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252 Resp: 2992
Ion Ratio Lower Upper
252 100
253 29.0 20.1 30.1
125 18.6 7.2 10.8#



#36
Benzo(k)fluoranthene
Concen: 0.39 ng
RT: 23.277 min Scan# 2620
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Tgt Ion:252 Resp: 2847
Ion Ratio Lower Upper
252 100
253 29.8 19.8 29.8#
125 18.7 7.1 10.7#









#37
Benzo(a)pyrene
Concen: 0.37 ng
RT: 23.835 min Scan# 2783
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252 Resp: 2553
Ion Ratio Lower Upper
252 100
253 31.9 21.3 31.9#
125 22.7 8.3 12.5#



#38
Dibenzo(a,h)anthracene
Concen: 0.33 ng
RT: 26.327 min Scan# 3511
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12


Tgt Ion:278 Resp: 2761
Ion Ratio Lower Upper
278 100
139 21.9 9.3 13.9#
279 31.5 21.6 32.4





#39
Benzo(g,h,i)perylene
Concen: 0.29 ng
RT: 27.049 min Scan# 3722
Delta R.T. 0.000 min
Lab File: BM028303.D
Acq: 31 Dec 2020 13:12

Instrument :
BNA_M
ClientSampleId :
SSTDICCC0.4



Tgt Ion	Ratio	Lower	Upper
276	100		
277	28.7	19.8	29.8
138	27.0	11.7	17.5

