

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010121\
 Data File : BM028304.D
 Acq On : 31 Dec 2020 13:49
 Operator : JU/CG
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.8

Quant Time: Dec 31 16:11:12 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-SIM-BM010121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 31 16:09:58 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.143	152	531	0.40	ng	0.00
7) Naphthalene-d8	10.960	136	2142	0.40	ng	0.00
13) Acenaphthene-d10	14.764	164	1173	0.40	ng	0.00
19) Phenanthrene-d10	17.481	188	2379	0.40	ng	# 0.00
27) Chrysene-d12	21.601	240	2483	0.40	ng	#-0.01
34) Perylene-d12	23.938	264	2401	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.639	112	1501	0.96	ng	0.00
5) Phenol-d6	7.281	99	2042	1.07	ng	0.00
8) Nitrobenzene-d5	9.299	82	1588	1.01	ng	0.00
11) 2-Methylnaphthalene-d10	12.535	152	3045	0.82	ng	0.00
14) 2,4,6-Tribromophenol	16.237	330	485	0.55	ng	0.00
15) 2-Fluorobiphenyl	13.390	172	3983	0.84	ng	0.00
25) Fluoranthene-d10	19.487	212	5996	0.97	ng	0.00
29) Terphenyl-d14	20.072	244	5472	0.94	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.408	88	647	1.10	ng	93
3) n-Nitrosodimethylamine	3.762	42	954	1.52	ng	# 92
6) bis(2-Chloroethyl)ether	7.547	93	1466	1.05	ng	96
9) Naphthalene	11.010	128	5077	0.92	ng	98
10) Hexachlorobutadiene	11.289	225	972	0.74	ng	# 98
12) 2-Methylnaphthalene	12.606	142	3443	0.85	ng	94
16) Acenaphthylene	14.491	152	4203	0.89	ng	100
17) Acenaphthene	14.824	154	3144	0.87	ng	96
18) Fluorene	15.804	166	3970	0.87	ng	99
20) Hexachlorobenzene	16.801	284	1336	0.70	ng	93
21) Atrazine	19.516	200	1316	0.78	ng	# 36
22) Pentachlorophenol	17.141	266	513	0.67	ng	99
23) Phenanthrene	17.523	178	6043	0.86	ng	99
24) Anthracene	17.619	178	5265	0.83	ng	99
26) Fluoranthene	19.516	202	6711	0.89	ng	# 96
28) Pyrene	19.879	202	6706	0.84	ng	99
30) Benzo(a)anthracene	21.590	228	6578	0.84	ng	97
31) Chrysene	21.643	228	7013	0.89	ng	98
32) Bis(2-ethylhexyl)phtha...	21.495	149	3027	0.76	ng	95
33) Indeno(1,2,3-cd)pyrene	26.313	276	7591	0.62	ng	# 91
35) Benzo(b)fluoranthene	23.229	252	7269	0.87	ng	# 96
36) Benzo(k)fluoranthene	23.274	252	6815	0.88	ng	# 96
37) Benzo(a)pyrene	23.832	252	6208	0.83	ng	# 95
38) Dibenzo(a,h)anthracene	26.327	278	6651	0.74	ng	# 95
39) Benzo(g,h,i)perylene	27.049	276	6619	0.64	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010121\
Data File : BM028304.D
Acq On : 31 Dec 2020 13:49
Operator : JU/CG
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC0.8

Quant Time: Dec 31 16:11:12 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-SIM-BM010121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 31 16:09:58 2020
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 8.143 min Scan# 635
 Delta R.T. 0.000 min
 Lab File: BM028304.D
 Acq: 31 Dec 2020 13:49




Instrument :
 BNA_M
 ClientSampled :
 SSTDICC0.8

Tgt Ion:152 Resp: 531
 Ion Ratio Lower Upper
 152 100
 150 143.6 116.6 174.8
 115 67.0 52.9 79.3






#2
 1,4-Dioxane
 Concen: 1.10 ng
 RT: 3.408 min Scan# 47
 Delta R.T. -0.008 min
 Lab File: BM028304.D
 Acq: 31 Dec 2020 13:49

Tgt Ion: 88 Resp: 647
 Ion Ratio Lower Upper
 88 100
 58 83.3 59.4 89.2
 43 39.7 32.3 48.5







Abundance Scan 92 (3.770 min): BM028303.D\data.ms (-84)




#3
n-Nitrosodimethylamine
Concen: 1.52 ng
RT: 3.762 min Scan# 91
Delta R.T. -0.008 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Tgt Ion:	42	Resp:	954
Ion	Ratio	Lower	Upper
42	100		
74	87.2	64.4	96.6
44	6.8	2.8	4.2#




#4
2-Fluorophenol
Concen: 0.96 ng
RT: 5.639 min Scan# 324
Delta R.T. -0.008 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Tgt Ion:	112	Resp:	1501
Ion	Ratio	Lower	Upper
112	100		
64	65.9	49.5	74.3
63	36.2	32.0	48.0







Instrument :
BNA_M
ClientSampled :
SSTDIC0.8



#5
Phenol-d6
Concen: 1.07 ng
RT: 7.281 min Scan# 528
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49




Instrument :
BNA_M
ClientSampled :
SSTDICC0.8


Tgt Ion:	99	Resp:	2042
Ion	Ratio	Lower	Upper
99	100		
42	27.8	22.4	33.6
71	37.3	37.8	56.6#



#6
bis(2-Chloroethyl)ether
Concen: 1.05 ng
RT: 7.547 min Scan# 561
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Tgt Ion:	93	Resp:	1466
Ion	Ratio	Lower	Upper
93	100		
63	84.2	63.1	94.7
95	32.3	26.1	39.1









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.960 min Scan# 874
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8


Tgt Ion:	136	Resp:	2142
Ion	Ratio	Lower	Upper
136	100		
137	14.5	10.6	15.8
54	10.1	8.6	12.8
68	7.3	5.7	8.5



#8
Nitrobenzene-d5
Concen: 1.01 ng
RT: 9.299 min Scan# 743
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Tgt Ion:	82	Resp:	1588
Ion	Ratio	Lower	Upper
82	100		
128	37.1	30.6	46.0
54	71.4	48.3	72.5









#9
Naphthalene
Concen: 0.92 ng
RT: 11.010 min Scan# 878
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49




Instrument :
BNA_M
ClientSampleId :
SSTDICC0.8


Tgt Ion:	128	Resp:	5077
Ion Ratio	Lower	Upper	
128	100		
129	12.3	9.8	14.8
127	14.1	12.2	18.4



#10
Hexachlorobutadiene
Concen: 0.74 ng
RT: 11.289 min Scan# 900
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Tgt Ion:	225	Resp:	972
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	65.0	51.0	76.6









#11
2-Methylnaphthalene-d10
Concen: 0.82 ng
RT: 12.535 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49




Instrument :
BNA_M
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 3045
Ion Ratio Lower Upper
152 100
151 20.4 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.85 ng
RT: 12.606 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49


Tgt Ion:142 Resp: 3443
Ion Ratio Lower Upper
142 100
141 90.2 69.4 104.2
115 38.0 35.7 53.5








#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.764 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49


Instrument :
BNA_M
ClientSampled :
SSTDICC0.8





Tgt Ion:	164	Resp:	1173
Ion	Ratio	Lower	Upper
164	100		
162	100.8	80.6	120.8
160	47.9	39.5	59.3




#14
2,4,6-Tribromophenol
Concen: 0.55 ng
RT: 16.237 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49



Tgt Ion:	330	Resp:	485
Ion	Ratio	Lower	Upper
330	100		
332	96.7	0.0	0.0#
141	37.1	34.4	51.6









#15
2-Fluorobiphenyl
Concen: 0.84 ng
RT: 13.390 min Scan# 1291
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49




Instrument :
BNA_M
ClientSampleId :
SSTDICC0.8


Tgt Ion:	172	Resp:	3983
Ion Ratio	Lower	Upper	
172	100		
171	36.5	28.2	42.4
170	24.8	20.0	30.0



#16
Acenaphthylene
Concen: 0.89 ng
RT: 14.491 min Scan# 1400
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Tgt Ion:	152	Resp:	4203
Ion Ratio	Lower	Upper	
152	100		
151	19.8	15.8	23.8
153	13.8	10.6	16.0









#17
Acenaphthene
Concen: 0.87 ng
RT: 14.824 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8


Tgt Ion:154	Resp:	3144
Ion Ratio	Lower	Upper
154	100	
153	108.8	83.6 125.4
152	53.3	43.9 65.9



#18
Fluorene
Concen: 0.87 ng
RT: 15.804 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Tgt Ion:166	Resp:	3970
Ion Ratio	Lower	Upper
166	100	
165	97.1	78.6 118.0
167	13.2	10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.481 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8


Tgt Ion:188	Resp:	2379
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	13.6	5.0



#20
Hexachlorobenzene
Concen: 0.70 ng
RT: 16.801 min Scan# 1624
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Tgt Ion:284	Resp:	1336
Ion Ratio	Lower	Upper
284	100	
142	34.1	32.0
249	31.1	27.0









#21
Atrazine
Concen: 0.78 ng
RT: 19.516 min Scan# 2041
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8


Tgt Ion:200 Resp: 1316
Ion Ratio Lower Upper
200 100
173 0.0 22.8 34.2#
215 0.0 38.1 57.1#



#22
Pentachlorophenol
Concen: 0.67 ng
RT: 17.141 min Scan# 1656
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Tgt Ion:266 Resp: 513
Ion Ratio Lower Upper
266 100
264 63.4 50.9 76.3
268 65.3 51.5 77.3









#23
Phenanthrene
Concen: 0.86 ng
RT: 17.523 min Scan# 1692
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8


Tgt Ion:178	Resp:	6043
Ion Ratio	Lower	Upper
178	100	
176	18.3	14.8 22.2
179	15.2	12.4 18.6



#24
Anthracene
Concen: 0.83 ng
RT: 17.619 min Scan# 1701
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Tgt Ion:178	Resp:	5265
Ion Ratio	Lower	Upper
178	100	
176	17.8	14.8 22.2
179	15.4	12.5 18.7









#25
Fluoranthene-d10
Concen: 0.97 ng
RT: 19.487 min Scan# 2035
Delta R.T. -0.005 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8


Tgt Ion:212 Resp: 5996
Ion Ratio Lower Upper
212 100
106 12.3 6.8 10.2#
104 7.1 4.1 6.1#



#26
Fluoranthene
Concen: 0.89 ng
RT: 19.516 min Scan# 2041
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Tgt Ion:202 Resp: 6711
Ion Ratio Lower Upper
202 100
101 11.2 5.8 8.6#
203 17.5 13.7 20.5









#27
Chrysene-d12
Concen: 0.40 ng
RT: 21.601 min Scan# 2290
Delta R.T. -0.010 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8


Tgt Ion:240 Resp: 2483
Ion Ratio Lower Upper
240 100
120 18.8 5.3 7.9#
236 28.6 21.8 32.8



#28
Pyrene
Concen: 0.84 ng
RT: 19.879 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Tgt Ion:202 Resp: 6706
Ion Ratio Lower Upper
202 100
200 20.2 16.6 24.8
203 17.7 13.8 20.8









#29
 Terphenyl-d14
 Concen: 0.94 ng
 RT: 20.072 min Scan# 2146
 Delta R.T. 0.000 min
 Lab File: BM028304.D
 Acq: 31 Dec 2020 13:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:244 Resp: 5472
 Ion Ratio Lower Upper
 244 100
 212 6.0 7.3 10.9#
 122 4.7 7.3 10.9#



#30
 Benzo(a)anthracene
 Concen: 0.84 ng
 RT: 21.590 min Scan# 2289
 Delta R.T. 0.000 min
 Lab File: BM028304.D
 Acq: 31 Dec 2020 13:49

Tgt Ion:228 Resp: 6578
 Ion Ratio Lower Upper
 228 100
 226 26.2 22.3 33.5
 229 20.3 15.7 23.5





#31

Chrysene

Concen: 0.89 ng

RT: 21.643 min Scan# 2294

Delta R.T. 0.000 min

Lab File: BM028304.D

Acq: 31 Dec 2020 13:49

Tgt Ion:228 Resp: 7013

Ion Ratio Lower Upper

228 100

226 28.7 24.2 36.2





229 20.4 15.7 23.5

Instrument :

BNA_M

ClientSampleId :

SSTDIC0.8



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.76 ng

RT: 21.495 min Scan# 2280

Delta R.T. 0.000 min

Lab File: BM028304.D

Acq: 31 Dec 2020 13:49




Tgt Ion:149 Resp: 3027


Ion Ratio Lower Upper

149 100

167 26.1 23.2 34.8

279 4.3 3.9 5.9









#33
Indeno(1,2,3-cd)pyrene
Concen: 0.62 ng
RT: 26.313 min Scan# 3507
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8


Tgt Ion:276 Resp: 7591
Ion Ratio Lower Upper
276 100
138 26.5 13.3 19.9#
277 25.5 20.2 30.4



#34
Perylene-d12
Concen: 0.40 ng
RT: 23.938 min Scan# 2813
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Tgt Ion:264 Resp: 2401
Ion Ratio Lower Upper
264 100
260 25.5 19.8 29.6
265 70.7 51.4 77.0









#35
Benzo(b)fluoranthene
Concen: 0.87 ng
RT: 23.229 min Scan# 2606
Delta R.T. -0.003 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49




Instrument :
BNA_M
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 7269
Ion Ratio Lower Upper
252 100
253 25.1 20.1 30.1
125 13.8 7.2 10.8#



#36
Benzo(k)fluoranthene
Concen: 0.88 ng
RT: 23.274 min Scan# 2619
Delta R.T. -0.003 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Tgt Ion:252 Resp: 6815
Ion Ratio Lower Upper
252 100
253 24.9 19.8 29.8
125 14.0 7.1 10.7#









#37
Benzo(a)pyrene
Concen: 0.83 ng
RT: 23.832 min Scan# 2782
Delta R.T. -0.003 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8


Tgt Ion:252 Resp: 6208
Ion Ratio Lower Upper
252 100
253 26.3 21.3 31.9
125 16.6 8.3 12.5#



#38
Dibenzo(a,h)anthracene
Concen: 0.74 ng
RT: 26.327 min Scan# 3511
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49


Tgt Ion:278 Resp: 6651
Ion Ratio Lower Upper
278 100
139 18.5 9.3 13.9#
279 26.9 21.6 32.4





#39
Benzo(g,h,i)perylene
Concen: 0.64 ng
RT: 27.049 min Scan# 3722
Delta R.T. 0.000 min
Lab File: BM028304.D
Acq: 31 Dec 2020 13:49

Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8



Tgt Ion:276 Resp: 6619
Ion Ratio Lower Upper
276 100
277 26.1 19.8 29.8
138 24.5 11.7 17.5#

