

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010121\
 Data File : BM028323.D
 Acq On : 01 Jan 2021 01:52
 Operator : JU/CG
 Sample : L5123-17MSD
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 TT180D1-20201214MSD

Quant Time: Jan 01 03:00:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-SIM-BM010121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 31 16:18:03 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.135	152	664	0.40	ng	0.00
7) Naphthalene-d8	10.959	136	2731	0.40	ng	0.00
13) Acenaphthene-d10	14.763	164	1429	0.40	ng	0.00
19) Phenanthrene-d10	17.480	188	2872	0.40	ng	# 0.00
27) Chrysene-d12	21.600	240	2897	0.40	ng	#-0.01
34) Perylene-d12	23.931	264	2771	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.638	112	217	0.10	ng	0.00
5) Phenol-d6	7.273	99	172	0.06	ng	0.00
8) Nitrobenzene-d5	9.299	82	678	0.27	ng	0.00
11) 2-Methylnaphthalene-d10	12.531	152	1233	0.26	ng	0.00
14) 2,4,6-Tribromophenol	16.237	330	190	0.25	ng	0.00
15) 2-Fluorobiphenyl	13.390	172	1628	0.27	ng	0.00
25) Fluoranthene-d10	19.486	212	3045	0.34	ng	0.00
29) Terphenyl-d14	20.061	244	2848	0.40	ng	-0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.407	88	1859	1.99	ng	# 86
3) n-Nitrosodimethylamine	3.762	42	167	0.12	ng	# 93
6) bis(2-Chloroethyl)ether	7.547	93	714	0.33	ng	96
9) Naphthalene	10.997	128	2318	0.29	ng	96
10) Hexachlorobutadiene	11.289	225	354	0.23	ng	# 96
12) 2-Methylnaphthalene	12.602	142	1551	0.29	ng	96
16) Acenaphthylene	14.481	152	1874	0.29	ng	99
17) Acenaphthene	14.824	154	1359	0.29	ng	95
18) Fluorene	15.804	166	1771	0.30	ng	99
20) Hexachlorobenzene	16.800	284	595	0.30	ng	94
21) Atrazine	19.516	200	717	0.37	ng	# 36
22) Pentachlorophenol	17.130	266	534	0.66	ng	99
23) Phenanthrene	17.523	178	3120	0.35	ng	99
24) Anthracene	17.619	178	2682	0.34	ng	98
26) Fluoranthene	19.516	202	3664	0.37	ng	# 96
28) Pyrene	19.873	202	3708	0.37	ng	98
30) Benzo(a)anthracene	21.590	228	3766	0.39	ng	96
31) Chrysene	21.643	228	3854	0.39	ng	96
32) Bis(2-ethylhexyl)phtha...	21.494	149	2305	0.51	ng	96
33) Indeno(1,2,3-cd)pyrene	26.310	276	4210	0.39	ng	# 90
35) Benzo(b)fluoranthene	23.226	252	3961	0.38	ng	# 89
36) Benzo(k)fluoranthene	23.270	252	3758	0.39	ng	# 88
37) Benzo(a)pyrene	23.828	252	3285	0.37	ng	# 88
38) Dibenzo(a,h)anthracene	26.316	278	3622	0.38	ng	# 88
39) Benzo(g,h,i)perylene	27.039	276	3694	0.39	ng	# 86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010121\
 Data File : BM028323.D
 Acq On : 01 Jan 2021 01:52
 Operator : JU/CG
 Sample : L5123-17MSD
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 TT180D1-20201214MSD

Quant Time: Jan 01 03:00:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-SIM-BM010121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 31 16:18:03 2020
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 8.135 min Scan# 634
 Delta R.T. -0.008 min
 Lab File: BM028323.D
 Acq: 01 Jan 2021 01:52




Instrument :
 BNA_M
 ClientSampleId :
 TT180D1-20201214MSD


Tgt Ion:152 Resp: 664
 Ion Ratio Lower Upper
 152 100
 150 152.1 116.6 174.8
 115 71.0 52.9 79.3



#2
 1,4-Dioxane
 Concen: 1.99 ng
 RT: 3.407 min Scan# 47
 Delta R.T. -0.008 min
 Lab File: BM028323.D
 Acq: 01 Jan 2021 01:52

Tgt Ion: 88 Resp: 1859
 Ion Ratio Lower Upper
 88 100
 58 89.2 59.4 89.2#
 43 45.5 32.3 48.5









#3
n-Nitrosodimethylamine
Concen: 0.12 ng
RT: 3.762 min Scan# 91
Delta R.T. -0.008 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:	42	Resp:	167
Ion	Ratio	Lower	Upper
42	100		
74	86.2	64.4	96.6
44	0.0	2.8	4.2#



#4
2-Fluorophenol
Concen: 0.10 ng
RT: 5.638 min Scan# 324
Delta R.T. -0.008 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:	112	Resp:	217
Ion	Ratio	Lower	Upper
112	100		
64	66.4	49.5	74.3
63	35.9	32.0	48.0









#5
Phenol-d6
Concen: 0.06 ng
RT: 7.273 min Scan# 527
Delta R.T. -0.008 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion: 99	Resp: 172
Ion Ratio	Lower Upper
99 100	
42 25.6	22.4 33.6
71 41.9	37.8 56.6



#6
bis(2-Chloroethyl)ether
Concen: 0.33 ng
RT: 7.547 min Scan# 561
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion: 93	Resp: 714
Ion Ratio	Lower Upper
93 100	
63 83.9	63.1 94.7
95 32.5	26.1 39.1









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.959 min Scan# 874
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:	136	Resp:	2731
Ion	Ratio	Lower	Upper
136	100		
137	13.7	10.6	15.8
54	8.9	8.6	12.8
68	6.6	5.7	8.5



#8
Nitrobenzene-d5
Concen: 0.27 ng
RT: 9.299 min Scan# 743
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:	82	Resp:	678
Ion	Ratio	Lower	Upper
82	100		
128	43.9	30.6	46.0
54	67.2	48.3	72.5









#9
Naphthalene
Concen: 0.29 ng
RT: 10.997 min Scan# 877
Delta R.T. -0.013 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:128 Resp: 2318
Ion Ratio Lower Upper
128 100
129 14.1 9.8 14.8
127 16.7 12.2 18.4



#10
Hexachlorobutadiene
Concen: 0.23 ng
RT: 11.289 min Scan# 900
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:225 Resp: 354
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 66.7 51.0 76.6









#11
2-Methylnaphthalene-d10
Concen: 0.26 ng
RT: 12.531 min Scan# 1146
Delta R.T. -0.004 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:152 Resp: 1233
Ion Ratio Lower Upper
152 100
151 21.3 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.29 ng
RT: 12.602 min Scan# 1162
Delta R.T. -0.004 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:142 Resp: 1551
Ion Ratio Lower Upper
142 100
141 89.8 69.4 104.2
115 40.2 35.7 53.5









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.763 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:	164	Resp:	1429
Ion	Ratio	Lower	Upper
164	100		
162	100.1	80.6	120.8
160	46.6	39.5	59.3



#14
2,4,6-Tribromophenol
Concen: 0.25 ng
RT: 16.237 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:	330	Resp:	190
Ion	Ratio	Lower	Upper
330	100		
332	94.7	0.0	0.0#
141	39.5	34.4	51.6









#15
2-Fluorobiphenyl
Concen: 0.27 ng
RT: 13.390 min Scan# 1291
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:172	Resp:	1628
Ion Ratio	Lower	Upper
172	100	
171	37.7	28.2 42.4
170	25.8	20.0 30.0



#16
Acenaphthylene
Concen: 0.29 ng
RT: 14.481 min Scan# 1399
Delta R.T. -0.010 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:152	Resp:	1874
Ion Ratio	Lower	Upper
152	100	
151	20.3	15.8 23.8
153	13.3	10.6 16.0









#17
Acenaphthene
Concen: 0.29 ng
RT: 14.824 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:154	Resp:	1359
Ion Ratio	Lower	Upper
154	100	
153	111.8	83.6 125.4
152	54.9	43.9 65.9



#18
Fluorene
Concen: 0.30 ng
RT: 15.804 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:166	Resp:	1771
Ion Ratio	Lower	Upper
166	100	
165	98.8	78.6 118.0
167	14.3	10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.480 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:188 Resp: 2872
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.7 5.0 7.4#



#20
Hexachlorobenzene
Concen: 0.30 ng
RT: 16.800 min Scan# 1624
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:284 Resp: 595
Ion Ratio Lower Upper
284 100
142 35.8 32.0 48.0
249 31.1 27.0 40.6









#21
Atrazine
Concen: 0.37 ng
RT: 19.516 min Scan# 2041
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:200 Resp: 717
Ion Ratio Lower Upper
200 100
173 0.0 22.8 34.2#
215 0.0 38.1 57.1#



#22
Pentachlorophenol
Concen: 0.66 ng
RT: 17.130 min Scan# 1655
Delta R.T. -0.011 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:266 Resp: 534
Ion Ratio Lower Upper
266 100
264 62.7 50.9 76.3
268 63.9 51.5 77.3









#23
Phenanthrene
Concen: 0.35 ng
RT: 17.523 min Scan# 1692
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:	178	Resp:	3120
Ion Ratio	Lower	Upper	
178	100		
176	18.3	14.8	22.2
179	16.4	12.4	18.6



#24
Anthracene
Concen: 0.34 ng
RT: 17.619 min Scan# 1701
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:	178	Resp:	2682
Ion Ratio	Lower	Upper	
178	100		
176	17.6	14.8	22.2
179	14.9	12.5	18.7









#25
Fluoranthene-d10
Concen: 0.34 ng
RT: 19.486 min Scan# 2035
Delta R.T. -0.005 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:212 Resp: 3045
Ion Ratio Lower Upper
212 100
106 12.6 6.8 10.2#
104 7.9 4.1 6.1#



#26
Fluoranthene
Concen: 0.37 ng
RT: 19.516 min Scan# 2041
Delta R.T. 0.000 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:202 Resp: 3664
Ion Ratio Lower Upper
202 100
101 12.1 5.8 8.6#
203 17.2 13.7 20.5









#27
Chrysene-d12
Concen: 0.40 ng
RT: 21.600 min Scan# 2290
Delta R.T. -0.011 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:240 Resp: 2897
Ion Ratio Lower Upper
240 100
120 13.0 5.3 7.9#
236 27.2 21.8 32.8



#28
Pyrene
Concen: 0.37 ng
RT: 19.873 min Scan# 2113
Delta R.T. -0.005 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:202 Resp: 3708
Ion Ratio Lower Upper
202 100
200 20.2 16.6 24.8
203 18.4 13.8 20.8









#29
 Terphenyl-d14
 Concen: 0.40 ng
 RT: 20.061 min Scan# 2145
 Delta R.T. -0.011 min
 Lab File: BM028323.D
 Acq: 01 Jan 2021 01:52




Instrument :
 BNA_M
 ClientSampleId :
 TT180D1-20201214MSD


Tgt Ion:244	Resp:	2848
Ion Ratio	Lower	Upper
244	100	
212	9.2	7.3 10.9
122	11.4	7.3 10.9#



#30
 Benzo(a)anthracene
 Concen: 0.39 ng
 RT: 21.590 min Scan# 2289
 Delta R.T. 0.000 min
 Lab File: BM028323.D
 Acq: 01 Jan 2021 01:52

Tgt Ion:228	Resp:	3766
Ion Ratio	Lower	Upper
228	100	
226	26.0	22.3 33.5
229	21.5	15.7 23.5





#31

Chrysene

Concen: 0.39 ng

RT: 21.643 min Scan# 2294

Delta R.T. 0.000 min

Lab File: BM028323.D

Acq: 01 Jan 2021 01:52





Tgt Ion:228 Resp: 3854

Ion Ratio Lower Upper

228 100

226 28.6 24.2 36.2

229 22.0 15.7 23.5



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.51 ng

RT: 21.494 min Scan# 2280

Delta R.T. 0.000 min

Lab File: BM028323.D

Acq: 01 Jan 2021 01:52




Tgt Ion:149 Resp: 2305


Ion Ratio Lower Upper

149 100

167 26.4 23.2 34.8

279 4.5 3.9 5.9









#33
Indeno(1,2,3-cd)pyrene
Concen: 0.39 ng
RT: 26.310 min Scan# 3506
Delta R.T. -0.003 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:276 Resp: 4210
Ion Ratio Lower Upper
276 100
138 27.2 13.3 19.9#
277 25.1 20.2 30.4



#34
Perylene-d12
Concen: 0.40 ng
RT: 23.931 min Scan# 2811
Delta R.T. -0.007 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:264 Resp: 2771
Ion Ratio Lower Upper
264 100
260 25.2 19.8 29.6
265 65.1 51.4 77.0









#35
Benzo(b)fluoranthene
Concen: 0.38 ng
RT: 23.226 min Scan# 2605
Delta R.T. -0.007 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:252 Resp: 3961
Ion Ratio Lower Upper
252 100
253 27.9 20.1 30.1
125 18.3 7.2 10.8#



#36
Benzo(k)fluoranthene
Concen: 0.39 ng
RT: 23.270 min Scan# 2618
Delta R.T. -0.007 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:252 Resp: 3758
Ion Ratio Lower Upper
252 100
253 28.3 19.8 29.8
125 17.9 7.1 10.7#









#37
Benzo(a)pyrene
Concen: 0.37 ng
RT: 23.828 min Scan# 2781
Delta R.T. -0.007 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:252	Resp:	3285
Ion Ratio	Lower	Upper
252	100	
253	29.6	21.3 31.9
125	21.2	8.3 12.5#



#38
Dibenzo(a,h)anthracene
Concen: 0.38 ng
RT: 26.316 min Scan# 3508
Delta R.T. -0.010 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Tgt Ion:278	Resp:	3622
Ion Ratio	Lower	Upper
278	100	
139	22.3	9.3 13.9#
279	29.6	21.6 32.4





#39
Benzo(g,h,i)perylene
Concen: 0.39 ng
RT: 27.039 min Scan# 3719
Delta R.T. -0.010 min
Lab File: BM028323.D
Acq: 01 Jan 2021 01:52

Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD

Tgt Ion:276 Resp: 3694
Ion Ratio Lower Upper
276 100
277 27.8 19.8 29.8
138 26.1 11.7 17.5#

