

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010121\
 Data File : BM028308.D
 Acq On : 31 Dec 2020 16:14
 Operator : JU/CG
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM010121

Quant Time: Dec 31 16:44:28 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-SIM-BM010121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 31 16:18:03 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	114	0.00
2	1,4-Dioxane	0.400	0.356	11.0	111	0.00
3	n-Nitrosodimethylamine	0.400	0.341	14.8	108	0.00
4 S	2-Fluorophenol	0.400	0.358	10.5	115	0.00
5 S	Phenol-d6	0.400	0.356	11.0	114	0.00
6	bis(2-Chloroethyl)ether	0.400	0.363	9.3	115	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	113	0.00
8 S	Nitrobenzene-d5	0.400	0.354	11.5	114	0.00
9	Naphthalene	0.400	0.353	11.8	113	0.00
10	Hexachlorobutadiene	0.400	0.361	9.8	115	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.351	12.3	113	0.00
12	2-Methylnaphthalene	0.400	0.350	12.5	113	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.400	0.310	22.5	99	0.00
15 S	2-Fluorobiphenyl	0.400	0.367	8.3	113	0.00
16	Acenaphthylene	0.400	0.338	15.5	112	0.00
17	Acenaphthene	0.400	0.347	13.3	112	0.00
18	Fluorene	0.400	0.349	12.8	112	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	109	0.00
20	Hexachlorobenzene	0.400	0.357	10.8	109	0.00
21	Atrazine	0.400	0.338	15.5	110	0.00
22	Pentachlorophenol	0.400	0.301	24.8	101	0.00
23	Phenanthrene	0.400	0.350	12.5	110	0.00
24	Anthracene	0.400	0.341	14.8	110	0.00
25 SURR	Fluoranthene-d10	0.400	0.339	15.3	108	0.00
26	Fluoranthene	0.400	0.336	16.0	109	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	104	0.00
28	Pyrene	0.400	0.355	11.3	107	0.00
29 S	Terphenyl-d14	0.400	0.305	23.8	85	-0.01
30	Benzo(a)anthracene	0.400	0.348	13.0	102	0.00
31	Chrysene	0.400	0.350	12.5	105	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.345	13.8	103	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.342	14.5	100	0.00
34 I	Perylene-d12	0.400	0.400	0.0	100	0.00
35	Benzo(b)fluoranthene	0.400	0.371	7.3	102	0.00
36	Benzo(k)fluoranthene	0.400	0.370	7.5	101	0.00
37 C	Benzo(a)pyrene	0.400	0.368	8.0	102	0.00
38	Dibenzo(a,h)anthracene	0.400	0.366	8.5	100	0.00
39	Benzo(g,h,i)perylene	0.400	0.371	7.3	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0