

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010121\
 Data File : BM028311.D
 Acq On : 31 Dec 2020 18:38
 Operator : JU/CG
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 31 19:10:12 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-SIM-BM010121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 31 16:18:03 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	111	0.00
2	1,4-Dioxane	0.400	0.387	3.3	118	0.00
3	n-Nitrosodimethylamine	0.400	0.353	11.8	109	0.00
4 S	2-Fluorophenol	0.400	0.353	11.8	111	0.00
5 S	Phenol-d6	0.400	0.352	12.0	110	0.00
6	bis(2-Chloroethyl)ether	0.400	0.364	9.0	112	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	111	0.00
8 S	Nitrobenzene-d5	0.400	0.352	12.0	112	0.00
9	Naphthalene	0.400	0.355	11.3	112	0.00
10	Hexachlorobutadiene	0.400	0.356	11.0	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.351	12.3	112	0.00
12	2-Methylnaphthalene	0.400	0.351	12.3	112	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.400	0.329	17.8	105	0.00
15 S	2-Fluorobiphenyl	0.400	0.363	9.3	113	0.00
16	Acenaphthylene	0.400	0.336	16.0	112	-0.01
17	Acenaphthene	0.400	0.349	12.8	113	0.00
18	Fluorene	0.400	0.353	11.8	114	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	114	0.00
20	Hexachlorobenzene	0.400	0.351	12.3	112	0.00
21	Atrazine	0.400	0.344	14.0	118	0.00
22	Pentachlorophenol	0.400	0.350	12.5	123	0.00
23	Phenanthrene	0.400	0.349	12.8	114	0.00
24	Anthracene	0.400	0.337	15.8	114	0.00
25 SURR	Fluoranthene-d10	0.400	0.342	14.5	115	0.00
26	Fluoranthene	0.400	0.339	15.3	115	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	116	0.00
28	Pyrene	0.400	0.340	15.0	114	0.00
29 S	Terphenyl-d14	0.400	0.391	2.3	121	-0.01
30	Benzo(a)anthracene	0.400	0.342	14.5	112	0.00
31	Chrysene	0.400	0.353	11.8	117	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.339	15.3	113	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.356	11.0	116	0.00
34 I	Perylene-d12	0.400	0.400	0.0	115	0.00
35	Benzo(b)fluoranthene	0.400	0.371	7.3	118	0.00
36	Benzo(k)fluoranthene	0.400	0.360	10.0	114	0.00
37 C	Benzo(a)pyrene	0.400	0.371	7.3	118	0.00
38	Dibenzo(a,h)anthracene	0.400	0.365	8.8	115	0.00
39	Benzo(g,h,i)perylene	0.400	0.369	7.8	115	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0