

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010121\
 Data File : BM028325.D
 Acq On : 01 Jan 2021 03:04
 Operator : JU/CG
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 01 03:35:54 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-SIM-BM010121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 31 16:18:03 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	123	0.00
2	1,4-Dioxane	0.400	0.365	8.8	123	0.00
3	n-Nitrosodimethylamine	0.400	0.363	9.3	124	0.00
4 S	2-Fluorophenol	0.400	0.355	11.3	123	0.00
5 S	Phenol-d6	0.400	0.356	11.0	123	0.00
6	bis(2-Chloroethyl)ether	0.400	0.365	8.8	125	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	123	0.00
8 S	Nitrobenzene-d5	0.400	0.353	11.8	124	0.00
9	Naphthalene	0.400	0.356	11.0	124	-0.01
10	Hexachlorobutadiene	0.400	0.358	10.5	124	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.353	11.8	125	0.00
12	2-Methylnaphthalene	0.400	0.351	12.3	124	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.400	0.342	14.5	119	0.00
15 S	2-Fluorobiphenyl	0.400	0.365	8.8	123	0.00
16	Acenaphthylene	0.400	0.338	15.5	122	0.00
17	Acenaphthene	0.400	0.359	10.3	127	0.00
18	Fluorene	0.400	0.346	13.5	122	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	119	0.00
20	Hexachlorobenzene	0.400	0.354	11.5	118	0.00
21	Atrazine	0.400	0.339	15.3	121	0.00
22	Pentachlorophenol	0.400	0.313	21.8	115	-0.01
23	Phenanthrene	0.400	0.351	12.3	120	0.00
24	Anthracene	0.400	0.342	14.5	120	-0.01
25 SURR	Fluoranthene-d10	0.400	0.340	15.0	119	0.00
26	Fluoranthene	0.400	0.336	16.0	119	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	115	-0.01
28	Pyrene	0.400	0.352	12.0	117	0.00
29 S	Terphenyl-d14	0.400	0.406	-1.5	124	-0.01
30	Benzo(a)anthracene	0.400	0.346	13.5	112	-0.01
31	Chrysene	0.400	0.355	11.3	117	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.350	12.5	115	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.357	10.8	115	0.00
34 I	Perylene-d12	0.400	0.400	0.0	114	0.00
35	Benzo(b)fluoranthene	0.400	0.370	7.5	117	0.00
36	Benzo(k)fluoranthene	0.400	0.361	9.8	114	0.00
37 C	Benzo(a)pyrene	0.400	0.370	7.5	117	0.00
38	Dibenzo(a,h)anthracene	0.400	0.368	8.0	115	-0.01
39	Benzo(g,h,i)perylene	0.400	0.368	8.0	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0