

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010121\
 Data File : BM028301.D
 Acq On : 31 Dec 2020 12:00
 Operator : JU/CG
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.1

Quant Time: Dec 31 16:10:36 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-SIM-BM010121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 31 16:09:58 2020
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.143	152	544	0.40	ng	0.00
7) Naphthalene-d8	10.959	136	2235	0.40	ng	0.00
13) Acenaphthene-d10	14.763	164	1202	0.40	ng	0.00
19) Phenanthrene-d10	17.480	188	2409	0.40	ng	# 0.00
27) Chrysene-d12	21.611	240	2439	0.40	ng	# 0.00
34) Perylene-d12	23.938	264	2420	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.646	112	205	0.13	ng	0.00
5) Phenol-d6	7.281	99	273	0.14	ng	0.00
8) Nitrobenzene-d5	9.299	82	205	0.12	ng	0.00
11) 2-Methylnaphthalene-d10	12.535	152	396	0.10	ng	0.00
14) 2,4,6-Tribromophenol	16.237	330	56	0.06	ng	0.00
15) 2-Fluorobiphenyl	13.390	172	519	0.11	ng	0.00
25) Fluoranthene-d10	19.486	212	741	0.12	ng	0.00
29) Terphenyl-d14	20.071	244	670	0.12	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.416	88	79	0.13	ng	# 77
6) bis(2-Chloroethyl)ether	7.547	93	187	0.13	ng	93
9) Naphthalene	11.010	128	665	0.12	ng	# 81
10) Hexachlorobutadiene	11.289	225	129	0.09	ng	# 99
12) 2-Methylnaphthalene	12.606	142	438	0.10	ng	95
16) Acenaphthylene	14.491	152	492	0.10	ng	98
17) Acenaphthene	14.824	154	378	0.10	ng	92
18) Fluorene	15.804	166	491	0.11	ng	96
20) Hexachlorobenzene	16.800	284	163	0.08	ng	95
21) Atrazine	19.516	200	157	0.09	ng	# 36
23) Phenanthrene	17.523	178	753	0.11	ng	99
24) Anthracene	17.619	178	615	0.10	ng	97
26) Fluoranthene	19.516	202	810	0.11	ng	# 96
28) Pyrene	19.878	202	825	0.10	ng	98
30) Benzo(a)anthracene	21.590	228	832	0.11	ng	# 91
31) Chrysene	21.643	228	907	0.12	ng	# 91
32) Bis(2-ethylhexyl)phtha...	21.494	149	402	0.10	ng	96
33) Indeno(1,2,3-cd)pyrene	26.310	276	949	0.08	ng	# 91
35) Benzo(b)fluoranthene	23.229	252	931	0.11	ng	# 54
36) Benzo(k)fluoranthene	23.277	252	848	0.11	ng	# 53
37) Benzo(a)pyrene	23.835	252	750	0.10	ng	# 43
38) Dibenzo(a,h)anthracene	26.327	278	824	0.09	ng	# 55
39) Benzo(g,h,i)perylene	27.046	276	841	0.08	ng	# 64


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010121\
Data File : BM028301.D
Acq On : 31 Dec 2020 12:00
Operator : JU/CG
Sample : SSTDIC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDIC0.1


Quant Time: Dec 31 16:10:36 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-SIM-BM010121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 31 16:09:58 2020
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 8.143 min Scan# 635
Delta R.T. -0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00


Instrument :
BNA_M
ClientSampleId :
SSTDIC0.1





Tgt Ion:152 Resp: 544
Ion Ratio Lower Upper
152 100
150 144.1 116.6 174.8
115 68.0 52.9 79.3




#2
1,4-Dioxane
Concen: 0.13 ng
RT: 3.416 min Scan# 48
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00



Tgt Ion: 88 Resp: 79
Ion Ratio Lower Upper
88 100
58 100.0 59.4 89.2#
43 46.8 32.3 48.5









#4
2-Fluorophenol
Concen: 0.13 ng
RT: 5.646 min Scan# 325
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.1


Tgt Ion:	112	Resp:	205
Ion	Ratio	Lower	Upper
112	100		
64	61.5	49.5	74.3
63	35.1	32.0	48.0



#5
Phenol-d6
Concen: 0.14 ng
RT: 7.281 min Scan# 528
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt Ion:	99	Resp:	273
Ion	Ratio	Lower	Upper
99	100		
42	31.9	22.4	33.6
71	37.0	37.8	56.6









#6
bis(2-Chloroethyl)ether
Concen: 0.13 ng
RT: 7.547 min Scan# 561
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.1


Tgt Ion: 93	Resp: 187
Ion Ratio	Lower Upper
93 100	
63 86.6	63.1 94.7
95 34.2	26.1 39.1



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.959 min Scan# 874
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt	Ion:136	Resp:	2235
Ion	Ratio	Lower	Upper
136	100		
137	14.3	10.6	15.8
54	10.8	8.6	12.8
68	8.1	5.7	8.5









#8
Nitrobenzene-d5
Concen: 0.12 ng
RT: 9.299 min Scan# 743
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.1


Tgt Ion: 82 Resp: 205
Ion Ratio Lower Upper
82 100
128 55.1 30.6 46.0#
54 81.0 48.3 72.5#



#9
Naphthalene
Concen: 0.12 ng
RT: 11.010 min Scan# 878
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt Ion: 128 Resp: 665
Ion Ratio Lower Upper
128 100
129 20.5 9.8 14.8#
127 22.6 12.2 18.4#









#10
Hexachlorobutadiene
Concen: 0.09 ng
RT: 11.289 min Scan# 900
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDICC0.1


Tgt Ion:225 Resp: 129
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.3 51.0 76.6



#11
2-Methylnaphthalene-d10
Concen: 0.10 ng
RT: 12.535 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt Ion:152 Resp: 396
Ion Ratio Lower Upper
152 100
151 20.2 16.7 25.1









#12
2-Methylnaphthalene
Concen: 0.10 ng
RT: 12.606 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.1


Tgt Ion:	142	Resp:	438
Ion	Ratio	Lower	Upper
142	100		
141	90.9	69.4	104.2
115	48.0	35.7	53.5



#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.763 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt Ion:	164	Resp:	1202
Ion	Ratio	Lower	Upper
164	100		
162	100.2	80.6	120.8
160	48.9	39.5	59.3









#14
2,4,6-Tribromophenol
Concen: 0.06 ng
RT: 16.237 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDICC0.1


Tgt Ion:330 Resp: 56
Ion Ratio Lower Upper
330 100
332 105.4 0.0 0.0#
141 46.4 34.4 51.6



#15
2-Fluorobiphenyl
Concen: 0.11 ng
RT: 13.390 min Scan# 1291
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt Ion:172 Resp: 519
Ion Ratio Lower Upper
172 100
171 45.4 28.2 42.4#
170 33.1 20.0 30.0#









#16
Acenaphthylene
Concen: 0.10 ng
RT: 14.491 min Scan# 1400
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.1


Tgt Ion:152	Resp:	492
Ion Ratio	Lower	Upper
152	100	
151	21.3	15.8 23.8
153	13.0	10.6 16.0



#17
Acenaphthene
Concen: 0.10 ng
RT: 14.824 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt Ion:154	Resp:	378
Ion Ratio	Lower	Upper
154	100	
153	116.1	83.6 125.4
152	56.3	43.9 65.9









#18
Fluorene
Concen: 0.11 ng
RT: 15.804 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.1


Tgt Ion:166 Resp: 491
Ion Ratio Lower Upper
166 100
165 94.5 78.6 118.0
167 14.3 10.7 16.1



#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.480 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt Ion:188 Resp: 2409
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 14.6 5.0 7.4#









#20
Hexachlorobenzene
Concen: 0.08 ng
RT: 16.800 min Scan# 1624
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.1


Tgt Ion:284	Resp:	163
Ion Ratio	Lower	Upper
284	100	
142	35.0	32.0 48.0
249	33.1	27.0 40.6



#21
Atrazine
Concen: 0.09 ng
RT: 19.516 min Scan# 2041
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt Ion:200	Resp:	157
Ion Ratio	Lower	Upper
200	100	
173	0.0	22.8 34.2#
215	0.0	38.1 57.1#









#23
Phenanthrene
Concen: 0.11 ng
RT: 17.523 min Scan# 1692
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.1


Tgt Ion:178 Resp: 753
Ion Ratio Lower Upper
178 100
176 18.5 14.8 22.2
179 16.5 12.4 18.6



#24
Anthracene
Concen: 0.10 ng
RT: 17.619 min Scan# 1701
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt Ion:178 Resp: 615
Ion Ratio Lower Upper
178 100
176 17.1 14.8 22.2
179 16.7 12.5 18.7









#25
Fluoranthene-d10
Concen: 0.12 ng
RT: 19.486 min Scan# 2035
Delta R.T. -0.005 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.1


Tgt Ion:212 Resp: 741
Ion Ratio Lower Upper
212 100
106 12.6 6.8 10.2#
104 7.8 4.1 6.1#



#26
Fluoranthene
Concen: 0.11 ng
RT: 19.516 min Scan# 2041
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt Ion:202 Resp: 810
Ion Ratio Lower Upper
202 100
101 11.0 5.8 8.6#
203 17.4 13.7 20.5









#27
Chrysene-d12
Concen: 0.40 ng
RT: 21.611 min Scan# 2291
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.1


Tgt Ion:240 Resp: 2439
Ion Ratio Lower Upper
240 100
120 10.3 5.3 7.9#
236 26.5 21.8 32.8



#28
Pyrene
Concen: 0.10 ng
RT: 19.878 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt Ion:202 Resp: 825
Ion Ratio Lower Upper
202 100
200 20.0 16.6 24.8
203 18.1 13.8 20.8









#29
 Terphenyl-d14
 Concen: 0.12 ng
 RT: 20.071 min Scan# 2146
 Delta R.T. 0.000 min
 Lab File: BM028301.D
 Acq: 31 Dec 2020 12:00




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:244 Resp: 670
 Ion Ratio Lower Upper
 244 100
 212 13.3 7.3 10.9#
 122 12.8 7.3 10.9#



#30
 Benzo(a)anthracene
 Concen: 0.11 ng
 RT: 21.590 min Scan# 2289
 Delta R.T. 0.000 min
 Lab File: BM028301.D
 Acq: 31 Dec 2020 12:00

Tgt Ion:228 Resp: 832
 Ion Ratio Lower Upper
 228 100
 226 30.9 22.3 33.5
 229 25.7 15.7 23.5#





#31

Chrysene

Concen: 0.12 ng

RT: 21.643 min Scan# 2294

Delta R.T. 0.000 min

Lab File: BM028301.D

Acq: 31 Dec 2020 12:00

Tgt Ion:228 Resp: 907

Ion Ratio Lower Upper

228 100

226 34.1 24.2 36.2





229 25.5 15.7 23.5#

Instrument :

BNA_M

ClientSampleId :

SSTDIC0.1



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.10 ng

RT: 21.494 min Scan# 2280

Delta R.T. 0.000 min

Lab File: BM028301.D

Acq: 31 Dec 2020 12:00




Tgt Ion:149 Resp: 402


Ion Ratio Lower Upper

149 100

167 26.4 23.2 34.8

279 5.5 3.9 5.9









#33
Indeno(1,2,3-cd)pyrene
Concen: 0.08 ng
RT: 26.310 min Scan# 3506
Delta R.T. -0.003 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDICC0.1


Tgt Ion:276 Resp: 949
Ion Ratio Lower Upper
276 100
138 26.0 13.3 19.9#
277 25.2 20.2 30.4



#34
Perylene-d12
Concen: 0.40 ng
RT: 23.938 min Scan# 2813
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt Ion:264 Resp: 2420
Ion Ratio Lower Upper
264 100
260 25.5 19.8 29.6
265 72.4 51.4 77.0









#35
Benzo(b)fluoranthene
Concen: 0.11 ng
RT: 23.229 min Scan# 2606
Delta R.T. -0.003 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.1


Tgt Ion:252 Resp: 931
Ion Ratio Lower Upper
252 100
253 43.9 20.1 30.1#
125 34.3 7.2 10.8#



#36
Benzo(k)fluoranthene
Concen: 0.11 ng
RT: 23.277 min Scan# 2620
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt Ion:252 Resp: 848
Ion Ratio Lower Upper
252 100
253 43.7 19.8 29.8#
125 35.5 7.1 10.7#









#37
Benzo(a)pyrene
Concen: 0.10 ng
RT: 23.835 min Scan# 2783
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00




Instrument :
BNA_M
ClientSampleId :
SSTDIC0.1


Tgt Ion:252 Resp: 750
Ion Ratio Lower Upper
252 100
253 49.7 21.3 31.9#
125 42.8 8.3 12.5#



#38
Dibenzo(a,h)anthracene
Concen: 0.09 ng
RT: 26.327 min Scan# 3511
Delta R.T. 0.000 min
Lab File: BM028301.D
Acq: 31 Dec 2020 12:00

Tgt Ion:278 Resp: 824
Ion Ratio Lower Upper
278 100
139 31.4 9.3 13.9#
279 48.8 21.6 32.4#





#39

Benzo(g,h,i)perylene


Concen: 0.08 ng

RT: 27.046 min Scan# 3721

Delta R.T. -0.003 min

Lab File: BM028301.D

Acq: 31 Dec 2020 12:00





Tgt Ion: 276 Resp: 841

Ion Ratio Lower Upper

276 100

277 37.5 19.8 29.8#

138 36.6 11.7 17.5#



Instrument :

BNA_M

ClientSampleId :

SSTDIC0.1