

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010121\
 Data File : BM028329.D
 Acq On : 01 Jan 2021 09:22
 Operator : JU/CG
 Sample : L5123-17MSD
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 TT180D1-20201214MSD

Quant Time: Jan 01 22:56:52 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-SIM-BM010121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 31 16:18:03 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.135	152	706	0.40	ng	0.00
7) Naphthalene-d8	10.959	136	2864	0.40	ng	0.00
13) Acenaphthene-d10	14.764	164	1522	0.40	ng	0.00
19) Phenanthrene-d10	17.481	188	3071	0.40	ng	# 0.00
27) Chrysene-d12	21.601	240	3091	0.40	ng	#-0.01
34) Perylene-d12	23.927	264	2993	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.638	112	230	0.10	ng	0.00
5) Phenol-d6	7.273	99	178	0.05	ng	0.00
8) Nitrobenzene-d5	9.299	82	705	0.27	ng	0.00
11) 2-Methylnaphthalene-d10	12.531	152	1304	0.26	ng	0.00
14) 2,4,6-Tribromophenol	16.237	330	198	0.25	ng	0.00
15) 2-Fluorobiphenyl	13.390	172	1743	0.27	ng	0.00
25) Fluoranthene-d10	19.486	212	3253	0.34	ng	0.00
29) Terphenyl-d14	20.061	244	3027	0.40	ng	-0.01
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.407	88	1926	1.94	ng	87
3) n-Nitrosodimethylamine	3.762	42	178	0.12	ng	# 96
6) bis(2-Chloroethyl)ether	7.547	93	750	0.33	ng	96
9) Naphthalene	10.997	128	2452	0.30	ng	97
10) Hexachlorobutadiene	11.289	225	376	0.24	ng	# 99
12) 2-Methylnaphthalene	12.602	142	1635	0.29	ng	96
16) Acenaphthylene	14.481	152	2078	0.31	ng	99
17) Acenaphthene	14.824	154	1487	0.30	ng	96
18) Fluorene	15.794	166	1898	0.30	ng	99
20) Hexachlorobenzene	16.790	284	640	0.30	ng	94
21) Atrazine	19.511	200	772	0.37	ng	# 36
22) Pentachlorophenol	17.130	266	564	0.65	ng	97
23) Phenanthrene	17.523	178	3360	0.35	ng	99
24) Anthracene	17.608	178	2882	0.34	ng	98
26) Fluoranthene	19.511	202	3904	0.37	ng	# 96
28) Pyrene	19.874	202	4005	0.38	ng	98
30) Benzo(a)anthracene	21.590	228	4018	0.39	ng	96
31) Chrysene	21.632	228	4124	0.39	ng	99
32) Bis(2-ethylhexyl)phtha...	21.494	149	2547	0.52	ng	95
33) Indeno(1,2,3-cd)pyrene	26.306	276	4584	0.39	ng	# 90
35) Benzo(b)fluoranthene	23.222	252	4372	0.39	ng	# 91
36) Benzo(k)fluoranthene	23.267	252	4089	0.39	ng	# 89
37) Benzo(a)pyrene	23.825	252	3583	0.38	ng	# 88
38) Dibenzo(a,h)anthracene	26.313	278	3978	0.39	ng	# 89
39) Benzo(g,h,i)perylene	27.035	276	3987	0.39	ng	# 87


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010121\
Data File : BM028329.D
Acq On : 01 Jan 2021 09:22
Operator : JU/CG
Sample : L5123-17MSD
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD

Quant Time: Jan 01 22:56:52 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-SIM-BM010121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 31 16:18:03 2020
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 8.135 min Scan# 634
 Delta R.T. -0.008 min
 Lab File: BM028329.D
 Acq: 01 Jan 2021 09:22




Instrument :
 BNA_M
 ClientSampleId :
 TT180D1-20201214MSD


Tgt Ion: 152 Resp: 706
 Ion Ratio Lower Upper
 152 100
 150 148.7 116.6 174.8
 115 66.9 52.9 79.3



#2
 1,4-Dioxane
 Concen: 1.94 ng
 RT: 3.407 min Scan# 47
 Delta R.T. -0.008 min
 Lab File: BM028329.D
 Acq: 01 Jan 2021 09:22

Tgt Ion: 88 Resp: 1926
 Ion Ratio Lower Upper
 88 100
 58 88.8 59.4 89.2
 43 44.3 32.3 48.5









#3
n-Nitrosodimethylamine
Concen: 0.12 ng
RT: 3.762 min Scan# 91
Delta R.T. -0.008 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:	42	Resp:	178
Ion Ratio	Lower	Upper	
42	100		
74	83.7	64.4	96.6
44	0.0	2.8	4.2#



#4
2-Fluorophenol
Concen: 0.10 ng
RT: 5.638 min Scan# 324
Delta R.T. -0.008 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Tgt Ion:	112	Resp:	230
Ion Ratio	Lower	Upper	
112	100		
64	68.7	49.5	74.3
63	37.0	32.0	48.0









#5
Phenol-d6
Concen: 0.05 ng
RT: 7.273 min Scan# 527
Delta R.T. -0.008 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:	99	Resp:	178
Ion	Ratio	Lower	Upper
99	100		
42	27.0	22.4	33.6
71	43.8	37.8	56.6



#6
bis(2-Chloroethyl)ether
Concen: 0.33 ng
RT: 7.547 min Scan# 561
Delta R.T. 0.000 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Tgt Ion:	93	Resp:	750
Ion	Ratio	Lower	Upper
93	100		
63	84.3	63.1	94.7
95	32.5	26.1	39.1









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.959 min Scan# 874
Delta R.T. 0.000 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:	136	Resp:	2864
Ion	Ratio	Lower	Upper
136	100		
137	13.8	10.6	15.8
54	8.6	8.6	12.8
68	6.5	5.7	8.5



#8
Nitrobenzene-d5
Concen: 0.27 ng
RT: 9.299 min Scan# 743
Delta R.T. 0.000 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Tgt Ion:	82	Resp:	705
Ion	Ratio	Lower	Upper
82	100		
128	44.2	30.6	46.0
54	65.1	48.3	72.5









#9
Naphthalene
Concen: 0.30 ng
RT: 10.997 min Scan# 877
Delta R.T. -0.013 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:	128	Resp:	2452
Ion Ratio	Lower	Upper	
128	100		
129	13.9	9.8	14.8
127	16.2	12.2	18.4



#10
Hexachlorobutadiene
Concen: 0.24 ng
RT: 11.289 min Scan# 900
Delta R.T. 0.000 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Tgt Ion:	225	Resp:	376
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.4	51.0	76.6









#11
2-Methylnaphthalene-d10
Concen: 0.26 ng
RT: 12.531 min Scan# 1146
Delta R.T. -0.004 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:152 Resp: 1304
Ion Ratio Lower Upper
152 100
151 20.3 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.29 ng
RT: 12.602 min Scan# 1162
Delta R.T. -0.004 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Tgt Ion:142 Resp: 1635
Ion Ratio Lower Upper
142 100
141 88.7 69.4 104.2
115 39.4 35.7 53.5









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.764 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:	164	Resp:	1522
Ion	Ratio	Lower	Upper
164	100		
162	99.4	80.6	120.8
160	46.1	39.5	59.3



#14
2,4,6-Tribromophenol
Concen: 0.25 ng
RT: 16.237 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Tgt Ion:	330	Resp:	198
Ion	Ratio	Lower	Upper
330	100		
332	95.5	0.0	0.0#
141	37.4	34.4	51.6









#15
2-Fluorobiphenyl
Concen: 0.27 ng
RT: 13.390 min Scan# 1291
Delta R.T. 0.000 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:172	Resp:	1743
Ion Ratio	Lower	Upper
172	100	
171	37.3	28.2 42.4
170	25.0	20.0 30.0



#16
Acenaphthylene
Concen: 0.31 ng
RT: 14.481 min Scan# 1399
Delta R.T. -0.010 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Tgt Ion:152	Resp:	2078
Ion Ratio	Lower	Upper
152	100	
151	19.3	15.8 23.8
153	12.9	10.6 16.0









#17
Acenaphthene
Concen: 0.30 ng
RT: 14.824 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:	154	Resp:	1487
Ion	Ratio	Lower	Upper
154	100		
153	109.5	83.6	125.4
152	53.0	43.9	65.9



#18
Fluorene
Concen: 0.30 ng
RT: 15.794 min Scan# 1529
Delta R.T. -0.010 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Tgt Ion:	166	Resp:	1898
Ion	Ratio	Lower	Upper
166	100		
165	99.0	78.6	118.0
167	14.1	10.7	16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.481 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:188 Resp: 3071
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.1 5.0 7.4#



#20
Hexachlorobenzene
Concen: 0.30 ng
RT: 16.790 min Scan# 1623
Delta R.T. -0.011 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Tgt Ion:284 Resp: 640
Ion Ratio Lower Upper
284 100
142 34.8 32.0 48.0
249 32.0 27.0 40.6









#21
Atrazine
Concen: 0.37 ng
RT: 19.511 min Scan# 2040
Delta R.T. -0.005 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:200 Resp: 772
Ion Ratio Lower Upper
200 100
173 0.0 22.8 34.2#
215 0.0 38.1 57.1#



#22
Pentachlorophenol
Concen: 0.65 ng
RT: 17.130 min Scan# 1655
Delta R.T. -0.011 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Tgt Ion:266 Resp: 564
Ion Ratio Lower Upper
266 100
264 63.8 50.9 76.3
268 68.1 51.5 77.3









#23
Phenanthrene
Concen: 0.35 ng
RT: 17.523 min Scan# 1692
Delta R.T. 0.000 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:	178	Resp:	3360
Ion Ratio	Lower	Upper	
178	100		
176	18.2	14.8	22.2
179	16.0	12.4	18.6



#24
Anthracene
Concen: 0.34 ng
RT: 17.608 min Scan# 1700
Delta R.T. -0.011 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Tgt Ion:	178	Resp:	2882
Ion Ratio	Lower	Upper	
178	100		
176	17.6	14.8	22.2
179	14.7	12.5	18.7









#25
Fluoranthene-d10
Concen: 0.34 ng
RT: 19.486 min Scan# 2035
Delta R.T. -0.005 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:212 Resp: 3253
Ion Ratio Lower Upper
212 100
106 12.7 6.8 10.2#
104 7.6 4.1 6.1#



#26
Fluoranthene
Concen: 0.37 ng
RT: 19.511 min Scan# 2040
Delta R.T. -0.005 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Tgt Ion:202 Resp: 3904
Ion Ratio Lower Upper
202 100
101 11.9 5.8 8.6#
203 17.2 13.7 20.5





#27

Chrysene-d12

Concen: 0.40 ng

RT: 21.601 min Scan# 2290

Delta R.T. -0.011 min

Lab File: BM028329.D

Acq: 01 Jan 2021 09:22

Tgt Ion:240 Resp: 3091

Ion Ratio Lower Upper

240 100

120 11.1 5.3 7.9#





236 26.9 21.8 32.8

Instrument :

BNA_M

ClientSampleId :

TT180D1-20201214MSD



#28

Pyrene

Concen: 0.38 ng

RT: 19.874 min Scan# 2113

Delta R.T. -0.005 min

Lab File: BM028329.D

Acq: 01 Jan 2021 09:22




Tgt Ion:202 Resp: 4005


Ion Ratio Lower Upper

202 100

200 20.0 16.6 24.8

203 18.1 13.8 20.8









#29
 Terphenyl-d14
 Concen: 0.40 ng
 RT: 20.061 min Scan# 2145
 Delta R.T. -0.011 min
 Lab File: BM028329.D
 Acq: 01 Jan 2021 09:22




Instrument :
 BNA_M
 ClientSampleId :
 TT180D1-20201214MSD


Tgt Ion:244 Resp: 3027
 Ion Ratio Lower Upper
 244 100
 212 8.2 7.3 10.9
 122 9.7 7.3 10.9



#30
 Benzo(a)anthracene
 Concen: 0.39 ng
 RT: 21.590 min Scan# 2289
 Delta R.T. 0.000 min
 Lab File: BM028329.D
 Acq: 01 Jan 2021 09:22

Tgt Ion:228 Resp: 4018
 Ion Ratio Lower Upper
 228 100
 226 26.2 22.3 33.5
 229 21.9 15.7 23.5





#31

Chrysene

Concen: 0.39 ng

RT: 21.632 min Scan# 2293

Delta R.T. -0.011 min

Lab File: BM028329.D

Acq: 01 Jan 2021 09:22

Tgt Ion:228 Resp: 4124

Ion Ratio Lower Upper

228 100

226 30.4 24.2 36.2





229 20.6 15.7 23.5

Instrument :

BNA_M

ClientSampleId :

TT180D1-20201214MSD



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.52 ng

RT: 21.494 min Scan# 2280

Delta R.T. 0.000 min

Lab File: BM028329.D

Acq: 01 Jan 2021 09:22




Tgt Ion:149 Resp: 2547


Ion Ratio Lower Upper

149 100

167 25.9 23.2 34.8

279 4.4 3.9 5.9









#33
Indeno(1,2,3-cd)pyrene
Concen: 0.39 ng
RT: 26.306 min Scan# 3505
Delta R.T. -0.007 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:276 Resp: 4584
Ion Ratio Lower Upper
276 100
138 27.5 13.3 19.9#
277 25.3 20.2 30.4



#34
Perylene-d12
Concen: 0.40 ng
RT: 23.927 min Scan# 2810
Delta R.T. -0.010 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Tgt Ion:264 Resp: 2993
Ion Ratio Lower Upper
264 100
260 25.0 19.8 29.6
265 62.6 51.4 77.0









#35
Benzo(b)fluoranthene
Concen: 0.39 ng
RT: 23.222 min Scan# 2604
Delta R.T. -0.010 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:252 Resp: 4372
Ion Ratio Lower Upper
252 100
253 27.5 20.1 30.1
125 17.1 7.2 10.8#



#36
Benzo(k)fluoranthene
Concen: 0.39 ng
RT: 23.267 min Scan# 2617
Delta R.T. -0.010 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Tgt Ion:252 Resp: 4089
Ion Ratio Lower Upper
252 100
253 27.8 19.8 29.8
125 17.3 7.1 10.7#









#37
Benzo(a)pyrene
Concen: 0.38 ng
RT: 23.825 min Scan# 2780
Delta R.T. -0.010 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22




Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD


Tgt Ion:252 Resp: 3583
Ion Ratio Lower Upper
252 100
253 29.5 21.3 31.9
125 20.6 8.3 12.5#



#38
Dibenzo(a,h)anthracene
Concen: 0.39 ng
RT: 26.313 min Scan# 3507
Delta R.T. -0.014 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22


Tgt Ion:278 Resp: 3978
Ion Ratio Lower Upper
278 100
139 22.2 9.3 13.9#
279 29.1 21.6 32.4





#39
Benzo(g,h,i)perylene
Concen: 0.39 ng
RT: 27.035 min Scan# 3718
Delta R.T. -0.014 min
Lab File: BM028329.D
Acq: 01 Jan 2021 09:22

Instrument :
BNA_M
ClientSampleId :
TT180D1-20201214MSD



Tgt Ion	Ratio	Lower	Upper
276	100		
277	27.0	19.8	29.8
138	26.1	11.7	17.5#

