


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010422\
 Data File : BM033620.D
 Acq On : 04 Jan 2022 11:46
 Operator : CG/JU
 Sample : DFTPP034
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP034

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM010422.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jan 04 15:58:35 2022



AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2471

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	45.7	112802	PASS
68	69	0.00	2	2.0	2152	PASS
69	198	0.00	100	43.2	106493	PASS
70	69	0.00	2	0.5	565	PASS
127	198	10	80	50.0	123320	PASS
197	198	0.00	2	0.6	1365	PASS
198	198	100	100	100.0	246613	PASS
199	198	5	9	6.9	16960	PASS
275	198	10	60	27.2	67037	PASS
365	198	1	100	4.4	10828	PASS
441	198	0.01	100	11.7	28899	PASS
442	198	50	100	75.7	186752	PASS
443	442	15	24	19.0	35469	PASS

m/z	Abundance
39.00	270.0
40.00	966.0
41.10	325.0
43.00	221.0
44.00	2189.0
45.00	179.0
48.90	352.0
50.00	214.0
51.00	401.0
53.10	167.0
55.00	180.0
55.90	160.0
65.00	231.0
73.00	543.0
74.00	185.0
74.90	210.0
76.10	171.0
76.90	436.0
78.20	256.0
79.00	247.0
79.90	178.0
84.00	329.0
85.90	228.0
91.10	193.0
93.00	529.0
95.80	224.0
103.90	178.0
105.10	298.0
107.90	180.0
121.10	204.0
132.90	218.0
135.00	176.0
147.20	246.0
149.00	356.0
191.00	349.0
193.10	832.0
207.10	2890.0
208.00	534.0
209.00	374.0
221.10	220.0
281.10	492.0

Instrument :
BNA_M
ClientSampleId :
DFTPP034

m/z	Abundance
36.00	191.0
37.00	707.0
38.10	1713.0
39.10	10018.0
40.00	1281.0
41.10	585.0
42.90	398.0
44.00	1745.0
45.10	412.0
47.00	169.0
49.00	1588.0
50.10	35240.0
51.10	136832.0
52.10	7079.0
53.10	469.0
55.10	659.0
56.00	4250.0
57.00	9273.0
58.00	393.0
59.10	233.0
61.00	2198.0
62.10	2090.0
63.00	5824.0
64.10	877.0
65.00	2569.0
66.10	182.0
67.00	340.0
68.00	2705.0
69.00	129720.0
69.90	567.0
73.00	2161.0
74.00	13383.0
75.00	20984.0
76.10	7608.0
77.10	143232.0
78.10	9933.0
79.00	10332.0
80.00	7849.0
81.00	10383.0
82.00	2680.0
83.10	2295.0
84.00	448.0
85.00	1681.0
86.00	3375.0
87.00	1621.0
88.00	394.0
89.10	209.0
91.00	2311.0
92.00	3132.0
93.00	17832.0
94.00	1159.0
94.90	274.0
96.00	822.0
97.00	342.0
98.00	14019.0
99.00	11170.0
100.00	979.0
101.00	5898.0
102.00	422.0
103.00	2306.0
104.00	3840.0
105.00	3984.0
106.00	1345.0
107.00	44552.0
108.00	7616.0
109.10	1276.0
110.00	77040.0
111.00	12429.0
112.10	1221.0
112.80	383.0
115.10	276.0
116.00	2277.0
117.00	39816.0
118.00	3063.0
118.90	343.0
120.00	343.0
120.90	365.0
122.00	3532.0
123.00	5126.0
124.00	2037.0
125.00	2015.0
126.10	821.0
127.00	142528.0
128.00	11405.0
129.00	59072.0
130.00	4988.0
130.90	943.0
131.90	554.0
133.00	493.0
134.00	1821.0
135.00	4313.0

Instrument :
BNA_M
ClientSampleId :
DFTPP034

136.00	1759.0
137.00	2883.0
137.90	631.0
139.00	381.0
140.00	799.0
141.00	7705.0
142.00	2963.0
143.00	1644.0
144.00	579.0
145.00	421.0
146.00	1537.0
147.00	4100.0
148.00	9963.0
149.00	2568.0
150.00	454.0
151.10	843.0
151.90	516.0
153.00	2666.0
154.10	2012.0
155.00	4325.0
156.10	6434.0
157.00	1062.0
158.00	1641.0
159.00	1008.0
160.00	2843.0
161.10	3903.0
162.00	853.0
163.00	319.0
164.00	421.0
165.00	2958.0
166.00	2419.0
167.00	15635.0
168.00	8739.0
169.00	1349.0
170.10	586.0
170.90	707.0
172.00	1471.0
173.00	1952.0
174.00	3153.0
175.10	5883.0
176.10	1601.0
177.00	3117.0
178.00	1186.0
179.00	11509.0
180.00	8445.0
181.00	3327.0
182.00	449.0
183.00	359.0
184.00	983.0
185.00	5130.0
186.10	41512.0
187.00	12354.0
188.10	1398.0
189.00	3049.0
190.00	628.0
191.00	1209.0
192.00	3295.0
193.00	4530.0
194.00	956.0
195.10	466.0
196.00	6925.0
197.10	1535.0
198.00	267008.0
199.00	18840.0
200.00	1620.0
201.60	1021.0
203.00	2406.0
204.00	11689.0
205.10	18200.0
206.10	75480.0
207.10	11805.0
208.00	3492.0
209.00	971.0
210.00	1574.0
211.10	2999.0
212.00	524.0
213.10	379.0
215.00	951.0
216.00	1918.0
217.00	22568.0
218.00	2620.0
219.00	186.0
220.00	506.0
221.10	9231.0
221.90	2265.0
223.00	4330.0
224.10	41072.0
225.10	10581.0
226.10	1320.0
227.00	20016.0
228.00	2781.0
229.00	4111.0
230.00	450.0
231.10	1480.0
232.10	231.0

Instrument :
BNA_M
ClientSampleId :
DFTPP034

233.00	282.0
234.00	876.0
235.00	1303.0
236.00	843.0
237.00	1548.0
238.00	194.0
239.00	850.0
240.00	539.0
241.00	1031.0
242.00	2043.0
243.10	2077.0
244.10	29592.0
245.10	4405.0
246.00	7407.0
247.00	1337.0
247.90	264.0
249.00	1233.0
250.90	350.0
252.00	481.0
253.10	627.0
254.10	1303.0
255.00	167488.0
256.00	23880.0
257.00	2317.0
258.00	10635.0
259.10	1625.0
260.00	290.0
260.90	303.0
263.10	166.0
264.00	266.0
265.00	4044.0
265.90	750.0
267.00	326.0
270.00	188.0
270.90	445.0
272.00	583.0
273.00	4602.0
274.00	12660.0
275.00	71072.0
276.00	8949.0
277.10	5718.0
278.00	1460.0
279.00	329.0
281.10	544.0
281.90	419.0
283.00	620.0
284.00	523.0
285.10	1033.0
286.10	173.0
289.10	195.0
290.00	290.0
292.10	394.0
292.90	1496.0
293.90	276.0
295.00	513.0
296.00	23312.0
297.00	2977.0
297.90	207.0
300.90	189.0
302.10	441.0
303.00	2764.0
304.00	695.0
308.00	376.0
309.00	204.0
310.00	246.0
313.10	199.0
314.00	1124.0
315.00	3090.0
316.00	1301.0
317.10	271.0
321.00	650.0
322.10	337.0
323.10	6684.0
324.10	1014.0
326.10	150.0
327.00	1734.0
328.10	755.0
332.00	511.0
333.00	774.0
334.10	4244.0
335.00	1037.0
336.00	169.0
339.80	182.0
341.10	872.0
342.10	175.0
346.00	1294.0
346.80	160.0
351.00	151.0
352.00	2143.0
353.00	1220.0
354.00	2400.0
355.00	655.0
359.10	200.0
365.00	9956.0
366.00	1627.0

Instrument :
BNA_M
ClientSampleId :
DFTPP034

370.10	242.0
371.00	521.0
372.00	3451.0
373.00	774.0
373.90	182.0
383.00	1185.0
384.00	231.0
390.10	436.0
391.00	277.0
401.00	226.0
402.10	1653.0
403.00	2180.0
404.00	926.0
421.00	1627.0
422.00	1597.0
423.00	10224.0
424.10	2841.0
425.00	224.0
441.10	25136.0
442.10	166272.0
443.00	32272.0
444.10	2830.0
445.00	231.0

Instrument :
BNA_M
ClientSampleId :
DFTPP034

m/z	Abundance
37.00	673.0
38.10	1589.0
39.10	8962.0
40.00	1425.0
41.10	477.0
42.10	184.0
43.00	225.0
44.00	1884.0
44.90	461.0
49.10	1192.0
50.10	33648.0
51.10	131200.0
52.10	6705.0
53.00	357.0
55.00	818.0
56.00	3711.0
57.00	8626.0
57.90	430.0
59.00	325.0
61.10	1469.0
62.10	2059.0
63.00	5291.0
64.00	743.0
65.10	2569.0
65.90	269.0
67.00	317.0
68.00	2353.0
69.00	121296.0
70.00	673.0
73.10	1700.0
74.10	13080.0
75.00	20032.0
76.10	6730.0
77.10	133376.0
78.10	9095.0
79.00	10562.0
80.00	7404.0
81.00	10466.0
82.00	2282.0
83.00	2205.0
84.00	476.0
85.00	1875.0
86.00	3219.0
87.00	1522.0
88.10	449.0
89.00	237.0
91.00	2416.0
92.00	2548.0
93.00	17024.0
94.10	1118.0
95.00	316.0
96.00	964.0
96.90	248.0
97.20	284.0
98.00	13635.0
99.00	10626.0
100.00	1028.0
101.00	5907.0
101.80	339.0
103.00	2205.0
104.00	3824.0
105.00	3412.0
106.10	1172.0
107.00	43664.0
108.00	7338.0
109.10	1340.0
110.00	74104.0
111.00	10946.0
112.00	1155.0
113.00	662.0
114.80	284.0
116.10	2383.0
117.00	40368.0
118.00	2861.0
118.90	449.0
120.00	697.0
120.90	308.0
122.00	3046.0
123.00	4393.0
124.00	1883.0
125.00	1987.0
126.10	847.0
127.00	142272.0
128.00	10285.0
129.00	58880.0
130.00	5192.0
131.00	925.0
131.90	461.0
133.00	376.0
133.90	1910.0
135.00	4959.0

Instrument :
BNA_M
ClientSampleId :
DFTPP034

136.00	1888.0
137.00	2704.0
138.00	560.0
138.90	328.0
140.10	844.0
141.00	7988.0
142.10	2522.0
143.00	1869.0
143.90	649.0
144.90	407.0
146.00	1702.0
147.00	4635.0
148.00	9681.0
149.10	2260.0
150.00	546.0
150.90	665.0
151.90	630.0
153.00	2288.0
154.00	2002.0
155.00	4616.0
156.00	6781.0
157.10	1356.0
158.00	1816.0
159.00	1232.0
159.90	2337.0
161.00	3440.0
162.00	1181.0
162.90	286.0
163.90	366.0
165.00	2947.0
166.00	2238.0
167.00	16480.0
168.00	8621.0
169.00	1298.0
169.90	775.0
171.00	751.0
172.00	1537.0
173.00	2339.0
174.00	3262.0
175.00	6489.0
176.00	1668.0
177.00	3418.0
178.00	1093.0
179.00	11949.0
180.10	8395.0
181.00	3924.0
181.90	689.0
183.00	305.0
184.00	953.0
185.00	5879.0
186.00	43240.0
187.00	12692.0
188.00	981.0
189.00	3475.0
189.90	654.0
191.00	1435.0
192.00	3424.0
193.00	4320.0
194.00	870.0
195.00	428.0
196.00	8319.0
197.10	1627.0
198.00	291200.0
199.00	19472.0
200.00	1759.0
201.50	1164.0
203.00	2543.0
204.00	12638.0
205.00	20024.0
206.10	80936.0
207.10	12661.0
208.00	3676.0
209.00	1287.0
210.00	1088.0
211.00	3417.0
211.70	481.0
213.10	303.0
215.00	1201.0
216.00	1723.0
217.00	22832.0
218.00	2877.0
218.90	218.0
220.20	541.0
221.10	9367.0
221.90	2808.0
223.00	5363.0
224.10	45088.0
225.10	11916.0
226.10	1177.0
227.00	21928.0
228.00	2855.0
229.00	4287.0
230.10	655.0
231.00	1404.0
232.10	196.0

Instrument :
BNA_M
ClientSampleId :
DFTPP034

233.00	445.0
234.00	1405.0
235.00	1468.0
236.00	833.0
237.10	1565.0
237.90	198.0
239.00	965.0
239.90	563.0
241.00	1324.0
242.00	2481.0
243.10	2445.0
244.10	33200.0
245.10	4784.0
246.10	7633.0
247.00	1404.0
247.90	391.0
249.00	1261.0
250.00	280.0
250.90	410.0
252.00	443.0
253.00	665.0
254.10	1549.0
255.00	181504.0
256.10	26352.0
257.00	1936.0
258.00	11561.0
259.00	1751.0
259.80	332.0
261.00	351.0
264.00	466.0
265.00	5015.0
266.00	717.0
267.10	177.0
270.10	261.0
271.00	389.0
272.10	574.0
273.00	5530.0
274.00	13965.0
275.00	76392.0
276.00	10726.0
277.00	8673.0
278.00	1426.0
278.90	258.0
281.00	473.0
282.00	467.0
283.10	846.0
284.00	632.0
285.10	1334.0
286.10	263.0
288.90	347.0
289.90	288.0
290.90	221.0
292.10	499.0
293.00	1808.0
293.90	424.0
295.10	757.0
296.00	26560.0
297.00	3366.0
298.00	157.0
301.00	397.0
302.00	407.0
303.00	2798.0
304.10	779.0
308.00	402.0
308.90	267.0
310.00	179.0
313.10	182.0
314.00	1104.0
315.00	2982.0
316.10	1314.0
317.10	244.0
321.00	921.0
321.90	383.0
323.10	7811.0
324.00	1367.0
325.10	240.0
326.10	232.0
327.00	1631.0
328.00	814.0
328.80	173.0
332.10	614.0
333.00	1115.0
334.00	4955.0
335.00	1261.0
341.00	693.0
342.00	247.0
346.00	1567.0
347.10	549.0
351.00	250.0
352.00	2650.0
353.10	1637.0
354.10	2366.0
355.00	645.0
359.00	268.0
365.00	12135.0

Instrument :
BNA_M
ClientSampleId :
DFTPP034

366.00	1851.0
367.00	163.0
370.00	253.0
371.00	694.0
372.10	4412.0
373.00	925.0
383.00	1195.0
383.90	241.0
390.00	494.0
391.00	277.0
392.20	275.0
401.00	257.0
402.00	1561.0
403.00	2365.0
404.00	751.0
404.90	263.0
414.90	173.0
421.10	2125.0
422.00	1818.0
423.00	12842.0
424.10	2960.0
424.90	377.0
440.10	238.0
441.10	34920.0
442.10	216064.0
443.10	40560.0
444.00	3881.0

Instrument :
BNA_M
ClientSampleId :
DFTPP034

m/z	Abundance
37.10	271.0
38.00	754.0
39.10	5424.0
40.00	923.0
41.10	460.0
42.10	196.0
43.00	309.0
44.00	1923.0
45.00	285.0
49.10	1039.0
50.00	18168.0
51.10	71576.0
52.10	3817.0
53.00	235.0
55.00	597.0
56.00	2477.0
57.00	5131.0
58.00	237.0
59.80	159.0
61.00	1022.0
62.00	1057.0
63.00	2798.0
64.00	389.0
65.00	1331.0
66.80	166.0
68.00	1398.0
69.00	68464.0
70.00	455.0
73.10	1332.0
74.00	7408.0
75.00	11430.0
76.00	3945.0
77.10	76424.0
78.10	4756.0
79.00	5595.0
80.00	4331.0
81.00	6148.0
82.00	1672.0
83.10	1537.0
84.00	380.0
85.00	1081.0
85.90	1682.0
87.00	922.0
87.90	404.0
89.10	168.0
91.00	2063.0
92.00	1627.0
93.00	9877.0
94.00	602.0
95.00	190.0
96.00	658.0
96.90	267.0
98.00	7825.0
99.00	5677.0
100.00	523.0
101.00	3651.0
102.10	215.0
103.00	1004.0
104.00	2469.0
105.00	1975.0
106.00	721.0
107.00	26368.0
108.10	4010.0
109.10	737.0
110.00	45024.0
111.00	6662.0
112.00	968.0
113.00	374.0
115.00	233.0
116.00	1275.0
117.00	23928.0
118.00	1588.0
119.00	275.0
120.00	318.0
121.00	239.0
122.00	1774.0
123.00	2654.0
124.00	1214.0
125.00	1278.0
126.20	756.0
127.00	85160.0
128.00	6598.0
129.00	33848.0
130.00	2714.0
131.00	394.0
132.00	360.0
133.00	446.0
134.00	1207.0
135.00	2868.0
136.00	1068.0
137.00	1635.0

Instrument :
BNA_M
ClientSampleId :
DFTPP034

138.00	318.0
139.00	285.0
139.90	422.0
141.00	4863.0
142.00	1950.0
143.00	996.0
144.00	251.0
145.00	333.0
146.00	781.0
147.00	2804.0
148.00	6631.0
149.00	1672.0
150.10	420.0
151.30	527.0
152.90	1315.0
154.00	1141.0
155.00	2542.0
156.10	4386.0
157.00	645.0
158.00	922.0
159.00	608.0
160.10	1212.0
161.00	2179.0
162.00	664.0
162.80	312.0
164.00	376.0
165.00	1944.0
166.00	1327.0
167.00	9040.0
168.00	5066.0
169.10	735.0
169.80	240.0
171.10	500.0
172.00	946.0
173.00	1019.0
174.00	1917.0
175.10	3692.0
176.00	798.0
177.00	2362.0
178.00	666.0
179.00	7510.0
180.00	5493.0
181.00	2619.0
182.00	405.0
183.00	185.0
184.00	737.0
185.00	4023.0
186.00	28328.0
187.10	7652.0
188.00	768.0
189.00	1696.0
189.90	454.0
191.00	1282.0
192.00	2425.0
193.00	2663.0
194.00	561.0
194.90	399.0
196.00	4876.0
197.00	932.0
198.00	181632.0
199.00	12569.0
200.00	1245.0
201.60	793.0
203.00	1496.0
204.00	7631.0
205.00	11884.0
206.10	53160.0
207.10	8350.0
208.00	2380.0
209.00	810.0
209.90	713.0
211.00	1964.0
212.10	207.0
213.00	207.0
214.90	624.0
216.00	1034.0
217.00	15840.0
218.00	2102.0
219.00	219.0
221.00	5633.0
221.90	1831.0
223.00	3283.0
224.10	30472.0
225.10	7960.0
226.00	812.0
227.00	14791.0
228.00	1761.0
229.00	2739.0
230.00	506.0
231.10	1365.0
232.00	183.0
233.00	213.0
233.90	1021.0
235.00	958.0
236.00	700.0

Instrument :
BNA_M
ClientSampleId :
DFTPP034

237.00	1333.0
239.00	604.0
240.00	300.0
241.00	754.0
242.00	1540.0
243.10	1654.0
244.10	22080.0
245.00	3045.0
246.00	5436.0
246.90	1040.0
249.00	796.0
249.90	190.0
250.90	263.0
252.00	228.0
253.10	367.0
254.20	989.0
255.00	124448.0
256.00	17264.0
257.00	1651.0
258.00	8235.0
259.00	1220.0
259.90	184.0
261.00	174.0
264.00	150.0
265.00	3358.0
266.00	514.0
271.00	211.0
271.90	238.0
273.00	3456.0
274.00	9570.0
275.00	53648.0
276.00	6685.0
277.00	5856.0
278.00	1114.0
279.10	342.0
281.00	472.0
282.00	290.0
283.00	686.0
284.00	633.0
285.00	917.0
285.90	167.0
288.90	225.0
290.00	153.0
291.90	278.0
293.00	1346.0
293.90	485.0
295.10	352.0
296.00	18432.0
297.00	2482.0
300.80	199.0
301.80	234.0
303.00	2359.0
303.80	477.0
308.00	305.0
308.90	163.0
310.00	154.0
313.10	180.0
314.00	1012.0
315.00	2587.0
316.00	1257.0
317.20	176.0
320.90	498.0
321.90	217.0
323.10	5716.0
324.00	943.0
326.00	200.0
326.90	1081.0
328.00	719.0
332.10	502.0
333.00	659.0
334.10	3565.0
335.00	893.0
341.10	646.0
342.10	349.0
346.00	1587.0
346.80	249.0
352.00	2053.0
353.00	1255.0
354.00	1800.0
355.00	676.0
365.00	10393.0
366.00	1067.0
370.00	202.0
371.10	416.0
372.10	3273.0
373.00	759.0
383.00	788.0
384.10	288.0
390.00	328.0
391.10	366.0
392.00	207.0
401.00	306.0
402.00	1306.0
403.10	1365.0
404.00	1030.0

Instrument :
BNA_M
ClientSampleId :
DFTPP034

405.00	177.0
421.00	1533.0
422.00	1434.0
423.10	10222.0
424.00	2019.0
425.20	248.0
441.10	26640.0
442.10	177920.0
443.10	33576.0
444.10	2960.0

Instrument :
BNA_M
ClientSampleId :
DFTPP034