

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010423\
 Data File : BM038362.D
 Acq On : 04 Jan 2023 09:51
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020082

Quant Time: Jan 05 01:31:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM010423.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 01:28:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.998	152	42940	20.000	ng/u1	0.00
20) Naphthalene-d8	10.810	136	174059	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.633	164	112227	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.374	188	268705	20.000	ng/u1	0.00
79) Chrysene-d12	21.545	240	312714	20.000	ng/u1	0.00
88) Perylene-d12	23.974	264	384606	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.363	96	10347	9.017	ng/uL	0.00
4) Pyridine-d5	3.799	84	75248	24.345	ng/u1	0.00
7) Phenol-d5	7.157	99	79586	22.535	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	58452	24.777	ng/u1	0.00
11) 2-Chlorophenol-d4	7.528	132	60045	22.636	ng/u1	0.00
15) 4-Methylphenol-d8	8.710	113	57125	20.693	ng/u1	0.00
21) Nitrobenzene-d5	9.169	128	26547	20.113	ng/u1	0.00
24) 2-Nitrophenol-d4	9.892	143	30496	19.962	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	58143	19.225	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	74682	19.006	ng/u1	0.00
46) Dimethylphthalate-d6	14.039	166	165433	19.663	ng/u1	0.00
49) Acenaphthylene-d8	14.327	160	194136	19.846	ng/u1	0.00
54) 4-Nitrophenol-d4	14.833	143	27774	18.693	ng/u1	0.00
60) Fluorene-d10	15.621	176	150414	20.213	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	34471	17.927	ng/u1	0.00
73) Anthracene-d10	17.474	188	248247	19.762	ng/u1	0.00
81) Pyrene-d10	19.756	212	319170	18.937	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.815	264	370346	19.451	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.399	88	12571	10.563	ng/uL	97
5) Pyridine	3.816	79	82742	25.523	ng/u1	95
6) Benzaldehyde	7.140	77	35515	25.989	ng/u1	94
8) Phenol	7.187	94	85761	23.375	ng/u1	92
10) Bis(2-Chloroethyl)ether	7.422	93	71533	23.852	ng/u1	94
12) 2-Chlorophenol	7.557	128	62061	22.923	ng/u1	98
13) 2-Methylphenol	8.445	108	54955	20.547	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.581	45	903	0.220	ng/u1#	52
16) Acetophenone	8.834	105	101385	21.307	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.810	70	56605	23.007	ng/u1	99
18) 4-Methylphenol	8.775	108	60489	20.727	ng/u1	87
19) Hexachloroethane	9.081	117	27659	21.266	ng/u1	89
22) Nitrobenzene	9.216	77	82981	21.536	ng/u1	98
23) Isophorone	9.739	82	145333	21.195	ng/u1	99
25) 2-Nitrophenol	9.928	139	30278	18.998	ng/u1	99
26) 2,4-Dimethylphenol	9.981	107	71670	20.675	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.222	93	84015	20.357	ng/u1	99
29) 2,4-Dichlorophenol	10.463	162	57301	20.032	ng/u1	96
30) Naphthalene	10.863	128	180868	20.121	ng/u1	99
32) 4-Chloroaniline	10.981	127	75795	18.800	ng/u1	100
33) Hexachlorobutadiene	11.139	225	42438	18.047	ng/u1	94
34) Caprolactam	11.763	113	15231	18.858	ng/u1	96
35) 4-Chloro-3-methylphenol	12.098	107	56980	19.591	ng/u1	94
36) 2-Methylnaphthalene	12.469	142	115907	19.227	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010423\
 Data File : BM038362.D
 Acq On : 04 Jan 2023 09:51
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD020082


Quant Time: Jan 05 01:31:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM010423.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 01:28:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.686	142	119753	19.429	ng/ul#	97
39) 1,2,4,5-Tetrachloroben...	12.833	216	76774	18.635	ng/ul	98
40) Hexachlorocyclopentadiene	12.804	237	48280	16.635	ng/ul	94
41) 2,4,6-Trichlorophenol	13.075	196	49090	19.288	ng/ul	97
42) 2,4,5-Trichlorophenol	13.145	196	51985	19.458	ng/ul	97
43) 1,1'-Biphenyl	13.469	154	162468	19.415	ng/ul	100
44) 2-Chloronaphthalene	13.516	162	133205	19.552	ng/ul	99
45) 2-Nitroaniline	13.722	65	42322	21.411	ng/ul	94
47) Dimethylphthalate	14.086	163	164353	19.428	ng/ul	99
48) 2,6-Dinitrotoluene	14.216	165	31690	19.422	ng/ul	91
50) Acenaphthylene	14.357	152	207259	20.525	ng/ul	98
51) 3-Nitroaniline	14.545	138	26272	18.026	ng/ul	95
52) Acenaphthene	14.698	153	146158	20.461	ng/ul	97
53) 2,4-Dinitrophenol	14.757	184	21161	17.376	ng/ul	90
55) 4-Nitrophenol	14.851	109	27108	18.791	ng/ul	93
56) Dibenzofuran	15.027	168	205587	20.295	ng/ul	100
57) 2,4-Dinitrotoluene	14.998	165	47516	20.234	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.257	232	46626	19.054	ng/ul#	94
59) Diethylphthalate	15.445	149	173894	20.215	ng/ul	98
61) Fluorene	15.674	166	172638	20.167	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.668	204	93559	20.087	ng/ul	93
63) 4-Nitroaniline	15.704	138	27613	20.695	ng/ul#	80
66) 4,6-Dinitro-2-methylph...	15.763	198	33074	17.897	ng/ul	90
67) N-Nitrosodiphenylamine	15.880	169	141158	19.214	ng/ul	99
68) 4-Bromophenyl-phenylether	16.563	248	53971	18.006	ng/ul	96
69) Hexachlorobenzene	16.674	284	61568	17.402	ng/ul	98
70) Atrazine	16.833	200	54685	17.394	ng/ul	97
71) Pentachlorophenol	17.021	266	37070	16.970	ng/ul	99
72) Phenanthrene	17.415	178	274642	19.313	ng/ul	99
74) Anthracene	17.510	178	285337	19.677	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.433	216	74187	17.389	ng/ul	98
76) Pentachlorobenzene	14.945	250	73820	17.953	ng/ul	95
77) Carbazole	17.780	167	248442	19.000	ng/ul	98
78) Di-n-butylphthalate	18.327	149	321310	20.499	ng/ul	99
80) Fluoranthene	19.421	202	349470	18.144	ng/ul	100
82) Pyrene	19.786	202	386015	18.666	ng/ul	99
83) Butylbenzylphthalate	20.674	149	154973	19.634	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.462	252	120761	18.231	ng/ul	99
85) Benzo(a)anthracene	21.527	228	423715	19.833	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.439	149	247673	20.358	ng/ul	98
87) Chrysene	21.580	228	398852	19.377	ng/ul	100
89) Di-n-octyl phthalate	22.368	149	467810	18.558	ng/ul	100
90) Benzo(b)fluoranthene	23.233	252	471208	19.562	ng/ul	98
91) Benzo(k)fluoranthene	23.280	252	450974	18.774	ng/ul	99
93) Benzo(a)pyrene	23.868	252	428067	20.048	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.515	276	560264	20.575	ng/ul	99
95) Dibenzo(a,h)anthracene	26.527	278	467707	20.764	ng/ul	97
96) Benzo(g,h,i)perylene	27.303	276	464868	20.960	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 05 01:31:59 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM010423.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 05 01:28:53 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 10.563 ng/uL
RT: 3.399 min Scan# 30
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion: 88 Resp: 12571
Ion Ratio Lower Upper
88 100
43 34.9 30.6 45.8
58 74.1 60.2 90.4



#5
Pyridine
Concen: 25.523 ng/uL
RT: 3.816 min Scan# 107
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion: 79 Resp: 82742
Ion Ratio Lower Upper
79 100
52 81.2 69.6 104.4
51 41.9 34.8 52.2









#6
Benzaldehyde
Concen: 25.989 ng/ul
RT: 7.140 min Scan# 61
Delta R.T. -0.006 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion: 77 Resp: 35515
Ion Ratio Lower Upper
77 100
105 83.8 64.5 96.7
106 82.4 60.6 90.8



#8
Phenol
Concen: 23.375 ng/ul
RT: 7.187 min Scan# 680
Delta R.T. -0.006 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion: 94 Resp: 85761
Ion Ratio Lower Upper
94 100
65 33.7 29.4 44.2
66 40.9 38.0 57.0









#10
Bis(2-Chloroethyl)ether
Concen: 23.852 ng/ul
RT: 7.422 min Scan# 715
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion: 93 Resp: 71533
Ion Ratio Lower Upper
93 100
63 81.0 69.6 104.4
95 30.3 26.6 40.0



#12
2-Chlorophenol
Concen: 22.923 ng/ul
RT: 7.557 min Scan# 743
Delta R.T. -0.006 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:128 Resp: 62061
Ion Ratio Lower Upper
128 100
64 58.7 47.3 70.9
130 30.1 25.7 38.5





#13

2-Methylphenol

Concen: 20.547 ng/ul

RT: 8.445 min Scan# 895

Delta R.T. -0.006 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Instrument :

BNA_M

ClientSampleId :





SSTD020082

Tgt Ion:108 Resp: 54955

Ion Ratio Lower Upper

108 100

107 90.8 70.3 105.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 0.220 ng/ul

RT: 8.581 min Scan# 917

Delta R.T. 0.047 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51




Tgt Ion: 45 Resp: 903


Ion Ratio Lower Upper

45 100

77 37.8 12.2 18.2#

79 25.3 8.6 13.0#









#16
Acetophenone
Concen: 21.307 ng/ul
RT: 8.834 min Scan# 960
Delta R.T. 0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:105 Resp: 101385
Ion Ratio Lower Upper
105 100
77 84.7 71.9 107.9
51 33.2 29.1 43.7



#17
N-Nitroso-di-n-propylamine
Concen: 23.007 ng/ul
RT: 8.810 min Scan# 956
Delta R.T. -0.006 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion: 70 Resp: 56605
Ion Ratio Lower Upper
70 100
42 67.0 53.0 79.6
101 8.8 7.2 10.8
130 15.7 12.2 18.4









#18
4-Methylphenol
Concen: 20.727 ng/ul
RT: 8.775 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:108 Resp: 60489
Ion Ratio Lower Upper
108 100
107 122.0 86.4 129.6



#19
Hexachloroethane
Concen: 21.266 ng/ul
RT: 9.081 min Scan# 1002
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:117 Resp: 27659
Ion Ratio Lower Upper
117 100
201 115.3 79.8 119.6
199 66.2 50.3 75.5









#22
Nitrobenzene
Concen: 21.536 ng/ul
RT: 9.216 min Scan# 1102
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion: 77 Resp: 82981
Ion Ratio Lower Upper
77 100
123 34.2 26.9 40.3
65 14.0 12.1 18.1



#23
Isophorone
Concen: 21.195 ng/ul
RT: 9.739 min Scan# 1114
Delta R.T. 0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion: 82 Resp: 145333
Ion Ratio Lower Upper
82 100
95 6.6 5.4 8.0
138 14.5 11.1 16.7









#25
2-Nitrophenol
Concen: 18.998 ng/ul
RT: 9.928 min Scan# 1146
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:139 Resp: 30278
Ion Ratio Lower Upper
139 100
65 71.3 55.9 83.9
109 36.4 29.0 43.4



#26
2,4-Dimethylphenol
Concen: 20.675 ng/ul
RT: 9.981 min Scan# 1155
Delta R.T. -0.006 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:107 Resp: 71670
Ion Ratio Lower Upper
107 100
121 46.3 35.2 52.8
122 73.7 58.6 87.8





#27

Bis(2-Chloroethoxy)methane

Concen: 20.357 ng/ul

RT: 10.222 min Scan# 1196

Delta R.T. -0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Instrument :

BNA_M

ClientSampleId :

SSTD020082





Tgt Ion: 93 Resp: 84015

Ion Ratio Lower Upper

93 100

95 32.2 25.4 38.2

123 10.1 7.5 11.3



#29

2,4-Dichlorophenol

Concen: 20.032 ng/ul

RT: 10.463 min Scan# 1237

Delta R.T. 0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51




Tgt Ion:162 Resp: 57301


Ion Ratio Lower Upper

162 100

164 66.6 51.8 77.8

98 33.7 30.6 45.8









#30
Naphthalene
Concen: 20.121 ng/ul
RT: 10.863 min Scan# 1305
Delta R.T. -0.006 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:128 Resp: 180868
Ion Ratio Lower Upper
128 100
129 11.7 8.6 12.8
127 13.0 10.5 15.7



#32
4-Chloroaniline
Concen: 18.800 ng/ul
RT: 10.981 min Scan# 1325
Delta R.T. 0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:127 Resp: 75795
Ion Ratio Lower Upper
127 100
129 32.7 26.2 39.2





#33

Hexachlorobutadiene

Concen: 18.047 ng/ul

RT: 11.139 min Scan# 11

Delta R.T. -0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Instrument :

BNA_M

ClientSampleId :

SSTD020082





Tgt Ion:225 Resp: 42438

Ion Ratio Lower Upper

225 100

223 59.7 52.6 78.8

227 63.6 48.2 72.4



#34

Caprolactam

Concen: 18.858 ng/ul

RT: 11.763 min Scan# 1458

Delta R.T. -0.012 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51




Tgt Ion:113 Resp: 15231


Ion Ratio Lower Upper

113 100

55 203.1 161.5 242.3

56 147.0 125.8 188.8









#35
4-Chloro-3-methylphenol
Concen: 19.591 ng/ul
RT: 12.098 min Scan# 1515
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:107 Resp: 56980
Ion Ratio Lower Upper
107 100
144 24.5 18.1 27.1
142 76.8 57.4 86.2



#36
2-Methylnaphthalene
Concen: 19.227 ng/ul
RT: 12.469 min Scan# 1578
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:142 Resp: 115907
Ion Ratio Lower Upper
142 100
141 89.3 73.2 109.8





#37

1-Methylnaphthalene

Concen: 19.429 ng/ul

RT: 12.686 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Instrument :

BNA_M

ClientSampleId :

SSTD020082





Tgt Ion:142 Resp: 119753

Ion Ratio Lower Upper

142 100

141 92.4 75.9 113.9

116 4.0 4.2 6.2#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.635 ng/ul

RT: 12.833 min Scan# 1640

Delta R.T. 0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Tgt Ion:216 Resp: 76774




Ion Ratio Lower Upper


216 100

214 76.1 62.0 93.0

179 17.9 15.6 23.4

108 17.1 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 16.635 ng/ul
RT: 12.804 min Scan# 1635
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:237 Resp: 48280
Ion Ratio Lower Upper
237 100
235 58.4 51.0 76.4
272 13.0 10.2 15.4



#41
2,4,6-Trichlorophenol
Concen: 19.288 ng/ul
RT: 13.075 min Scan# 1681
Delta R.T. 0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:196 Resp: 49090
Ion Ratio Lower Upper
196 100
198 94.0 77.7 116.5
200 29.0 25.0 37.4









#42
2,4,5-Trichlorophenol
Concen: 19.458 ng/ul
RT: 13.145 min Scan# 1693
Delta R.T. 0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:196 Resp: 51985
Ion Ratio Lower Upper
196 100
198 93.0 76.6 115.0
200 30.7 25.4 38.2



#43
1,1'-Biphenyl
Concen: 19.415 ng/ul
RT: 13.469 min Scan# 1748
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:154 Resp: 162468
Ion Ratio Lower Upper
154 100
153 41.0 32.6 49.0
76 14.2 11.6 17.4









#44
2-Chloronaphthalene
Concen: 19.552 ng/ul
RT: 13.516 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:162 Resp: 133205
Ion Ratio Lower Upper
162 100
164 32.8 26.6 39.8
127 36.2 29.7 44.5



#45
2-Nitroaniline
Concen: 21.411 ng/ul
RT: 13.722 min Scan# 1791
Delta R.T. -0.006 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion: 65 Resp: 42322
Ion Ratio Lower Upper
65 100
92 64.1 47.6 71.4
138 86.6 64.6 96.8





#47

Dimethylphthalate

Concen: 19.428 ng/ul

RT: 14.086 min Scan# 1853

Delta R.T. -0.006 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Instrument :

BNA_M

ClientSampleId :

SSTD020082





Tgt Ion:163 Resp: 164353

Ion Ratio Lower Upper

163 100

194 4.2 3.0 4.6

164 9.7 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 19.422 ng/ul

RT: 14.216 min Scan# 1875

Delta R.T. 0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51




Tgt Ion:165 Resp: 31690


Ion Ratio Lower Upper

165 100

89 72.3 66.2 99.2

121 22.5 18.9 28.3





#50

Acenaphthylene

Concen: 20.525 ng/ul

RT: 14.357 min Scan# 1899

Delta R.T. 0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Instrument :

BNA_M

ClientSampleId :

SSTD020082





Tgt Ion:152 Resp: 207259

Ion Ratio Lower Upper

152 100

151 19.1 16.0 24.0

153 13.3 10.2 15.4



#51

3-Nitroaniline

Concen: 18.026 ng/ul

RT: 14.545 min Scan# 1931

Delta R.T. -0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51




Tgt Ion:138 Resp: 26272


Ion Ratio Lower Upper

138 100

108 9.7 8.6 12.8

92 132.4 111.4 167.2









#52
Acenaphthene
Concen: 20.461 ng/ul
RT: 14.698 min Scan# 1957
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:153 Resp: 146158
Ion Ratio Lower Upper
153 100
152 47.1 40.1 60.1
154 84.0 69.0 103.6



#53
2,4-Dinitrophenol
Concen: 17.376 ng/ul
RT: 14.757 min Scan# 1967
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:184 Resp: 21161
Ion Ratio Lower Upper
184 100
63 74.3 69.1 103.7
154 63.8 54.5 81.7









#55
4-Nitrophenol
Concen: 18.791 ng/ul
RT: 14.851 min Scan# 1983
Delta R.T. 0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:109 Resp: 27108
Ion Ratio Lower Upper
109 100
139 110.5 79.1 118.7
65 140.0 108.6 162.8



#56
Dibenzofuran
Concen: 20.295 ng/ul
RT: 15.027 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:168 Resp: 205587
Ion Ratio Lower Upper
168 100
139 37.8 30.5 45.7









#57
2,4-Dinitrotoluene
Concen: 20.234 ng/ul
RT: 14.998 min Scan# 2008
Delta R.T. -0.006 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:165 Resp: 47516
Ion Ratio Lower Upper
165 100
63 60.1 46.9 70.3
182 1.8 1.3 1.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.054 ng/ul
RT: 15.257 min Scan# 2052
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:232 Resp: 46626
Ion Ratio Lower Upper
232 100
131 48.7 34.4 51.6
130 1.7 2.2 3.4#
166 29.7 22.8 34.2





#59

Diethylphthalate

Concen: 20.215 ng/ul

RT: 15.445 min Scan# 2084

Delta R.T. 0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Instrument :

BNA_M

ClientSampleId :

SSTD020082





Tgt Ion:149 Resp: 173894

Ion Ratio Lower Upper

149 100

177 22.1 17.0 25.4

150 11.0 9.8 14.8



#61

Fluorene

Concen: 20.167 ng/ul

RT: 15.674 min Scan# 2123

Delta R.T. -0.006 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51




Tgt Ion:166 Resp: 172638


Ion Ratio Lower Upper

166 100

165 100.9 78.7 118.1

167 12.9 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 20.087 ng/ul
RT: 15.668 min Scan# 2122
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:204 Resp: 93559
Ion Ratio Lower Upper
204 100
206 30.3 26.8 40.2
141 48.4 42.9 64.3



#63
4-Nitroaniline
Concen: 20.695 ng/ul
RT: 15.704 min Scan# 2128
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:138 Resp: 27613
Ion Ratio Lower Upper
138 100
92 53.4 50.2 75.2
108 86.1 89.3 133.9#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 17.897 ng/ul
 RT: 15.763 min Scan# 2138
 Delta R.T. -0.000 min
 Lab File: BM038362.D
 Acq: 04 Jan 2023 09:51




Instrument :
 BNA_M
 ClientSampleId :
 SST020082


Tgt Ion:198 Resp: 33074
 Ion Ratio Lower Upper
 198 100
 182 3.9 3.0 4.6
 77 25.0 24.8 37.2



#67
 N-Nitrosodiphenylamine
 Concen: 19.214 ng/ul
 RT: 15.880 min Scan# 2158
 Delta R.T. -0.006 min
 Lab File: BM038362.D
 Acq: 04 Jan 2023 09:51

Tgt Ion:169 Resp: 141158
 Ion Ratio Lower Upper
 169 100
 168 70.0 56.4 84.6
 167 37.9 30.5 45.7









#68
4-Bromophenyl-phenylether
Concen: 18.006 ng/ul
RT: 16.563 min Scan# 211
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:248 Resp: 53971
Ion Ratio Lower Upper
248 100
250 98.7 76.7 115.1
141 58.4 50.3 75.5



#69
Hexachlorobenzene
Concen: 17.402 ng/ul
RT: 16.674 min Scan# 2293
Delta R.T. 0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:284 Resp: 61568
Ion Ratio Lower Upper
284 100
142 35.7 30.1 45.1
249 31.8 25.0 37.4









#70
Atrazine
Concen: 17.394 ng/ul
RT: 16.833 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:200 Resp: 54685
Ion Ratio Lower Upper
200 100
173 24.1 18.6 27.8
215 48.7 41.1 61.7



#71
Pentachlorophenol
Concen: 16.970 ng/ul
RT: 17.021 min Scan# 2352
Delta R.T. 0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:266 Resp: 37070
Ion Ratio Lower Upper
266 100
264 64.7 51.6 77.4
268 65.0 50.9 76.3





#72

Phenanthrene

Concen: 19.313 ng/ul

RT: 17.415 min Scan# 2419

Delta R.T. -0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Instrument :

BNA_M

ClientSampleId :

SSTD020082





Tgt Ion:178 Resp: 274642

Ion Ratio Lower Upper

178 100

179 15.6 12.2 18.4

176 19.6 15.4 23.2



#74

Anthracene

Concen: 19.677 ng/ul

RT: 17.510 min Scan# 2435

Delta R.T. 0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51




Tgt Ion:178 Resp: 285337


Ion Ratio Lower Upper

178 100

179 14.8 12.2 18.4

176 18.5 15.5 23.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 17.389 ng/uL
RT: 13.433 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:216 Resp: 74187
Ion Ratio Lower Upper
216 100
214 77.6 60.2 90.4
218 46.9 38.2 57.4



#76
Pentachlorobenzene
Concen: 17.953 ng/uL
RT: 14.945 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:250 Resp: 73820
Ion Ratio Lower Upper
250 100
248 61.3 50.3 75.5
252 60.6 53.2 79.8





#77

Carbazole

Concen: 19.000 ng/ul

RT: 17.780 min Scan# 2481

Delta R.T. -0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Instrument :

BNA_M

ClientSampleId :

SSTD020082





Tgt Ion:167 Resp: 248442

Ion Ratio Lower Upper

167 100

166 21.1 17.8 26.8

139 12.1 10.2 15.2



#78

Di-n-butylphthalate

Concen: 20.499 ng/ul

RT: 18.327 min Scan# 2574

Delta R.T. 0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51




Tgt Ion:149 Resp: 321310


Ion Ratio Lower Upper

149 100

150 9.0 7.4 11.0

104 5.7 4.8 7.2









#80
Fluoranthene
Concen: 18.144 ng/ul
RT: 19.421 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:202 Resp: 349470
Ion Ratio Lower Upper
202 100
101 9.0 7.0 10.6
100 7.3 5.8 8.8



#82
Pyrene
Concen: 18.666 ng/ul
RT: 19.786 min Scan# 2822
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:202 Resp: 386015
Ion Ratio Lower Upper
202 100
101 10.8 8.6 13.0
100 9.7 7.4 11.0





#83

Butylbenzylphthalate

Concen: 19.634 ng/ul

RT: 20.674 min Scan# 2973

Delta R.T. 0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Instrument :

BNA_M

ClientSampleId :

SSTD020082





Tgt Ion:149 Resp: 154973

Ion Ratio Lower Upper

149 100

91 77.1 65.0 97.6

206 19.6 16.0 24.0



#84

3,3'-Dichlorobenzidine

Concen: 18.231 ng/ul

RT: 21.462 min Scan# 3107

Delta R.T. -0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51




Tgt Ion:252 Resp: 120761


Ion Ratio Lower Upper

252 100

254 64.6 52.1 78.1

126 7.6 6.4 9.6





#85

Benzo(a)anthracene

Concen: 19.833 ng/ul

RT: 21.527 min Scan# 3118

Delta R.T. -0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Instrument :

BNA_M

ClientSampleId :

SSTD020082





Tgt Ion:228 Resp: 423715

Ion Ratio Lower Upper

228 100

229 19.2 15.6 23.4

226 27.0 21.1 31.7



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.358 ng/ul

RT: 21.439 min Scan# 3103

Delta R.T. -0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51




Tgt Ion:149 Resp: 247673


Ion Ratio Lower Upper

149 100

167 23.6 19.7 29.5

279 3.2 2.8 4.2









#87
Chrysene
Concen: 19.377 ng/ul
RT: 21.580 min Scan# 3127
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:228 Resp: 398852
Ion Ratio Lower Upper
228 100
226 29.7 24.0 36.0
229 19.4 15.4 23.0



#89
Di-n-octyl phthalate
Concen: 18.558 ng/ul
RT: 22.368 min Scan# 3261
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:149 Resp: 467810









#90
Benzo(b)fluoranthene
Concen: 19.562 ng/ul
RT: 23.233 min Scan# 3408
Delta R.T. -0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51




Instrument :
BNA_M
ClientSampleId :
SSTD020082


Tgt Ion:252 Resp: 471208
Ion Ratio Lower Upper
252 100
253 21.1 17.8 26.6
125 8.2 6.6 10.0



#91
Benzo(k)fluoranthene
Concen: 18.774 ng/ul
RT: 23.280 min Scan# 3416
Delta R.T. 0.000 min
Lab File: BM038362.D
Acq: 04 Jan 2023 09:51

Tgt Ion:252 Resp: 450974
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 8.5 6.3 9.5





#93

Benzo(a)pyrene

Concen: 20.048 ng/ul

RT: 23.868 min Scan# 3516

Delta R.T. 0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Instrument :

BNA_M

ClientSampleId :

SSTD020082





Tgt Ion:252 Resp: 428067

Ion Ratio Lower Upper

252 100

253 21.5 17.7 26.5

125 9.4 7.4 11.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.575 ng/ul

RT: 26.515 min Scan# 3966

Delta R.T. 0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51




Tgt Ion:276 Resp: 560264


Ion Ratio Lower Upper

276 100

138 20.6 15.7 23.5

277 25.4 20.2 30.4





#95

Dibenzo(a,h)anthracene

Concen: 20.764 ng/ul

RT: 26.527 min Scan# 3968

Delta R.T. 0.000 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Instrument :

BNA_M

ClientSampleId :

SSTD020082





Tgt Ion:278 Resp: 467707

Ion Ratio Lower Upper

278 100

139 13.5 10.0 15.0

279 21.7 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 20.960 ng/ul

RT: 27.303 min Scan# 4100

Delta R.T. 0.006 min

Lab File: BM038362.D

Acq: 04 Jan 2023 09:51

Tgt Ion:276 Resp: 464868

Ion Ratio Lower Upper

276 100

138 16.7 13.8 20.6

277 22.1 18.6 28.0

